



American Telecommunications Certification Body Inc.
6731 Whittier Ave, McLean, VA 22101

April 29, 2005

RE: Medtronic Inc.

FCC ID: LF537751

After a review of the submitted information, I have a few comments on the above referenced Application.

- 1) Earlier information regarding 175 kHz mentioned that the peak to average difference was 8 dB, while the latest corrected report mentions 3 dB. Please confirm the correct ratio.

Timothy R. Johnson
Examining Engineer

[mailto: tjohnson@AmericanTCB.com](mailto:tjohnson@AmericanTCB.com)

The items indicated above must be submitted before processing can continue on the above referenced application. Failure to provide the requested information may result in application termination. Correspondence should be considered part of the permanent submission and may be viewed from the Internet after a Grant of Equipment Authorization is issued.

Please do not respond to this correspondence using the email reply button. In order for your response to be processed expeditiously, you must submit your documents through the AmericanTCB.com website. Also, please note that partial responses increase processing time and should not be submitted.

Any questions about the content of this correspondence should be directed to the sender.