

TÜV America Inc.
1775 Old Highway 8 NW
Suite 104
New Brighton, MN 55112

Phone: (651) 631-2487
Fax: (651) 638-0285
E-mail: info@tuvam.com
www.TUVamerica.com



29 April 2005

RE: Medtronic Inc

FCC ID: LF537751

In response to 29 April 2005 ATCB comments on the above referenced application.

The peak to average difference of 3 dB is correct for FCC ID: LF37751.

Please let us know if you require any further information.

Thank you.

A handwritten signature in black ink that reads "Joel T. Schneider". The signature is written in a cursive, flowing style.

Joel T Schneider
Senior Engineer