

Medtronic

November 23, 2016

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

To whom it may concern.

Reference: ODIN External Neurostimulator model 4NR003
FCC ID: LF54NR003

I, the undersigned, as the authorized signatory for Medtronic, Inc. hereby apply to seek original grant for ODIN External NeuroStimulator model 4NR003. Model 4NR003 is a External Neurostimulator for deep brain long term clinical investigation. It contains a WIFI radio operating in the 2.4-2.4835GHz range with less than 10 dBm e.i.r.p.. The ENS model 4NR003 is compliant to FCC rule parts 15.247.

Thank you in advance, should you have any questions please feel free to contact me.



Guillaume Girard
EMC and RF Regulatory Compliance
Technical Fellow
Medtronic, Inc.
Tel 763-526-0652
g.guillaume@medtronic.com