


TCB  
BABT Product Service  
4855 Patrick Henry Drive, Building 6  
Santa Clara CA 95054

Declaration of Compliance to RF Exposure Limits for Humans

The 3031A, 7434A and 7438 Programmer complies with the RF exposure limits for humans as called out in RSS-102. The transmitters are used to transmit data relating to Implantable Device Programming. It is exempt from RF evaluation based on its operating frequency of 175 kHz, and output power of 4 microwatts based on  $TP = (FS \times D)^2 / 30 \times 1.64$ ,  $(1412 \mu V/m \times 10)^2 / 30 \times 1.64$ . This would be less than the 200 milliwatt requirement for a portable device.

A handwritten signature in black ink that reads "Joel T. Schneider". The signature is written in a cursive style with a large initial 'J' and a long, sweeping underline.

Joel T Schneider

TUV Product Service