

TÜV PRODUCT SERVICE INC  
19333 WILD MOUNTAIN ROAD  
TAYLORS FALLS, MINNESOTA 55084-1786

ENGINEERING STATEMENT

For Certification of

Medtronic Incorporated – Neurological Division  
710 Medtronic Parkway  
Minneapolis MN 55432-5604

N'VISION Hand Held Clinical Programmer

Model No: 8840


FCC ID: LF58840

I am an Electromagnetic Compatibility Engineer for TÜV PRODUCT SERVICE INC, Taylors Falls, Minnesota. My education and experience are a matter of record with the Federal Communications Commission.

TÜV PRODUCT SERVICE INC has been authorized by Clifton W Owens, Vice President and General Manager, of Medtronic Incorporated – Neurological Division, to make Certification measurements on the 8840 Hand Held Clinical Programmer. These tests were made by me or under my supervision at our Wild River Laboratory which is on file with the FCC under Paragraph 2.948 of the FCC Rules.

Test data required by the TCB is included in this report.

It is submitted that the above equipment meets the FCC Part 15, Subpart C (15.209) Requirements and Certification is hereby requested.



Joel T. Schneider  
Chief Engineer  
Minnesota Wild River Lab  
TÜV Product Service  
Jschneider@tuvam.com  
Dated: 13 September 2001