

10CM



Medtronic

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- [1] this device may not cause harmful interference and
- [2] this device must accept any interference received, including interference that may cause undesired operation.



Manufacturer
Medtronic, Inc.
 Minneapolis, MN 55432
 USA



E215693

Medical Equipment
 With Respect to Electric Shock,
 Fire and Mechanical Hazards Only
 In Accordance with UL60601-1
 CAN/CSA C22.2 No. 601.1



Authorized Representative
in the European Community
Medtronic, B.V.
 Earl Bakkenstraat 10
 6422 PJ Heerlen
 The Netherlands

Input: 15V 4.2A max.



Rx Only

FCC ID: LF58880CW



8880CW

Only use the adapter
 Sinpro Electronics Co. Ltd.
 Type: MPU63-106
 Input: 100-240Vac~, 47-63Hz. 1.62-0.72A
 Output: 15 Vdc .4.2A max

IC : 3408D-MICA101

10CM



Medtronic

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- [1] this device may not cause harmful interference and
- [2] this device must accept any interference received, including interference that may cause undesired operation.



Manufacturer
Medtronic, Inc.
 Minneapolis, MN 55432
 USA



E215693

Medical Equipment
 With Respect to Electric Shock,
 Fire and Mechanical Hazards Only
 In Accordance with UL60601-1
 CAN/CSA C22.2 No. 601.1



Authorized Representative
in the European Community
Medtronic, B.V.
 Earl Bakkenstraat 10
 6422 PJ Heerlen
 The Netherlands

Input: 15V 4.2A max.



Rx Only

FCC ID: LF58880CW



8880CW

Only use the adapter
 Sinpro Electronics Co. Ltd.
 Type: MPU63-106
 Input: 100-240Vac~, 47-63Hz. 1.62-0.72A
 Output: 15 Vdc .4.2A max

IC : 3408D-MICA101

5CM

Label Sample & Location

