

Medtronic

December 17, 2016

Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road
Columbia, Maryland 21046

To whom it may concern.

Reference: INS model 97715 and B35300R under LF597715

I, the undersigned, as the authorized signatory for Medtronic, Inc. hereby apply to seek original grant for Implantable Neurostimulator (INS) Model(s): 97715, and B35300R. These model contained a radio operating in the 402-405MHz band with less than 25 μ W e.r.p.

Model difference is based on therapy delivery firmware. All radio hardware, firmware and antenna are identical.

Thank you in advance, should you have any questions please feel free to contact me.



Guillaume Girard
EMC and RF Regulatory Compliance
Technical Fellow
Medtronic, Inc.
Tel 763-526-0652
g.guillaume@medtronic.com