


1. THIS SPECIFICATION IS CONTROLLED BY CRHF PACKAGE ENGINEERING. THIS SPECIFICATION IS INTENDED TO BE THE FINAL LABEL DESIGN OUTPUT GENERATED FROM A COMBINATION OF LABEL PARTS WITHIN THE ASSOCIATED PDM BOM.
2. THESE LABEL IMAGES ARE THE COMBINATION OF THE TEMPLATE FILE, MODEL DATA RECORD, SAMPLE TRIGGER DATA AND REPRESENTATIVE VIEWS OF THE LABEL DIE LINES.
3. PRINTING REQUIREMENTS - THE LABELS ARE TO BE PRINTED USING A 300 DPI ZEBRA THERMAL TRANSFER PRINTER.
4. THE LOT OR SERIAL NUMBER, USE BY DATE, DATE OF MANUFACTURE, PIN NUMBER, ADDRESS CODE, GTIN, PRODUCT NUMBER (REF) AND BAR CODES ARE SHOWN FOR EXAMPLE ONLY. THESE FIELDS ARE VARIABLE AND ARE TO BE SUPPLIED BY MANUFACTURING AT THE TIME OF PRINTING.
5. SPACING AND TEXT WRAPPING SHOWN IN THE IMAGES MAY DIFFER SLIGHTLY FROM ACTUAL PRINTED LABEL, AS LONG AS CONTENT OF PRINTED VERSION IS ACCURATE. THE IMAGES INCLUDE REPRESENTATION OF THE PRE-PRINTED LABEL STOCK. BRANDING, PART NUMBER BAR CODES (2D OR OTHER) AND OTHER PRE-PRINTED AREAS OR ADDITIONAL STICKERS MAY CHANGE AND ARE ACCEPTABLE.
6. THE LABEL PRINTED IN MANUFACTURING MUST BE AT AN EQUAL OR LATER MODEL DATA AND TEMPLATE REVISION AS SHOWN ON THE IMAGE.
7. (FOR LABEL SET ONLY) THE TRIGGER DATA SHOWN ON THE LABEL SET IMAGE MAY NOT MATCH THE PRINTED LABEL.

SPEC TAB	MFG ADDRESS
A85058-001	Juncos, Puerto Rico, USA
A85058-002	Tolochenaz, Switzerland

DeviceImageSpec_02

 Medtronic MEDTRONIC, INC. MINNEAPOLIS, MINNESOTA USA	TITLE LABEL IMAGE - AZURE XT W1DR01 US WIRELESS				
	MEDTRONIC CONFIDENTIAL THIS DOCUMENT IS THE PROPERTY OF MEDTRONIC, INC. AND MUST BE ACCOUNTED FOR. INFORMATION HEREON IS CONFIDENTIAL. DO NOT REPRODUCE IT, REVEAL IT TO UNAUTHORIZED PERSONS OR SEND IT OUTSIDE MEDTRONIC WITHOUT PROPER AUTHORIZATION.				
	LABEL IMAGE SPEC TEMPLATE: A83320 REV C				
	DATE 26-Oct-2016	SCALE N/A	ENGINEERING GLENN BIDDICK		
MODEL W1DR01	Size B	PART NUMBER A85058	REV A	SHEET 1 of 3	

Medtronic **Bluetooth**

Azure™ XT DR MRI SureScan™

Pacemaker (dual chamber, RA, RV)

(Model) **W1DR01**

SN **WWW000000W**

2000-00-00
YYYY-MM-DD

REF **W1DR01**

OAE-DDDR IS-1 MR conditional

ACCELEROMETER

Burst+: RA Ramp: RA

Package contents: 1x Implantable device Torque wrench Product documentation

AAI <=> DDD

Upper tracking rate and lower rate: \uparrow 130 bpm, \downarrow 60 bpm

Amplitude and pulse width: RA: \uparrow 3.5 V, \leftrightarrow 0.4 ms

Sensitivity: RA: 0.3 mV

Amplitude and pulse width: RV: \uparrow 3.5 V, \leftrightarrow 0.4 ms

Sensitivity: RV: 0.9 mV

PVARP: AUTO

A-V interval (paced and sensed): \rightarrow 180 ms, \rightarrow 150 ms

Pacing polarity: RA, RV

Therapies: AT, AF

Adaptive

Rx Only

FCC ID: LFSBLEIMPLANT

STERILE EO

55 °C 131 °F

-18 °C 0 °F

Medtronic, Inc., 710 Medtronic Parkway, Minneapolis, MN 55432 USA
Manufactured in: Tolochenaz, Switzerland
www.medtronic.com

M968371A001 A - M967969A001 A

Shelf Box Label

Medtronic OAE-DDDR **2x IS-1**

Azure™ XT DR MRI SureScan™ **W1DR01**

REF **W1DR01**

SN **WWW000000W**

2000-00-00
YYYY-MM-DD

(01)000000000000(17)000000(21)WWW000000W

M968371A001 A - M967969A001 A

End Label 1

Medtronic **Bluetooth**

Azure™ XT DR MRI SureScan™

Pacemaker (dual chamber, RA, RV)

(Model) **W1DR01**

SN **WWW000000W**

2000-00-00
YYYY-MM-DD

REF **W1DR01**

OAE-DDDR IS-1 MR conditional

ACCELEROMETER

Burst+: RA Ramp: RA

Package contents: 1x Implantable device Torque wrench

AAI <=> DDD

Upper tracking rate and lower rate: \uparrow 130 bpm, \downarrow 60 bpm

Amplitude and pulse width: RA: \uparrow 3.5 V, \leftrightarrow 0.4 ms

Sensitivity: RA: 0.3 mV

Amplitude and pulse width: RV: \uparrow 3.5 V, \leftrightarrow 0.4 ms

Sensitivity: RV: 0.9 mV

PVARP: AUTO

A-V interval (paced and sensed): \rightarrow 180 ms, \rightarrow 150 ms

Pacing polarity: RA, RV

Therapies: AT, AF

Adaptive

Rx Only

FCC ID: LFSBLEIMPLANT

STERILE EO

55 °C 131 °F

-18 °C 0 °F

Medtronic, Inc., 710 Medtronic Parkway, Minneapolis, MN 55432 USA
Manufactured in: Tolochenaz, Switzerland
www.medtronic.com

M968371A001 A - M967969A001 A

Sterile Pack Label

Medtronic OAE-DDDR **2x IS-1**

Azure™ XT DR MRI SureScan™ **W1DR01**

• BlueSync™ cardiac monitoring system

REF **W1DR01**

SN **WWW000000W**

2000-00-00
YYYY-MM-DD

M968371A001 A - M967969A001 A

PIN: W000WWW00W

End Label 2

M968371A001 A - M967969A001 A

2000-00-00
YYYY-MM-DD

Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W
Medtronic W1DR01	Azure™ XT DR MRI SureScan™ SN WWW000000W

Medtronic **Bluetooth**

Azure™ XT DR MRI SureScan™ W1DR01

Implantation date: 2000-00-00
YYYY-MM-DD

REF **W1DR01**

SN **WWW000000W**

Medtronic **Bluetooth**

Azure™ XT DR MRI SureScan™ W1DR01

Implantation date: 2000-00-00
YYYY-MM-DD

SN **WWW000000W**

Medtronic **Bluetooth**

Azure™ XT DR MRI SureScan™ W1DR01

Implantation date: 2000-00-00
YYYY-MM-DD

SN **WWW000000W**

FYR Labels

SPEC TAB
A85058-002

NOTE: IMAGES NOT TO SCALE
LABEL IMAGE SPEC TEMPLATE: A83320 REV C

PART NUMBER	REV	SHEET
A85058	A	3 of 3