


1. The label images shown are a representation of printed labels using the EBOM label stock and label software files with the label system database and manufacturing-provided data.
2. Manufacturing printed content:
  - a. Printers shall be configured to enable printability of all content.
  - b. Printed content shall be positioned on each individual label stock approximately as shown.
  - c. Size and position of printed content may vary.
  - d. Printed content shall be indelible.
  - e. No parts of a letter, whole letter, symbol or graphic shall be blurred, missing or broken which impairs legibility when viewed without magnification.
3. Data provided by manufacturing:
  - a. Serial Number shall be ten alpha-numeric characters.
  - b. Use by Date to populate the indicated format of YYYY-MM-DD.
  - c. Date of Manufacture to populate the indicated format of YYYY-MM-DD and be determined from the date of initial battery connection.
  - d. PIN number shall be ten alpha-numeric characters.
  - e. REF number
  - f. "Manufactured in:" text per the Specification Tabulation table.
4. The GS1-128 bar code on End Label 1 shall contain:
  - a. Encoded bar code data in the following order:
    - i. Application Identifier 01 followed by the GTIN (also known as UPN)
    - ii. Application Identifier 17 followed by the Use by Date in format YYMMDD
    - iii. Application Identifier 21 followed by the Serial Number
  - b. Human readable text below the bar code in the following order:
    - i. (01) followed by GTIN
    - ii. (17) followed by Use by Date in format YYMMDD
    - iii. (21) followed by Serial Number
5. Data matrix codes shown on the indicated labels (Shelf Box, Sterile Pack, End Label 1, End Label 3, FYR [upper left corner]) are provided to facilitate manufacturing. Encoded information is at the discretion of the manufacturing site.
6. The FYR Labels shall contain:
  - a. Encoded data:
    - i. REF Number for the REF Code 39 bar code
    - ii. Serial Number for the SN Code 39 bar code
    - iii. Serial Number for the SN data matrix code
  - b. Human readable text:
    - i. REF Number above the REF Code 39 bar code
    - ii. Serial Number above the SN Code 39 bar code
    - iii. Serial Number below the SN data matrix code
7. The label printed in manufacturing may be at an equal or later model data and template revision as shown on each label.

Specification Tabulation	"Manufactured in:" Text
A94231-001	Tolochenaz, Switzerland

DeviceImageSpecEnd3\_05

 MEDTRONIC, INC. MINNEAPOLIS, MINNESOTA USA  MEDTRONIC CONFIDENTIAL THIS DOCUMENT IS THE PROPERTY OF MEDTRONIC, INC. AND MUST BE ACCOUNTED FOR. INFORMATION HEREON IS CONFIDENTIAL. DO NOT REPRODUCE IT, REVEAL IT TO UNAUTHORIZED PERSONS OR SEND IT OUTSIDE MEDTRONIC WITHOUT PROPER AUTHORIZATION.	TITLE				
	LABEL IMAGE SPEC-COBALT XT HF QUAD CRTD MRI DTPA2QQ OUS MKT				
	LABEL IMAGE SPEC TEMPLATE: A83320 REV E				
	DATE	SCALE	ENGINEERING		
12-Sep-2019	N/A	TIM WIEGAND			
MODEL	Size	PART NUMBER	REV	SHEET	
DTPA2QQ	B	A94231	A	1 of 2	

**Medtronic** Bluetooth

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™

# DTPA2QQ

SN WWW000000W

2000-00-00  
YYYY-MM-DD

REF DTPA2QQ

IS-1 IS4-LLLL DF4-LLHH

DDE-DDDR A LV

ACTIVE CAN™

1x

DDD

↑ 130 min-1  
↓ 50 min-1

0.3 mV

0.3 mV

→ 130 ms\*  
→ 100 ms\*

↑ 3.5 V\*  
→ 0.4 ms\*

↑ 3.5 V\*  
→ 0.4 ms\*

↑ 3.5 V\*  
→ 0.4 ms

LV1 / LV2

LV->RV\*  
0 ms\*

OFF

OFF

OFF

OFF

OFF

ON

FCC ID: LF5BLEMPLANT4  
IC: 3408D-BLEIMPLANT4

Medtronic B.V.  
Earl Bakkenstraat 10  
6422 PJ Heerlen  
The Netherlands

www.medtronic.com/manuals

Medtronic, Inc., 710 Medtronic Parkway, Minneapolis, MN 55432 USA  
Manufactured in: Tolochenaz, Switzerland  
Product of Switzerland

STERILE EO

CE0123

M984919A001\_A - M984991A001\_A

Shelf Box Label

**Medtronic**

DDE-DDDR 1x IS-1 BI 1x IS4-LLLL 1x DF4-LLHH

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™ DTPA2QQ

REF DTPA2QQ

SN WWW000000W

2000-00-00  
YYYY-MM-DD

REF DTPA2QQ

01,0000000000000000(17,000000)21WWW000000W

M984919A001\_A - M984991A001\_A

End Label 1

**Medtronic** Bluetooth

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™

# DTPA2QQ

SN WWW000000W

2000-00-00  
YYYY-MM-DD

REF DTPA2QQ

IS-1 IS4-LLLL DF4-LLHH

DDE-DDDR A LV

ACTIVE CAN™

DDD

↑ 130 min-1  
↓ 50 min-1

0.3 mV

0.3 mV

→ 130 ms\*  
→ 100 ms\*

↑ 3.5 V\*  
→ 0.4 ms\*

↑ 3.5 V\*  
→ 0.4 ms\*

↑ 3.5 V\*  
→ 0.4 ms

LV1 / LV2

LV->RV\*  
0 ms\*

OFF

OFF

OFF

OFF

ON

FCC ID: LF5BLEMPLANT4  
IC: 3408D-BLEIMPLANT4

www.medtronic.com/manuals

Medtronic, Inc., 710 Medtronic Parkway, Minneapolis, MN 55432 USA  
Manufactured in: Tolochenaz, Switzerland

STERILE EO

CE0123

M984919A001\_A - M984991A001\_A

Sterile Pack Label

**Medtronic**

DDE-DDDR 1x IS-1 BI 1x IS4-LLLL 1x DF4-LLHH

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™ DTPA2QQ

REF DTPA2QQ

SN WWW000000W

2000-00-00  
YYYY-MM-DD

REF DTPA2QQ

01,0000000000000000(17,000000)21WWW000000W

M984919A001\_A - M984991A001\_A

End Label 2

M984919A001\_A - M984991A001\_A

2000-00-00  
YYYY-MM-DD

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

Medtronic DTPA2QQ Cobalt XT HF™ Quad CRT-D MRI SureScan™ SN WWW000000W

FYR Labels

**Medtronic** Bluetooth

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™ DTPA2QQ

Implementation date: 2000-00-00  
YYYY-MM-DD

REF DTPA2QQ

SN WWW000000W

**Medtronic** Bluetooth

**DF4** Cobalt XT HF™ Quad CRT-D MRI SureScan™ DTPA2QQ

Implementation date: 2000-00-00  
YYYY-MM-DD

SN WWW000000W

**Medtronic**

SN

WWW 000 000 W

en CRT-D	it CRT-D	pt CRT-D	sk CRT-D	hr CRT-D	pt-br CRT-D
fr CRT-D	no CRT-D	hu CRT-D	tr CRT-D	et CRT-D	fi CRT-D
de CRT-D	sv CRT-D	pl CRT-D	ro CRT-D	lt CRT-D	
es CRT-D	el CRT-D	cs CRT-D	bg CRT-D	lv CRT-D	
nl CRT-D	da CRT-D	ru CRT-D	sr CRT-D	sl CRT-D	

M984919A001\_A - M984991A001\_A

End Label 3

DeviceImageSpecEnd3\_05

TITLE LABEL IMAGE SPEC-COBALT XT HF QUAD CRTD MRI DTPA2QQ OUS MKT		
PART NUMBER A94231	REV A	SHEET 2 of 2

**SPEC TAB**  
**A94231-001**

NOTE: IMAGES NOT TO SCALE  
LABEL IMAGE SPEC TEMPLATE: A83320 REV E