

Smart Technology. Delivered.™

Test Report # 317245B

Equipment Under Test:	Patient Assistant
Test Date(s):	11/8/17 – 11/13/17
Prepared for:	ATTN: Ana Santos Medtronic Inc 710 Medtronic Parkway NE Minneapolis MN 55432

Report Issued by: Shane Dock, EMC Engineer

Signature: Shane Inck

Date: 2/27/2018

Report Reviewed by: Adam Alger, Quality Systems Engineer		
Signature: Mum O Age	Date: 12/4/2017	

Report Constructed by: Shane Dock, EMC Engineer

Signature:

Shame Dock

Date: 1/9/2018

This test report may not be reproduced, except in full, without written approval of Laird Technologies, Inc.

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 1 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



CONTENTS

С	ontents		2
	Laird T	echnologies Test Services in Review3	;
1	Test	Report Summary4	ŀ
2	Clie	nt Information5	;
	2.1	Equipment Under Test (EUT) Information5	;
	2.2	Product Description5	;
	2.3	Modifications Incorporated for Compliance5	;
	2.4	Deviations and Exclusions from Test Specifications6	;
	2.5	Additional Information6	;
3	Refe	erences6	;
4	Unc	ertainty Summary7	7
5	Test	Data8	}
	5.1	Antenna Port Conducted Emissions	3
	5.2	Radiated Emissions11	_
6	Revi	sion History15	,

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 2 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



Laird Technologies Test Services in Review

The Laird Technologies, Inc. laboratory located at W66 N220 Commerce Court Cedarburg, Wisconsin, 53012 USA is recognized through the following organizations:



A2LA – American Association for Laboratory Accreditation Accreditation based on ISO/IEC 17025: 2005 with Electrical (EMC) Scope A2LA Certificate Number: 1255.01 Scope of accreditation includes all test methods listed herein, unless otherwise noted.



Federal Communications Commission (FCC) – USA

Accredited recognition of two 3 meter Semi-Anechoic Chambers Accredited Test Firm Registration Number: 953492



Innovation, Science and Economic Development Canada

ISED Site listing of two 3 meter Semi-Anechoic Chambers based on RSS-GEN – Issue 4 File Number: IC 3088A-2 File Number: IC 3088A-3

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 3 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



1 TEST REPORT SUMMARY

During **11/8/17 – 11/13/17** the Equipment Under Test (EUT), **Patient Assistant**, as provided by **Medtronic Inc.** was tested to the following requirements:

Requirement	Description	Specification	Method	Result
FCC: 15.209 IC: ICES-003, RSS-210, RSS GEN	Radiated Emissions	9 kHz - 30 MHz	ANSI C63.10	Pass
FCC: 15.207 IC: RSS-GEN 8.8, RSS- 210	AC Power Line Conducted Emissions	0.150-30 MHz	ANSI C63.10	N/A (Battery Powered EUT)

Notice:

The results relate only to the item tested and described in this report. Any modifications made to the equipment under test after the specified test date(s) may invalidate the data herein.

If the resulting measurement margin is seen to be within the uncertainty value, as listed in this report, the possibility exists that this unit may not meet the required limit specification if subsequently tested.

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 4 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



2 CLIENT INFORMATION

Company Name	Medtronic Inc	
Contact Person	Keely Wagner (Plexus)	
Address	Medtronic Inc 710 Medtronic Parkway NE Minneapolis MN 55432	

2.1 Equipment Under Test (EUT) Information

The following information has been supplied by the client

Product Name	Patient Assistant
Model Number	PA97000
Serial Number	MPB000031A
FCC / IC ID	FCC: LF5PA97000 IC: 3408D-PA97000

2.2 Product Description

The Activator uses a magnetic dual tone wake up signal which functions as a Tissue Conductance Communication (TCC) sting signal to initiate ICM BLE fast-advertising facilitating BLE connection. After BLE connection a symptom mark command is sent to the Device and the Device responds with an acknowledgement. A successful symptom mark is indicated by an audible tone and an illuminated green LED. The patient assistant activates the data management features in the LINQ II implanted device to initiate recording of cardiac event data in the implanted device memory.

Use of the Patient Assistant

R			
Press the symptom Blue light will start flashing		Patient holds PA over LINQ II device until the success tone and light occur.	Tone sounds, green light illuminates

2.3 Modifications Incorporated for Compliance

None noted at time of test

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 5 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



- 2.4 Deviations and Exclusions from Test Specifications None noted at time of test
- 2.5 Additional Information

EUT emissions peaked out in three orientations. Unit programmed via button presses. Unit configured to run with a 3 VDC Bench Supply.

3 REFERENCES

Publication	Edition	Date
CFR 47 Part 15	-	2017
ANSI C63.10	-	2013
RSS-247	2	2017
RSS GEN	4	2014
RSS-210	9	2016

Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 6 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



4 UNCERTAINTY SUMMARY

Using the guidance of the following publications the calculated measurement uncertainty represents an expanded uncertainty expressed at approximately the 95 % confidence level, using a coverage factor of k = 2.

References	Version / Date
CISPR 16-4-1	Ed. 2 (2009-02)
CISPR 16-4-2	Ed. 2 (2011-06)
CISPR 32	Ed. 1 (2012-01)
ANSI C63.23	2012
A2LA P103	February 4, 2016
A2LA P103c	August 10, 2015
ETSI TR 100-028	V1.3.1 (2001-03)

Measurement Type	Configuration	Uncertainty ±	
Radiated Emissions	Biconical Antenna	5.0 dB	
Radiated Emissions	Log Periodic Antenna	5.3 dB	
Radiated Emissions	Horn Antenna	4.7 dB	
AC Line Conducted Emissions	Artificial Mains Network	3.4 dB	
Telecom Conducted Emissions Asymmetric Artificial Network		4.9 dB	
Disturbance Power Emissions Absorbing Clamp		4.1 dB	
Radiated Immunity	3 Volts/meter	2.2 dB	
Conducted Immunity	CDN/EM/BCI	2.4/3.5/3.4 dB	
EFT Burst/Surge	Peak pulse voltage	164 volts	
ESD Immunity 15 kV level		1377 Volts	

Parameter	ETSI U.C. ±	U.C. ±
Radio Frequency, from F0	1x10 ⁻⁷	0.55x10 ⁻⁷
Occupied Channel Bandwidth	5 %	2 %
RF conducted Power (Power Meter)	1.5 dB	1.2 dB
RF conducted emissions (Spectrum Analyzer)	3.0 dB	1.7 dB
All emissions, radiated	6.0 dB	5.3 dB
Temperature	1° C	0.65° C
Humidity	5 %	2.9 %
Supply voltages	3 %	1 %

Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 7 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



5 TEST DATA

5.1 Antenna Port Conducted Emissions

Description of	The direct measurement of emissions at the antenna port of the EUT is achieved by use of a RF connection to a spectrum analyzer or power meter.
Measurement	The cable and attenuator factors are loaded into the analyzer or power meter allowing for direct measurement readings without the need for further corrections.
Example Calculations	Measurement (dBm) + Cable factor (dB) + External Attenuator (dB) = Corrected Reading (dBm) Margin (dB) = Limit (dBm) – Corrected Reading (dBm)

Block Diagram



Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 8 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



5.1.1 Antenna Port Conducted Emissions – Bandwidth

Operator	Shane Dock
Test Date	11/13/17
Location	Conducted RF Measurement Area
Temp. / R.H.	71/42%
Requirement	OBW: FCC: 2.1049 IC: RSS-GEN 6.6
Method	ANSI C63.10 Section 6.9.2

Test Parameters

Frequency	175 kHz
-----------	---------

Tables

Value	Tx Mode
99% BW	84.1
(kHz)	

Instrumentation

Smar	Laird Smart Technology. Delivered.							
	Date :	3-Nov-2017	Test	LF Tx			Job :	C-2837
PE: Shane Dock		Customer : Plexus			Quote : 317245			
			-		-	-		
No	. Asset	Description	Manufacturer	Model	Serial	Cal Date	Cal Due Date	Equipment Status
1	EE 960088	Analyzer - EMI Receiver	Agilent	N9038A	MY 51210138	3/2/2017	3/2/2018	Active Calibration
2	AA 960006	Active Loop Antenna	EMCO	6502	9205-2753	8/28/2017	8/28/2019	Active Calibration

Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 9 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



Plots



Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 10 of 15	Model: PA97000
Job: C-2837	-	Serial: See Section 2



5.2 Radiated Emissions

	The frequency spectrum is investigated for intentional and / or unintentional signals emanating from the EUT by use of a standardized test site and measurement antenna.
Description of Measurement	The antenna, cable, pre-amp, and other necessary measurement system correction factors are loaded onto the EMI receiver / spectrum analyzer when the measurements are performed allowing the data to be gathered and reported as corrected values.
	The maximum emissions from the EUT are determined by turn-table azimuth rotation (360°) and scanning of the measurement antenna. Maximized levels are noted at degree values of azimuth, measurement antenna height, and measurement antenna polarity.
	Measurement (dBμV) + Cable factor (dB) + Other (dB) + Antenna Factor (dB/m) = Corrected Reading (dBμV/m)
Example Calculations	Margin (dB) = Limit (dB μ V/m) - Corrected Reading (dB μ V/m) Example at 4000 MHz:
	Average Limit = 20 log (500) = 54 dB μ V/m Margin = 54 dB μ V/m - 50.8 dB μ V/m = 3.2 dB

Block Diagram



Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 11 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



5.2.1 Radiated Emissions

Operator	Shane Dock
Test Date	11/8/17
Location	Chamber 5
Temp. / R.H.	70/40%
Requirement	FCC Part 15.209 RSS-GEN Section 7.1.2
Method	ANSI C63.10 Sections 6.4

Limits:

Frequency (kHz)	150	200
Field Strength at 300 m (μV/m)	16.0	12.0
Field Strength at 300 m (dBµV/m)	24.1	21.6

Test Parameters

Frequency	9 kHz - 30 MHz
Distance	3 meters
EUT	Unit tested and measured in three orientations.
Notes	Unit tested and measured with the loop antenna in three polarizations
Example Calculation	Limit (dB μ V/m) at 300m = 2400 / f(kHz) 16 = 2400/ f(kHz) Raw Data + Antenna Factor + Cable Factor = Reported Data 19.77 dB μ V + 12.50 dB/m + 0.93 dB = 38.80 dB μ V/m Corrected FS Value (150 kHz) = Measured Value – (40*log(limit distance/measurement distance) -14.5 = 65.5 – 20 log(300/3) Corrected FS Value (200 kHz) = Measured Value – (40*log([1/f(kHz)]/measurement distance)) – 20*log(limit distance/[1/f(kHz)]) -15.8 = 62.2 – 40 log (238.9/3) – 20 log (300/238.9)

Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 12 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



Instrumentation

Smar	aird							
	Date :	3-Nov-2017	Test:	LF Tx			Job :	C-2837
	PE:	Shane Dock	Customer :	Plexus			Quote	317245
		-						
NO.	Asset	Description	Manufacturer	Model	Serial	Cal Date	Cal Due Date	Equipment Status
1	EE 960085	Analyzer - EMI Receiver	Agilent	N9038A	MY 51210148	5/12/2017	5/12/2018	Active Calibration
2	AA 960006	Active Loop Antenna	EMCO	6502	9205-2753	8/28/2017	8/28/2019	Active Calibration

Table

Frequency (kHz)	Azimuth (degrees)	EUT Orientation	Antenna Polarity	Average Measure- ment at 3m (dBµV/m)	Corrected Average Measure- ment at 300m (dBµV/m)	Average Limit at 300m (dBμV/m)	Average Margin (dB)
150	0	V	V	65.5	-14.5	24.1	38.6
150	188.75	Н	V	65.7	-14.3	24.1	38.4
150	261.75	F	V	55.5	-24.5	24.1	48.6
150	76.75	V	S	61.3	-18.7	24.1	42.8
150	106.25	Н	S	55.8	-24.2	24.1	48.3
150	210.75	V	F	55.7	-24.3	24.1	48.4
150	0	Н	F	55.1	-24.9	24.1	49.0
150	256.5	F	F	56.9	-23.1	24.1	47.2
200	197.75	V	V	62.2	-15.8	21.6	37.4
200	204.75	Н	V	62.4	-15.6	21.6	37.2
200	270.25	F	V	54.0	-24.0	21.6	45.6
200	276.25	V	S	59.1	-18.9	21.6	40.5
200	93.25	Н	S	58.9	-19.1	21.6	40.7
200	200	V	F	53.5	-24.5	21.6	46.1
200	185.5	Н	F	52.4	-25.6	21.6	47.2
200	148.75	F	F	53.3	-24.7	21.6	46.3

Company: Medtronic Inc.		Name: Patient Assistant	
Report: 317245B	Page 13 of 15	Model: PA97000	
Job: C-2837		Serial: See Section 2	



Plots (Worst Case Shown)



Company: Medtronic Inc.		Name: Patient Assistant
Report: 317245B	Page 14 of 15	Model: PA97000
Job: C-2837		Serial: See Section 2



6 **REVISION HISTORY**

Version	Date	Notes	Person
V0	11/21/17	Rough Draft	Shane Dock
V1	11/28/17	Revised Draft	Shane Dock
V2	1/9/18	Final Draft	Shane Dock
V3	2/27/18	Updated Draft	Shane Dock

END OF REPORT

Company: Medtronic Inc.	Page 15 of 15	Name: Patient Assistant
Report: 317245B		Model: PA97000
Job: C-2837		Serial: See Section 2