RF TEST REPORT



Report No.: SL18041906-MED-029-FCC-IC Rev 1.0 Supersede Report No.: SL18041906-MED-029-FCC-IC

Applicant		Medtronic Inc.		
Product Name		Indra		
Model No.		TM91		
Product Description		Restorative Therapy Ultra Low Power Active Medical Implant Peripheral (UP-AMI-P)		
Test Standard		47 CFR FCC Part 95I RSS-Gen, RSS-243 Issue 3 Feb 2010		
Test Method	:	ANSI/TIA/EIA-603-D:2010 RSS-Gen, RSS-243 Issue 3 Feb 2010		
FCC ID	;	LF5TM91		
IC ID		3408D-TM91		
Dates of test		05/14/2018 to 05/18/2018		
Issue Date		03/21/2019		
Test Result		⊠ Pass ☐ Fail		
Equipment complied with the specification [X] Equipment did not comply with the specification []				
This Test Report is Issu	ed	Under the Authority of:		
was and				
Cip	her	Chen Ge		
Test Er	ngin			
Test re	sult	This test report may be reproduced in full only presented in this test report is applicable to the tested sample only		

Issued By:
SIEMIC Laboratories
775 Montague Expressway, Milpitas, CA 95035





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	2 of 26

Laboratory Introduction

SIEMIC, headquartered in the heart of Silicon Valley, with superior facilities in US and Asia, is one of the leading independent testing and certification facilities providing customers with one-stop shop services for Compliance Testing and Global Certifications.



In addition to testing and certification, SIEMIC provides initial design reviews and compliance management throughout a project. Our extensive experience with China, Asia Pacific, North America, European, and International compliance requirements, assures the fastest, most cost effective way to attain regulatory compliance for the global markets.

Accreditations for Conformity Assessment

Country/Region	Accreditation Body	Scope	
USA	FCC, A2LA	EMC, RF/Wireless, Telecom	
Canada	IC, A2LA, NIST	EMC, RF/Wireless, Telecom	
Taiwan	BSMI, NCC, NIST	EMC, RF, Telecom, Safety	
Hong Kong	OFTA, NIST	RF/Wireless, Telecom	
Australia	NATA, NIST	EMC, RF, Telecom, Safety	
Korea	KCC/RRA, NIST	EMI, EMS, RF, Telecom, Safety	
Japan	VCCI, JATE, TELEC, RFT	EMI, RF/Wireless, Telecom	
Mexico	NOM, COFETEL, Caniety	EMC, RF/Wireless, Telecom, Safety	
Europe	A2LA, NIST	EMC, RF, Telecom, Safety	
Israel	MOC, NIST	EMC, RF, Telecom, Safety	

Accreditations for Product Certifications

Country	Accreditation Body	Scope	
USA	FCC TCB, NIST	EMC, RF, Telecom	
Canada	IC FCB, NIST	EMC, RF, Telecom	
Singapore	iDA, NIST	EMC, RF, Telecom	
EU	NB	EMC & R&TTE Directive	
Japan	MIC (RCB 208)	RF, Telecom	
Hong Kong	OFTA (US002)	RF, Telecom	

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	3 of 26

CONTENTS

1	F	REPOR	FREVISION HISTORY	4
2			TIVE SUMMARY	
- 3			MER INFORMATION	
4			TE INFORMATION	
5			CATION	
6			ORMATION	
•	- 6.1		Description	
	6.2		o Description	
	6.3		test modes/configuration Description	
7	5		RTING EQUIPMENT/SOFTWARE AND CABLING DESCRIPTION	
	7.1	Supp	orting Equipment	8
	7.2		ing Description	
	7.3	Test	Software Description	8
8	7	TEST S	JMMARY	9
9	N	MEASU	REMENT UNCERTAINTY	10
10		MEA	SUREMENTS, EXAMINATION AND DERIVED RESULTS	12
	10.	1 E	RP of the Fundamental Emission	12
	10.2	2 R	adiated Measurements	14
	1	10.2.1	Radiated Measurements 30MHz to 1GHz	14
	1	10.2.2	Radiated Spurious Emissions above 1GHz	
	1	10.2.3	Frequency Stability/Error	19
		10.2.4	Occupied bandwidth	
ΑN	NE	X A. T	EST INSTRUMENT	24
ΑN	NE	X B. SI	EMIC ACCREDITATION	25



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	4 of 26

Report Revision History

Report No.	Report Version	Description	Issue Date
SL18041906-MED-029-FCC-IC	None	Original	05/29/2018
SL18041906-MED-029-FCC-IC Rev 1.0	1.0	Update	03/21/2019





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	5 of 26

2 **Executive Summary**

The purpose of this test program was to demonstrate compliance of following product

Company: Medtronic, Inc.

Product: Indra Model: TM91

against the current Stipulated Standards. The specified model product stated above has demonstrated compliance with the Stipulated Standard listed on 1st page.

3 Customer information

Applicant Name	:	Medtronic, Inc.
Applicant Address 1	:	8200 Coral Sea St. NE, Mounds View, MN 55112
Applicant Address 2	:	7000 Central Ave. NE, Minneapolis, MN 55432
Manufacturer Name	:	Medtronic, Inc.
Manufacturer Address 1	:	8200 Coral Sea St. NE, Mounds View, MN 55112
Manufacturer Address 2	:	7000 Central Ave. NE, Minneapolis, MN 55432

4 Test site information

Lab performing tests	:	SIEMIC Laboratories	
Lab Address	:	775 Montague Expressway, Milpitas, CA 95035	
FCC Test Site No.	:	540430	
IC Test Site No.	:	4842D	
VCCI Test Site No.	:	A0133	

5 Modification

Index	Item	Description	Note
-	-	-	-

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	6 of 26

EUT Information

<u>6.1</u> **EUT Description**

Product Name	:	Indra	
Model No.	:	TM91	
Trade Name	:	Medtronic	
Serial No.	:	NPE000046N	
Input Power	:	3.7 Vdc from Battery, and 5.0Vdc USB charge	
Product Hardware version	:	1.8	
Product Software version	:	1.8	
Radio Hardware version	:	1.8	
Radio Software version	:	1.8	
Product Radio Test firmware	:	NRP1025-37052	
Date of EUT received	:	04/28/2018	
Equipment Class/ Category	:	MICS	
Working Frequencies	:	402-405MHz	

Radio Description 6.2

Specifications for Radio:

Radio Type	MICS
Operating Frequency	402-405 MHz
Modulation	FSK
Channel Spacing	300 KHz
Antenna Type	Integral Loop Antenna
Antenna Gain	-8.94dBi
N0. Of RF Channels	10
Antenna Connector Type	N/A





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	7 of 26

EUT test modes/configuration Description

Mode	Note
RF test	Set the EUT to transmit continuously in different test modes and channels.

Note: None

Test Item	Operating mode	Tested antenna port	Test frequencies	
Antenna Requirement	N/A	-	-	
Effective Radiated Power of the Fundamental Emission	Continuous Transmit	-	402.15 MHz (Low Channel 1),404.85 MHz (High Channel 10)	
Radiated Measurements 30MHz to 1GHz	Continuous Transmit	-	402.15 MHz (Low Channel 1),404.85 MHz (High Channel 10)	
Radiated Measurements above 1GHz	Continuous Transmit	-	402.15 MHz (Low	
Frequency Stability	Continuous Transmit	-	Channel 1), 403.35MHz (Mid Channel 5), 404.85	
Occupied Bandwidth	Continuous Transmit	-	MHz (High Channel 10)	

Note: Only radiated measurements were performed during the test.

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088 Visit us at: www.siemic.com; Follow us at:



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	8 of 26

7 Supporting Equipment/Software and cabling Description

7.1 Supporting Equipment

Item	Support Equipment Description	Model	Serial Number	Manufacturer	Notes
1	LAPTOP	LATITUDE E6220	N/A	DELL	N/A

7.2 Cabling Description

Item	Connection Start		Connection S	top	Length / shieldin	g Info	Note
item	From	I/O Port	То	I/O Port	Length (m)	Shielding	Note
N/A	N/A	N/A	N/A	N/A	N/A	N/A	N/A

7.3 Test Software Description

Test Item	Software	Description
RF Testing	BLUE MAGIC 1.95	Provided by manufacturer to set EUT in continuous
		mode

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	9 of 26

Test Summary

Test Item		Test standard		Test Method/Procedure	Pass / Fail
Effective Radiated Power of the	FCC	47 CFR §95.2567	FCC	ANSI/TIA/EIA-603-D:2010	□ Pass □ N/A
Fundamental Emission	IC	RSS-243	IC	RSS-243 5.4	
Radiated Spurious Emissions	FCC	47 CFR §95.2569, §95.2579	FCC	ANSI/TIA/EIA-603-D:2010	⊠ Pass □ N/A
	IC	RSS-243	IC	RSS-243 5.5	
Frequency Stability	FCC	47 CFR §95.2565	FCC	ANSI/TIA/EIA-603-D:2010	⊠ Pass □ N/A
requestey etablisty	IC	RSS-243	IC	RSS-243 3.3 b & 5.3	
Occupied Bandwidth	FCC	47 CFR §95.2573	FCC	ANSI/TIA/EIA-603-D:2010	⊠ Pass □ N/A
Оссиреи Ванимиин	IC	RSS-Gen	IC	RSS Gen 4.6, RSS-293 5.1	
Remark	 All measurement uncertainties are not taken into consideration for all presented test result. The applicant shall ensure frequency stability by showing that an emission is maintained within the band of operation under all normal operating conditions as specified in the user's manual. 				s maintained





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	10 of 26

9 Measurement Uncertainty

9.1 Conducted Emissions

The test is to measure the conducted emissions to the mains port of the EUT.

Some error sources that can contribute to the total uncertainty:

- Uncertainty of the receiver
- Uncertainty of the LISN
- Uncertainty of cables
- Uncertainty due to the mismatches
- Etc. see the below table for details

Source of Uncertainty	Value	Probability	Division	Sensitivity	Expanded
	(dB)	Distribution		Coefficient	Uncertainty
Receiver Reading	0.12	Rectangular	1.732	1	0.069284
Cable Insertion Loss	0.21	Normal	2	1	0.105
Filter Insertion Loss	0.25	Normal	2	1	0.125
LISN Insertion Loss	0.40	Normal	2	1	0.20
Receiver CW accuracy	0.5	Rectangular	1.732	1	0.2886836
Pulse Amplitude	1.5	Rectangular	1.732	1	0.86605081
Response					
PRF Response	1.5	Rectangular	1.732	1	0.86605081
Mismatch LISN -	0.25	U-Shape	1.414	1	0.1768033
Receiver					
LISN Impedance	2.5	Triangular	2.449	1	1.0208248
Combined Standard Unce	1.928133				
Expanded Uncertainty (F	(=2)				3.856266

The total derived measurement uncertainty is +/- 3.86 dB.

9.2 Radiated Emissions (30MHz to 1GHz)

The test is to measure the radiated emissions of the EUT.

Some error sources that can contribute to the total uncertainty:

- Uncertainty of the receiver
- Uncertainty of the antenna
- Uncertainty of cables
- Uncertainty due to the mismatches
- NSA Calibration
- Etc., details see the below table

Source of Uncertainty	Value Probability (dB) Distribution		Division	Sensitivity Coefficient	Expanded Uncertainty	
Receiver Reading	0.12	Rectangular	1.732	1	0.069284	
Cable Insertion Loss	0.21	Normal	2	1	0.105	
Filter Insertion Loss	0.25	Normal	2	1	0.125	
Antenna Factor	0.65	Normal	2	1	0.325	
Receiver CW accuracy	0.5	Rectangular	1.732	1	0.2886836	
Pulse Amplitude Response	1.5	Rectangular	1.732	1	0.86605081	
PRF Response	1.5	Rectangular	1.732	1	0.86605081	
Mismatch Filter - Receiver	0.25	U-Shape	1.414	1	0.1768033	
NSA Calibration	4.0	U-Shape	1.414	1	2.8288543	
Combined Standard Uncertaint	3.0059131					
Expanded Uncertainty (K=2)					6.0118262	

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	11 of 26

The total derived measurement uncertainty is +/- 6.00 dB.

9.3 Radiated Emissions (1GHz to 60GHz)

The test is to measure the radiated emissions of the EUT.

Some error sources that can contribute to the total uncertainty:

- Uncertainty of the receiver
- Uncertainty of the antenna
- Uncertainty of cables
- Uncertainty due to the mismatches
- VSWR Calibration
- Etc., details see the below table

Source of Uncertainty	Value (dB)	Value Probability Distribution		Sensitivity Coefficient	Expanded Uncertainty				
Receiver Reading	0.12	Rectangular	1.732	1	0.0692840				
Cable Insertion Loss	0.21	Normal	2	1	0.1050000				
Filter Insertion Loss	0.25	Normal	2	1	0.1250000				
Antenna Factor	0.65	Normal	2	1	0.3250000				
Receiver CW accuracy	0.5	Rectangular 1.732		1	0.2886836				
Pulse Amplitude Response	1.5	Rectangular	1.732	1	0.8660508				
PRF Response	1.5	Rectangular	1.732	1	0.8660508				
Mismatch Filter - Receiver	0.25	U-Shape	1.414	1	0.1768033				
VSWR Calibration	2.0	U-Shape	1.414	1	1.4144272				
Combined Standard Uncertain	4.2363								
Expanded Uncertainty (K=2	Expanded Uncertainty (K=2)								

The total derived measurement uncertainty is +/- 8.47 dB.

9.4 RF conducted measurement

The test is to measure the RF output power from the EUT.

Some error sources that can contribute to the total uncertainty:

- Uncertainty of the Reference Level Uncertainty
- Uncertainty of variable attenuators
- Uncertainty of cables
- Uncertainty due to the mismatches

Source of Uncertainty	Value (dB)	Probability Distribution	Division	Sensitivity Coefficient	Expanded Uncertainty
Reference Level	0.12	Rectangular	1.732	1	0.069284
Cable Insertion Loss	0.21	Normal	2	1	0.105
Attenuator	0.25	Normal	2	1	0.125
Mismatch	0.25	U-Shape	1.414	1	0.1768033
Combined Standard Unce	0.476087				
Expanded Uncertainty (K=2)				0.952174

The total derived measurement uncertainty is +/- 0.95 dB.



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	12 of 26

10 Measurements, examination and derived results

10.1 EIRP of the Fundamental Emission

Requirement(s):

Spec	Requirement	Applicable
	Each MedRadio transmitter type must be designed such that the MedRadio equivalent isotropically radiated power (M-EIRP) does not exceed the limits in this section. Compliance with these limits must be determined as set forth in §95.2569.	
	(a) Transmitters subject to frequency monitoring—401-406 MHz. For MedRadio transmitters that are not excepted under §95.2559(b) from the frequency monitoring requirements of §95.2559(a):	
	(1) The M-EIRP within any 300 kHz bandwidth within the 402-405 MHz band must not exceed 25 microwatts.	
	(2) The M-EIRP within any 100 kHz bandwidth within the 401-402 MHz or 405-406 MHz bands must not exceed 25 microwatts.	
	(b) Transmitters excepted from frequency monitoring—401-402 MHz and 405-406 MHz. For MedRadio transmitters that are excepted under §95.2559(b)(2) or (3) from the frequency monitoring requirements of §95.2559(a):	
FCC 95.2567; &	(1) The M-EIRP of any transmitter operating in the 401-401.85 MHz or 405-406 MHz bands must not exceed 250 nanowatts in any 100 kHz bandwidth.	
RSS-243 3.1, 5.4	(2) The M-EIRP of any transmitter operating in the 401.85-402 MHz band must not exceed 25 microwatts in any 150 kHz bandwidth.	
	(c) Transmitters excepted from frequency monitoring—403.65 MHz. For MedRadio transmitters that are excepted under §95.2559(b)(4) from the frequency monitoring requirements of §95.2559(a), the M-EIRP must not exceed 100 nanowatts in the 300 kHz bandwidth centered at 403.65 MHz.	
	(d) Transmitters—other frequency bands. For MedRadio transmitters operating in the 413-419 MHz, 426-432 MHz, 438-444 MHz, or 451-457 MHz bands:	
	(1) The peak M-EIRP over the frequency bands of operation must not exceed the lesser of zero dBm (1 mW) or 10 log (B)-7.782 dBm, where B is the MedRadio 20 dB emission bandwidth in megahertz.	
	(2) The peak power spectral density must not exceed 800 microwatts per megahertz in any one megahertz band.	
	(e) Transmitters—2360-2390 MHz band. For MedRadio transmitters operating in the 2360-2390 MHz band, the M-EIRP over the bands of operation must not exceed the lesser of zero dBm (1 mW) or 10 log (B) dBm, where B is the MedRadio 20 dB emission bandwidth	



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	13 of 26

	in megahertz.		
	(f) Transmitters—2390-2400 MHz band. For MedRadio transmitters—2390-2400 MHz band. For MedRadio transmitters—1390-2400 MHz band. For MedRadio transmitters—2390-2400 MHz bands of operation must not or 16 + 10 log (B) dBm, where B is the MedRadio 20 dB of the following transmitters—2390-2400 MHz bands of operation must not or 16 + 10 log (B) dBm, where B is the MedRadio 20 dB of the following transmitters—2390-2400 MHz bands of operation must not or 16 + 10 log (B) dBm, where B is the MedRadio 20 dB of the following transmitters—2390-2400 MHz bands of the following transmitters—239	exceed the lesser of 13	dBm (20 mW)
Test Setup	Semi Anechoic Chamber Radio Absorbing Material Ground Plane	Antenna	Spectrum Analyzer
Procedure	The EUT was switched on and allowed to warm up to the test was carried out at the selected frequency portion of the emissions, was carried out by rote and adjusting the antenna height in the following mars. Vertical or horizontal polarisation (whicheve the EUT) was chosen. The EUT was then rotated to the direction of the control of th	ints obtained from the El ating the EUT, changing ner: er gave the higher emissi that gave the maximum e	JT characterisation. the antenna polarization, on level over a full rotation of emission.
Test Date	05/14/2018 to 05/18/2018	Environmental conditions	Temperature 23°C Relative Humidity 48% Atmospheric Pressure 1026mbar
Result	⊠ Pass □ Fail		
Test Data ⊠	Yes (See below)		

Test Plot ⊠ Yes (See below) \square N/A

Test was done by Cipher at 10 meter chamber.

EIRP - 402.15 MHz

	Indicated Test Antenna					Substituted									
Frequency Raw (MHz) (dBuV/m)	Degree	Degree	Degree	Degree	Degree	Degree	Height	Polarity	Frequency	Level	Ant Gain	Cable Loss	Absolute Level	Limit	Margin
(IVITIZ)	(ubuv/iii)		(cm)		(MHz)	(dBm)	(dBi)	(dB)	(dBm)	(dBm)	(dB)				
402.15	81.74	36	157	V	402.2	-26.28	0	1	-27.28	-16	-11.28				
402.15	79.72	185	157	Н	402.2	-29.45	0	1	-30.45	-16	-14.45				

EIRP - 404.85MHz

	Indicated Test Antenna				Substituted							
Frequency (MHz)	Raw	Degree	Height	Polarity	Frequency	Polarity Frequency	Frequency Level	Level Ant Gain	Cable Loss	Absolute Level	Limit	Margin
	(dBuV/m)		(cm)		(MHz)	(dBm)	(dBi)	(dB)	(dBm)	(dBm)	(dB)	
404.85	80.75	36	157	V	404.9	-27.27	0	1	-28.27	-16	-12.27	
404.85	79.1	185	157	Н	404.9	-30.07	0	1	-31.07	-16	-15.07	



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	14 of 26

10.2 Radiated Measurements

10.2.1 Radiated Measurements 30MHz to 1GHz

Spec	Requirement			Applicable				
	(a) Field strength limits. The field strengths of unwanted emissions from each MedRadio transmitter type, measured at a distance of 3 meters, must not exceed the field strength limits shown in the table in this paragraph for the indicated frequency ranges, if the frequencies of these emissions are:							
	(1) More than 250 kHz outside the 402-405 MHz band);	of the 402-405 MHz band (for	devices designed to opera	te in				
		of either the 401-402 MHz or 40 -402 MHz or 405-406 MHz ban		rices				
FCC 95.2579 and RSS-	(3) In the 406.000-406.100 MHz band (for devices designed to operate in the 401-402 MHz or 405-406 MHz bands); or							
243 5.5	(4) More than 2.5 MHz outside of the 413-419 MHz, 426-432 MHz, 438-444 MHz or 451-457 MHz bands (for devices designed to operate in these four bands).							
	(5) More than 2.5 MHz outside of the 2360-2400 MHz band (for devices designed to operate in the 2360-2400 MHz band).							
	Frequency range (MHz)	Field Strength (uV/m)	Measurement Distance (meter)					
	30 – 88	100	3					
	88 – 216	150	3					
	216 960	200	3					
	Above 960	500	3					
		Semi Anechoic Chamber	A A A A A A A A A A A A A A A A A A A					
Test Setup	Radio Absorbing M	EUT 3m	intenna I dam	trum Analyzer				



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	15 of 26

Procedure	 The EUT was switched on and allowed to warm up to its normal operating condition. The test was carried out at the selected frequency points obtained from the EUT characterisation. Maximization of the emissions, was carried out by rotating the EUT, changing the antenna polarization, and adjusting the antenna height in the following manner: Vertical or horizontal polarisation (whichever gave the higher emission level over a full rotation of the EUT) was chosen. The EUT was then rotated to the direction that gave the maximum emission. Finally, the antenna height was adjusted to the height that gave the maximum emission. A Quasi-peak measurement was then made for that frequency point. Steps 2 and 3 were repeated for the next frequency point, until all selected frequency points were measured. 								
Test Date	05/14/2018 to 05/18/2018	Environmental conditions	Temperature 23°C Relative Humidity 48% Atmospheric Pressure 1026mbar						
Remark	-								
Result	⊠ Pass ☐ Fail								
Test Data ≥	☐ Yes (See below) ☐ N/A								
Test Plot ≥	☐ Yes (See below) ☐ N/A								

Test was done by Cipher at 10 meter chamber.




Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	16 of 26

Radiated Emission Test Results (Below 1GHz)

Test specification:	Radiated Emissions	Radiated Emissions					
Mains Power:	3.2 VDC						
Tested by:	Cipher	Result:	☑ Pass☐ Fail				
Test Date:	05/14/2018 to 05/18/2018						
Remarks:	402.15 MHz	'	'				

f=30MHz - 1000MHz plot and 3 meter distance measurement



f=30MHz - 1000MHz and 3 meter distance measurement

Frequency MHz	Raw dBµV/m	Cable Loss	AF dB	Level dBµV/m	Measurement Type	Pol	Hgt cm	Azt Deg	Limit dBµV/m	Margin dB	Pass /Fail
884.78	21.1	15.92	-14.27	22.76	Quasi Max	V	219	82	46	-23.24	Pass
276.72	22.34	13.15	-23.11	12.37	Quasi Max	V	101	298	46	-33.63	Pass
207.46	21.37	12.69	-25.53	8.53	Quasi Max	V	134	80	43.5	-34.97	Pass

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	17 of 26

10.2.2 Radiated Spurious Emissions above 1GHz

Requirement(s):

Spec	Requirement	Applicable
FCC 95.2579 and RSS- 243 5.5	For non-restricted band, In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB or 30dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, determined by the measurement method on output power to be used. □ 20 dB down □ 30 dB down	
Test Setup	Semi Anechoic Chamber Radio Absorbing Material Tut I.5m Antenna Ground Plane	um Analyzer
Procedure	 The EUT was switched on and allowed to warm up to its normal operating condition. The test was carried out at the selected frequency points obtained from the EUT chara Maximization of the emissions, was carried out by rotating the EUT, changing the ante and adjusting the antenna height in the following manner: Vertical or horizontal polarisation (whichever gave the higher emission level rotation of the EUT) was chosen. The EUT was then rotated to the direction that gave the maximum emission c. Finally, the antenna height was adjusted to the height that gave the maximum An average measurement was then made for that frequency point. Steps 2 and 3 were repeated for the next frequency point, until all selected frequency measured. 	enna polarization, over a full . m emission.
Remark	The EUT was scanned up to 6 GHz. Both horizontal and vertical polarities were investigated. show only the worst case which is vertical.	The results
Result	⊠ Pass □ Fail	

Test Data \boxtimes Yes (See below) \square N/A

Test Plot ☐ Yes (See below) ☐ N/A

Test was done by Cipher at 3m chamber.



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	18 of 26

Radiated Emission-3 meter distance Measurements Test Results (Above 1GHz)

Above 1GHz – 402.15 MHz

Frequency MHz	Raw dBuV	Cable Loss	AF dB	Level dBuV/m	Measurement Type	Pol	Hgt cm	Azt Deg	Limit dBuV/m	Margin dB	Pass /Fail
3780.64	38.85	11.37	1.40	51.62	Peak Max	V	124	52	74	-22.38	Pass
2653.54	40.61	9.30	-4.89	45.01	Peak Max	V	115	291	74	-28.99	Pass
1635.43	43.40	7.21	-11.82	38.79	Peak Max	V	120	81	74	-35.21	Pass
3780.64	25.93	11.37	1.40	38.70	Average Max	V	124	52	54	-15.30	Pass
2653.54	27.52	9.30	-4.89	31.92	Average Max	V	115	291	54	-22.08	Pass
1635.43	30.34	7.21	-11.82	25.73	Average Max	V	120	81	54	-28.27	Pass

Above 1GHz -403.35 MHz

Frequency MHz	Raw dBuV	Cable Loss	AF dB	Level dBuV/m	Measurement Type	Pol	Hgt cm	Azt Deg	Limit dBuV/m	Margin dB	Pass /Fail
1873.38	43.08	7.70	-9.01	41.77	Peak Max	٧	148	264	74	-32.23	Pass
4873.57	38.92	13.2	3.76	55.88	Peak Max	V	138	87	74	-18.12	Pass
3797.79	39.58	11.41	1.55	52.54	Peak Max	V	183	15	74	-21.46	Pass
1873.38	29.88	7.70	-9.01	28.57	Average Max	V	148	264	54	-25.43	Pass
4873.57	25.04	13.2	3.76	42.01	Average Max	V	138	87	54	-11.99	Pass
3797.79	25.82	11.41	1.55	38.78	Average Max	V	183	15	54	-15.22	Pass

Above 1GHz – 404.85 MHz

Frequency MHz	Raw dBuV	Cable Loss	AF dB	Level dBuV/m	Measurement Type	Pol	Hgt cm	Azt Deg	Limit dBuV/m	Margin dB	Pass /Fail
4715.91	40.12	13.03	2.86	56.01	Peak Max	V	101	202	74	-17.99	Pass
1877.62	44.18	7.70	-8.89	42.99	Peak Max	V	194	145	74	-31.01	Pass
2495.18	41.57	8.91	-6.20	44.28	Peak Max	V	186	213	74	-29.72	Pass
4715.91	24.95	13.03	2.86	40.84	Average Max	V	101	202	54	-13.16	Pass
1877.62	30.10	7.70	-8.89	28.91	Average Max	V	194	145	54	-25.09	Pass
2495.18	28.25	8.91	-6.20	30.96	Average Max	V	186	213	54	-23.04	Pass

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	19 of 26

10.2.3 Frequency Stability/Error

Requirement(s):

Spec	Requirement			Applicable
FCC 95.2565; RSS-243 3.3 b & 5.3	Each MedRadio transmitter type m ±100 ppm of the operating frequer this section. Frequency stability test temperature range. (a) 25 °C to 45 °C in the case of m (b) 0 °C to 55 °C in the case of Mebody-worn	ncy over the applicable temper sting shall be performed over t nedical implant transmitters; ar	rature range set forth in the appropriate	
Test Setup	EUT Environmental Chamber 1. The EUT was set up inside a 2. The EUT was placed in the control of the EUT was placed in the control of the EUT was placed in the control of the EUT was placed in the	in environmental chamber.	wer Meter	
Procedure	Frequency Stability was measured analyzer. The spectrum analyzer monitor when varying the voltage.			
Test Date	05/14/2018 to 05/18/2018	Environmental conditions	Temperature Relative Humidity Atmospheric Pressure	23°C 48% 1026mbar
Remark	None			
Result	⊠ Pass □ Fail			

Test Data ⊠ Yes (See below)		Ν	1/	A	١
-----------------------------	--	---	----	---	---

Test Plot ☐ Yes (See below) ☐ N/A

Test was done by Cipher at RF test site.



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	20 of 26

Test Result for Frequency Stability

Channel 1: 402.15 MHz

Chamile 1. 402.13 Willz			
Test Conditions	Measured Frequency Condition MHz	Frequency Error PPM	Limit PPM
T (+25°C)	402.149524	-1.18	+/-100
T max (55°C)	402.149523	-1.19	+/-100
T min (-10°C)	402.150358	0.89	+/-100

Channel 5: 403.35 MHz

Test Conditions	Measured Frequency Condition MHz	Frequency Error PPM	Limit PPM
T (+25°C)	403.34618	-9.47	+/-100
T max (55°C)	403.350836	2.07	+/-100
T min (-10°C)	403.35024	0.60	+/-100

Channel 10: 404.85 MHz

Test Conditions	Measured Frequency Condition MHz	Frequency Error PPM	Limit PPM
T (+25°C)	404.849116	-2.18	+/-100
T max (55°C)	404.849115	-2.19	+/-100
T min (-10°C)	404.850023	0.06	+/-100

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088

Visit us at: www.siemic.com: Follow us at:





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	21 of 26

10.2.4 Occupied bandwidth

Requirement(s):

Spec	Requirement			Applicable
FCC 95.2573, RSS Gen 4.6, RSS-293 5.1	(a) For MedRadio transmitters opera bandwidth is 300 kHz. Such transmit during a MedRadio communications half duplex communications provided in a MedRadio communications sess	ters must not use more than 30 session. This provision does red that the total bandwidth of all	0 kHz of bandwidth (total) not preclude full duplex or	
Test Setup	Support Units Turn T	3 m	p antenna m height	
Procedure	To measure conducted, a an external antenna was to	and allowed to warm up to its r SMA cable was used to replacused to detect EUT transmission Occupied Bandwidth of EUT to	ce the EUT antenna. To me on signal.	
Test Date	05/14/2018 to 05/18/2018	Environmental conditions	Temperature Relative Humidity Atmospheric Pressure	23°C 48% 1025mbar
Remark	-			
Result	⊠ Pass □ Fail			

Test Data ☐ Yes (See	e below) 🖂 N/A
----------------------	----------------

Test Plot ⊠ Yes (See below) □ N/A


Test was done by Cipher at 10 meter chamber.



Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	22 of 26


Test results:

402 MHz



Frequency (MHz)	Occupied Bandwidth (KHz)
402.15	57.96

403 MHz



Frequency (MHz)	Occupied Bandwidth (KHz)	
403.35	58.26	

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	23 of 26

405 MHz



Frequency (MHz)	Occupied Bandwidth (KHz)
404.85	58.21

775 Montague Expressway, Milpitas, CA 95035, USA • Phone: (+1) 408 526 1188 • Facsimile (+1) 408 526 1088

Visit us at: www.siemic.com: Follow us at:





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	24 of 26

Annex A. TEST INSTRUMENT

Instrument	Model	Serial #	Cal Date	Cal Cycle	Cal Due	In use
Radiated Emissions						
R & S Receiver	ESIB 40	100179	06/08/2017	1 Year	06/08/2018	~
Pre-Amplifier (1-26.5GHz)	8449B	3008A00715	03/30/2018	1 Year	03/30/2019	~
Preamplifier (100KHz-7GHz)	LPA-6-30	11140711	02/10/2018	1 Year	02/10/2019	~
ETS-Lingren Loop Antenna	6512	49120	07/14/2017	1 Year	07/14/2018	~
Bi-Log antenna (30MHz~2GHz)	JB1	A030702	07/08/2017	1 Year	07/08/2018	~
Horn Antenna (18-40 GHz)	AH-840	101013	07/15/2017	1 Year	07/15/2018	~
Spectrum Analyzer	N9010A	10SL0219	08/02/2017	1 Year	08/02/2018	~
Agilent Signal Generator	N5182A	MY47071065	04/12/2018	1 Year	04/12/2019	~
Test Equity Environment Chamber	1007H	61201	07/21/2017	1 Year	07/21/2018	~





Test report	SL18041906-MED-029-FCC-IC Rev 1.0
Page	25 of 26

Annex B. SIEMIC Accreditation

Accreditations	Document	Scope / Remark
ISO 17025 (A2LA)		Please see the documents for the detailed scope
ISO Guide 65 (A2LA)		Please see the documents for the detailed scope
TCB Designation		A1, A2, A3, A4, B1, B2, B3, B4, C
FCC DoC Accreditation	7	FCC Declaration of Conformity Accreditation
FCC Site Registration	7	3 meter site
FCC Site Registration	7	10 meter site
IC Site Registration		3 meter site
IC Site Registration	A	10 meter site
EU NB		Radio & Telecommunications Terminal Equipment: EN45001 – EN ISO/IEC 17025
		Electromagnetic Compatibility: EN45001 – EN ISO/IEC 17025
Singapore iDA CB(Certification Body)	包包	Phase I, Phase II
Vietnam MIC CAB Accreditation		Please see the document for the detailed scope
	7	(Phase II) OFCA Foreign Certification Body for Radio and Telecom
Hong Kong OFCA	7	(Phase I) Conformity Assessment Body for Radio and Telecom
	7	Radio: Scope A – All Radio Standard Specification in Category I
Industry Canada CAB	7	Telecom: CS-03 Part I, II, V, VI, VII, VIII



Test report SL18041906-MED-029-FCC-IC Rev 1.0 26 of 26 Page

Japan Recognized Certification Body Designation	包包	Radio: A1. Terminal equipment for purpose of calling Telecom: B1. Specified radio equipment specified in Article 38-2, Paragraph 1, Item 1 of the Radio Law	
		EMI: KCC Notice 2008-39, RRL Notice 2008-3: CA Procedures for EMI KN22: Test Method for EMI EMS: KCC Notice 2008-38, RRL Notice 2008-4: CA Procedures for EMS KN24, KN61000-4-2, -4-3, -4-4, -4-5, -4-6, -4-8, -4-11: Test Method for EMS	
Korea CAB Accreditation		Radio: RRL Notice 2008-26, RRL Notice 2008-2, RRL Notice 2008-10, RRL Notice 2007-49, RRL Notice 2007-20, RRL Notice 2007-21, RRL Notice 2007-80, RRL Notice 2004-68	
		Telecom: President Notice 20664, RRL Notice 2007-30, RRL Notice 2008-7 with attachments 1, 3, 5, 6; President Notice 20664, RRL Notice 2008-7 with attachment 4	
Taiwan NCC CAB Recognition		LP0002, PSTN01, ADSL01, ID0002, IS6100, CNS14336, PLMN07, PLMN01, PLMN08	
Taiwan BSMI CAB Recognition	Z	CNS 13438	
Japan VCCI		R-3083: Radiation 3 meter site C-3421: Main Ports Conducted Interference Measurement T-1597: Telecommunication Ports Conducted Interference Measurements	
	Ē	EMC: AS/NZS CISPR 11, AS/NZS CISPR 14.1, AS/NZS CISPR22, AS/NZS 61000.6.3, AS/NZS 61000.6.4	
Australia CAB Recognition		Radio communications: AS/NZS 4281, AS/NZS 4268, AS/NZS 4280.1, AS/NZS 4280.2, AS/NZS 4295, AS/NZS 4582, AS/NZS 4583, AS/NZS 4769.1, AS/NZS 4769.2, AS/NZS 4770, AS/NZS 4771	
		Telecommunications: AS/ACIF S002:05, AS/ACIF S003:06, AS/ACIF S004:06 AS/ACIF S006:01, AS/ACIF S016:01, AS/ACIF S031:01, AS/ACIF S038:01, AS/ACIF S040:01, AS/ACIF S041:05, AS/ACIF S043.2:06, AS/ACIF S60950.1	
Australia NATA Recognition	₺	AS/ACIF S002, AS/ACIF S003, AS/ACIF S004, AS/ACIF S006, AS/ACIF S016,AS/ACIF S031, AS/ACIF S038, AS/ACIF S040, AS/ACIF S041, AS/ACIF S043.2	