

TP/PPS/MAS/CertS
Kare Oksanen

September 22th, 2007

1. Please note that the power relationship between the SAR testing and the EMC testing needs to be of the same type (i.e. either conducted power or radiated power). Please note that the SAR report appears to show radiated power but the EMC report shows conducted power. Please provide either the radiated power measured for EMC or the conducted power measure during SAR so that a comparison may be made between the two. Please note that if conducted power is used then the difference between the SAR and EMC report should not be greater than 0.5dB.

2. Please provide the RSS102 annex A and B exhibit.

[A: I have uploaded both missing documents now to your server.](#)