

Federal Communications Commission
Authorization and Evaluation Division

Regarding application for certification of FCC ID: N9Y0086.

The following clarification has been requested by the TCB:

The module, part# 100-0086, will always be populated to support both the 900MHz ISM band and the 600MHz WMTS band, and only one band at a time will be enabled.

Warren Bendler
Engineering Coordinator
Cleveland Medical Devices, Inc.
4415 Euclid Ave. 4th Floor
Cleveland, Ohio 44103
216-619-5902