

Test Laboratory: Compliance Certification Services

File Name: [D2450V2 SN 706_13.8 mW.da4](#)

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:706

Program: System Performance Check at 2450 MHz

Ambient Temperature: 25.7 deg C; Liquid Temperature: 24.7 deg C

Communication System: CW - 2450 MHz; Frequency: 2450 MHz;Duty Cycle: 1:1

Medium: Head 2450 MHz ($\sigma = 1.8239$ mho/m, $\epsilon_r = 39.0162$, $\rho = 1000$ kg/m³)

Phantom section: Flat Section

DASY4 Configuration:

- Probe: ET3DV6 - SN1577; ConvF(5.1, 5.1, 5.1); Calibrated: 2/7/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
Sensor-Surface: 4mm (Mechanical And Optical Surface Detection)
- Electronics: DAE3 Sn427; Calibrated: 2/4/2003
- Phantom: SAM 1; Type: SAM 1; Serial: 1185
- Measurement SW: DASY4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

d=10mm, Pin=250mW/Area Scan (7x7x1): Measurement grid: dx=15mm, dy=15mm

d=10mm, Pin=250mW/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm



Peak SAR (extrapolated) = 28.6 W/kg

SAR(1 g) = 13.8 mW/g; SAR(10 g) = 6.31 mW/g

Reference Value = 96.3 V/m

Power Drift = 0.006 dB

Maximum value of SAR = 15.2 mW/g



0 dB = 15.2mW/g