

Test Laboratory: The name of your organization
 File Name: [D2450V2SN706_Probe 3021_052004.da4](#)

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:706
Program Name: System Performance Check at 2450 MHz
Ambient Temp.: 24.0 deg. C; Liquid Temp.: 23.0 deg. C


Communication System: CW - 2450 MHz; Frequency: 2450 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 2450$ MHz; $\sigma = 1.88$ mho/m; $\epsilon_r = 39.6$; $\rho = 1000$ kg/m³
 Phantom section: Flat Section

DASY4 Configuration:

- Probe: ES3DV2 - SN3021; ConvF(4.5, 4.5, 4.5); Calibrated: 7/29/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn500; Calibrated: 12/23/2003
- Phantom: SAM 1; Type: SAM 1; Serial: 1185
- Measurement SW: DASY4, V4.2 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 112

d=10mm, Pin=250mW/Area Scan (7x7x1): Measurement grid: dx=15mm, dy=15mm
 Reference Value = 91.9 V/m; Power Drift = -0.0 dB
 Maximum value of SAR (measured) = 15.2 mW/g

d=10mm, Pin=250mW/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm
 Reference Value = 91.9 V/m; Power Drift = -0.0 dB
 Maximum value of SAR (measured) = 15 mW/g
 Peak SAR (extrapolated) = 28.6 W/kg
SAR(1 g) = 13.5 mW/g; SAR(10 g) = 6.25 mW/g




0 dB = 15mW/g

Test Laboratory: The name of your organization
File Name: [D2450V2SN706_Probe 3021_052004.da4](#)

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:706
Program Name: System Performance Check at 2450 MHz

Communication System: CW - 2450 MHz; Frequency: 2450 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2450$ MHz; $\sigma = 1.88$ mho/m; $\epsilon_r = 39.6$; $\rho = 1000$ kg/m³
Phantom section: Flat Section

d=10mm, Pin=250mW/Z Scan (1x1x51): Measurement grid: dx=20mm, dy=20mm, dz=2mm
Reference Value = 91.9 V/m; Power Drift = -0.0 dB
Maximum value of SAR (measured) = 15.2 mW/g



Test Laboratory: Compliance Certification Services

File Name: [D2450V2SN748_ES3DV2SN3023_072904.da4](#)

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:748

Program Name: System Performance Check at 2450 MHz

Ambient Temp.: 25.0 deg. C; Liquid Temp.: 24.0 deg. C

Communication System: CW - 2450 MHz; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 2450$ MHz; $\sigma = 1.92$ mho/m; $\epsilon_r = 52.2$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

DASY4 Configuration:

- Probe: EX3DV3 - SN3531; ConvF(8.32, 8.32, 8.32); Calibrated: 7/18/2004

- Sensor-Surface: 4mm (Mechanical Surface Detection)

- Electronics: DAE3 Sn500; Calibrated: 12/23/2003

- Phantom: SAM 2; Type: SAM 2; Serial: 1050

- Measurement SW: DASY4, V4.2 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 112

d=10mm, Pin=250mW/Area Scan (7x7x1): Measurement grid: dx=15mm, dy=15mm

Reference Value = 87.3 V/m; Power Drift = -0.1 dB

Maximum value of SAR (measured) = 14.3 mW/g


d=10mm, Pin=250mW/Zoom Scan (5x5x7)/Cube 0: Measurement grid: dx=7.5mm, dy=7.5mm, dz=5mm

Reference Value = 87.3 V/m; Power Drift = -0.1 dB

Maximum value of SAR (measured) = 13.9 mW/g

Peak SAR (extrapolated) = 24.4 W/kg

SAR(1 g) = 12.4 mW/g; SAR(10 g) = 5.9 mW/g



0 dB = 13.9mW/g

Test Laboratory: Compliance Certification Services

File Name: [D2450V2SN748_ES3DV2SN3023_072904.da4](#)

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN:748

Program Name: System Performance Check at 2450 MHz

Communication System: CW - 2450 MHz; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium parameters used: $f = 2450$ MHz; $\sigma = 1.92$ mho/m; $\epsilon_r = 52.2$; $\rho = 1000$ kg/m³

Phantom section: Flat Section

d=10mm, Pin=250mW/Z Scan (1x1x51): Measurement grid: dx=20mm, dy=20mm, dz=2mm

Reference Value = 87.3 V/m; Power Drift = -0.1 dB

Maximum value of SAR (measured) = 15.4 mW/g

