

Dec 20, 2013
Federal Communications Commission
Authorization and Evaluation Division
7435 Oakland Mills Road Columbia,
MD 21046

Attn: OET Dept.

Ref: FCC Class II Permissive change for FCC ID: NMTBPU321
Applicant: **IDT Technology Limited**

Dear Examiner,

This is to request a Class II Permissive change for FCC ID: NMTBPU321 originally granted on 10/25/2012.

The change under this application is to add following accessories:

1. Add power adapter;
2. Add other model name BPU329 based on BPU321;
3. Change product name from Blood Pressure Meter into Blood Pressure Monitor

The change under this application is adding such power adapter as seen in external photos, Hence following test items should be retest.

AC power line conducted emission

Radiated spurious emission(9KHz-1GHz)

I attest that adding pwer adapter will not affect the RF function out of the grant conditions.

Sincerely yours,

IDT Technology Limited_



Alex chan/ project manager

Fax: +86 755 29799021

Tel: +86 755 29701028

E-mail: achan.ld@idthk.com