

Response to TCB Findings

1. The schematics of the RF module (M1-MINI) from SkyeTek is required. The schematic diagram supplied shows the module as a block only, without any details of the components on it.

Supplied.

2. The operational description mentions a data sheet for the RF module above. Please supply this document as well.

Supplied.

3. Pg 13 of the report states that the frequency range inspected is above 1MHz. Please clarify.

Test data starts at 1 MHz; since the lowest frequency generated in the device is 4.00 MHz, this data should be sufficient according to 15.33(a).

4. Please specify the highest frequency generated or used in the device.

Customer states 13.56 MHz.