

EB-TX210A / EB-TX220A
(FCC ID NWJ10A002A)
Dipole system validation

Table of Contents

| | |
|------------------------------|---|
| 1. Overview | 2 |
| 2. 900MHz system validation | 3 |
| 3. 1900MHz system validation | 4 |


1. Overview

Plot data in this report is validation of dipole antenna. Measurement is based on the document ('Dipole Validation Kit') attached.

2. 900MHz system validation



3. 1900MHz system validation



**Schmid & Partner
Engineering AG**

Staffelstrasse 8, 8045 Zurich, Switzerland, Telefon +41 1 280 08 60, Fax +41 1 280 08 64

DASY

Dipole Validation Kit

Type : D1800V2

Ser.: 205

Manufactured: January 1997
Calibrated: February 1997

1. Measurement Conditions

The measurements were performed with the the new generic twin phantom (shell thickness 2mm) filled with brain simulating sugar solution of the following electrical parameters at 1800 MHz:

| | | |
|------------------------|-------------------|-------|
| Relative Dielectricity | 40.3 | ± 5% |
| Conductivity | 1.69 mho/m | ± 10% |

The DASY2 System (Software version 2.3d) with a dosimetric E-field probe ET3DV4 (SN:1025) was used for the measurements. The Conversion Factor (probe parameter) for the probe was 4.5.

With the Head Phantom, the dipole feedpoint was positioned below the ear hole marking. The dipole orientations used were in a horizontal plane parallel and normal to the body axis. The standard measuring distance was 20mm from dipole centre to the solution surface. The accurate distance positioning was done by using the included distance holder.

This measuring point is not very critical for SAR measurements. The measured variations are:

| | |
|--|---------------------|
| Horizontal shift from/to phantom nose | < 2% for 5mm shift |
| Horizontal shift to bottom/top of head | < 5% for 5mm shift |
| Positioning angle in horizontal plane | < 2% for ±10° shift |

The repeatability of SAR-measurements with normally careful positioning should be better than 5%. The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. The variations from using different coarse grid orientations and spacings or from using the 4x4x7 fine cube were all within 3% of the assessed SAR-value.

2. SAR Measurement

Standard SAR-measurements were performed with the head phantom according to the measurement conditions described in section 1. The results (see figure) have been normalised to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

- SAR in flat phantom:

| | |
|--|------------------|
| SAR at surface (extrapolated): | 25.0mW/g |
| averaged over 1 cm ³ (1 g) of tissue: | 13.7 mW/g |
| averaged over 10 cm ³ (10 g) of tissue: | 7.50 mW/g |

- Dipole normal to body axis (horizontal):

| | |
|--|------------------|
| SAR at surface (extrapolated): | 28.5 mW/g |
| averaged over 1 cm ³ (1 g) of tissue: | 14.6 mW/g |
| averaged over 10 cm ³ (10 g) of tissue: | 8.37 mW/g |

- Dipole parallel to body axis (vertical):

| | |
|--|------------------|
| SAR at surface (extrapolated): | 23.9 mW/g |
| averaged over 1 cm ³ (1 g) of tissue: | 12.6 mW/g |
| averaged over 10 cm ³ (10 g) of tissue: | 7.30 mW/g |

If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are as follows (see also Application Note 4: SAR Sensitivities):

- SAR at the phantom surface:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.73 \qquad \frac{dSAR / SAR}{d\sigma / \sigma} = +0.90$$

- SAR averaged over a cube of 1cm³:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.54 \qquad \frac{dSAR / SAR}{d\sigma / \sigma} = +0.51$$

- SAR averaged over a cube of 10cm³:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.41 \qquad \frac{dSAR / SAR}{d\sigma / \sigma} = +0.23$$

- Penetration depth:

$$\frac{d\delta / \delta}{d\epsilon / \epsilon} = +0.46 \qquad \frac{d\delta / \delta}{d\sigma / \sigma} = -0.96$$

3. Dipole Impedances

The impedances were measured at the SMA-connector with a network analyser and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay: **1.2394 ns** (one direction)
Transmission factor: **0.979** (voltage transmission, one direction)

- Dipole impedance at Flat phantom:

| Distance from solution [mm] | Re{Z} [Ω] | Im{Z} [Ω] | Return Loss [dB] |
|--------------------------------|--------------|--------------|---------------------|
| 10 | 50.9 | 6.3 | -24.5 |
| 15 | 49.2 | -0.3 | -41.3 |
| 20 | 55.3 | 0.9 | -26.2 |
| 25 | 64.6 | 2.4 | -18.2 |
| 30 | 74.3 | 1.1 | -14.6 |
| 35 | 83.6 | -3.5 | -12.3 |
| 40 | 90.4 | -9.0 | -11.0 |
| 45 | 95.4 | -18.2 | -9.9 |
| 50 | 96.0 | -26.5 | -9.3 |

- Dipole impedance at Head phantom:

| Distance from solution [mm] | Re{Z} [Ω] | Im{Z} [Ω] | Return Loss [dB] |
|--------------------------------|--------------|--------------|---------------------|
| 10 | 42.7 | -8.5 | -18.7 |
| 15 | 44.1 | -8.9 | -19.3 |
| 20 | 51.6 | -7.1 | -23.3 |
| 25 | 59.7 | -7.5 | -19.4 |
| 30 | 68.3 | -11.2 | -15.3 |
| 35 | 74.8 | -16.7 | -12.9 |
| 40 | 78.8 | -23.8 | -11.3 |

- Dipole in free space 71.2 -30.0 -11.0

4. Handling

The dipole is made of standard semirigid coaxial cable. The centre conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.


Do not apply excessive force to the dipole arms, because they might bend. If the dipole arms have to be bent back, take care to release stress to the soldered connections near the feedpoint; they might come off.

After prolonged use with 40W radiated power, only a very slight warming of the dipole near the feedpoint can be measured.

D1800V2 sn:205 / Flat Phantom / Pin = 1W / d = 20 mm
 $\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\mu = 1.00$ [g/cm³]

SAR [mW/g] Max: 15.1

SAR (1g): 13.7 [mW/g] SAR (10g): 7.50 [mW/g]




D1800V2 sn:205 / Head Phantom / horizontal / Pin = 1W / d = 20mm

$\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\rho = 1.00$ [g/cm³]

SAR [mW/g] Max: 16.20

SAR (1g): 14.6 [mW/g] SAR (10g): 8.37 [mW/g]




D1800V2 sn:205 / Head Phantom / vertical / Pin = 1W / d = 20mm

$\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\rho = 1.00$ [g/cm³]

SAR [mW/g] Max: 14.25

SAR (1g): 12.6 [mW/g] SAR (10g): 7.30 [mW/g]




D1800V2 SN:205

S11

Flat phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution




D1800V2 SN:205

S11

Head phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



CHI 1→1 1 U FS L: 49.391 0 1: -556264 m0 158.84 pF
1 800.000 000 MHz

Cor
E
De1


D1800V2 SN:205

S11

Flat phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



START 1 400.000 000 MHz STOP 2 200.000 000 MHz

CHI 1→1 1 U FS 1

(a)

Cor
E
Del


D1800V2 SN:205

S11

Head phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



START 1 400.000 000 MHZ

STOP 2 200.000 000 MHZ

**Schmid & Partner
Engineering AG**

Staffelstrasse 8, 8045 Zurich, Switzerland, Telefon +41 1 280 08 60, Fax +41 1 280 08 64

DASY

Dipole Validation Kit

Type : D1800V2

Ser.: 205

Manufactured: January 1997
Calibrated: February 1997

1. Measurement Conditions

The measurements were performed with the the new generic twin phantom (shell thickness 2mm) filled with brain simulating sugar solution of the following electrical parameters at 1800 MHz:

| | | |
|------------------------|-------------------|------------|
| Relative Dielectricity | 40.3 | $\pm 5\%$ |
| Conductivity | 1.69 mho/m | $\pm 10\%$ |

The DASY2 System (Software version 2.3d) with a dosimetric E-field probe ET3DV4 (SN:1025) was used for the measurements. The Conversion Factor (probe parameter) for the probe was 4.5.

With the Head Phantom, the dipole feedpoint was positioned below the ear hole marking. The dipole orientations used were in a horizontal plane parallel and normal to the body axis. The standard measuring distance was 20mm from dipole centre to the solution surface. The accurate distance positioning was done by using the included distance holder.

This measuring point is not very critical for SAR measurements. The measured variations are:

| | |
|--|-------------------------------|
| Horizontal shift from/to phantom nose | < 2% for 5mm shift |
| Horizontal shift to bottom/top of head | < 5% for 5mm shift |
| Positioning angle in horizontal plane | < 2% for $\pm 10^\circ$ shift |

The repeatability of SAR-measurements with normally careful positioning should be better than 5%. The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. The variations from using different coarse grid orientations and spacings or from using the 4x4x7 fine cube were all within 3% of the assessed SAR-value.

2. SAR Measurement

Standard SAR-measurements were performed with the head phantom according to the measurement conditions described in section 1. The results (see figure) have been normalised to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

- SAR in flat phantom:

| | |
|--|------------------|
| SAR at surface (extrapolated): | 25.0mW/g |
| averaged over 1 cm ³ (1 g) of tissue: | 13.7 mW/g |
| averaged over 10 cm ³ (10 g) of tissue: | 7.50 mW/g |

- Dipole normal to body axis (horizontal):

SAR at surface (extrapolated): **28.5 mW/g**

averaged over 1 cm³ (1 g) of tissue: **14.6 mW/g**

averaged over 10 cm³ (10 g) of tissue: **8.37 mW/g**

- Dipole parallel to body axis (vertical):

SAR at surface (extrapolated): **23.9 mW/g**

averaged over 1 cm³ (1 g) of tissue: **12.6 mW/g**

averaged over 10 cm³ (10 g) of tissue: **7.30 mW/g**

If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are as follows (see also Application Note 4: SAR Sensitivities):

- SAR at the phantom surface:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.73$$

$$\frac{dSAR / SAR}{d\sigma / \sigma} = +0.90$$

- SAR averaged over a cube of 1cm³:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.54$$

$$\frac{dSAR / SAR}{d\sigma / \sigma} = +0.51$$

- SAR averaged over a cube of 10cm³:

$$\frac{dSAR / SAR}{d\epsilon / \epsilon} = -0.41$$

$$\frac{dSAR / SAR}{d\sigma / \sigma} = +0.23$$

- Penetration depth:

$$\frac{d\delta / \delta}{d\epsilon / \epsilon} = +0.46$$

$$\frac{d\delta / \delta}{d\sigma / \sigma} = -0.96$$

3. Dipole Impedances

The impedances were measured at the SMA-connector with a network analyser and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay: **1.2394 ns** (one direction)
Transmission factor: **0.979** (voltage transmission, one direction)

- Dipole impedance at Flat phantom:

| Distance from solution [mm] | Re{Z} [Ω] | Im{Z} [Ω] | Return Loss [dB] |
|--------------------------------|--------------|--------------|---------------------|
| 10 | 50.9 | 6.3 | -24.5 |
| 15 | 49.2 | -0.3 | -41.3 |
| 20 | 55.3 | 0.9 | -26.2 |
| 25 | 64.6 | 2.4 | -18.2 |
| 30 | 74.3 | 1.1 | -14.6 |
| 35 | 83.6 | -3.5 | -12.3 |
| 40 | 90.4 | -9.0 | -11.0 |
| 45 | 95.4 | -18.2 | -9.9 |
| 50 | 96.0 | -26.5 | -9.3 |

- Dipole impedance at Head phantom:

| Distance from solution [mm] | Re{Z} [Ω] | Im{Z} [Ω] | Return Loss [dB] |
|--------------------------------|--------------|--------------|---------------------|
| 10 | 42.7 | -8.5 | -18.7 |
| 15 | 44.1 | -8.9 | -19.3 |
| 20 | 51.6 | -7.1 | -23.3 |
| 25 | 59.7 | -7.5 | -19.4 |
| 30 | 68.3 | -11.2 | -15.3 |
| 35 | 74.8 | -16.7 | -12.9 |
| 40 | 78.8 | -23.8 | -11.3 |

- Dipole in free space 71.2 -30.0 -11.0

4. Handling

The dipole is made of standard semirigid coaxial cable. The centre conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

Do not apply excessive force to the dipole arms, because they might bend. If the dipole arms have to be bent back, take care to release stress to the soldered connections near the feedpoint; they might come off.


After prolonged use with 40W radiated power, only a very slight warming of the dipole near the feedpoint can be measured.

D1800V2 sn:205 / Flat Phantom / Pin = 1W / d = 20 mm

$\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\mu = 1.00$ [g/cm³]

SAR [mW/g] Max: 15.1

SAR (1g): 13.7 [mW/g] SAR (10g): 7.50 [mW/g]




D1800V2 sn:205 / Head Phantom / horizontal / Pin = 1W / d = 20mm

$\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\rho = 1.00$ [g/cm³]

SAR [mW/g] Max: 16.20

SAR (1g): 14.6 [mW/g] SAR (10g): 8.37 [mW/g]




D1800V2 sn:205 / Head Phantom / vertical / Pin = 1W / d = 20mm

$\sigma = 1.69$ [mho/m] $\epsilon_r = 40.3$ $\rho = 1.00$ [g/cm³]

SAR [mW/g] Max: 14.25

SAR (1g): 12.6 [mW/g] SAR (10g): 7.30 [mW/g]




D1800V2 SN:205

S11

Flat phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution




D1800V2 SN:205

S11

Head phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



CHI 1→1 1 U FS L: 49.391 0 1: -556264 m0 158.84 pF
1 800.000 000 MHz

Cor
E
De1


D1800V2 SN:205

S11

Flat phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



START 1 400.000 000 MHz STOP 2 200.000 000 MHz

CHI 1→1 1 U FS 1

(a)

Cor
E
Del


D1800V2 SN:205

S11

Head phantom with
brain simulating
solution

- 1: d = 10mm (red)
- 2: d = 15mm (green)
- 3: d = 20mm (blue)
- 4: d = 30mm (violet)

d=distance from dipole
center to solution



START 1 400.000 000 MHZ

STOP 2 200.000 000 MHZ

**Schmid & Partner
Engineering AG**

Staffelstrasse 8, 8045 Zurich, Switzerland, Telefon +41 1 280 08 60, Fax +41 1 280 08 64

Probe ET3DV5

SN:1303

| | |
|-------------------|------------------|
| Manufactured: | October 10, 1996 |
| Last calibration: | October 15, 1998 |
| Recalibrated: | February 1, 2000 |

Calibrated for System DASY2

ET3DV5 SN:1303

DASY2 - Parameters of Probe: ET3DV5 SN:1303

Sensitivity in Free Space

| | |
|-------|--|
| NormX | 1.34 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | 1.26 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | 1.41 $\mu\text{V}/(\text{V}/\text{m})^2$ |

Diode Compression

| | |
|-------|-------|
| DCP X | 82 mV |
| DCP Y | 82 mV |
| DCP Z | 82 mV |


Sensitivity in Tissue Simulating Liquid

| | | | | |
|--------------|-----------------|-----------------------------|--|------|
| Brain | 450 MHz | $\epsilon_r = 48 \pm 5\%$ | $\sigma = 0.50 \pm 10\% \text{ mho/m}$ | |
| ConvF X | 6.17 | extrapolated | Boundary effect: | |
| ConvF Y | 6.17 | extrapolated | Alpha | 0.07 |
| ConvF Z | 6.17 | extrapolated | Depth | 5.42 |
| Brain | 900 MHz | $\epsilon_r = 42.5 \pm 5\%$ | $\sigma = 0.86 \pm 10\% \text{ mho/m}$ | |
| ConvF X | 5.83 | $\pm 7\%$ (k=2) | Boundary effect: | |
| ConvF Y | 5.83 | $\pm 7\%$ (k=2) | Alpha | 0.22 |
| ConvF Z | 5.83 | $\pm 7\%$ (k=2) | Depth | 4.33 |
| Brain | 1500 MHz | $\epsilon_r = 41 \pm 5\%$ | $\sigma = 1.32 \pm 10\% \text{ mho/m}$ | |
| ConvF X | 5.37 | interpolated | Boundary effect: | |
| ConvF Y | 5.37 | interpolated | Alpha | 0.42 |
| ConvF Z | 5.37 | interpolated | Depth | 2.89 |
| Brain | 1800 MHz | $\epsilon_r = 41 \pm 5\%$ | $\sigma = 1.69 \pm 10\% \text{ mho/m}$ | |
| ConvF X | 5.15 | $\pm 7\%$ (k=2) | Boundary effect: | |
| ConvF Y | 5.15 | $\pm 7\%$ (k=2) | Alpha | 0.52 |
| ConvF Z | 5.15 | $\pm 7\%$ (k=2) | Depth | 2.16 |


Sensor Offset

| | | |
|----------------------------|----------------|----|
| Probe Tip to Sensor Center | 2.7 | mm |
| Optical Surface Detection | 1.95 \pm 0.2 | mm |


Receiving Pattern (ϕ), $\theta = 0^\circ$



ET3DV5 SN:1303




Isotropy Error (ϕ), $\theta = 0^\circ$



ET3DV5 SN:1303



Frequency Response of E-Field

(TEM-Cell:ifi110, Waveguide R22, R26)




ET3DV5 SN:1303

Dynamic Range f(SAR_{brain}) (TEM-Cell:ifi110)



Conversion Factor Assessment



Receiving Pattern (ϕ) (in brain tissue, z = 5 mm)

