


**BODY-BOTTOM**



**BODY LIQUID DEPTH**



**HEAD LIQUID DEPTH**

**APPENDIX C: PROBE CALIBRATION CERTIFICATE**



**COMOSAR E-Field Probe Calibration Report**

Ref : ACR.266.2.14.SATU.A

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION**  
**133 JIANG-JUN ROAD, JIANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE**  
**NANJING 211153 PR CHINA**  
**SATIMO COMOSAR DOSIMETRIC E-FIELD PROBE**  
**SERIAL NO.: SN 35/11 EP131**

**Calibrated at SATIMO US**  
**2105 Barrett Park Dr. - Kennesaw, GA 30144**



**09/22/2014**

*Summary:*

This document presents the method and results from an accredited COMOSAR Dosimetric E-Field Probe calibration performed in SATIMO USA using the CALISAR / CALIBAIR test bench, for use with a SATIMO COMOSAR system only. All calibration results are traceable to national metrology institutions.



|                      | <i>Name</i>   | <i>Function</i> | <i>Date</i> | <i>Signature</i>     |
|----------------------|---------------|-----------------|-------------|----------------------|
| <i>Prepared by :</i> | Jérôme LUC    | Product Manager | 9/23/2014   | <i>JS</i>            |
| <i>Checked by :</i>  | Jérôme LUC    | Product Manager | 9/23/2014   | <i>JS</i>            |
| <i>Approved by :</i> | Kim RUTKOWSKI | Quality Manager | 9/23/2014   | <i>Kim Rutkowski</i> |

|                       | <i>Customer Name</i>                              |
|-----------------------|---|
| <i>Distribution :</i> | Inventec Appliances<br>(Jiangning)<br>Corporation |

| <i>Issue</i> | <i>Date</i> | <i>Modifications</i> |
|--------------|-------------|----------------------|
| A            | 9/23/2014   | Initial release      |
|              |             |                      |
|              |             |                      |
|              |             |                      |



---

TABLE OF CONTENTS

|     |                                       |   |
|-----|---------------------------------------|---|
| 1   | Device Under Test .....               | 4 |
| 2   | Product Description .....             | 4 |
| 2.1 | General Information .....             | 4 |
| 3   | Measurement Method .....              | 4 |
| 3.1 | Linearity .....                       | 4 |
| 3.2 | Sensitivity .....                     | 5 |
| 3.3 | Lower Detection Limit .....           | 5 |
| 3.4 | Isotropy .....                        | 5 |
| 3.5 | Boundary Effect .....                 | 5 |
| 4   | Measurement Uncertainty .....         | 5 |
| 5   | Calibration Measurement Results ..... | 6 |
| 5.1 | Sensitivity in air .....              | 6 |
| 5.2 | Linearity .....                       | 7 |
| 5.3 | Sensitivity in liquid .....           | 7 |
| 5.4 | Isotropy .....                        | 8 |
| 6   | List of Equipment .....               | 9 |



**1 DEVICE UNDER TEST**

| Device Under Test                        |   |
|--|---|
| Device Type                              | COMOSAR DOSIMETRIC E FIELD PROBE  |
| Manufacturer                             | Satimo  |
| Model                                    | SSE5  |
| Serial Number                            | SN 35/11 EP131  |
| Product Condition (new / used)           | Used  |
| Frequency Range of Probe                 | 0.7 GHz-3GHz  |
| Resistance of Three Dipoles at Connector | Dipole 1: R1=0.999 MΩ<br>Dipole 2: R2=1.244 MΩ<br>Dipole 3: R3=1.253 MΩ |

A yearly calibration interval is recommended.

**2 PRODUCT DESCRIPTION**

**2.1 GENERAL INFORMATION**

Satimo's COMOSAR E field Probes are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards.



**Figure 1 – Satimo COMOSAR Dosimetric E field Dipole**

|  |        |
|--|--------|
| Probe Length                               | 330 mm |
| Length of Individual Dipoles               | 4.5 mm |
| Maximum external diameter                  | 8 mm   |
| Probe Tip External Diameter                | 5 mm   |
| Distance between dipoles / probe extremity | 2.7 mm |

**3 MEASUREMENT METHOD**

The IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards provide recommended practices for the probe calibrations, including the performance characteristics of interest and methods by which to assess their affect. All calibrations / measurements performed meet the fore mentioned standards.

**3.1 LINEARITY**

The evaluation of the linearity was done in free space using the waveguide, performing a power sweep to cover the SAR range 0.01W/kg to 100W/kg.



### 3.2 SENSITIVITY

The sensitivity factors of the three dipoles were determined using a two step calibration method (air and tissue simulating liquid) using waveguides as outlined in the standards.

### 3.3 LOWER DETECTION LIMIT

The lower detection limit was assessed using the same measurement set up as used for the linearity measurement. The required lower detection limit is 10 mW/kg.

### 3.4 ISOTROPY

The axial isotropy was evaluated by exposing the probe to a reference wave from a standard dipole with the dipole mounted under the flat phantom in the test configuration suggested for system validations and checks. The probe was rotated along its main axis from 0 - 360 degrees in 15 degree steps. The hemispherical isotropy is determined by inserting the probe in a thin plastic box filled with tissue-equivalent liquid, with the plastic box illuminated with the fields from a half wave dipole. The dipole is rotated about its axis (0°–180°) in 15° increments. At each step the probe is rotated about its axis (0°–360°).

### 3.5 BOUNDARY EFFECT

The boundary effect is defined as the deviation between the SAR measured data and the expected exponential decay in the liquid when the probe is oriented normal to the interface. To evaluate this effect, the liquid filled flat phantom is exposed to fields from either a reference dipole or waveguide. With the probe normal to the phantom surface, the peak spatial average SAR is measured and compared to the analytical value at the surface.

## 4 MEASUREMENT UNCERTAINTY

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty associated with an E-field probe calibration using the waveguide technique. All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2, traceable to the Internationally Accepted Guides to Measurement Uncertainty.

| Uncertainty analysis of the probe calibration in waveguide |                       |                          |            |    |                          |
|--|-----------------------|--------------------------|------------|----|--------------------------|
| ERROR SOURCES  | Uncertainty value (%) | Probability Distribution | Divisor    | ci | Standard Uncertainty (%) |
| Incident or forward power                                  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Reflected power  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Liquid conductivity  | 5.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.887%                   |
| Liquid permittivity  | 4.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.309%                   |
| Field homogeneity  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Field probe positioning                                    | 5.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.887%                   |
| Field probe linearity                                      | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |



|   |  |  |  |  |        |
|---|--|--|--|--|--------|
| Combined standard uncertainty                       |  |  |  |  | 5.831% |
| Expanded uncertainty<br>95 % confidence level k = 2 |  |  |  |  | 12.0%  |

## 5 CALIBRATION MEASUREMENT RESULTS

| Calibration Parameters |       |
|------------------------|-------|
| Liquid Temperature     | 21 °C |
| Lab Temperature        | 21 °C |
| Lab Humidity           | 45 %  |


### 5.1 SENSITIVITY IN AIR

| Normx dipole<br>1 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normy dipole<br>2 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normz dipole<br>3 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) |
|---|---|---|
| 4.98  | 6.07  | 5.22  |

| DCP dipole 1<br>(mV) | DCP dipole 2<br>(mV) | DCP dipole 3<br>(mV) |
|----------------------|----------------------|----------------------|
| 96                   | 93                   | 99                   |

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain H-field value using the formula:


$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$





5.2 LINEARITY

Linearity



Linearity:  $\pm 1.45\%$  ( $\pm 0.06\text{dB}$ )

5.3 SENSITIVITY IN LIQUID

| Liquid | Frequency<br>(MHz +/-<br>100MHz) | Permittivity | Epsilon (S/m) | ConvF |
|--------|----------------------------------|--------------|---------------|-------|
| HL1800 | 1800                             | 41.31        | 1.38          | 6.99  |
| BL1800 | 1800                             | 53.27        | 1.51          | 7.27  |
| HL1900 | 1900                             | 41.09        | 1.42          | 7.69  |
| BL1900 | 1900                             | 54.20        | 1.54          | 7.95  |
| HL2000 | 2000                             | 39.72        | 1.43          | 7.15  |
| BL2000 | 2000                             | 53.91        | 1.53          | 7.35  |
| HL2000 | 2000                             | 39.72        | 1.43          | 7.30  |
| BL2000 | 2000                             | 53.91        | 1.53          | 7.47  |
| HL2450 | 2450                             | 39.05        | 1.77          | 7.22  |
| BL2450 | 2450                             | 52.97        | 1.93          | 7.46  |
| HL2600 | 2600                             | 38.35        | 1.92          | 7.08  |
| BL2600 | 2600                             | 51.81        | 2.19          | 7.32  |

LOWER DETECTION LIMIT: 9mW/kg






5.4 ISOTROPY

HL1800 MHz

- Axial isotropy: 0.04 dB
- Hemispherical isotropy: 0.07 dB





6 LIST OF EQUIPMENT

| Equipment Summary Sheet       |                      |                    |   |   |
|-------------------------------|----------------------|--------------------|---|---|
| Equipment Description         | Manufacturer / Model | Identification No. | Current Calibration Date                      | Next Calibration Date                         |
| Flat Phantom                  | Satimo               | SN-20/09-SAM71     | Validated. No cal required.                   | Validated. No cal required.                   |
| COMOSAR Test Bench            | Version 3            | NA                 | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer              | Rhode & Schwarz ZVA  | SN100132           | 02/2013                                       | 02/2016                                       |
| Reference Probe               | Satimo               | EP 94 SN 37/08     | 10/2013                                       | 10/2014                                       |
| Multimeter                    | Keithley 2000        | 1188656            | 12/2013                                       | 12/2016                                       |
| Signal Generator              | Agilent E4438C       | MY49070581         | 12/2013                                       | 12/2016                                       |
| Amplifier                     | Aethercomm           | SN 046             | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                   | HP E4418A            | US38261498         | 12/2013                                       | 12/2016                                       |
| Power Sensor                  | HP ECP-E26A          | US37181460         | 12/2013                                       | 12/2016                                       |
| Directional Coupler           | Narda 4216-20        | 01386              | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Waveguide                     | Mega Industries      | 069Y7-158-13-712   | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide Transition          | Mega Industries      | 069Y7-158-13-701   | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide Termination         | Mega Industries      | 069Y7-158-13-701   | Validated. No cal required.                   | Validated. No cal required.                   |
| Temperature / Humidity Sensor | Control Company      | 11-661-9           | 8/2012  | 8/2015  |



## COMOSAR E-Field Probe Calibration Report

Ref : ACR.266.1.14.SATU.A

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION**  
133 JIANG-JUN ROAD, JIANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE  
NANJING 211153 PR CHINA  
**SATIMO COMOSAR DOSIMETRIC E-FIELD PROBE**  
SERIAL NO.: SN 18/11 EP121

Calibrated at SATIMO US  
2105 Barrett Park Dr. - Kennesaw, GA 30144



09/22/2014

*Summary:*

This document presents the method and results from an accredited COMOSAR Dosimetric E-Field Probe calibration performed in SATIMO USA using the CALISAR / CALIBAIR test bench, for use with a SATIMO COMOSAR system only. All calibration results are traceable to national metrology institutions.



|                      | <i>Name</i>   | <i>Function</i> | <i>Date</i> | <i>Signature</i>     |
|----------------------|---------------|-----------------|-------------|----------------------|
| <i>Prepared by :</i> | Jérôme LUC    | Product Manager | 9/23/2014   | <i>JS</i>            |
| <i>Checked by :</i>  | Jérôme LUC    | Product Manager | 9/23/2014   | <i>JS</i>            |
| <i>Approved by :</i> | Kim RUTKOWSKI | Quality Manager | 9/23/2014   | <i>Kim Rutkowski</i> |

|                       | <i>Customer Name</i>                              |
|-----------------------|---|
| <i>Distribution :</i> | Inventec Appliances<br>(Jiangning)<br>Corporation |

| <i>Issue</i> | <i>Date</i> | <i>Modifications</i> |
|--------------|-------------|----------------------|
| A            | 9/23/2014   | Initial release      |
|              |             |                      |
|              |             |                      |
|              |             |                      |



---

TABLE OF CONTENTS

|     |                                       |   |
|-----|---------------------------------------|---|
| 1   | Device Under Test .....               | 4 |
| 2   | Product Description .....             | 4 |
| 2.1 | General Information .....             | 4 |
| 3   | Measurement Method .....              | 4 |
| 3.1 | Linearity .....                       | 4 |
| 3.2 | Sensitivity .....                     | 5 |
| 3.3 | Lower Detection Limit .....           | 5 |
| 3.4 | Isotropy .....                        | 5 |
| 3.5 | Boundary Effect .....                 | 5 |
| 4   | Measurement Uncertainty .....         | 5 |
| 5   | Calibration Measurement Results ..... | 6 |
| 5.1 | Sensitivity in air .....              | 6 |
| 5.2 | Linearity .....                       | 7 |
| 5.3 | Sensitivity in liquid .....           | 7 |
| 5.4 | Isotropy .....                        | 8 |
| 6   | List of Equipment .....               | 9 |



**1 DEVICE UNDER TEST**

| Device Under Test                        |   |
|--|---|
| Device Type                              | COMOSAR DOSIMETRIC E FIELD PROBE  |
| Manufacturer                             | Satimo  |
| Model                                    | SSE5  |
| Serial Number                            | SN 18/11 EP121  |
| Product Condition (new / used)           | Used  |
| Frequency Range of Probe                 | 0.7 GHz-3GHz  |
| Resistance of Three Dipoles at Connector | Dipole 1: R1=1.399 MΩ<br>Dipole 2: R2=1.338 MΩ<br>Dipole 3: R3=1.165 MΩ |

A yearly calibration interval is recommended.

**2 PRODUCT DESCRIPTION**

**2.1 GENERAL INFORMATION**

Satimo's COMOSAR E field Probes are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards.



**Figure 1 – Satimo COMOSAR Dosimetric E field Dipole**

|  |        |
|--|--------|
| Probe Length                               | 330 mm |
| Length of Individual Dipoles               | 4.5 mm |
| Maximum external diameter                  | 8 mm   |
| Probe Tip External Diameter                | 5 mm   |
| Distance between dipoles / probe extremity | 2.7 mm |

**3 MEASUREMENT METHOD**

The IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards provide recommended practices for the probe calibrations, including the performance characteristics of interest and methods by which to assess their affect. All calibrations / measurements performed meet the fore mentioned standards.

**3.1 LINEARITY**

The evaluation of the linearity was done in free space using the waveguide, performing a power sweep to cover the SAR range 0.01W/kg to 100W/kg.



### 3.2 SENSITIVITY

The sensitivity factors of the three dipoles were determined using a two step calibration method (air and tissue simulating liquid) using waveguides as outlined in the standards.

### 3.3 LOWER DETECTION LIMIT

The lower detection limit was assessed using the same measurement set up as used for the linearity measurement. The required lower detection limit is 10 mW/kg.

### 3.4 ISOTROPY

The axial isotropy was evaluated by exposing the probe to a reference wave from a standard dipole with the dipole mounted under the flat phantom in the test configuration suggested for system validations and checks. The probe was rotated along its main axis from 0 - 360 degrees in 15 degree steps. The hemispherical isotropy is determined by inserting the probe in a thin plastic box filled with tissue-equivalent liquid, with the plastic box illuminated with the fields from a half wave dipole. The dipole is rotated about its axis (0°–180°) in 15° increments. At each step the probe is rotated about its axis (0°–360°).

### 3.5 BOUNDARY EFFECT

The boundary effect is defined as the deviation between the SAR measured data and the expected exponential decay in the liquid when the probe is oriented normal to the interface. To evaluate this effect, the liquid filled flat phantom is exposed to fields from either a reference dipole or waveguide. With the probe normal to the phantom surface, the peak spatial average SAR is measured and compared to the analytical value at the surface.

## 4 MEASUREMENT UNCERTAINTY

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty associated with an E-field probe calibration using the waveguide technique. All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2, traceable to the Internationally Accepted Guides to Measurement Uncertainty.

| Uncertainty analysis of the probe calibration in waveguide |                       |                          |            |    |                          |
|--|-----------------------|--------------------------|------------|----|--------------------------|
| ERROR SOURCES  | Uncertainty value (%) | Probability Distribution | Divisor    | ci | Standard Uncertainty (%) |
| Incident or forward power                                  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Reflected power  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Liquid conductivity  | 5.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.887%                   |
| Liquid permittivity  | 4.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.309%                   |
| Field homogeneity  | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |
| Field probe positioning                                    | 5.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 2.887%                   |
| Field probe linearity                                      | 3.00%                 | Rectangular              | $\sqrt{3}$ | 1  | 1.732%                   |

Page: 5/9

*This document shall not be reproduced, except in full or in part, without the written approval of SATIMO. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of SATIMO.*



|   |  |  |  |  |        |
|---|--|--|--|--|--------|
| Combined standard uncertainty                       |  |  |  |  | 5.831% |
| Expanded uncertainty<br>95 % confidence level k = 2 |  |  |  |  | 12.0%  |

## 5 CALIBRATION MEASUREMENT RESULTS

| Calibration Parameters |       |
|------------------------|-------|
| Liquid Temperature     | 21 °C |
| Lab Temperature        | 21 °C |
| Lab Humidity           | 45 %  |


### 5.1 SENSITIVITY IN AIR

| Normx dipole<br>1 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normy dipole<br>2 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) | Normz dipole<br>3 ( $\mu\text{V}/(\text{V}/\text{m})^2$ ) |
|---|---|---|
| 5.55  | 5.30  | 4.95  |

| DCP dipole 1<br>(mV) | DCP dipole 2<br>(mV) | DCP dipole 3<br>(mV) |
|----------------------|----------------------|----------------------|
| 96                   | 93                   | 98                   |

Calibration curves  $e_i=f(V)$  ( $i=1,2,3$ ) allow to obtain H-field value using the formula:

$$E = \sqrt{E_1^2 + E_2^2 + E_3^2}$$




Page: 6/9

*This document shall not be reproduced, except in full or in part, without the written approval of SATIMO. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of SATIMO.*





5.2 LINEARITY



Linearity:  $\pm 1.98\%$  ( $\pm 0.09\text{dB}$ )

5.3 SENSITIVITY IN LIQUID

| Liquid | Frequency (MHz +/- 100MHz) | Permittivity | Epsilon (S/m) | ConvF |
|--------|----------------------------|--------------|---------------|-------|
| HL450  | 450                        | 43.90        | 0.87          | 8.61  |
| BL450  | 450                        | 58.63        | 0.98          | 8.86  |
| HL750  | 750                        | 42.06        | 0.89          | 7.18  |
| BL750  | 750                        | 56.57        | 0.99          | 7.45  |
| HL850  | 835                        | 42.81        | 0.89          | 7.73  |
| BL850  | 835                        | 53.46        | 0.96          | 8.03  |
| HL900  | 900                        | 42.47        | 0.96          | 7.40  |
| BL900  | 900                        | 56.69        | 1.08          | 7.64  |
| HL1450 | 1450                       | 39.54        | 1.18          | 7.23  |
| BL1450 | 1450                       | 54.74        | 1.34          | 7.38  |


LOWER DETECTION LIMIT: 7mW/kg



5.4 ISOTROPY

HL900 MHz

- Axial isotropy: 0.04 dB
- Hemispherical isotropy: 0.06 dB





6 LIST OF EQUIPMENT

| Equipment Summary Sheet       |                      |                    |   |   |
|-------------------------------|----------------------|--------------------|---|---|
| Equipment Description         | Manufacturer / Model | Identification No. | Current Calibration Date                      | Next Calibration Date                         |
| Flat Phantom                  | Satimo               | SN-20/09-SAM71     | Validated. No cal required.                   | Validated. No cal required.                   |
| COMOSAR Test Bench            | Version 3            | NA                 | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer              | Rhode & Schwarz ZVA  | SN100132           | 02/2013                                       | 02/2016                                       |
| Reference Probe               | Satimo               | EP 94 SN 37/08     | 10/2013                                       | 10/2014                                       |
| Multimeter                    | Keithley 2000        | 1188656            | 12/2013                                       | 12/2016                                       |
| Signal Generator              | Agilent E4438C       | MY49070581         | 12/2013                                       | 12/2016                                       |
| Amplifier                     | Aethercomm           | SN 046             | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                   | HP E4418A            | US38261498         | 12/2013                                       | 12/2016                                       |
| Power Sensor                  | HP ECP-E26A          | US37181460         | 12/2013                                       | 12/2016                                       |
| Directional Coupler           | Narda 4216-20        | 01386              | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Waveguide                     | Mega Industries      | 069Y7-158-13-712   | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide Transition          | Mega Industries      | 069Y7-158-13-701   | Validated. No cal required.                   | Validated. No cal required.                   |
| Waveguide Termination         | Mega Industries      | 069Y7-158-13-701   | Validated. No cal required.                   | Validated. No cal required.                   |
| Temperature / Humidity Sensor | Control Company      | 11-661-9           | 8/2012  | 8/2015  |

**APPENDIX D: DIPOLE CALIBRATION REPORT**

**Inventec Appliances(Jiangning) Corporation  
Testing Laboratory**

**Report NO: JZ-1607-0075**

**SAR Reference Dipole Calibration Report**

**FREQUENCY: 835 MHZ**

**SERIAL NO.:SN 39/09 DIPC117**

|  |
|--|
| <p><b>INVENTEC APPLIANCES (JIANGNING)<br/>CORPORATION TESTING LABORATORY</b></p> <p><b>133 JIANG-JUN ROAD, JIANGNING ECONOMIC AND<br/>TECHNOLOGICAL DEVELOPMENT ZONE<br/>NANJING 211153 PR CHINA</b></p> |
| <p>Calibrated at INVENTEC</p> <p>07/01/2016</p>  |

|              | Name         | Function        | Date       | Signature    |
|--------------|--------------|-----------------|------------|--------------|
| Tested By:   | Zhang shuqin | Test Engineer   | 07/01/2016 | Zhang Shuqin |
| Reviewed By: | Ji jianlin   | Manager         | 07/01/2016 | Ji Jianlin   |
| Approved By: | Xu chunxiu   | Quality Manager | 07/01/2016 | Xu Chunxiu   |

| Issue | Date       | Modifications   |
|-------|------------|-----------------|
| A     | 07/01/2016 | Initial release |
|       |            |                 |

**TABLE OF CONTENTS**

|     |   |    |
|-----|---|----|
| 1.  | INTRODUCTION .....                            | 4  |
| 2.  | DEVICE UNDER TEST .....                       | 4  |
| 3.  | PRODUCT DESCRIPTION .....                     | 4  |
| 3.1 | GENERAL INFORMATION .....                     | 4  |
| 4.  | MEASUREMENT METHOD .....                      | 4  |
| 4.1 | RETURN LOSS REQUIREMENTS .....                | 4  |
| 4.2 | MECHANICAL REQUIREMENTS .....                 | 4  |
| 5.  | MEASUREMENT UNCERTAINTY .....                 | 5  |
| 5.1 | RETURN LOSS .....                             | 5  |
| 5.2 | DIMENSION MEASUREMENT .....                   | 5  |
| 5.3 | VALIDATION MEASUREMENT .....                  | 5  |
| 6.  | CALIBRATION MEASUREMENT RESULTS .....         | 6  |
| 6.1 | RETURN LOSS .....                             | 6  |
| 6.2 | MECHANICAL DIMENSIONS .....                   | 6  |
| 7.  | VALIDATION MEASUREMENT .....                  | 7  |
| 7.1 | HEAD LIQUID MEASUREMENT .....                 | 7  |
| 7.2 | SAR MEASUREMENT RESULT WITH HEAD LIQUID ..... | 7  |
| 7.3 | BODY LIQUID MEASUREMENT .....                 | 9  |
| 7.4 | SAR MEASUREMENT RESULT WITH BODY LIQUID ..... | 9  |
| 8.  | LIST OF EQUIPMENT .....                       | 11 |

## 1. INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2. DEVICE UNDER TEST

| Device Under Test              |                                  |
|--------------------------------|----------------------------------|
| Device Type                    | COMOSAR 835 MHz REFERENCE DIPOLE |
| Manufacturer                   | Satimo                           |
| Model                          | SID 835                          |
| Serial Number                  | SN 39/09 DIPC117                 |
| Product Condition (new / used) | Used                             |

## 3. PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

Satimo's COMOSAR Validation Dipoles are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards. The product is designed for use with the COMOSAR test bench only.



Figure 1 – Satimo COMOSAR Validation Dipole

## 4. MEASUREMENT METHOD

The IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE Std. 1528 and CEI/IEC 62209 standards specify the mechanical components and dimensions of the validation dipoles, with the dimensions frequency and phantom shell

thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness.

## 5. MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.1 dB                              |

### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 3-300       | 0.05mm                         |

### 5.3 VALIDATION MEASUREMENT


The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 16.19 %              |
| 10 g        | 15.86 %              |



## 6. CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance ( $\Omega$ ) |                |
|-----------------|------------------|------------------|------------------------|----------------|
|                 |                  |                  | Real                   | Imaginary part |
| 835             | -27.82           | -20              | 49.14                  | -31.63         |

### 6.2 MECHANICAL DIMENSIONS

| Frequency MHz | L mm       |          | h mm       |          | d mm      |          |
|---------------|------------|----------|------------|----------|-----------|----------|
|               | required   | measured | required   | measured | required  | measured |
| 300           | 420.0 ±1 % |          | 250.0 ±1 % |          | 6.35 ±1 % |          |
| 450           | 290.0 ±1 % |          | 166.7 ±1 % |          | 6.35 ±1 % |          |
| 750           | 176.0 ±1 % |          | 100.0 ±1 % |          | 6.35 ±1 % |          |
| 835           | 161.0 ±1 % | Pass     | 89.8 ±1 %  | Pass     | 3.6 ±1 %  | Pass     |
| 900           | 149.0 ±1 % |          | 83.3 ±1 %  |          | 3.6 ±1 %  |          |
| 1450          | 89.1 ±1 %  |          | 51.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1500          | 80.5 ±1 %  |          | 50.0 ±1 %  |          | 3.6 ±1 %  |          |
| 1640          | 79.0 ±1 %  |          | 45.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1750          | 75.2 ±1 %  |          | 42.9 ±1 %  |          | 3.6 ±1 %  |          |
| 1800          | 72.0 ±1 %  |          | 41.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1900          | 68.0 ±1 %  |          | 39.5 ±1 %  |          | 3.6 ±1 %  |          |
| 1950          | 66.3 ±1 %  |          | 38.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2000          | 64.5 ±1 %  |          | 37.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2100          | 61.0 ±1 %  |          | 35.7 ±1 %  |          | 3.6 ±1 %  |          |
| 2300          | 55.5 ±1 %  |          | 32.6 ±1 %  |          | 3.6 ±1 %  |          |
| 2450          | 51.5 ±1 %  |          | 30.4 ±1 %  |          | 3.6 ±1 %  |          |
| 2600          | 48.5 ±1 %  |          | 28.8 ±1 %  |          | 3.6 ±1 %  |          |
| 3000          | 41.5 ±1 %  |          | 25.0 ±1 %  |          | 3.6 ±1 %  |          |
| 3500          | 37.0 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |
| 3700          | 34.7 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |

## 7. VALIDATION MEASUREMENT

The IEEE Std. 1528, OET 65 Bulletin C and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity( $\epsilon'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|--------------------------------------|----------|-------------------------------|----------|
|                  | required                             | measured | required                      | measured |
| 300              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 450              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 750              | 41.9 $\pm$ 5 %                       |          | 0.89 $\pm$ 5 %                |          |
| 835              | 41.5 $\pm$ 5 %                       | Pass     | 0.90 $\pm$ 5 %                | Pass     |
| 900              | 41.5 $\pm$ 5 %                       |          | 0.97 $\pm$ 5 %                |          |
| 1450             | 40.5 $\pm$ 5 %                       |          | 1.20 $\pm$ 5 %                |          |
| 1500             | 40.4 $\pm$ 5 %                       |          | 1.23 $\pm$ 5 %                |          |
| 1640             | 40.2 $\pm$ 5 %                       |          | 1.31 $\pm$ 5 %                |          |
| 1750             | 40.1 $\pm$ 5 %                       |          | 1.37 $\pm$ 5 %                |          |
| 1800             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1900             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1950             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2000             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2100             | 39.8 $\pm$ 5 %                       |          | 1.49 $\pm$ 5 %                |          |
| 2300             | 39.5 $\pm$ 5 %                       |          | 1.67 $\pm$ 5 %                |          |
| 2450             | 39.2 $\pm$ 5 %                       |          | 1.80 $\pm$ 5 %                |          |
| 2600             | 39.0 $\pm$ 5 %                       |          | 1.96 $\pm$ 5 %                |          |
| 3000             | 38.5 $\pm$ 5 %                       |          | 2.40 $\pm$ 5 %                |          |
| 3500             | 37.9 $\pm$ 5 %                       |          | 2.91 5 %                      |          |



### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |  |
|---|--|
| Software                                  | OPENSAR V4   |
| Phantom                                   | SN 40/14 SAM117  |
| Probe                                     | SN 17/14 EP220   |
| Liquid                                    | Head Liquid Values: $\epsilon'$ : 41.0 $\sigma$ : 0.90 |
| Distance between dipole center and liquid | 15.0 mm  |
| Area scan resolution                      | dx=8mm/dy=8mm  |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                    |

|                    |         |
|--------------------|---------|
| Frequency          | 835 MHz |
| Input power        | 20 dBm  |
| Liquid Temperature | 22 °C   |
| Lab Temperature    | 22 °C   |
| Lab Humidity       | 45%     |

| Frequency MHz | 1 g SAR (W/kg/W) |             | 10 g SAR (W/kg/W) |             |
|---------------|------------------|-------------|-------------------|-------------|
|               | required         | measured    | required          | measured    |
| 300           | 2.85             |             | 1.94              |             |
| 450           | 4.85             |             | 3.06              |             |
| 750           | 8.49             |             | 5.55              |             |
| 835           | 9.56             | 9.82(0.98)) | 6.22              | 6.34(0.63)) |
| 900           | 10.9             |             | 6.99              |             |
| 1450          | 29               |             | 16                |             |
| 1500          | 30.5             |             | 16.8              |             |
| 1640          | 34.2             |             | 18.4              |             |
| 1750          | 36.4             |             | 19.3              |             |
| 1800          | 38.4             |             | 20.1              |             |
| 1900          | 39.7             |             | 20.5              |             |
| 1950          | 40.5             |             | 20.9              |             |
| 2000          | 41.1             |             | 21.1              |             |
| 2100          | 43.6             |             | 21.9              |             |
| 2300          | 48.7             |             | 23.3              |             |
| 2450          | 52.4             |             | 24                |             |
| 2600          | 55.3             |             | 24.6              |             |
| 3000          | 63.8             |             | 25.7              |             |
| 3500          | 67.1             |             | 25                |             |



7.3 BODY LIQUID MEASUREMENT


| Frequency<br>MHz | Relative permittivity( $\epsilon'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|--------------------------------------|----------|-------------------------------|----------|
|                  | required                             | measured | required                      | measured |
| 300              | 58.2 $\pm$ 5 %                       |          | 0.92 $\pm$ 5 %                |          |
| 450              | 56.7 $\pm$ 5 %                       |          | 0.94 $\pm$ 5 %                |          |
| 750              | 55.5 $\pm$ 5 %                       |          | 0.96 $\pm$ 5 %                |          |
| 835              | 55.2 $\pm$ 5 %                       | Pass     | 0.97 $\pm$ 5 %                | Pass     |
| 900              | 55.0 $\pm$ 5 %                       |          | 1.05 $\pm$ 5 %                |          |
| 915              | 55.0 $\pm$ 5 %                       |          | 1.06 $\pm$ 5 %                |          |
| 1450             | 54.0 $\pm$ 5 %                       |          | 1.30 $\pm$ 5 %                |          |
| 1610             | 53.8 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1800             | 53.3 $\pm$ 5 %                       |          | 1.52 $\pm$ 5 %                |          |
| 1900             | 53.3 $\pm$ 5 %                       |          | 1.52 $\pm$ 5 %                |          |
| 2000             | 53.3 $\pm$ 5 %                       |          | 1.52 $\pm$ 5 %                |          |
| 2100             | 53.2 $\pm$ 5 %                       |          | 1.62 $\pm$ 5 %                |          |
| 2450             | 52.7 $\pm$ 5 %                       |          | 1.95 $\pm$ 5 %                |          |
| 2600             | 52.5 $\pm$ 5 %                       |          | 2.16 $\pm$ 5 %                |          |
| 3000             | 52.0 $\pm$ 5 %                       |          | 2.73 $\pm$ 5 %                |          |
| 3500             | 51.3 $\pm$ 5 %                       |          | 3.31 $\pm$ 5 %                |          |
| 5200             | 49.0 $\pm$ 5 %                       |          | 5.30 $\pm$ 5 %                |          |
| 5300             | 48.9 $\pm$ 5 %                       |          | 5.42 $\pm$ 5 %                |          |
| 5400             | 48.7 $\pm$ 5 %                       |          | 5.53 $\pm$ 5 %                |          |
| 5500             | 48.6 $\pm$ 5 %                       |          | 5.65 $\pm$ 5 %                |          |
| 5600             | 48.5 $\pm$ 5 %                       |          | 5.77 $\pm$ 5 %                |          |
| 5800             | 48.2 $\pm$ 5 %                       |          | 6.00 $\pm$ 5 %                |          |

7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |   |
|---|---|
| Software                                  | OPENSAR V4  |
| Phantom                                   | SN 40/14 SAM117                                       |
| Probe                                     | SN 17/14 EP220  |
| Liquid                                    | Body Liquid Values: $\epsilon'$ :54.1 $\sigma$ : 0.99 |
| Distance between dipole center and liquid | 15.0 mm   |
| Area scan resolution                      | dx=8mm/dy=8mm   |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                   |
| Frequency                                 | 835 MHz   |
| Input power                               | 20 dBm  |
| Liquid Temperature                        | 22 °C   |
| Lab Temperature                           | 22 °C   |
| Lab Humidity                              | 45 %  |

| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | measured         | measured          |
| 835           | 10.05(1.01)      | 6.56(0.66))       |



## 8. LIST OF EQUIPMENT

| Equipment Summary Sheet         |                      |                    |                             |                             |
|---------------------------------|----------------------|--------------------|-----------------------------|-----------------------------|
| Equipment Description           | Manufacturer / Model | Identification No. | Current Calibration Date    | Next Calibration Date       |
| Flat Phantom                    | Satimo               | SN 40/14<br>SAM117 | Validated. No cal required. | Validated. No cal required. |
| COMOSAR Test Bench              | Version 3            | NA                 | Validated. No cal required. | Validated. No cal required. |
| Network Analyzer                | Agilent              | 8753E              | 2015/11/27                  | 2016/11/28                  |
| Reference Probe                 | Satimo               | EP220<br>SN 17/14  | 2015/10/01                  | 2016/10/01                  |
| Multimeter                      | Keithley             | MiltiMeter2000     | 2016/02/27                  | 2017/02/28                  |
| Signal Generator                | Agilent              | E4432B             | 2016/04/09                  | 2017/04/08                  |
| Power Meter                     | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Power Sensor                    | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Temperature and Humidity Sensor | JM                   | JM222              | 2016/06/06                  | 2017/06/05                  |

## **SAR Reference Dipole Calibration Report**

**FREQUENCY: 1800 MHZ**

**SERIAL NO.:SN 39/09 DIPF119**

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION TESTING LABORATORY**

**133 JIANG-JUN ROAD, JI ANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE  
NANJING 211153 PR CHINA**

Calibrated at INVENTEC

07/01/2016

|              | Name         | Function        | Date       | Signature    |
|--------------|--------------|-----------------|------------|--------------|
| Tested By:   | Zhang shuqin | Test Engineer   | 07/01/2016 | Zhang Shuqin |
| Reviewed By: | Ji jianlin   | Manager         | 07/01/2016 | Ji Jianlin   |
| Approved By: | Xu chunxiu   | Quality Manager | 07/01/2016 | Xu Chunxiu   |

| Issue | Date       | Modifications   |
|-------|------------|-----------------|
| A     | 07/01/2016 | Initial release |
|       |            |                 |



**TABLE OF CONTENTS**

|     |   |    |
|-----|---|----|
| 1.  | INTRODUCTION .....                            | 4  |
| 2.  | DEVICE UNDER TEST .....                       | 4  |
| 3.  | PRODUCT DESCRIPTION .....                     | 4  |
| 3.1 | GENERAL INFORMATION.....                      | 4  |
| 4.  | MEASUREMENT METHOD .....                      | 4  |
| 4.1 | RETURN LOSS REQUIREMENTS .....                | 4  |
| 4.2 | MECHANICAL REQUIREMENTS .....                 | 4  |
| 5.  | MEASUREMENT UNCERTAINTY .....                 | 5  |
| 5.1 | RETURN LOSS .....                             | 5  |
| 5.2 | DIMENSION MEASUREMENT .....                   | 5  |
| 5.3 | VALIDATION MEASUREMENT .....                  | 5  |
| 6.  | CALIBRATION MEASUREMENT RESULTS .....         | 6  |
| 6.1 | RETURN LOSS .....                             | 6  |
| 6.2 | MECHANICAL DIMENSIONS .....                   | 6  |
| 7.  | VALIDATION MEASUREMENT .....                  | 7  |
| 7.1 | HEAD LIQUID MEASUREMENT .....                 | 7  |
| 7.2 | SAR MEASUREMENT RESULT WITH HEAD LIQUID ..... | 7  |
| 7.3 | BODY LIQUID MEASUREMENT .....                 | 9  |
| 7.4 | SAR MEASUREMENT RESULT WITH BODY LIQUID ..... | 9  |
| 8.  | LIST OF EQUIPMENT .....                       | 11 |

## 1. INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2. DEVICE UNDER TEST

| Device Under Test              |                                   |
|--------------------------------|-----------------------------------|
| Device Type                    | COMOSAR 1800 MHz REFERENCE DIPOLE |
| Manufacturer                   | Satimo                            |
| Model                          | SID1800                           |
| Serial Number                  | SN 39/09 DIPF119                  |
| Product Condition (new / used) | Used                              |

## 3. PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

Satimo's COMOSAR Validation Dipoles are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards. The product is designed for use with the COMOSAR test bench only.



Figure 1 – Satimo COMOSAR Validation Dipole

## 4. MEASUREMENT METHOD

The IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE Std. 1528 and CEI/IEC 62209 standards specify the mechanical components and

dimensions of the validation dipoles, with the dimensions frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness.

## **5. MEASUREMENT UNCERTAINTY**

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### **5.1 RETURN LOSS**

The following uncertainties apply to the return loss measurement:

| <b>Frequency band</b> | <b>Expanded Uncertainty on Return Loss</b> |
|-----------------------|--|
| 400-6000MHz           | 0.1 dB                                     |

### **5.2 DIMENSION MEASUREMENT**

The following uncertainties apply to the dimension measurements:

| <b>Length (mm)</b> | <b>Expanded Uncertainty on Length</b> |
|--------------------|---------------------------------------|
| 3-300              | 0.05mm                                |


### **5.3 VALIDATION MEASUREMENT**

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| <b>Scan Volume</b> | <b>Expanded Uncertainty</b> |
|--------------------|-----------------------------|
| 1 g                | 16.19 %                     |
| 10 g               | 15.86 %                     |

## 6. CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance ( $\Omega$ ) |                |
|-----------------|------------------|------------------|------------------------|----------------|
|                 |                  |                  | Real                   | Imaginary part |
| 1800            | -29.04           | -20              | 49.22                  | 25.83          |

### 6.2 MECHANICAL DIMENSIONS

| Frequency MHz | L mm       |          | h mm       |          | d mm      |          |
|---------------|------------|----------|------------|----------|-----------|----------|
|               | required   | measured | required   | measured | required  | measured |
| 300           | 420.0 ±1 % |          | 250.0 ±1 % |          | 6.35 ±1 % |          |
| 450           | 290.0 ±1 % |          | 166.7 ±1 % |          | 6.35 ±1 % |          |
| 750           | 176.0 ±1 % |          | 100.0 ±1 % |          | 6.35 ±1 % |          |
| 835           | 161.0 ±1 % |          | 89.8 ±1 %  |          | 3.6 ±1 %  |          |
| 900           | 149.0 ±1 % |          | 83.3 ±1 %  |          | 3.6 ±1 %  |          |
| 1450          | 89.1 ±1 %  |          | 51.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1500          | 80.5 ±1 %  |          | 50.0 ±1 %  |          | 3.6 ±1 %  |          |
| 1640          | 79.0 ±1 %  |          | 45.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1750          | 75.2 ±1 %  |          | 42.9 ±1 %  |          | 3.6 ±1 %  |          |
| 1800          | 72.0 ±1 %  | pass     | 41.7 ±1 %  | pass     | 3.6 ±1 %  | pass     |
| 1900          | 68.0 ±1 %  |          | 39.5 ±1 %  |          | 3.6 ±1 %  |          |
| 1950          | 66.3 ±1 %  |          | 38.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2000          | 64.5 ±1 %  |          | 37.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2100          | 61.0 ±1 %  |          | 35.7 ±1 %  |          | 3.6 ±1 %  |          |
| 2300          | 55.5 ±1 %  |          | 32.6 ±1 %  |          | 3.6 ±1 %  |          |
| 2450          | 51.5 ±1 %  |          | 30.4 ±1 %  |          | 3.6 ±1 %  |          |
| 2600          | 48.5 ±1 %  |          | 28.8 ±1 %  |          | 3.6 ±1 %  |          |
| 3000          | 41.5 ±1 %  |          | 25.0 ±1 %  |          | 3.6 ±1 %  |          |
| 3500          | 37.0 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |
| 3700          | 34.7 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |

## 7. VALIDATION MEASUREMENT

The IEEE Std. 1528, OET 65 Bulletin C and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

### 7.1 HEAD LIQUID MEASUREMENT


| Frequency<br>MHz | Relative permittivity( $\epsilon'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|--------------------------------------|----------|-------------------------------|----------|
|                  | required                             | measured | required                      | measured |
| 300              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 450              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 750              | 41.9 $\pm$ 5 %                       |          | 0.89 $\pm$ 5 %                |          |
| 835              | 41.5 $\pm$ 5 %                       |          | 0.90 $\pm$ 5 %                |          |
| 900              | 41.5 $\pm$ 5 %                       |          | 0.97 $\pm$ 5 %                |          |
| 1450             | 40.5 $\pm$ 5 %                       |          | 1.20 $\pm$ 5 %                |          |
| 1500             | 40.4 $\pm$ 5 %                       |          | 1.23 $\pm$ 5 %                |          |
| 1640             | 40.2 $\pm$ 5 %                       |          | 1.31 $\pm$ 5 %                |          |
| 1750             | 40.1 $\pm$ 5 %                       |          | 1.37 $\pm$ 5 %                |          |
| 1800             | 40.0 $\pm$ 5 %                       | pass     | 1.40 $\pm$ 5 %                | pass     |
| 1900             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1950             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2000             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2100             | 39.8 $\pm$ 5 %                       |          | 1.49 $\pm$ 5 %                |          |
| 2300             | 39.5 $\pm$ 5 %                       |          | 1.67 $\pm$ 5 %                |          |
| 2450             | 39.2 $\pm$ 5 %                       |          | 1.80 $\pm$ 5 %                |          |
| 2600             | 39.0 $\pm$ 5 %                       |          | 1.96 $\pm$ 5 %                |          |
| 3000             | 38.5 $\pm$ 5 %                       |          | 2.40 $\pm$ 5 %                |          |
| 3500             | 37.9 $\pm$ 5 %                       |          | 2.91 5 %                      |          |

### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |  |
|---|--|
| Software                                  | OPENSAR V4   |
| Phantom                                   | SN 40/14 SAM117  |
| Probe                                     | SN 17/14 EP220   |
| Liquid                                    | Head Liquid Values: $\epsilon'$ : 39.4 $\sigma$ : 1.39 |
| Distance between dipole center and liquid | 10.0 mm  |
| Area scan resolution                      | dx=8mm/dy=8mm  |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                    |

| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | measured         | measured          |
| 1800          | 36.86(3.69)      | 19.69(1.97)       |




7.3 BODY LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity(r') |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|---------------------------|----------|-------------------------------|----------|
|                  | required                  | measured | required                      | measured |
| 300              | 58.2 ±5 %                 |          | 0.92 ±5 %                     |          |
| 450              | 56.7 ±5 %                 |          | 0.94 ±5 %                     |          |
| 750              | 55.5 ±5 %                 |          | 0.96 ±5 %                     |          |
| 835              | 55.2 ±5 %                 |          | 0.97 ±5 %                     |          |
| 900              | 55.0 ±5 %                 |          | 1.05 ±5 %                     |          |
| 915              | 55.0 ±5 %                 |          | 1.06 ±5 %                     |          |
| 1450             | 54.0 ±5 %                 |          | 1.30 ±5 %                     |          |
| 1610             | 53.8 ±5 %                 |          | 1.40 ±5 %                     |          |
| 1800             | 53.3 ±5 %                 | Pass     | 1.52 ±5 %                     | Pass     |
| 1900             | 53.3 ±5 %                 |          | 1.52 ±5 %                     |          |
| 2000             | 53.3 ±5 %                 |          | 1.52 ±5 %                     |          |
| 2100             | 53.2 ±5 %                 |          | 1.62 ±5 %                     |          |
| 2450             | 52.7 ±5 %                 |          | 1.95 ±5 %                     |          |
| 2600             | 52.5 ±5 %                 |          | 2.16 ±5 %                     |          |
| 3000             | 52.0 ±5 %                 |          | 2.73 ±5 %                     |          |
| 3500             | 51.3 ±5 %                 |          | 3.31 ±5 %                     |          |
| 5200             | 49.0 ±5 %                 |          | 5.30 ±5 %                     |          |
| 5300             | 48.9 ±5 %                 |          | 5.42 ±5 %                     |          |
| 5400             | 48.7 ±5 %                 |          | 5.53 ±5 %                     |          |
| 5500             | 48.6 ±5 %                 |          | 5.65 ±5 %                     |          |
| 5600             | 48.5 ±5 %                 |          | 5.77 ±5 %                     |          |
| 5800             | 48.2 ±5 %                 |          | 6.00 ±5 %                     |          |

7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|   |   |
|---|---|
| Software                                  | OPENSAR V4                                  |
| Phantom                                   | SN 40/14 SAM117                             |
| Probe                                     | SN 17/14 EP220                              |
| Liquid                                    | Body Liquid Values: eps' :54.1 sigma : 1.50 |
| Distance between dipole center and liquid | 10.0 mm                                     |
| Area scan resolution                      | dx=8mm/dy=8mm                               |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                         |
| Frequency                                 | 1800 MHz                                    |
| Input power                               | 20 dBm                                      |
| Liquid Temperature                        | 22 °C                                       |
| Lab Temperature                           | 22 °C                                       |
| Lab Humidity                              | 45 %  |

| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | <b>measured</b>  | <b>measured</b>   |
| 1800          | 36.86(3.69)      | 19.69(1.97)       |





**8. LIST OF EQUIPMENT**

| Equipment Summary Sheet         |                             |                           |                                 |                              |
|---------------------------------|-----------------------------|---------------------------|---------------------------------|------------------------------|
| <b>Equipment Description</b>    | <b>Manufacturer / Model</b> | <b>Identification No.</b> | <b>Current Calibration Date</b> | <b>Next Calibration Date</b> |
| Flat Phantom                    | Satimo                      | SN 40/14<br>SAM117        | Validated. No cal required.     | Validated. No cal required.  |
| COMOSAR Test Bench              | Version 3                   | NA                        | Validated. No cal required.     | Validated. No cal required.  |
| Network Analyzer                | Agilent                     | 8753E                     | 2015/11/27                      | 2016/11/28                   |
| Reference Probe                 | Satimo                      | EP220<br>SN 17/14         | 2015/10/01                      | 2016/10/01                   |
| Multimeter                      | Keithley                    | MiltiMeter2000            | 2016/02/27                      | 2017/02/28                   |
| Signal Generator                | Agilent                     | E4432B                    | 2016/04/09                      | 2017/04/08                   |
| Power Meter                     | R & S                       | NRP-Z23                   | 2016/06/17                      | 2017/06/16                   |
| Power Sensor                    | R & S                       | NRP-Z23                   | 2016/06/17                      | 2017/06/16                   |
| Temperature and Humidity Sensor | JM                          | JM222                     | 2016/06/06                      | 2017/06/05                   |

## **SAR Reference Dipole Calibration Report**

**FREQUENCY: 1900 MHZ**

**SERIAL NO.:SN 39/09 DIPG120**

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION TESTING LABORATORY**

**133 JIANG-JUN ROAD, JI ANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE  
NANJING 211153 PR CHINA**

Calibrated at INVENTEC

07/01/2016

|              | Name         | Function        | Date       | Signature    |
|--------------|--------------|-----------------|------------|--------------|
| Tested By:   | Zhang shuqin | Test Engineer   | 07/01/2016 | Zhang Shuqin |
| Reviewed By: | Ji jianlin   | Manager         | 07/01/2016 | Ji Jianlin   |
| Approved By: | Xu chunxiu   | Quality Manager | 07/01/2016 | Xu chunxiu   |

| Issue | Date       | Modifications   |
|-------|------------|-----------------|
| A     | 07/01/2016 | Initial release |
|       |            |                 |

**TABLE OF CONTENTS**

|     |   |    |
|-----|---|----|
| 1.  | INTRODUCTION .....                            | 4  |
| 2.  | DEVICE UNDER TEST .....                       | 4  |
| 3.  | PRODUCT DESCRIPTION .....                     | 4  |
| 3.1 | GENERAL INFORMATION .....                     | 4  |
| 4.  | MEASUREMENT METHOD .....                      | 4  |
| 4.1 | RETURN LOSS REQUIREMENTS .....                | 4  |
| 4.2 | MECHANICAL REQUIREMENTS .....                 | 4  |
| 5.  | MEASUREMENT UNCERTAINTY .....                 | 5  |
| 5.1 | RETURN LOSS .....                             | 5  |
| 5.2 | DIMENSION MEASUREMENT .....                   | 5  |
| 5.3 | VALIDATION MEASUREMENT .....                  | 5  |
| 6.  | CALIBRATION MEASUREMENT RESULTS .....         | 6  |
| 6.1 | RETURN LOSS .....                             | 6  |
| 6.2 | MECHANICAL DIMENSIONS .....                   | 6  |
| 7.  | VALIDATION MEASUREMENT .....                  | 7  |
| 7.1 | HEAD LIQUID MEASUREMENT .....                 | 7  |
| 7.2 | SAR MEASUREMENT RESULT WITH HEAD LIQUID ..... | 7  |
| 7.3 | BODY LIQUID MEASUREMENT .....                 | 9  |
| 7.4 | SAR MEASUREMENT RESULT WITH BODY LIQUID ..... | 9  |
| 8.  | LIST OF EQUIPMENT .....                       | 11 |

## 1. INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2. DEVICE UNDER TEST

| Device Under Test              |                                   |
|--------------------------------|-----------------------------------|
| Device Type                    | COMOSAR 1900 MHz REFERENCE DIPOLE |
| Manufacturer                   | Satimo                            |
| Model                          | SID1900                           |
| Serial Number                  | SN 39/09 DIPG120                  |
| Product Condition (new / used) | Used                              |

## 3. PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

Satimo's COMOSAR Validation Dipoles are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards. The product is designed for use with the COMOSAR test bench only.



Figure 1 – Satimo COMOSAR Validation Dipole

## 4. MEASUREMENT METHOD

The IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE Std. 1528 and CEI/IEC 62209 standards specify the mechanical components and

dimensions of the validation dipoles, with the dimensions frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness.

## **5. MEASUREMENT UNCERTAINTY**

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### **5.1 RETURN LOSS**

The following uncertainties apply to the return loss measurement:

| <b>Frequency band</b> | <b>Expanded Uncertainty on Return Loss</b> |
|-----------------------|--|
| 400-6000MHz           | 0.1 dB                                     |

### **5.2 DIMENSION MEASUREMENT**

The following uncertainties apply to the dimension measurements:

| <b>Length (mm)</b> | <b>Expanded Uncertainty on Length</b> |
|--------------------|---------------------------------------|
| 3-300              | 0.05mm                                |


### **5.3 VALIDATION MEASUREMENT**

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| <b>Scan Volume</b> | <b>Expanded Uncertainty</b> |
|--------------------|-----------------------------|
| 1 g                | 16.19 %                     |
| 10 g               | 15.86 %                     |

## 6. CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance( $\Omega$ ) |                |
|-----------------|------------------|------------------|-----------------------|----------------|
|                 |                  |                  | Real                  | Imaginary part |
| 1900            | -23.74           | -20              | 45.66                 | -15.41         |

### 6.2 MECHANICAL DIMENSIONS

| Frequency MHz | L mm       |          | h mm       |          | d mm      |          |
|---------------|------------|----------|------------|----------|-----------|----------|
|               | required   | measured | required   | measured | required  | measured |
| 300           | 420.0 ±1 % |          | 250.0 ±1 % |          | 6.35 ±1 % |          |
| 450           | 290.0 ±1 % |          | 166.7 ±1 % |          | 6.35 ±1 % |          |
| 750           | 176.0 ±1 % |          | 100.0 ±1 % |          | 6.35 ±1 % |          |
| 835           | 161.0 ±1 % |          | 89.8 ±1 %  |          | 3.6 ±1 %  |          |
| 900           | 149.0 ±1 % |          | 83.3 ±1 %  |          | 3.6 ±1 %  |          |
| 1450          | 89.1 ±1 %  |          | 51.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1500          | 80.5 ±1 %  |          | 50.0 ±1 %  |          | 3.6 ±1 %  |          |
| 1640          | 79.0 ±1 %  |          | 45.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1750          | 75.2 ±1 %  |          | 42.9 ±1 %  |          | 3.6 ±1 %  |          |
| 1800          | 72.0 ±1 %  |          | 41.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1900          | 68.0 ±1 %  | pass     | 39.5 ±1 %  | pass     | 3.6 ±1 %  | pass     |
| 1950          | 66.3 ±1 %  |          | 38.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2000          | 64.5 ±1 %  |          | 37.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2100          | 61.0 ±1 %  |          | 35.7 ±1 %  |          | 3.6 ±1 %  |          |
| 2300          | 55.5 ±1 %  |          | 32.6 ±1 %  |          | 3.6 ±1 %  |          |
| 2450          | 51.5 ±1 %  |          | 30.4 ±1 %  |          | 3.6 ±1 %  |          |
| 2600          | 48.5 ±1 %  |          | 28.8 ±1 %  |          | 3.6 ±1 %  |          |
| 3000          | 41.5 ±1 %  |          | 25.0 ±1 %  |          | 3.6 ±1 %  |          |
| 3500          | 37.0 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |
| 3700          | 34.7 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |

## 7. VALIDATION MEASUREMENT

The IEEE Std. 1528, OET 65 Bulletin C and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity(r') |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|---------------------------|----------|-------------------------------|----------|
|                  | required                  | measured | required                      | measured |
| 300              | 45.3 $\pm$ 5 %            |          | 0.87 $\pm$ 5 %                |          |
| 450              | 45.3 $\pm$ 5 %            |          | 0.87 $\pm$ 5 %                |          |
| 750              | 41.9 $\pm$ 5 %            |          | 0.89 $\pm$ 5 %                |          |
| 835              | 41.5 $\pm$ 5 %            |          | 0.90 $\pm$ 5 %                |          |
| 900              | 41.5 $\pm$ 5 %            |          | 0.97 $\pm$ 5 %                |          |
| 1450             | 40.5 $\pm$ 5 %            |          | 1.20 $\pm$ 5 %                |          |
| 1500             | 40.4 $\pm$ 5 %            |          | 1.23 $\pm$ 5 %                |          |
| 1640             | 40.2 $\pm$ 5 %            |          | 1.31 $\pm$ 5 %                |          |
| 1750             | 40.1 $\pm$ 5 %            |          | 1.37 $\pm$ 5 %                |          |
| 1800             | 40.0 $\pm$ 5 %            |          | 1.40 $\pm$ 5 %                |          |
| 1900             | 40.0 $\pm$ 5 %            | pass     | 1.40 $\pm$ 5 %                | pass     |
| 1950             | 40.0 $\pm$ 5 %            |          | 1.40 $\pm$ 5 %                |          |
| 2000             | 40.0 $\pm$ 5 %            |          | 1.40 $\pm$ 5 %                |          |
| 2100             | 39.8 $\pm$ 5 %            |          | 1.49 $\pm$ 5 %                |          |
| 2300             | 39.5 $\pm$ 5 %            |          | 1.67 $\pm$ 5 %                |          |
| 2450             | 39.2 $\pm$ 5 %            |          | 1.80 $\pm$ 5 %                |          |
| 2600             | 39.0 $\pm$ 5 %            |          | 1.96 $\pm$ 5 %                |          |
| 3000             | 38.5 $\pm$ 5 %            |          | 2.40 $\pm$ 5 %                |          |
| 3500             | 37.9 $\pm$ 5 %            |          | 2.91 5 %                      |          |

### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID


The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |  |
|---|--|
| Software                                  | OPENSAR V4                                   |
| Phantom                                   | SN 40/14 SAM117                              |
| Probe                                     | SN 17/14 EP220                               |
| Liquid                                    | Head Liquid Values: eps' : 38.9 sigma : 1.38 |
| Distance between dipole center and liquid | 10.0 mm                                      |
| Area scan resolution                      | dx=8mm/dy=8mm                                |



|                      |                     |
|----------------------|---------------------|
| Zoon Scan Resolution | dx=8mm/dy=8m/dz=5mm |
| Frequency            | 1900 MHz            |
| Input power          | 20 dBm              |
| Liquid Temperature   | 22 °C               |
| Lab Temperature      | 22 °C               |
| Lab Humidity         | 45%                 |

| Frequency MHz | 1 g SAR (W/kg/W) |             | 10 g SAR (W/kg/W) |             |
|---------------|------------------|-------------|-------------------|-------------|
|               | required         | measured    | required          | measured    |
| 300           | 2.85             |             | 1.94              |             |
| 450           | 4.85             |             | 3.06              |             |
| 750           | 8.49             |             | 5.55              |             |
| 835           | 9.56             |             | 6.22              |             |
| 900           | 10.9             |             | 6.99              |             |
| 1450          | 29               |             | 16                |             |
| 1500          | 30.5             |             | 16.8              |             |
| 1640          | 34.2             |             | 18.4              |             |
| 1750          | 36.4             |             | 19.3              |             |
| 1800          | 38.4             |             | 20.1              |             |
| 1900          | 39.7             | 41.62(4.16) | 20.5              | 21.36(2.14) |
| 1950          | 40.5             |             | 20.9              |             |
| 2000          | 41.1             |             | 21.1              |             |
| 2100          | 43.6             |             | 21.9              |             |
| 2300          | 48.7             |             | 23.3              |             |
| 2450          | 52.4             |             | 24                |             |
| 2600          | 55.3             |             | 24.6              |             |
| 3000          | 63.8             |             | 25.7              |             |
| 3500          | 67.1             |             | 25                |             |




7.3 BODY LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity( $\epsilon'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|--------------------------------------|----------|-------------------------------|----------|
|                  | required                             | measured | required                      | measured |
| 300              | 58.2 $\pm$ 5 %                       |          | 0.92 $\pm$ 5 %                |          |
| 450              | 56.7 $\pm$ 5 %                       |          | 0.94 $\pm$ 5 %                |          |
| 750              | 55.5 $\pm$ 5 %                       |          | 0.96 $\pm$ 5 %                |          |
| 835              | 55.2 $\pm$ 5 %                       |          | 0.97 $\pm$ 5 %                |          |
| 900              | 55.0 $\pm$ 5 %                       |          | 1.05 $\pm$ 5 %                |          |
| 915              | 55.0 $\pm$ 5 %                       |          | 1.06 $\pm$ 5 %                |          |
| 1450             | 54.0 $\pm$ 5 %                       |          | 1.30 $\pm$ 5 %                |          |
| 1610             | 53.8 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1800             | 53.3 $\pm$ 5 %                       |          | 1.52 $\pm$ 5 %                |          |
| 1900             | 53.3 $\pm$ 5 %                       | Pass     | 1.52 $\pm$ 5 %                | Pass     |
| 2000             | 53.3 $\pm$ 5 %                       |          | 1.52 $\pm$ 5 %                |          |
| 2100             | 53.2 $\pm$ 5 %                       |          | 1.62 $\pm$ 5 %                |          |
| 2450             | 52.7 $\pm$ 5 %                       |          | 1.95 $\pm$ 5 %                |          |
| 2600             | 52.5 $\pm$ 5 %                       |          | 2.16 $\pm$ 5 %                |          |
| 3000             | 52.0 $\pm$ 5 %                       |          | 2.73 $\pm$ 5 %                |          |
| 3500             | 51.3 $\pm$ 5 %                       |          | 3.31 $\pm$ 5 %                |          |
| 5200             | 49.0 $\pm$ 5 %                       |          | 5.30 $\pm$ 5 %                |          |
| 5300             | 48.9 $\pm$ 5 %                       |          | 5.42 $\pm$ 5 %                |          |
| 5400             | 48.7 $\pm$ 5 %                       |          | 5.53 $\pm$ 5 %                |          |
| 5500             | 48.6 $\pm$ 5 %                       |          | 5.65 $\pm$ 5 %                |          |
| 5600             | 48.5 $\pm$ 5 %                       |          | 5.77 $\pm$ 5 %                |          |
| 5800             | 48.2 $\pm$ 5 %                       |          | 6.00 $\pm$ 5 %                |          |

7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|   |   |
|---|---|
| Software                                  | OPENSAR V4  |
| Phantom                                   | SN 40/14 SAM117   |
| Probe                                     | SN 17/14 EP220  |
| Liquid                                    | Body Liquid Values: $\epsilon_s'$ :54.0 $\sigma$ : 1.53 |
| Distance between dipole center and liquid | 10.0 mm   |
| Area scan resolution                      | dx=8mm/dy=8mm   |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                     |
| Frequency                                 | 1900 MHz  |
| Input power                               | 20 dBm  |
| Liquid Temperature                        | 22 °C   |
| Lab Temperature                           | 22 °C   |
| Lab Humidity                              | 45 %  |

| Frequency MHz | 1 g SAR (W/kg/W)<br>measured | 10 g SAR (W/kg/W)<br>measured |
|---------------|------------------------------|-------------------------------|
| 1900          | 41.39(4.14)                  | 21.65(2.17)                   |



## 8. LIST OF EQUIPMENT

| Equipment Summary Sheet         |                      |                    |                             |                             |
|---------------------------------|----------------------|--------------------|-----------------------------|-----------------------------|
| Equipment Description           | Manufacturer / Model | Identification No. | Current Calibration Date    | Next Calibration Date       |
| Flat Phantom                    | Satimo               | SN 40/14<br>SAM117 | Validated. No cal required. | Validated. No cal required. |
| COMOSAR Test Bench              | Version 3            | NA                 | Validated. No cal required. | Validated. No cal required. |
| Network Analyzer                | Agilent              | 8753E              | 2015/11/27                  | 2016/11/28                  |
| Reference Probe                 | Satimo               | EP220<br>SN 17/14  | 2015/10/01                  | 2016/10/01                  |
| Multimeter                      | Keithley             | MiltiMeter2000     | 2016/02/27                  | 2017/02/28                  |
| Signal Generator                | Agilent              | E4432B             | 2016/04/09                  | 2017/04/08                  |
| Power Meter                     | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Power Sensor                    | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Temperature and Humidity Sensor | JM                   | JM222              | 2016/06/06                  | 2017/06/05                  |

## **SAR Reference Dipole Calibration Report**

**FREQUENCY: 2000 MHZ**

**SERIAL NO.:SN 39/09 DIP1121**

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION TESTING LABORATORY**

**133 JIANG-JUN ROAD, JI ANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE  
NANJING 211153 PR CHINA**

Calibrated at INVENTEC

07/01/2016

|              | Name         | Function        | Date       | Signature    |
|--------------|--------------|-----------------|------------|--------------|
| Tested By:   | Zhang shuqin | Test Engineer   | 07/01/2016 | Zhang Shuqin |
| Reviewed By: | Ji jianlin   | Manager         | 07/01/2016 | Ji Jianlin   |
| Approved By: | Xu chunxiu   | Quality Manager | 07/01/2016 | Xu Chunxiu   |

| Issue | Date       | Modifications   |
|-------|------------|-----------------|
| A     | 07/01/2016 | Initial release |
|       |            |                 |

**TABLE OF CONTENTS**

|     |   |    |
|-----|---|----|
| 1.  | INTRODUCTION .....                            | 4  |
| 2.  | DEVICE UNDER TEST .....                       | 4  |
| 3.  | PRODUCT DESCRIPTION .....                     | 4  |
| 3.1 | GENERAL INFORMATION.....                      | 4  |
| 4.  | MEASUREMENT METHOD .....                      | 4  |
| 4.1 | RETURN LOSS REQUIREMENTS .....                | 4  |
| 4.2 | MECHANICAL REQUIREMENTS .....                 | 4  |
| 5.  | MEASUREMENT UNCERTAINTY .....                 | 5  |
| 5.1 | RETURN LOSS .....                             | 5  |
| 5.2 | DIMENSION MEASUREMENT.....                    | 5  |
| 5.3 | VALIDATION MEASUREMENT .....                  | 5  |
| 6.  | CALIBRATION MEASUREMENT RESULTS.....          | 6  |
| 6.1 | RETURN LOSS .....                             | 6  |
| 6.2 | MECHANICAL DIMENSIONS .....                   | 6  |
| 7.  | VALIDATION MEASUREMENT .....                  | 7  |
| 7.1 | HEAD LIQUID MEASUREMENT .....                 | 7  |
| 7.2 | SAR MEASUREMENT RESULT WITH HEAD LIQUID.....  | 7  |
| 7.3 | BODY LIQUID MEASUREMENT .....                 | 9  |
| 7.4 | SAR MEASUREMENT RESULT WITH BODY LIQUID ..... | 9  |
| 8.  | LIST OF EQUIPMENT .....                       | 11 |

## 1. INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2. DEVICE UNDER TEST

| Device Under Test              |                                   |
|--------------------------------|-----------------------------------|
| Device Type                    | COMOSAR 2000 MHz REFERENCE DIPOLE |
| Manufacturer                   | Satimo                            |
| Model                          | SID2000                           |
| Serial Number                  | SN 39/09 DIPI121                  |
| Product Condition (new / used) | Used                              |

## 3. PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

Satimo's COMOSAR Validation Dipoles are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards. The product is designed for use with the COMOSAR test bench only.



Figure 1 – Satimo COMOSAR Validation Dipole

## 4. MEASUREMENT METHOD

The IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE Std. 1528 and CEI/IEC 62209 standards specify the mechanical components and



dimensions of the validation dipoles, with the dimensions frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness.

## **5. MEASUREMENT UNCERTAINTY**

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### **5.1 RETURN LOSS**

The following uncertainties apply to the return loss measurement:

| <b>Frequency band</b> | <b>Expanded Uncertainty on Return Loss</b> |
|-----------------------|--|
| 400-6000MHz           | 0.1 dB                                     |

### **5.2 DIMENSION MEASUREMENT**

The following uncertainties apply to the dimension measurements:

| <b>Length (mm)</b> | <b>Expanded Uncertainty on Length</b> |
|--------------------|---------------------------------------|
| 3-300              | 0.05mm                                |


### **5.3 VALIDATION MEASUREMENT**

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| <b>Scan Volume</b> | <b>Expanded Uncertainty</b> |
|--------------------|-----------------------------|
| 1 g                | 16.19 %                     |
| 10 g               | 15.86 %                     |

## 6. CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance( $\Omega$ ) |                |
|-----------------|------------------|------------------|-----------------------|----------------|
|                 |                  |                  | Real                  | Imaginary part |
| 2000            | -25.52           | -20              | 37.19                 | -17.74         |

### 6.2 MECHANICAL DIMENSIONS

| Frequency MHz | L mm       |          | h mm       |          | d mm      |          |
|---------------|------------|----------|------------|----------|-----------|----------|
|               | required   | measured | required   | measured | required  | measured |
| 300           | 420.0 ±1 % |          | 250.0 ±1 % |          | 6.35 ±1 % |          |
| 450           | 290.0 ±1 % |          | 166.7 ±1 % |          | 6.35 ±1 % |          |
| 750           | 176.0 ±1 % |          | 100.0 ±1 % |          | 6.35 ±1 % |          |
| 835           | 161.0 ±1 % |          | 89.8 ±1 %  |          | 3.6 ±1 %  |          |
| 900           | 149.0 ±1 % |          | 83.3 ±1 %  |          | 3.6 ±1 %  |          |
| 1450          | 89.1 ±1 %  |          | 51.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1500          | 80.5 ±1 %  |          | 50.0 ±1 %  |          | 3.6 ±1 %  |          |
| 1640          | 79.0 ±1 %  |          | 45.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1750          | 75.2 ±1 %  |          | 42.9 ±1 %  |          | 3.6 ±1 %  |          |
| 1800          | 72.0 ±1 %  |          | 41.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1900          | 68.0 ±1 %  |          | 39.5 ±1 %  |          | 3.6 ±1 %  |          |
| 1950          | 66.3 ±1 %  |          | 38.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2000          | 64.5 ±1 %  | pass     | 37.5 ±1 %  | pass     | 3.6 ±1 %  | pass     |
| 2100          | 61.0 ±1 %  |          | 35.7 ±1 %  |          | 3.6 ±1 %  |          |
| 2300          | 55.5 ±1 %  |          | 32.6 ±1 %  |          | 3.6 ±1 %  |          |
| 2450          | 51.5 ±1 %  |          | 30.4 ±1 %  |          | 3.6 ±1 %  |          |
| 2600          | 48.5 ±1 %  |          | 28.8 ±1 %  |          | 3.6 ±1 %  |          |
| 3000          | 41.5 ±1 %  |          | 25.0 ±1 %  |          | 3.6 ±1 %  |          |
| 3500          | 37.0 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |
| 3700          | 34.7 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |

## 7. VALIDATION MEASUREMENT

The IEEE Std. 1528, OET 65 Bulletin C and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity( $r'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|-------------------------------|----------|-------------------------------|----------|
|                  | required                      | measured | required                      | measured |
| 300              | 45.3 $\pm$ 5 %                |          | 0.87 $\pm$ 5 %                |          |
| 450              | 45.3 $\pm$ 5 %                |          | 0.87 $\pm$ 5 %                |          |
| 750              | 41.9 $\pm$ 5 %                |          | 0.89 $\pm$ 5 %                |          |
| 835              | 41.5 $\pm$ 5 %                |          | 0.90 $\pm$ 5 %                |          |
| 900              | 41.5 $\pm$ 5 %                |          | 0.97 $\pm$ 5 %                |          |
| 1450             | 40.5 $\pm$ 5 %                |          | 1.20 $\pm$ 5 %                |          |
| 1500             | 40.4 $\pm$ 5 %                |          | 1.23 $\pm$ 5 %                |          |
| 1640             | 40.2 $\pm$ 5 %                |          | 1.31 $\pm$ 5 %                |          |
| 1750             | 40.1 $\pm$ 5 %                |          | 1.37 $\pm$ 5 %                |          |
| 1800             | 40.0 $\pm$ 5 %                |          | 1.40 $\pm$ 5 %                |          |
| 1900             | 40.0 $\pm$ 5 %                |          | 1.40 $\pm$ 5 %                |          |
| 1950             | 40.0 $\pm$ 5 %                |          | 1.40 $\pm$ 5 %                |          |
| 2000             | 40.0 $\pm$ 5 %                | pass     | 1.40 $\pm$ 5 %                | pass     |
| 2100             | 39.8 $\pm$ 5 %                |          | 1.49 $\pm$ 5 %                |          |
| 2300             | 39.5 $\pm$ 5 %                |          | 1.67 $\pm$ 5 %                |          |
| 2450             | 39.2 $\pm$ 5 %                |          | 1.80 $\pm$ 5 %                |          |
| 2600             | 39.0 $\pm$ 5 %                |          | 1.96 $\pm$ 5 %                |          |
| 3000             | 38.5 $\pm$ 5 %                |          | 2.40 $\pm$ 5 %                |          |
| 3500             | 37.9 $\pm$ 5 %                |          | 2.91 5 %                      |          |




### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |  |
|---|--|
| Software                                  | OPENSAR V4   |
| Phantom                                   | SN 40/14 SAM117  |
| Probe                                     | SN 17/14 EP220   |
| Liquid                                    | Head Liquid Values: $\epsilon_p'$ : 38.5 $\sigma$ : 1.43 |
| Distance between dipole center and liquid | 10.0 mm  |
| Area scan resolution                      | dx=8mm/dy=8mm  |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                      |

|                    |          |
|--------------------|----------|
| Frequency          | 2000 MHz |
| Input power        | 20 dBm   |
| Liquid Temperature | 22 °C    |
| Lab Temperature    | 22 °C    |
| Lab Humidity       | 45%      |

| Frequency<br>MHz | 1 g SAR (W/kg/W) |             | 10 g SAR (W/kg/W) |             |
|------------------|------------------|-------------|-------------------|-------------|
|                  | required         | measured    | required          | measured    |
| 300              | 2.85             |             | 1.94              |             |
| 450              | 4.85             |             | 3.06              |             |
| 750              | 8.49             |             | 5.55              |             |
| 835              | 9.56             |             | 6.22              |             |
| 900              | 10.9             |             | 6.99              |             |
| 1450             | 29               |             | 16                |             |
| 1500             | 30.5             |             | 16.8              |             |
| 1640             | 34.2             |             | 18.4              |             |
| 1750             | 36.4             |             | 19.3              |             |
| 1800             | 38.4             |             | 20.1              |             |
| 1900             | 39.7             |             | 20.5              |             |
| 1950             | 40.5             |             | 20.9              |             |
| 2000             | 41.1             | 42.32(4.23) | 21.1              | 21.00(2.10) |
| 2100             | 43.6             |             | 21.9              |             |
| 2300             | 48.7             |             | 23.3              |             |
| 2450             | 52.4             |             | 24                |             |
| 2600             | 55.3             |             | 24.6              |             |
| 3000             | 63.8             |             | 25.7              |             |
| 3500             | 67.1             |             | 25                |             |




7.3 BODY LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity(r') |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|---------------------------|----------|-------------------------------|----------|
|                  | required                  | measured | required                      | measured |
| 300              | 58.2 ±5 %                 |          | 0.92 ±5 %                     |          |
| 450              | 56.7 ±5 %                 |          | 0.94 ±5 %                     |          |
| 750              | 55.5 ±5 %                 |          | 0.96 ±5 %                     |          |
| 835              | 55.2 ±5 %                 |          | 0.97 ±5 %                     |          |
| 900              | 55.0 ±5 %                 |          | 1.05 ±5 %                     |          |
| 915              | 55.0 ±5 %                 |          | 1.06 ±5 %                     |          |
| 1450             | 54.0 ±5 %                 |          | 1.30 ±5 %                     |          |
| 1610             | 53.8 ±5 %                 |          | 1.40 ±5 %                     |          |
| 1800             | 53.3 ±5 %                 |          | 1.52 ±5 %                     |          |
| 1900             | 53.3 ±5 %                 |          | 1.52 ±5 %                     |          |
| 2000             | 53.3 ±5 %                 | Pass     | 1.52 ±5 %                     | Pass     |
| 2100             | 53.2 ±5 %                 |          | 1.62 ±5 %                     |          |
| 2450             | 52.7 ±5 %                 |          | 1.95 ±5 %                     |          |
| 2600             | 52.5 ±5 %                 |          | 2.16 ±5 %                     |          |
| 3000             | 52.0 ±5 %                 |          | 2.73 ±5 %                     |          |
| 3500             | 51.3 ±5 %                 |          | 3.31 ±5 %                     |          |
| 5200             | 49.0 ±5 %                 |          | 5.30 ±5 %                     |          |
| 5300             | 48.9 ±5 %                 |          | 5.42 ±5 %                     |          |
| 5400             | 48.7 ±5 %                 |          | 5.53 ±5 %                     |          |
| 5500             | 48.6 ±5 %                 |          | 5.65 ±5 %                     |          |
| 5600             | 48.5 ±5 %                 |          | 5.77 ±5 %                     |          |
| 5800             | 48.2 ±5 %                 |          | 6.00 ±5 %                     |          |

7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|   |   |
|---|---|
| Software                                  | OPENSAR V4                                  |
| Phantom                                   | SN 40/14 SAM117                             |
| Probe                                     | SN 17/14 EP220                              |
| Liquid                                    | Body Liquid Values: eps' :53.3 sigma : 1.57 |
| Distance between dipole center and liquid | 10.0 mm                                     |
| Area scan resolution                      | dx=8mm/dy=8mm                               |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                         |
| Frequency                                 | 2000 MHz                                    |
| Input power                               | 20 dBm                                      |
| Liquid Temperature                        | 22 °C                                       |
| Lab Temperature                           | 22 °C                                       |
| Lab Humidity                              | 45 %  |

| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | <b>measured</b>  | <b>measured</b>   |
| 2000          | 42.20(4.22)      | 21.64(2.16)       |



**8. LIST OF EQUIPMENT**

| Equipment Summary Sheet         |                      |                    |                             |                             |
|---------------------------------|----------------------|--------------------|-----------------------------|-----------------------------|
| Equipment Description           | Manufacturer / Model | Identification No. | Current Calibration Date    | Next Calibration Date       |
| Flat Phantom                    | Satimo               | SN 40/14<br>SAM117 | Validated. No cal required. | Validated. No cal required. |
| COMOSAR Test Bench              | Version 3            | NA                 | Validated. No cal required. | Validated. No cal required. |
| Network Analyzer                | Agilent              | 8753E              | 2015/11/27                  | 2016/11/28                  |
| Reference Probe                 | Satimo               | EP220<br>SN 17/14  | 2015/10/01                  | 2016/10/01                  |
| Multimeter                      | Keithley             | MultiMeter2000     | 2016/02/27                  | 2017/02/28                  |
| Signal Generator                | Agilent              | E4432B             | 2016/04/09                  | 2017/04/08                  |
| Power Meter                     | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Power Sensor                    | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Temperature and Humidity Sensor | JM                   | JM222              | 2016/06/06                  | 2017/06/05                  |

## **SAR Reference Dipole Calibration Report**

**FREQUENCY: 2450 MHZ**

**SERIAL NO.:SN 39/09 DIPJ122**

**INVENTEC APPLIANCES (JIANGNING)  
CORPORATION TESTING LABORATORY**

**133 JIANG-JUN ROAD, JI ANGNING ECONOMIC AND  
TECHNOLOGICAL DEVELOPMENT ZONE  
NANJING 211153 PR CHINA**

Calibrated at INVENTEC

07/01/2016



|              | Name         | Function        | Date       | Signature    |
|--------------|--------------|-----------------|------------|--------------|
| Tested By:   | Zhang shuqin | Test Engineer   | 07/01/2016 | Zhang Shuqin |
| Reviewed By: | Ji jianlin   | Manager         | 07/01/2016 | Ji Jianlin   |
| Approved By: | Xu chunxiu   | Quality Manager | 07/01/2016 | Xu Chunxiu   |

| Issue | Date       | Modifications   |
|-------|------------|-----------------|
| A     | 07/01/2016 | Initial release |
|       |            |                 |

**TABLE OF CONTENTS**

|     |   |    |
|-----|---|----|
| 1.  | INTRODUCTION .....                            | 4  |
| 2.  | DEVICE UNDER TEST .....                       | 4  |
| 3.  | PRODUCT DESCRIPTION .....                     | 4  |
| 3.1 | GENERAL INFORMATION.....                      | 4  |
| 4.  | MEASUREMENT METHOD .....                      | 4  |
| 4.1 | RETURN LOSS REQUIREMENTS .....                | 4  |
| 4.2 | MECHANICAL REQUIREMENTS .....                 | 4  |
| 5.  | MEASUREMENT UNCERTAINTY .....                 | 5  |
| 5.1 | RETURN LOSS .....                             | 5  |
| 5.2 | DIMENSION MEASUREMENT .....                   | 5  |
| 5.3 | VALIDATION MEASUREMENT .....                  | 5  |
| 6.  | CALIBRATION MEASUREMENT RESULTS .....         | 6  |
| 6.1 | RETURN LOSS .....                             | 6  |
| 6.2 | MECHANICAL DIMENSIONS .....                   | 6  |
| 7.  | VALIDATION MEASUREMENT .....                  | 7  |
| 7.1 | HEAD LIQUID MEASUREMENT .....                 | 7  |
| 7.2 | SAR MEASUREMENT RESULT WITH HEAD LIQUID.....  | 7  |
| 7.3 | BODY LIQUID MEASUREMENT .....                 | 9  |
| 7.4 | SAR MEASUREMENT RESULT WITH BODY LIQUID ..... | 9  |
| 8.  | LIST OF EQUIPMENT .....                       | 11 |

## 1. INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards for reference dipoles used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2. DEVICE UNDER TEST

| Device Under Test              |                                   |
|--------------------------------|-----------------------------------|
| Device Type                    | COMOSAR 2450 MHz REFERENCE DIPOLE |
| Manufacturer                   | Satimo                            |
| Model                          | SID2450                           |
| Serial Number                  | SN 39/09 DIPJ122                  |
| Product Condition (new / used) | Used                              |

## 3. PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

Satimo's COMOSAR Validation Dipoles are built in accordance to the IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards. The product is designed for use with the COMOSAR test bench only.



Figure 1 – Satimo COMOSAR Validation Dipole

## 4. MEASUREMENT METHOD

The IEEE 1528, OET 65 Bulletin C and CEI/IEC 62209 standards provide requirements for reference dipoles used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The dipole used for SAR system validation measurements and checks must have a return loss of -20 dB or better. The return loss measurement shall be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE Std. 1528 and CEI/IEC 62209 standards specify the mechanical components and

dimensions of the validation dipoles, with the dimensions frequency and phantom shell thickness dependent. The COMOSAR test bench employs a 2 mm phantom shell thickness therefore the dipoles sold for use with the COMOSAR test bench comply with the requirements set forth for a 2 mm phantom shell thickness.

## 5. MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.1 dB                              |

### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 3-300       | 0.05mm                         |


### 5.3 VALIDATION MEASUREMENT

The guidelines outlined in the IEEE 1528, OET 65 Bulletin C, CENELEC EN50361 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 16.19 %              |
| 10 g        | 15.86 %              |

## 6. CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance( $\Omega$ ) |                |
|-----------------|------------------|------------------|-----------------------|----------------|
|                 |                  |                  | Real                  | Imaginary part |
| 2450            | -23.20           | -20              | 26.20                 | 8.63           |

### 6.2 MECHANICAL DIMENSIONS

| Frequency MHz | L mm       |          | h mm       |          | d mm      |          |
|---------------|------------|----------|------------|----------|-----------|----------|
|               | required   | measured | required   | measured | required  | measured |
| 300           | 420.0 ±1 % |          | 250.0 ±1 % |          | 6.35 ±1 % |          |
| 450           | 290.0 ±1 % |          | 166.7 ±1 % |          | 6.35 ±1 % |          |
| 750           | 176.0 ±1 % |          | 100.0 ±1 % |          | 6.35 ±1 % |          |
| 835           | 161.0 ±1 % |          | 89.8 ±1 %  |          | 3.6 ±1 %  |          |
| 900           | 149.0 ±1 % |          | 83.3 ±1 %  |          | 3.6 ±1 %  |          |
| 1450          | 89.1 ±1 %  |          | 51.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1500          | 80.5 ±1 %  |          | 50.0 ±1 %  |          | 3.6 ±1 %  |          |
| 1640          | 79.0 ±1 %  |          | 45.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1750          | 75.2 ±1 %  |          | 42.9 ±1 %  |          | 3.6 ±1 %  |          |
| 1800          | 72.0 ±1 %  |          | 41.7 ±1 %  |          | 3.6 ±1 %  |          |
| 1900          | 68.0 ±1 %  |          | 39.5 ±1 %  |          | 3.6 ±1 %  |          |
| 1950          | 66.3 ±1 %  |          | 38.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2000          | 64.5 ±1 %  |          | 37.5 ±1 %  |          | 3.6 ±1 %  |          |
| 2100          | 61.0 ±1 %  |          | 35.7 ±1 %  |          | 3.6 ±1 %  |          |
| 2300          | 55.5 ±1 %  |          | 32.6 ±1 %  |          | 3.6 ±1 %  |          |
| 2450          | 51.5 ±1 %  | pass     | 30.4 ±1 %  | pass     | 3.6 ±1 %  | pass     |
| 2600          | 48.5 ±1 %  |          | 28.8 ±1 %  |          | 3.6 ±1 %  |          |
| 3000          | 41.5 ±1 %  |          | 25.0 ±1 %  |          | 3.6 ±1 %  |          |
| 3500          | 37.0 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |
| 3700          | 34.7 ±1 %  |          | 26.4 ±1 %  |          | 3.6 ±1 %  |          |

## 7. VALIDATION MEASUREMENT

The IEEE Std. 1528, OET 65 Bulletin C and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity( $\epsilon'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|--------------------------------------|----------|-------------------------------|----------|
|                  | required                             | measured | required                      | measured |
| 300              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 450              | 45.3 $\pm$ 5 %                       |          | 0.87 $\pm$ 5 %                |          |
| 750              | 41.9 $\pm$ 5 %                       |          | 0.89 $\pm$ 5 %                |          |
| 835              | 41.5 $\pm$ 5 %                       |          | 0.90 $\pm$ 5 %                |          |
| 900              | 41.5 $\pm$ 5 %                       |          | 0.97 $\pm$ 5 %                |          |
| 1450             | 40.5 $\pm$ 5 %                       |          | 1.20 $\pm$ 5 %                |          |
| 1500             | 40.4 $\pm$ 5 %                       |          | 1.23 $\pm$ 5 %                |          |
| 1640             | 40.2 $\pm$ 5 %                       |          | 1.31 $\pm$ 5 %                |          |
| 1750             | 40.1 $\pm$ 5 %                       |          | 1.37 $\pm$ 5 %                |          |
| 1800             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1900             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 1950             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2000             | 40.0 $\pm$ 5 %                       |          | 1.40 $\pm$ 5 %                |          |
| 2100             | 39.8 $\pm$ 5 %                       |          | 1.49 $\pm$ 5 %                |          |
| 2300             | 39.5 $\pm$ 5 %                       |          | 1.67 $\pm$ 5 %                |          |
| 2450             | 39.2 $\pm$ 5 %                       | pass     | 1.80 $\pm$ 5 %                | pass     |
| 2600             | 39.0 $\pm$ 5 %                       |          | 1.96 $\pm$ 5 %                |          |
| 3000             | 38.5 $\pm$ 5 %                       |          | 2.40 $\pm$ 5 %                |          |
| 3500             | 37.9 $\pm$ 5 %                       |          | 2.91 5 %                      |          |


### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

|   |  |
|---|--|
| Software                                  | OPENSAR V4   |
| Phantom                                   | SN 40/14 SAM117  |
| Probe                                     | SN 17/14 EP220   |
| Liquid                                    | Head Liquid Values: $\epsilon_s'$ : 38.0 $\sigma$ : 1.84 |
| Distance between dipole center and liquid | 10.0 mm  |
| Area scan resolution                      | dx=8mm/dy=8mm  |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                                      |

|                    |          |
|--------------------|----------|
| Frequency          | 2450 MHz |
| Input power        | 20 dBm   |
| Liquid Temperature | 22 °C    |
| Lab Temperature    | 22 °C    |
| Lab Humidity       | 45%      |

| Frequency<br>MHz | 1 g SAR (W/kg/W) |             | 10 g SAR (W/kg/W) |             |
|------------------|------------------|-------------|-------------------|-------------|
|                  | required         | measured    | required          | measured    |
| 300              | 2.85             |             | 1.94              |             |
| 450              | 4.85             |             | 3.06              |             |
| 750              | 8.49             |             | 5.55              |             |
| 835              | 9.56             |             | 6.22              |             |
| 900              | 10.9             |             | 6.99              |             |
| 1450             | 29               |             | 16                |             |
| 1500             | 30.5             |             | 16.8              |             |
| 1640             | 34.2             |             | 18.4              |             |
| 1750             | 36.4             |             | 19.3              |             |
| 1800             | 38.4             |             | 20.1              |             |
| 1900             | 39.7             |             | 20.5              |             |
| 1950             | 40.5             |             | 20.9              |             |
| 2000             | 41.1             |             | 21.1              |             |
| 2100             | 43.6             |             | 21.9              |             |
| 2300             | 48.7             |             | 23.3              |             |
| 2450             | 52.4             | 55.19(5.52) | 24                | 24.96(2.50) |
| 2600             | 55.3             |             | 24.6              |             |
| 3000             | 63.8             |             | 25.7              |             |
| 3500             | 67.1             |             | 25                |             |



7.3 BODY LIQUID MEASUREMENT


| Frequency<br>MHz | Relative permittivity(r') |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|---------------------------|----------|-------------------------------|----------|
|                  | required                  | measured | required                      | measured |
| 300              | 58.2 $\pm$ 5 %            |          | 0.92 $\pm$ 5 %                |          |
| 450              | 56.7 $\pm$ 5 %            |          | 0.94 $\pm$ 5 %                |          |
| 750              | 55.5 $\pm$ 5 %            |          | 0.96 $\pm$ 5 %                |          |
| 835              | 55.2 $\pm$ 5 %            |          | 0.97 $\pm$ 5 %                |          |
| 900              | 55.0 $\pm$ 5 %            |          | 1.05 $\pm$ 5 %                |          |
| 915              | 55.0 $\pm$ 5 %            |          | 1.06 $\pm$ 5 %                |          |
| 1450             | 54.0 $\pm$ 5 %            |          | 1.30 $\pm$ 5 %                |          |
| 1610             | 53.8 $\pm$ 5 %            |          | 1.40 $\pm$ 5 %                |          |
| 1800             | 53.3 $\pm$ 5 %            |          | 1.52 $\pm$ 5 %                |          |
| 1900             | 53.3 $\pm$ 5 %            |          | 1.52 $\pm$ 5 %                |          |
| 2000             | 53.3 $\pm$ 5 %            |          | 1.52 $\pm$ 5 %                |          |
| 2100             | 53.2 $\pm$ 5 %            |          | 1.62 $\pm$ 5 %                |          |
| 2450             | 52.7 $\pm$ 5 %            | Pass     | 1.95 $\pm$ 5 %                | Pass     |
| 2600             | 52.5 $\pm$ 5 %            |          | 2.16 $\pm$ 5 %                |          |
| 3000             | 52.0 $\pm$ 5 %            |          | 2.73 $\pm$ 5 %                |          |
| 3500             | 51.3 $\pm$ 5 %            |          | 3.31 $\pm$ 5 %                |          |
| 5200             | 49.0 $\pm$ 5 %            |          | 5.30 $\pm$ 5 %                |          |
| 5300             | 48.9 $\pm$ 5 %            |          | 5.42 $\pm$ 5 %                |          |
| 5400             | 48.7 $\pm$ 5 %            |          | 5.53 $\pm$ 5 %                |          |
| 5500             | 48.6 $\pm$ 5 %            |          | 5.65 $\pm$ 5 %                |          |
| 5600             | 48.5 $\pm$ 5 %            |          | 5.77 $\pm$ 5 %                |          |
| 5800             | 48.2 $\pm$ 5 %            |          | 6.00 $\pm$ 5 %                |          |

7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|   |   |
|---|---|
| Software                                  | OPENSAR V4                                  |
| Phantom                                   | SN 40/14 SAM117                             |
| Probe                                     | SN 17/14 EP220                              |
| Liquid                                    | Body Liquid Values: eps' :51.7 sigma : 1.93 |
| Distance between dipole center and liquid | 10.0 mm                                     |
| Area scan resolution                      | dx=8mm/dy=8mm                               |
| Zoon Scan Resolution                      | dx=8mm/dy=8m/dz=5mm                         |
| Frequency                                 | 2450 MHz                                    |
| Input power                               | 20 dBm                                      |
| Liquid Temperature                        | 22 °C                                       |
| Lab Temperature                           | 22 °C                                       |
| Lab Humidity                              | 45 %  |



| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | <b>measured</b>  | <b>measured</b>   |
| 2450          | 52.85(5.29)      | 24.78(2.48)       |



**8. LIST OF EQUIPMENT**

| Equipment Summary Sheet         |                      |                    |                             |                             |
|---------------------------------|----------------------|--------------------|-----------------------------|-----------------------------|
| Equipment Description           | Manufacturer / Model | Identification No. | Current Calibration Date    | Next Calibration Date       |
| Flat Phantom                    | Satimo               | SN 40/14<br>SAM117 | Validated. No cal required. | Validated. No cal required. |
| COMOSAR Test Bench              | Version 3            | NA                 | Validated. No cal required. | Validated. No cal required. |
| Network Analyzer                | Agilent              | 8753E              | 2015/11/27                  | 2016/11/28                  |
| Reference Probe                 | Satimo               | EP220<br>SN 17/14  | 2015/10/01                  | 2016/10/01                  |
| Multimeter                      | Keithley             | MiltiMeter2000     | 2016/02/27                  | 2017/02/28                  |
| Signal Generator                | Agilent              | E4432B             | 2016/04/09                  | 2017/04/08                  |
| Power Meter                     | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Power Sensor                    | R & S                | NRP-Z23            | 2016/06/17                  | 2017/06/16                  |
| Temperature and Humidity Sensor | JM                   | JM222              | 2016/06/06                  | 2017/06/05                  |