



20150716 DIPOLE2450

RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



Frequency (MHz)	Return Loss (dB)	Requirement (dB)	Impedance(Ω)	
			Real	Imaginary part
2450	-24.55	-20	24.50	7.4


RETURN LOSS AND IMPEDANCE IN BODY LIQUID



Frequency (MHz)	Return Loss (dB)	Requirement (dB)	Impedance(Ω)	
			Real	Imaginary part
2450	-27.41	-20	25.51	6.2


20160701 DIPOLE2450

RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



Frequency (MHz)	Return Loss (dB)	Requirement (dB)	Impedance(Ω)	
			Real	Imaginary part
2450	-23.60	-20	26.20	8.6

RETURN LOSS AND IMPEDANCE IN BODY LIQUID



Frequency (MHz)	Return Loss (dB)	Requirement (dB)	Impedance(Ω)	
			Real	Imaginary part
2450	-24.24	-20	24.32	10.2