

To: "Timothy R. Johnson" <TRJ@adelphia.net>
Subject: RE: Questions 2,4 and 6

Hi Tim,

Please find the attached three files for Q2,4 and 6.
I am still waiting for Q 5 from BroMax If you can get from them?

Liming

>
> Question 2)
> I attach our performance statement for the dipoles supplied
> (IXS0212.doc).
>
> Question 4)
> I attach (SN0082cal2450.doc) a document setting out calibration
> factors to be used at 2450MHz until such time as this probe receives its
> next calibration.
>
> Question 6)
> The pulsed performance of Indexsar probes is covered in the original
> calibration document. This indicated, for probe 0082, that pulsed signals
> would be correctly measured using the calibration supplied for CW signals.
>
> <<SN0082cal2450.doc>> <<SN0016_1900.doc>> <<IXS0212 new.doc>>



[SN0082cal2450.doc](#)



[SN0016_1900.doc](#)



[IXS0212 new.doc](#)