



May 1, 2017

FCC Request for Additional information

Office of Engineering and Technology

FCC Filing : O8PCAPDH-2

Attn: Dusmantha Tennakoon

Dear Mr. Tennakoon,

In response to your question, please find an updated test report and test set up photos.

Question 1: Depth of the tissue liquid- 76cm deep .

We trust that the updated reports provide the Commission with the additional confirmation that the simulation fluid was prepared to the appropriate depth.

To avoid confusion in future filings we would like to receive from the Commission confirmation that our test set up is acceptable.

Thank you in advance.

Should you have further questions, please feel free to contact me directly at (408) 992-1632 or
janet.gavidia@covidien.com

Sincerely,

A handwritten signature in black ink, appearing to read "Janet H. Gavidia".

**Janet Gavidia
Medtronic**

Program Manager, Regulatory Affairs | RF
On Behalf of Given Imaging
Email:janet.gavidia@covidien.com