

3/21/2024

Federal Communications Commission  
7435 Oakland Mills Road  
Columbia, MD 21046


To whom it may concern:

Medtronic Minimed hereby requests a Permissive Change Class II of the radio equipment certification for Simpler certification number, FCC ID: OH2BLEGST1. This product differs from the originally approved product in the following manner:

- (a) Associated digital circuitry. –
  - Addition of two ESD diodes in RF circuit
  - Modified matching circuit components
- (b) Functional capabilities. – None
- (c) Antenna Characteristics. – None
- (d) Cosmetic differences – None
- (e) Other(?)
  - Addition of battery protection fuse

I do attest that the Field strength and/or RF Output readings remained the same or are lower than the originally approved product and that they did not increase therefore qualifying this product for this application type.

Sincerely,

A handwritten signature in black ink, appearing to read "Jonathan Tabalujan". The signature is stylized and written in a cursive-like font.

Jonathan Tabalujan  
Sr. Electrical Engineer - RF  
Medtronic Minimed

Note: Signature is subject to FCC's authorized individual policy KDB 852134; signature must be the named grantee or an authorized delegate.