

RADIO FREQUENCY EXPOSURE REPORT

REVISED TEST REPORT TO 102993-10

Device: NGP BLE Pump
Model: M994838A001*
(*See Appendix C for Manufacturer Declaration)

Report No.: 102993-10A

Date of issue: November 25, 2019

PREPARED FOR:

Medtronic MiniMed
18000 Devonshire Street, SS-32
Northridge, CA 91325

PREPARED BY:

M. Harrison
CKC Laboratories, Inc.
5046 Sierra Pines Drive
Mariposa, CA 95338

The test data contained in this report documents the observed testing parameters pertaining to and are relevant for only the sample equipment tested in the agreed upon operational mode(s) and configuration(s) as identified herein. Compliance assessment remains the client's responsibility. This report may not be used to claim product endorsement by A2LA or any government agencies. This test report has been authorized for release under quality control from CKC Laboratories, Inc.



Steve Behm
Director of Quality Assurance & Engineering Services
CKC Laboratories, Inc.

Revision History:

Revision A: Corrected the calculations for Power dBm and EIRP dBm in US and IC Exemption tables.

Purpose:

To demonstrate compliance with United States, Canada, Australia and/or European Union RF Exposure requirements for Portable equipment (devices used ≤ 20 cm from the body) with power output below exemption levels or Mobile equipment (devices used > 20 cm from the body) where Maximum Permissible Exposure (MPE) Calculations apply.

Device and Antenna Operating Configuration:

Device operating at maximum output power with continuous transmission of modulated data.

Test Procedure:

This equipment is evaluated in accordance with the guidelines set forth in KDB 447498 & ANSI C95.1 for the US, Health Canada Safety Code 6 & RSS 102 for Canada, ARPANSA RPS3 for AU and EN 62479 or EN 62311 for EU.

Other Considerations:

Report considers stand-alone configurations only. RF Exposure limits are calculated at the mid-point of each operating band.

Referenced Test Reports:

The following test reports were referenced in conjunction with this assessment:

102993-2
102993-3
102993-5A
102993-6A

RF Exposure Exemption Calculations

US Exemption

Power Reported is:		<input checked="" type="checkbox"/> Peak <input type="checkbox"/> Average				
Limit Used is:		<input checked="" type="checkbox"/> General Population <input type="checkbox"/> Occupational Exposure				
Operating Band MHz	Power dBm	Ant Type/Gain dBi	EIRP dBm	Sep. Dist mm	Exemption Limit ¹	Comments
2402-2480	-3.92	Chip / -0.5 dBi	-4.42	5	3	Highest power of conducted or EIRP used for calculations Calculated Ratio = 0.0

IC Exemption

Power Reported is:		<input checked="" type="checkbox"/> Peak <input type="checkbox"/> Average				
Limit Used is:		<input checked="" type="checkbox"/> General Population <input type="checkbox"/> Occupational Exposure				
Operating Band MHz	Power dBm	Ant Type/Gain dBi	EIRP dBm	Sep Dist mm	Exemption Limit	Comments
2402-2480	-3.92 = 0.4 mW	Chip / -0.5 dBi	-4.42	5	4mW	Highest power of conducted or EIRP used for calculations

AU Exemption

Power Reported is:		<input checked="" type="checkbox"/> Peak <input type="checkbox"/> Average				
Limit Used is:		<input checked="" type="checkbox"/> General Population <input type="checkbox"/> Occupational Exposure				
Operating Band MHz	Power dBm	Ant Type/Gain dBi	EIRP dBm	Exemption Limit	Comments	
2402-2480	-1.67 = 0.7mW	Chip / -0.5 dBi	-2.17	20 mW		

EU Exemption

Power Reported is:		<input checked="" type="checkbox"/> Peak <input type="checkbox"/> Average				
Limit Used is:		<input checked="" type="checkbox"/> General Population <input type="checkbox"/> Occupational Exposure				
Operating Band MHz	Power dBm	Ant Type/Gain dBi	EIRP dBm	Exemption Limit	Comments	
2402-2480	-1.67 = 0.7mW	Chip / -0.5 dBi	-2.17	20 mW		

¹ See Appendix A for calculation used. Assumes R=3 unless otherwise indicated.

Summary:

Exemptions:

Equipment demonstrating compliance by portable RF exposure exemption have been evaluated, without further testing, for use under either portable or mobile RF exposure configurations. Additional configurations including collocation or simultaneous transmission with other transmitters (including necessary separation distances) are subject to further assessment. Certain exemption criteria may define specific separation distances. It is assumed that the manufacturer shall design the equipment such that these minimum separation distance is met.

MPE Calculation Results:

Equipment demonstrating compliance with MPE calculations have been evaluated, without further testing, for use under mobile RF exposure configurations as identified herein. Additional configurations including collocation or simultaneous transmission with other transmitters (including necessary separation distances) are subject to further assessment. It is assumed that the manufacturer shall design the equipment such that the minimum separation distance of 20cm (or greater, as listed above) is met or that the manufacturer provides a protection guide (e.g. installation instructions) to the end user such that the antenna(s) may be installed in accordance with the manufacturer's instructions in such a manor to maintain the minimum separation distance.

General Comments:

The absorption and distribution of Electromagnetic energy in the body is a very complex phenomena that depends on the mass, shape and physiological condition of the body; the orientation of the body with respect to the fields; and, the electrical properties of the body and the environment. Variables that may play a substantial role in possible biological effects are those that characterize the environment (including but not limited to: ambient temperature, air velocity, relative humidity and body insulation); and those that characterize the individual (including but not limited to: age, gender, activity level and existing debilitation or disease). Because innumerable factors may interact to determine specific biological effects of exposure to electromagnetic fields, any protection guide should consider both intended and unintended operational environments and provide guidance for installation and use of the product such that proper separation distances can be maintained. (ANSI C95.1)

APPENDIX A - RF Exposure Limits

United States Compliance Requirements (1.1310):

RF Exposure Evaluation Limits Occupational / Controlled Exposure

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (mW/cm ²)	Averaging Time (minutes)
0.3-3.0	614	1.63	*(100)	6
3.0-30	1842 / f _{MHz}	4.89 / f _{MHz}	*(900 / f _{MHz} ²)	6
30-300	61.4	0.163	1	6
300-1500	---	---	f _{MHz} / 300	6
1500-100,000	---	---	5.0	6

RF Exposure Evaluation Limits General Population / Uncontrolled Exposure

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (mW/cm ²)	Averaging Time (minutes)
0.3-1.34	614	1.63	*(100)	30
1.34-30	824 / f _{MHz}	2.19 / f _{MHz}	*(180 / f _{MHz} ²)	30
30-300	27.5	0.073	0.2	30
300-1500	---	---	f _{MHz} / 1500	30
1500-100,000	---	---	1.0	30

* Plane wave equivalent power density

Limit is calculated based on the mid-band frequency used in the operating frequency range.

Exemption Limits for Stand-Alone SAR Evaluation:

In accordance with KDB 447498 D01

Max Output Power at Exemption Limit (mW)		
Frequency (MHz)	d ≤ 50mm	50mm < d ≤ 20cm
<100	$\frac{1}{2} \cdot \left(\frac{R \cdot 50}{\sqrt{0.1}} \right) \cdot \left(1 + \text{LOG} \left(\frac{100}{f_{\text{MHz}}} \right) \right)$	$\left(\frac{R \cdot 50}{\sqrt{0.1}} + (d - 50) \frac{100}{150} \right) \cdot \left(1 + \text{LOG} \left(\frac{100}{f_{\text{MHz}}} \right) \right)$
100-1500	$\left(\frac{R \cdot d}{\sqrt{f_{\text{GHz}}}} \right)$	$\left(\frac{R \cdot 50}{\sqrt{f_{\text{GHz}}}} + (d - 50) \frac{f_{\text{MHz}}}{150} \right)$
1500-6000		$\left(\frac{R \cdot 50}{\sqrt{f_{\text{GHz}}}} + (d - 50) \cdot 10 \right)$

R is the allowed ratio: 3 for 1-g SAR and 7.5 for 10-g extremity SAR.

d is distance in mm, rounded to the nearest mm.

Canadian Compliance Requirements (RSS-102):

***RF Exposure Evaluation Limits
Occupational / Controlled Exposure:***

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m²)	Averaging Time (minutes)
0.003-10	170	180	--	Instantaneous
0.1-10	--	1.6 / f _{MHz}	--	6
1.29-10	193 / f _{MHz} ^{0.5}	--	--	6
10-20	61.4	0.163	10	6
20-48	129.8 / f ^{0.5}	0.3444 / f _{MHz} ^{0.25}	44.72 / f _{MHz} ^{0.5}	6
48-100	49.33	0.1309	6.455	6
100-6000	15.60 f _{MHz} ^{0.25}	0.04138 f _{MHz} ^{0.25}	0.6455 f _{MHz} ^{0.5}	6
6000-15000	137	0.364	50	6
15000-150,000	137	0.364	50	616000 / f _{MHz} ^{1.2}
150,000-300,000	0.354 f _{MHz} ^{0.5}	9.40x10 ⁻⁴ f _{MHz} ^{0.5}	3.33x10 ⁻⁴ f _{MHz}	616000 / f _{MHz} ^{1.2}

***RF Exposure Evaluation Limits
General Population / Uncontrolled Exposure***

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m²)	Averaging Time (minutes)
0.003-10	83	90	--	Instantaneous
0.1-10	--	0.73 / f	--	6
1.1-10	87 / f _{MHz} ^{0.5}	--	--	6
10-20	27.46	0.0728	2	6
20-48	58.07 / f _{MHz} ^{0.25}	0.1540 / f _{MHz} ^{0.25}	8.944 / f _{MHz} ^{0.5}	6
48-300	22.06	0.05852	1.291	6
300-6000	3.142 f _{MHz} ^{0.3417}	0.008335 f _{MHz} ^{0.3417}	0.02619 f _{MHz} ^{0.6834}	6
6000-15000	61.4	0.163	10	6
15000-150,000	61.4	0.163	10	616000 / f _{MHz} ^{1.2}
150,000-300,000	0.158 f _{MHz} ^{0.5}	4.21x10 ⁻⁴ f _{MHz} ^{0.5}	6.67x10 ⁻⁵ f _{MHz}	616000 / f _{MHz} ^{1.2}


Exemption Limits for Stand-Alone SAR Evaluation:

Exemption Limits (mW) at Separation Distance (mm)										
Freq(MHz)	≤5	10	15	20	25	30	35	40	45	≥50
≤300	71	101	132	162	193	223	254	284	315	345
450	52	70	88	106	123	141	159	177	195	213
835	17	30	42	55	67	80	92	105	117	130
1900	7	10	18	34	60	99	153	225	316	431
2450	4	7	15	30	52	83	123	173	235	309
3500	2	6	16	32	55	86	124	170	225	290
5800	1	6	15	27	41	56	71	85	97	106

Exemption Limits for Stand-Alone Mobile Routine Evaluation:

Frequency (MHz)	RF Exposure Exemption Limit (mW)
<20	1000
20-48	$4490 / f_{\text{MHz}}^{0.5}$
48-300	600
300-6000	$13.1 f_{\text{MHz}}^{0.6834}$
≥6000	5000

General² Comparison of FCC and IC Exemption Limits for Stand Alone SAR Evaluation



² Non-Exhaustive

Australian Radiation Protection and Nuclear Safety Agency Requirements (ARPANSA):

**RF Exposure Evaluation Limits
Occupational / Controlled Exposure:**

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m ²)	Averaging Time (minutes)
0.1 – 1.0	614	1.63 / f _{MHz}	---	6
1.0-10	614 / f _{MHz}	1.63 / f _{MHz}	1000 / f _{MHz} ²	6
10-400	61.4	0.163	10	6
400-2000	3.07 * f _{MHz} ^{0.5}	0.00814 * f _{MHz} ^{0.5}	f _{MHz} / 40	6
2000-10,000	137	0.36	50	6
10,000 – 300,000	137	0.36	50	9.6x10 ⁴ / f _{MHz} ^{1.05}

**RF Exposure Evaluation Limits
General Population / Uncontrolled Exposure**

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m ²)	Averaging Time (minutes)
0.10-0.15	86.8	4.86	---	6
0.150-1.0	86.8	0.729 / f _{MHz}	--	6
1.0-10	86.8 / f _{MHz} ^{0.5}	0.729 / f _{MHz}	---	6
10-400	27.4	0.0729	2	6
400-2000	1.37 f _{MHz} ^{0.5}	0.00364* f _{MHz} ^{0.5}	f _{MHz} / 200	6
2000-10,000	61.4	0.163	10	6
10,000 – 300,000	61.4	0.163	10	9.6x10 ⁴ / f _{MHz} ^{1.05}

*Power density limit applicable >100MHz

Exemption Limits for Stand-Alone Evaluation:

Occupational Exposure: 100mW

Portable - General Public: 20mW

Mobile – General Public: Separation distance >20cm and power < 20mW or according to IEC 62479.

Or according to ARPANSA RPS3 Table S1

European Union Compliance Requirements (ICNIRP):

**RF Exposure Evaluation Limits
Occupational / Controlled Exposure:**

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m ²)	Averaging Time (minutes)
0.00082-0.065	610	24.4	---	6
0.065-1.0	610	1.6/f	---	6
1.0-10	610 / f _{MHz}	1.6/f	---	6
10-400	61	0.16	10	6
400-2000	3.0 * f _{MHz} ^{0.5}	0.008 * f _{MHz} ^{0.5}	f _{MHz} / 40	6
2000-300,000	137	0.36	50	6

**RF Exposure Evaluation Limits
General Population / Uncontrolled Exposure**

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (W/m ²)	Averaging Time (minutes)
0.003-0.150	87	5.0	---	6
0.150-1.0	87	0.73 / f _{MHz}	--	6
1.0-10	87 / f _{MHz} ^{0.5}	0.73 / f _{MHz}	---	6
10-400	28	0.073	2	6
400-2000	1.375 f _{MHz} ^{0.5}	0.0037* f _{MHz} ^{0.5}	f _{MHz} / 200	6
2000-300,000	61	0.16	10	6

*Power density limit applicable >100MHz

Exemption Limits³ for Stand-Alone SAR Evaluation:

Head / Body: 20mW

Extremity: 40mW

³ EN 62479 Annex A, General Public

Appendix B - References

1. ACMA Radiocommunications (Electromagnetic Radio – Human Exposure) Standard, 2014.
2. AS/NZS 2772.2, Radiofrequency fields – Principles and method of measurement and computation – 3 kHz to 300 GHz, 2011.
3. Australian Radiation Protection and Nuclear Safety Agency, ARPANSA RPS 3, Maximum Exposure Levels to Radiofrequency Fields 3 kHz to 300 GHz, 2016.
4. New Zealand Standard, NZS 2772.1, Radiofrequency Fields Part 1: Maximum Exposure Levels 3 kHz to 300 GHz, 2009.
5. Federal Communications Commission Knowledge Database (KDB) Publication 447498, “What are the RF exposure requirements and procedures for mobile and portable devices?” As in effect on the issue date of this report.
6. Title 47 Code of Federal Regulations, Part 1.1310, “Radiofrequency radiation exposure limits.” As in effect on the issue date of this report.
7. Title 47 Code of Federal Regulations, Part 2.1091, “Radiofrequency radiation exposure evaluation: mobile devices.” As in effect on the issue date of this report.
8. ANSI C95.1 (2005) IEEE Standard for Safety Level with Respect to Human Exposure to Radio Frequency Electromagnetic Fields, 3kHz to 300 GHz, 2005.
9. Health Canada Safety Code 6 Limits of Human Exposure to Radiofrequency Electromagnetic Energy in the Frequency Range from 3 kHz to 300 GHz, 2015.
10. Industry Canada GL-01 Guidelines for the Measurement of Radio Frequency Fields at Frequencies From 3 kHz to 300 GHz, Issue 3, March 2015.
11. Industry Canada RSS-102 Radio Frequency (RF) Exposure Compliance of Radiocommunication Apparatus (All Frequency Bands) Issue 5, March 2015.
12. EC Council Recommendation 1999/519/EC “On the limitation of exposure of the general public to electromagnetic fields (0Hz to 300GHz),” (1999).
13. European Committee for Electrotechnical Standardization. European Normative, EN 62311 Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz to 300 GHz), 2008.
14. European Committee for Electrotechnical Standardization. European Normative, EN 62479 Assessment of the compliance of low power electronic and electrical equipment with the basic restrictions related to human exposure to electromagnetic fields (10 MHz to 300 GHz), 2010.
15. International Commission on Non-Ionizing Radiation Protection. Guidelines for Limiting Exposure to Time-Varying Electric, Magnetic, and Electromagnetic Fields (up to 300 GHz). Health Physics 74 (4): 494-522; 1998.
16. International Commission on Non-Ionizing Radiation Protection Statement on the "Guidelines for limiting exposure to time-varying electric, magnetic and electromagnetic fields (up to 300 GHz). Health Physics 97(3):257-259, 2009.

Appendix C: Manufacturer Declaration

The following device/model has been tested by CKC Laboratories: **NGP BLE Pump/M994838A001**

The manufacturer declares the **M994838A001** represents the hardware of the NGP BLE Platform which may include a variety of different brand names and model numbers that offer different therapies. Users will be able to upgrade their pump to a more comprehensive model via a firmware update. This update will enable the user to access more therapy options without having to purchase a different pump. The firmware updates will not impact RF, EMC and Safety characteristics.

The manufacturer declares that the following additional models are identical electrically or any differences between them do not affect their RF and EMC characteristics, and therefore meets the level of testing equivalent to the tested models.

NGP BLE Pump	Brand Name	Configuration
<i>M994838A001</i>	<i>MiniMed 700</i>	<i>MMT-1800</i>
		<i>MMT-1801</i>
		<i>MMT-1805</i>
	<i>MiniMed 720G</i>	<i>MMT-1809</i>
		<i>MMT-1810</i>
		<i>MMT-1817</i>
		<i>MMT-1818</i>
	<i>MiniMed 740G</i>	<i>MMT-1811</i>
		<i>MMT-1812</i>
	<i>MiniMed 770G</i>	<i>MMT-1880</i>
		<i>MMT-1881</i>
		<i>MMT-1882</i>
	<i>MiniMed 780G</i>	<i>MMT-1884</i>
		<i>MMT-1885</i>
		<i>MMT-1886</i>
		<i>MMT-1887</i>

Note: The products identified in the table above have the same hardware but different software and firmware that provide various therapy delivery options. The insulin pumps are offered in different configurations to meet regional needs for the display of blood glucose values. The display of blood glucose values is a functionality of device software and is not related to hardware. The software and firmware do not affect product radio or electromagnetic compatibility performance or compliance. Hardware documentation such as schematics, block diagram, printed circuit board and component layouts are identical between these insulin pumps. Additional configuration identifiers (e.g., K) may be added to the base configuration number for inventory management purposes and intended to only represent different keypad overlay color.