

835MHz Brain Dipole Validation

SAM Phantom; Flat Section; Probe:ET3DV6 - SN1677; ConvF(6.70,6.70,6.70)

Med. Parameters 835 MHz Brain: $\sigma = 0.91$ mho/m $\epsilon_r = 40.0$ $\rho = 1.00$ g/cm³; Antenna Position -- Out; Crest Factor 1.0

SAR (1g): 2.34mW/g, SAR (10g): 1.60 mW/g

835MHz Brain Dipole Validation (D835V2 S/N: 406)

Frequency: 835 MHz; Antenna Input Power: 250 [mW]; Ambient Temp. = 22.3°C / Meas. Tissue Temp. = 22.2°C

PCTEST Brain Tissue Simulating Liquid [09/09/2002]

