




**MAC 2000**

**GE MEDICAL SYSTEMS INFORMATION TECHNOLOGIES, Inc.,**  
 8200 W, Tower Avenue, Milwaukee, WI, USA  
 100 - 240 V~, 50 - 60 Hz, 1.5A. Made in India  
 Rx Only




c 03







IP20



0459





LBL P/N: 2064019-001

GE Healthcare



**CS** ARTWORK PRODUCT LABEL  
 ENG MAC2000

Date: 16/AUG/2013	DUSHYANTHA
2064019-001	Issue 03
Scale 1:1	Sheet 1 of 1



MAC 2000  
 REF 2056333-001  
 SN SMT15324569PA  
 2015-08

MAC 2000  
 GE MEDICAL SYSTEMS INFORMATION TECHNOLOGIES, Inc.,  
 8200 W. Tower Avenue, Milwaukee, WI, USA  
 100 - 240 V~, 50 - 60 Hz, 1.5A. Made in India  
 Rx Only

IP20 CE 0459

正式 EJT 标签  
 设备名称: MAC 2000  
 设备编号: D-DCA-RD-000  
 设备负责人: 王恩明  
 PM 有效期: 2017.3.26

无需计量  
 No Calibration required  
 GE Healthcare systems

LABEL 1





FCC ID:OU5-MAC2000  
**This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.**  
 LBL P/N: 2106059-001

54 mm

27.5 mm

LABEL 2

ECG analysis system  
 FCC ID: OU5-MAC2000  
 IC : 4048B-MAC2000  
 CAN ICES-3 (B)/NMB-3(B)



   

H005 XX  
 LBL P/N:2106059-001

Note

1. use the blank label 2047695-001 as material
2. Color: black and white
3. Pick one label for different countries and regions as per the order requirements, the details as below table.

USA	Print the label 1 and stick on the device
EU	Print the label 2 and stick on the device
Turkey	Print the label 2 and stick on the device
Saudi Arabia	Print the label 2 and stick on the device
Kuwait	Print the label 2 and stick on the device
China	Print the label 2 and stick on the device
Serbia	Print the label 2 and stick on the device
Australia	Print the label 2 and stick on the device
Taiwan	Print the label 2 and stick on the device
Korea, Republic of	Print the label 3 and stick on the device

DO NOT SCALE DRAWING				DRAWN BY Zhang Haibo		DATE 13-Sep-2017		 <b>GE Healthcare</b>	
REV		CHANGE NO.		DATE		LIFECYCLE			
UNLESS OTHERWISE SPECIFIED: 1) REMOVE ALL BURRS AND SHARP EDGES. 2) DIMENSIONS ARE IN MILLIMETERS. X ±0.1 X.X ±0.1 X.XX ±0.01 X" ±0.5" 3) M PROFILE METRIC THREAD TOLERANCE CLASS 6H/6g PER ISO 965/1.				ESSENTIAL CHARACTERISTIC: NO		CONFIDENTIAL: THIS PRINT AND THE DESIGN THEREIN REPRESENTED ARE THE PROPERTY OF GE HEALTHCARE AND MAY NOT BE USED IN ANY MANNER DETRIMENTAL TO IT'S INTEREST. THIS PRINT HAS BEEN ISSUED WITH THE UNDERSTANDING THAT IT WILL NOT BE USED FOR MANUFACTURE FOR OTHERS, AND WILL BE RETURNED UPON REQUEST.		DRAWING NUMBER <b>2106059-001</b>	
MATERIAL		FINISH		ENG. APPROVED ELECTRONICALLY APPROVED		MODEL NAME		REV B	
--		--		--		--		SCALE 1:1	
CTQ TABLE QTY FLAG DESCRIPTION 0 # CTQ FLAG: MEF140 QUALITY PLAN REQUIRED.				INSPECTION KEY: ITEMS INDICATED MAY REQUIRE A SUPPLIER INSPECTION PLAN.		BALLOON KEY:  XX YY XX= FIND NUMBER, YY= QUANTITY			

Wireless label

MAC 2000  
[SN] SERIAL NUMBER

R12L	123456789012	W1FC	123456789012
E12L	123456789012	W1FM	123456789012
ERGO	123456789012	M100	123456789012
ME12	123456789012	M200	123456789012
T1P1	123456789012	PDFC	123456789012
M112	123456789012	CFRA	123456789012
MEHR	123456789012	CTDG	123456789012
M1HR	123456789012	BCRD	123456789012
RRAN	123456789012	ADTF	123456789012
LANM	123456789012	ADTL	123456789012
LANC	123456789012	SOMF	123456789012
MODC	123456789012	SOML	123456789012
MODM	123456789012	AOMF	123456789012
		AOML	123456789012
		FULL	123456789012

lbi p/n 2064479-001