GE – Marquette Medical Systems FCC ID #: OU5418500


ApexPro Transmitter output Spectrum

Board SN: NB1164290

The modulation index, Minimum shift keying (MSK), which is binary digital FM, has a modulation index of 0.5. The Base Band I & Q signals are Low Pass Filtered, keeping the constant envelope property with a Gaussian filter with B t of 0.3.

The 20db bandwidth is within the 12kHz channel limit.

Date: 03-01-00 Time: 10:09 AM



Date: 03-01-00 Time: 10:09 RM

