

FCC - TEST REPORT

Report Number : **64.790.19.06028.01** Date of Issue: 2020-02-24

Model : AKTIIA INIT I1

Product Type : Blood Pressure Monitor


Applicant : Guangdong Transtek Medical Electronics Co.,Ltd

Address : Zone A, No.105, Dongli Road, Torch Development District, 528437
: Zhongshan, Guangdong, China

Production Facility : Guangdong Transtek Medical Electronics Co.,Ltd

Address : Zone B, No.105, Dongli Road, Torch Development District, 528437
: Zhongshan, Guangdong, China

Test Result : **Positive** **Negative**



Total pages including Appendices : 31

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch is a subcontractor to TÜV SÜD Product Service GmbH according to the principles outlined in ISO 17025.

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch reports apply only to the specific samples tested under stated test conditions. Construction of the actual test samples has been documented. It is the manufacturer's responsibility to assure that additional production units of this model are manufactured with identical electrical and mechanical components. The manufacturer/importer is responsible to the Competent Authorities in Europe for any modifications made to the production units which result in non-compliance to the relevant regulations. TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch shall have no liability for any deductions, inferences or generalizations drawn by the client or others from TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch issued reports.

This report is the confidential property of the client. As a mutual protection to our clients, the public and ourselves, extracts from the test report shall not be reproduced except in full without our written approval.

1 Table of Contents

1	Table of Contents	2
2	Details about the Test Laboratory	3
3	Description of the Equipment Under Test	4
4	Summary of Test Standards	5
5	Summary of Test Results	6
6	General Remarks	7
7	Test Setups	8
8	Systems test configuration	9
9	Technical Requirement	10
9.1	Conducted peak output power	10
9.2	Power spectral density	13
9.3	6 dB Bandwidth and 99% Occupied Bandwidth	16
9.4	Spurious RF conducted emissions	20
9.5	Band edge	24
9.6	Spurious radiated emissions for transmitter	26
10	Test Equipment List	30
11	System Measurement Uncertainty	31

2 Details about the Test Laboratory

Details about the Test Laboratory

Test Site 1

Company name: TÜV SÜD Certification and Testing (China) Co., Ltd. Shenzhen Branch
Building 12 & 13, Zhiheng Wisdomland Business Park, Nantou Checkpoint
Road 2, Nanshan District
Shenzhen 518052
P.R. China

Telephone: 86 755 8828 6998
Fax: 86 755 8828 5299

FCC Registration No.: 514049

3 Description of the Equipment Under Test

Product: Blood Pressure Monitor

Model no.: AKTIIA INIT I1
FCC ID: OU9TMB1973BS
Options and accessories: N/A

Rating: Internal Battery:3.7VDC
DC input:5V
RF Transmission Frequency: 2402MHz-2480MHz
No. of Operated Channel: 40

Modulation: GFSK

Antenna Type: Internal Antenna

Antenna Gain: 0dBi

Description of the EUT: The Equipment Under Test (EUT) AKTIIA INIT I1 is Blood Pressure Monitor with Bluetooth operated at 2.402-2.48GHz

4 Summary of Test Standards

Test Standards	
FCC Part 15 Subpart C 10-1-2019 Edition	PART 15 - RADIO FREQUENCY DEVICES Subpart C - Intentional Radiators

All the test methods were according to KDB558074 v05r02 DTS Measurement Guidance and ANSI C63.10 (2013).

5 Summary of Test Results

Technical Requirements					
FCC Part 15 Subpart C					
Test Condition		Test Site	Test Result		
			Pass	Fail	N/A
§15.207	Conducted emission AC power port	---	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
§15.247 (b) (1)	Conducted peak output power	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.247(a)(1)	20dB bandwidth	---	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
§15.247(a)(1)	Carrier frequency separation	---	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
§15.247(a)(1)(iii)	Number of hopping frequencies	---	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
§15.247(a)(1)(iii)	Dwell Time	---	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>
§15.247(a)(2)	6dB bandwidth and 99% Occupied Bandwidth	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.247(e)	Power spectral density	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.247(d)	Spurious RF conducted emissions	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.247(d)	Band edge	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.247(d) & §15.209	Spurious radiated emissions for transmitter	Site 1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
§15.203	Antenna requirement	See note 2	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Note 1: N/A=Not Applicable.

Note 2: The EUT uses a Integrated antenna, which gain is 0dBi. In accordance to §15.203, it is considered sufficiently to comply with the provisions of this section.

6 General Remarks

This submittal(s) (test report) is intended for FCC ID: OU9TMB1973BS complies with Section 15.209, 15.247 of the FCC Part 15, Subpart C rules.

AKTIIA INIT I1 is Blood Pressure Monitor with Bluetooth 4.0. The TX and RX range is 2402MHz-2480MHz for 4.0.

SUMMARY:

All tests according to the regulations cited on page 5 were

■ - Performed

□ - Not Performed

The Equipment under Test

■ - **Fulfills** the general approval requirements.

□ - **Does not** fulfill the general approval requirements.

Sample Received Date: December 1, 2019

Testing Start Date: December 4, 2019

Testing End Date: December 6, 2019

- TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch -

Reviewed by:


Tony Liu

Prepared by:


Kevin Ouyang


Test by:


Louise Liu


7 Test Setups

7.1 Radiated test setups


Below 1GHz



Above 1GHz



7.2 Conducted RF test setups



8 Systems test configuration

Auxiliary Equipment Used during Test:

DESCRIPTION	MANUFACTURER	MODEL NO.(SHIELD)	S/N(LENGTH)
Notebook	Lenovo	X220	---
Mobile phone	Huawei	P10	---

Test software: nRFgo Studio, which used to control the EUT in continues transmitting mode.

The system was configured to channel 0, 19, and 39 for the test.

9 Technical Requirement

9.1 Conducted peak output power

Test Method

1. Use the following spectrum analyzer settings:
 RBW > the 6 dB bandwidth of the emission being measured, VBW \geq 3RBW, Span \geq 3RBW
 Sweep = auto, Detector function = peak, Trace = max hold.
2. Add a correction factor to the display.
3. Allow the trace to stabilize. Use the marker-to-peak function to set the marker to the peak of the emission. The indicated level is the peak output power.

Limits


According to §15.247 (b) (1), conducted peak output power limit as below:

Frequency Range MHz	Limit W	Limit dBm
2400-2483.5	≤ 1	≤ 30

Test result as below table


Frequency MHz	Conducted Peak Output Power dBm	Result
Bottom channel 2402MHz	-2.81	Pass
Middle channel 2440MHz	-2.10	Pass
Top channel 2480MHz	-0.98	Pass

Low channel 2402MHz




Date: 5 DEC 2019 19:02:27

Middle channel 2440MHz



Date: 5 DEC 2019 19:04:27

High channel 2480MHz



Date: 5 DEC 2019 19:06:14

9.2 Power spectral density

Test Method

This procedure shall be used if maximum peak conducted output power was used to demonstrate compliance:

1. Set analyzer center frequency to DTS channel center frequency. RBW=3kHz, VBW \geq 3RBW, Span=1.5 times DTS bandwidth, Detector=Peak, Sweep=auto, Trace= max hold.
2. Allow trace to fully stabilize, use the peak marker function to determine the maximum amplitude level within the RBW.
3. Repeat above procedures until other frequencies measured were completed.

Limit


Limit [dBm/3KHz]

≤ 8

Test result


Frequency MHz	Power spectral density dBm	Result
Bottom channel 2402MHz	-17.22	Pass
Middle channel 2440MHz	-15.36	Pass
Top channel 2480MHz	-14.96	Pass

Low channel 2402MHz




Date: 5.DEC.2019 19:02:33

Middle channel 2440MHz



Date: 5.DEC.2019 19:04:33

High channel 2480MHz



Date: 5.DEC.2019 19:06:20

9.3 6 dB Bandwidth and 99% Occupied Bandwidth

Test Method

1. Use the following spectrum analyzer settings:
RBW=100K, VBW \geq 3RBW, Sweep = auto, Detector function = peak, Trace = max hold
2. Use the automatic bandwidth measurement capability of an instrument, may be employed using the X dB bandwidth mode with X set to 6 dB, care shall be taken so that the bandwidth measurement is not influenced by any intermediate power nulls in the fundamental emission that might be \geq 6 dB.
3. Allow the trace to stabilize, record the X dB Bandwidth value.

Limit

Limit [kHz]


—————
≥500

Test result


Frequency MHz	6dB bandwidth kHz	99% bandwidth kHz	Result
Bottom channel 2402MHz	820	1966	Pass
Middle channel 2440MHz	840	1590	Pass
Top channel 2480MHz	684	1115	Pass

6 dB Bandwidth

Low channel 2402MHz




Date: 5 DEC. 2019 19:02:09



Date: 5 DEC. 2019 19:02:20

Middle channel 2440MHz




Date: 5.DEC.2019 19:04:09



Date: 5.DEC.2019 19:04:20

High channel 2480MHz



Date: 5.DEC.2019 19:05:56



Date: 5.DEC.2019 19:06:07

9.4 Spurious RF conducted emissions

Test Method


1. Establish a reference level by using the following procedure:
 - a. Set RBW=100 kHz. VBW \geq 3RBW. Detector =peak, Sweep time = auto couple, Trace mode = max hold.
 - b. Allow trace to fully stabilize, use the peak marker function to determine the maximum PSD level.
2. Use the maximum PSD level to establish the reference level.
 - a. Set the center frequency and span to encompass frequency range to be measured.
 - b. Use the peak marker function to determine the maximum amplitude level. Ensure that the amplitude of all unwanted emissions outside of the authorized frequency band (excluding restricted frequency bands) are attenuated by at least the minimum requirements, report the three highest emissions relative to the limit.
3. Repeat above procedures until other frequencies measured were completed.

Limit


Frequency Range MHz	Limit (dBc)
30-25000	-20

Spurious RF conducted emissions

2402MHz




Date: 5 DEC. 2019 19:02:58




Date: 5 DEC. 2019 19:03:10

2440MHz




Date: 5.DEC.2019 19:04:48




Date: 5.DEC.2019 19:05:00

2480MHz



Date: 5.DEC.2019 19:06:45



Date: 5.DEC.2019 19:06:57

9.5 Band edge

Test Method


- 1 Use the following spectrum analyzer settings:
Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 100 kHz, VBW \geq RBW, Sweep = auto, Detector function = peak, Trace = max hold.
- 2 Allow the trace to stabilize, use the peak and delta measurement to record the result.
- 3 The level displayed must comply with the limit specified in this Section.

Limit


Frequency Range MHz	Limit (dBc)
30-25000	-20

Band edge testing

2402MHz



2480MHz



9.6 Spurious radiated emissions for transmitter

Test Method

- 1: The EUT was placed on a turn table which is 1.5m above ground plane for above 1GHz and 0.8m above ground for below 1GHz at 3 meter chamber room for test. The table was rotated 360 degrees to determine the position of the highest radiation.
- 2: The EUT was set 3 meters away from the interference – receiving antenna, which was mounted on the top of a variable – height antenna tower.
- 3: The height of antenna is varied from one meter to four meters above the ground to determine the maximum value of the field strength. Both horizontal and vertical polarizations of the antenna are set to make the measurement.
- 4: For each suspected emission, the EUT was arranged to its worst case and then the antenna was tuned to heights from 1 meter to 4 meters and the rotatable table was turned from 0 degrees to 360 degrees to find the maximum reading.
- 5: Use the following spectrum analyzer settings According to C63.10:

For Above 1GHz

Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 1MHz, VBW \geq RBW for peak measurement and VBW = 10Hz for average measurement, Sweep = auto, Detector function = peak, Trace = max hold.

For Below 1GHz

Use the following spectrum analyzer settings:

Span = wide enough to capture the peak level of the in-band emission and all spurious
RBW = 100 KHz, VBW \geq RBW for peak measurement, Sweep = auto, Detector function = peak, Trace = max hold.

Note:

- 1: The resolution bandwidth and video bandwidth of test receiver/spectrum analyzer is 120 KHz for Quasi-peak detection (QP) at frequency below 1GHz.
- 2: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 3MHz for peak detection (PK) at frequency above 1GHz.
- 3: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 3MHz for RMS Average ((duty cycle < 98%) for Average detection (AV) at frequency above 1GHz, then the measurement results was added to a correction factor ($20\log(1/\text{duty cycle})$).
- 4: The resolution bandwidth of test receiver/spectrum analyzer is 1MHz and the video bandwidth is 10Hz (duty cycle > 98%) for Average detection (AV) at frequency above 1GHz.

Limit

The radio emission outside the operating frequency band shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power. Radiated emissions which fall in the restricted bands, as defined in section 15.205, must comply with the radiated emission limits specified in section 15.209.

Frequency MHz	Field Strength uV/m	Field Strength dB μ V/m	Detector
30-88	100	40	QP
88-216	150	43.5	QP
216-960	200	46	QP
960-1000	500	54	QP
Above 1000	500	54	AV
Above 1000	5000	74	PK

Spurious radiated emissions for transmitter

According to C63.10, if the peak (or quasi-peak) measured value complies with the average limit, it is unnecessary to perform an average measurement, so AV emission value did not show in below table if the peak value complies with average limit.

Transmitting spurious emission test result as below:

Below 1GHz

Frequency Band	Frequency	Emission Level	Polarization	Limit	Detector	Margin	Correct factor	Result
	MHz	dBuV/m		dBuV/m				
30-1000MHz	55.543333	21.64	H	40.00	QP	18.36	-24.7	Pass
	168.009444	22.80	H	43.50	QP	20.70	-30.8	Pass
	915.771667	33.94	H	46.00	QP	12.06	-15.7	Pass
	62.441111	24.65	V	40.00	QP	15.35	-26.1	Pass
	264.201111	24.76	V	46.00	QP	21.24	-26.8	Pass
	943.255000	35.59	V	46.00	QP	10.41	-15.8	Pass

Above 1GHz

Low channel 2402MHz Test Result

Frequency Band	Frequency	Emission Level	Polarization	Limit	Detector	Margin	Correct factor	Result
	MHz	dBuV/m		dBuV/m				
1000-25000MHz	2446.687500	47.76	H	74	PK	26.24	-6.5	Pass
	2462.062500	32.84	H	74	PK	41.16	-6.5	Pass
	4803.281250	42.90	H	74	PK	31.10	1.3	Pass
	7205.156250	42.18	H	74	PK	31.82	4.8	Pass
	2240.000000	33.98	V	74	PK	40.02	-7.1	Pass
	2415.875000	37.52	V	74	PK	36.48	-6.6	Pass
	4803.281250	42.58	V	74	PK	31.42	1.3	Pass
	7205.156250	44.24	V	74	PK	29.76	4.8	Pass

Middle channel 2440MHz Test Result

Frequency Band	Frequency	Emission Level	Polarization	Limit	Detector	Margin	Correct factor	Result
	MHz	dBuV/m		dBuV/m				
1000-25000MHz	1931.250000	27.91	H	74	PK	46.09	-8.3	Pass
	2462.500000	34.24	H	74	PK	39.76	-6.5	Pass
	4879.687500	49.63	H	74	PK	24.37	1.8	Pass
	7320.468750	45.01	H	74	PK	28.99	3.1	Pass
	2368.625000	30.12	V	74	PK	43.88	-6.7	Pass
	2461.812500	30.13	V	74	PK	43.87	-6.5	Pass
	4879.687500	44.04	V	74	PK	29.96	1.8	Pass
	7319.062500	48.89	V	74	PK	25.11	3.1	Pass

High channel 2480MHz Test Result

Frequency Band	Frequency	Emission Level	Polarization	Limit	Detector	Margin	Correct factor	Result
	MHz	dBuV/m		dBμV/m		dB	(dB)	
1000-25000MHz	2446.875000	37.29	H	74	PK	36.71	-6.5	Pass
	2575.187500	44.52	H	74	PK	29.48	-6.0	Pass
	4960.312500	41.52	H	74	PK	32.48	1.7	Pass
	7439.062500	44.53	H	74	PK	29.47	3.4	Pass
	2180.312500	31.91	V	74	PK	42.09	-7.3	Pass
	2446.562500	37.57	V	74	PK	36.43	-6.5	Pass
	4959.375000	42.44	V	74	PK	31.56	1.7	Pass
	7439.531250	47.48	V	74	PK	26.52	3.4	Pass

Remark:

- (1) Data of measurement within this frequency range shown "--" in the table above means the reading of emissions are the noise floor or attenuated more than 10dB below the permissible limits or the field strength is too small to be measured.
- (2) Level = Reading Level + Correction Factor
 Above 1GHz: Corrector factor = Antenna Factor + Cable Loss- Amplifier Gain
 Below 1GHz: Corrector factor = Antenna Factor + Cable Loss
 (The Reading level is recorded by software which is not shown in the sheet)

10 Test Equipment List

List of Test Instruments

Radiated Emission Test

Description	Manufacturer	Model no.	Serial no.	cal. due date
EMI Test Receiver	Rohde & Schwarz	ESR 26	101269	2020-6-28
Trilog Super Broadband Test Antenna	Schwarzbeck	VULB 9163	707	2020-8-20
Horn Antenna	Rohde & Schwarz	HF907	102294	2020-6-22
Pre-amplifier	Rohde & Schwarz	SCU 18	102230	2020-7-7
Signal Generator	Rohde & Schwarz	SMY01	839369/005	2020-6-28
Attenuator	Agilent	8491A	MY39264334	2020-6-28
3m Semi-anechoic chamber	TDK	9X6X6	----	2020-6-28
Test software	Rohde & Schwarz	EMC32	Version 9.15.00	N/A

RF conducted test

Description	Manufacturer	Model no.	Serial no.	cal. due date
Signal Analyzer	Rohde & Schwarz	FSV40	101030	2020-6-28
RF Switch Module	Rohde & Schwarz	OSP120/OSP	101226/100851	2020-6-28
Power Splitter	Weinschel	1580	SC319	2020-6-28
10dB Attenuator	Weinschel	4M-10	43152	2020-6-28
Test software	Rohde & Schwarz	EMC32	Version 10.38.00	N/A
Test software	Tonscend	System for BT/WIFI	Version 2.6	N/A

11 System Measurement Uncertainty

For a 95% confidence level, the measurement expanded uncertainties for defined systems, in accordance with the recommendations of ISO 17025 were:

System Measurement Uncertainty	
Test Items	Extended Uncertainty
Uncertainty for Radiated Spurious Emission 25MHz-3000MHz	Horizontal: 4.80dB; Vertical: 4.87dB;
Uncertainty for Radiated Spurious Emission 3000MHz-18000MHz	Horizontal: 4.59dB; Vertical: 4.58dB;
Uncertainty for Radiated Spurious Emission 18000MHz-40000MHz	Horizontal: 5.05dB; Vertical: 5.04dB;
Uncertainty for Conducted RF test with TS 8997	Power level test involved: 1.16dB Frequency test involved: 0.6×10 ⁻⁷ or 1%