

## Appendix C for KSCR221200236501

### Calibration Certificate

Object	Apply	No	Model	SN	Calibration Date
Dipole	<input type="checkbox"/>	1	CLA150	4025	2021/04/26
	<input type="checkbox"/>	2	D450V3	1103	2021/04/21
	<input type="checkbox"/>	3	D750V3	1188	2022/03/29
	<input checked="" type="checkbox"/>	4	D835V2	4d114	2022/03/31
	<input type="checkbox"/>	5	D900V2	1d079	2022/06/07
	<input checked="" type="checkbox"/>	6	D1800V2	2d170	2022/03/31
	<input checked="" type="checkbox"/>	7	D1900V2	5d136	2022/06/07
	<input type="checkbox"/>	8	D2000V2	1041	2022/06/06
	<input type="checkbox"/>	9	D2300V2	1096	2022/03/31
	<input checked="" type="checkbox"/>	10	D2450V2	817	2022/04/01
	<input checked="" type="checkbox"/>	11	D2600V2	1158	2022/03/31
	<input checked="" type="checkbox"/>	12	D5GHzV2	1095	2022/06/01
DAE	<input checked="" type="checkbox"/>	13	DAE4	1245	2022/05/30
Probe	<input checked="" type="checkbox"/>	14	EX3DV4	7346	2022/03/30



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No. 10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 [www.sgs.com](http://www.sgs.com)  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## 1 Dipole

### 1.1 CLA150 - SN 4025

<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zehrgassestrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: <b>SGS-CN (Auden)</b>      Certificate No.: <b>CLA150-4025_Apr21</b></p> <p style="text-align: right;">Accreditation No.: <b>SCS 0108</b></p> <hr/> <p style="text-align: center;"><b>CALIBRATION CERTIFICATE</b></p> <p>Object: <b>CLA150 - SN: 4025</b></p> <p>Calibration procedure(s): <b>QA CAL-15-V9 Calibration Procedure for SAR Validation Sources below 700 MHz</b></p> <p>Calibration date: <b>April 26, 2021</b></p> <p>The calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity &lt; 70%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 10476</td> <td>09-Apr-21 (No. 217-03201/03202)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 10364</td> <td>09-Apr-21 (No. 217-03201)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z01</td> <td>SN: 10365</td> <td>09-Apr-21 (No. 217-03202)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C22062 (20)</td> <td>09-Apr-21 (No. 217-03343)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch combination</td> <td>SN: 31095 / 0037</td> <td>09-Apr-21 (No. 217-03344)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3004 (DIE4)</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3007_Dec20)</td> <td>Dec-21</td> </tr> <tr> <td></td> <td>SN: 664</td> <td>26-Jun-20 (No. DMS4-656_Jun20)</td> <td>Jun-21</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4118B</td> <td>SN: G814282/74</td> <td>06-Apr-19 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Power sensor E4119A</td> <td>SN: MY41490/07</td> <td>06-Apr-19 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Power sensor E4119A</td> <td>SN: 000100210</td> <td>06-Apr-19 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>RF generator HP 8594B</td> <td>SN: U32842/01700</td> <td>04-Aug-19 (in house check Jun-20)</td> <td>in house check Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8363A</td> <td>SN: U54100/477</td> <td>31-Mar-19 (in house check Oct-20)</td> <td>in house check Oct-21</td> </tr> </tbody> </table> <p>Calibrated by: <b>Jeffrey Katzman</b>      Function: <b>Laboratory Technician</b>      Signature: <i>[Signature]</i></p> <p>Approved by: <b>Kate Polovic</b>      Technical Manager      <i>[Signature]</i></p> <p style="text-align: right;">Issued: April 26, 2021</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: CLA150-4025_Apr21      Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Power meter NRP	SN: 10476	09-Apr-21 (No. 217-03201/03202)	Apr-22	Power sensor NRP Z01	SN: 10364	09-Apr-21 (No. 217-03201)	Apr-22	Power sensor NRP Z01	SN: 10365	09-Apr-21 (No. 217-03202)	Apr-22	Reference 20 dB Attenuator	SN: C22062 (20)	09-Apr-21 (No. 217-03343)	Apr-22	Type-N mismatch combination	SN: 31095 / 0037	09-Apr-21 (No. 217-03344)	Apr-22	Reference Probe EX3004 (DIE4)	SN: 3877	30-Dec-20 (No. EX3007_Dec20)	Dec-21		SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Power meter E4118B	SN: G814282/74	06-Apr-19 (in house check Jun-20)	in house check Jun-22	Power sensor E4119A	SN: MY41490/07	06-Apr-19 (in house check Jun-20)	in house check Jun-22	Power sensor E4119A	SN: 000100210	06-Apr-19 (in house check Jun-20)	in house check Jun-22	RF generator HP 8594B	SN: U32842/01700	04-Aug-19 (in house check Jun-20)	in house check Jun-22	Network Analyzer Agilent E8363A	SN: U54100/477	31-Mar-19 (in house check Oct-20)	in house check Oct-21	<p>Calibration Laboratory of Schmid &amp; Partner Engineering AG Zehrgassestrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: <b>SCS 0108</b></p> <hr/> <p><b>Glossary:</b></p> <p>TSL: Issue simulating liquid ConvF: sensitivity in TSL / NORM x,y,z N/A: not applicable or not measured</p> <p><b>Calibration is Performed According to the Following Standards:</b></p> <ol style="list-style-type: none"> <li>IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013</li> <li>IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016</li> <li>IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010</li> <li>KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</li> </ol> <p><b>Additional Documentation:</b></p> <ol style="list-style-type: none"> <li>DASY4/5 System Handbook</li> </ol> <p><b>Methods Applied and Interpretation of Parameters:</b></p> <ul style="list-style-type: none"> <li><b>Measurement Conditions:</b> Further details are available from the Validation Report at the end of the certificate. All figures stated in this certificate are valid at the frequency indicated.</li> <li><b>Antenna Parameters with TSL:</b> The source is mounted in a touch configuration below the center marking of the flat phantom.</li> <li><b>Return Loss:</b> This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required.</li> <li><b>SAR measured:</b> SAR measured at the stated antenna input power.</li> <li><b>SAR normalized:</b> SAR as measured, normalized to an input power of 1 W at the antenna connector.</li> <li><b>SAR for nominal TSL parameters:</b> The measured TSL parameters are used to calculate the nominal SAR result.</li> </ul> <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: CLA150-4025_Apr21      Page 2 of 6</p>		
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																																																								
Power meter NRP	SN: 10476	09-Apr-21 (No. 217-03201/03202)	Apr-22																																																								
Power sensor NRP Z01	SN: 10364	09-Apr-21 (No. 217-03201)	Apr-22																																																								
Power sensor NRP Z01	SN: 10365	09-Apr-21 (No. 217-03202)	Apr-22																																																								
Reference 20 dB Attenuator	SN: C22062 (20)	09-Apr-21 (No. 217-03343)	Apr-22																																																								
Type-N mismatch combination	SN: 31095 / 0037	09-Apr-21 (No. 217-03344)	Apr-22																																																								
Reference Probe EX3004 (DIE4)	SN: 3877	30-Dec-20 (No. EX3007_Dec20)	Dec-21																																																								
	SN: 664	26-Jun-20 (No. DMS4-656_Jun20)	Jun-21																																																								
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																																																								
Power meter E4118B	SN: G814282/74	06-Apr-19 (in house check Jun-20)	in house check Jun-22																																																								
Power sensor E4119A	SN: MY41490/07	06-Apr-19 (in house check Jun-20)	in house check Jun-22																																																								
Power sensor E4119A	SN: 000100210	06-Apr-19 (in house check Jun-20)	in house check Jun-22																																																								
RF generator HP 8594B	SN: U32842/01700	04-Aug-19 (in house check Jun-20)	in house check Jun-22																																																								
Network Analyzer Agilent E8363A	SN: U54100/477	31-Mar-19 (in house check Oct-20)	in house check Oct-21																																																								
<p><b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1.</p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th>DASY Version</th> <th>DASY5</th> <th>V52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>ELN Flat Phantom</td> <td>Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td>Touch Position</td> <td></td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Graded Ratio = 1.4 (Z direction)</td> </tr> <tr> <td>Frequency</td> <td>150 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p><b>Head TSL parameters</b> The following parameters and calculations were applied.</p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>52.3</td> <td>0.75 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.75 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt; 0.5 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p><b>SAR result with Head TSL</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th>SAR averaged over 1 cm<sup>3</sup> (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>3.88 W/kg ± 18.4 % (k=2)</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <thead> <tr> <th>SAR averaged over 10 cm<sup>3</sup> (10 g) of Head TSL</th> <th>condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>2.59 W/kg ± 18.0 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21      Page 3 of 6</p>	DASY Version	DASY5	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm	EUT Positioning	Touch Position		Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)	Frequency	150 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %	Head TSL temperature change during test	< 0.5 °C	---	---	SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		SAR measured	1 W input power	3.90 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)	SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition		SAR measured	1 W input power	2.60 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)	<p><b>Appendix (Additional assessments outside the scope of SCS 0108)</b></p> <p><b>Antenna Parameters with Head TSL</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>47.9 Ω ± 1.5 Ω</td> </tr> <tr> <td>Return Loss</td> <td>-31.4 dB</td> </tr> </tbody> </table> <p><b>Additional EUT Data</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: small;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21      Page 4 of 6</p>	Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω	Return Loss	-31.4 dB	Manufactured by	SPEAG
DASY Version	DASY5	V52.10.4																																																									
Extrapolation	Advanced Extrapolation																																																										
Phantom	ELN Flat Phantom	Shell thickness: 2 ± 0.2 mm																																																									
EUT Positioning	Touch Position																																																										
Zoom Scan Resolution	dx, dy = 4.0 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)																																																									
Frequency	150 MHz ± 1 MHz																																																										
	Temperature	Permittivity	Conductivity																																																								
Nominal Head TSL parameters	22.0 °C	52.3	0.75 mho/m																																																								
Measured Head TSL parameters	(22.0 ± 0.2) °C	51.1 ± 6 %	0.75 mho/m ± 6 %																																																								
Head TSL temperature change during test	< 0.5 °C	---	---																																																								
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition																																																										
SAR measured	1 W input power	3.90 W/kg																																																									
SAR for nominal Head TSL parameters	normalized to 1W	3.88 W/kg ± 18.4 % (k=2)																																																									
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition																																																										
SAR measured	1 W input power	2.60 W/kg																																																									
SAR for nominal Head TSL parameters	normalized to 1W	2.59 W/kg ± 18.0 % (k=2)																																																									
Impedance, transformed to feed point	47.9 Ω ± 1.5 Ω																																																										
Return Loss	-31.4 dB																																																										
Manufactured by	SPEAG																																																										



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300    t(86-512) 5735 5888    f(86-512) 57370818    www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号    邮编: 215300    t(86-512) 57355888    f(86-512) 57370818    sgs.china@sgs.com

**DASY5 Validation Report for Head TSL**

Date: 26.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA150; Type: CLA150; Serial: CLA150 - SN: 4025

Communication System: UID 0 - CW; Frequency: 150 MHz  
 Medium parameters used:  $f = 150 \text{ MHz}$ ;  $\sigma = 0.76 \text{ S/m}$ ;  $\epsilon_0 = 51.1$ ;  $\rho = 1000 \text{ kg/m}^3$   
 Phantom section: Flat Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(12.51, 12.51, 12.51) @ 150 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA54 Sn654; Calibrated: 26.06.2020
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP.1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

**CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x8)Cube D:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm  
 Reference Value = 85.93 W/m; Power Drift = -0.02 dB  
 Peak SAR (extrapolated) = 7.36 W/kg  
**SAR(1 g) = 3.90 W/kg; SAR(10 g) = 2.60 W/kg**  
 Smallest distance from peaks to all points 3 dB below: Larger than measurement grid (> 30mm)  
 Ratio of SAR at M2 to SAR at M1 = 80.4%  
 Maximum value of SAR (measured) = 5.48 W/kg

0 dB = 5.48 W/kg = 7.39 dBW/kg

Certificate No: CLA150-4025\_Apr21 Page 6 of 6

Certificate No: CLA150-4025\_Apr21 Page 6 of 6

## 1.2 D450V3 - SN 1103

**Calibration Laboratory of Schmid & Partner Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103\_Apr21**

**CALIBRATION CERTIFICATE**

Object: **D450V3 - SN: 1103**

Calibration procedure(s): **QA CAL-15\_v9**  
 Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closest laboratory facility; environment temperature (22 ± 0.1) °C and humidity < 70%.

Calibration Equipment used (MPE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NRP	SN: 104778	09-Apr-21 (No. 217-03021/03030)	Apr-22
Power sensor NRP-291	SN: 102344	09-Apr-21 (No. 217-03021)	Apr-22
Reference 20 dB Attenuator	SN: CC2502 (200)	09-Apr-21 (No. 217-03343)	Apr-22
Type-N mission combination	SN: 310982 / 06327	09-Apr-21 (No. 217-03344)	Apr-22
Reference Probe E3030A	SN: 38177	30-Dec-20 (No. E30-2077_Decl2)	Dec-21
DAEA	SN: 654	05-Jan-20 (No. 0454-664_Jan20)	Jan-21

Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Power meter E4118B	SN: GB41200274	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4112A	SN: MY4149602	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Power sensor E4413A	SN: 00010010	06-Apr-16 (in house check Jun-20)	In house check Jun-22
RF generator HP 8648C	SN: US340601700	06-Apr-16 (in house check Jun-20)	In house check Jun-22
Network Analyzer Agilent E8358A	SN: US41980477	31-Mar-14 (in house check Oct-20)	In house check Oct-21

Calibrated by: **Christoph Leuber** (Function: Laboratory Technician)

Approved by: **Katja Polovic** (Function: Technical Manager)

Issued: April 23, 2021

Certificate No: D450V3-1103\_Apr21 Page 1 of 6

**Calibration Laboratory of Schmid & Partner Engineering AG**  
 Zeughausstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
 The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103\_Apr21**

**Glossary:**

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103\_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com).

No. 10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 | (86-512) 5735 5888 | (86-512) 57370818 | [www.sgs.com](http://www.sgs.com)  
 中国·江苏·昆山开发区伟业路10号 | 邮编: 215300 | (86-512) 57355888 | (86-512) 57370818 | [sgs.china@sgs.com](http://sgs.china@sgs.com)



**Measurement Conditions**  
DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V82.10.4
Extrapolation	Advanced Extrapolation	
Phantom	ELJ4 Flat Phantom	Shell thickness: 2 ± 0.2 mm
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	450 MHz ± 1 MHz	

**Head TSL parameters**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	43.5	0.57 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	43.1 ± 0.6 %	0.57 mho/m ± 0 %
Head TSL temperature change during test	< 0.5 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.14 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	4.56 W/kg ± 18.1 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	0.757 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	3.06 W/kg ± 17.6 % (k=2)

Certificate No: D450V3-1103\_Apr21 Page 3 of 6

**Appendix (Additional assessments outside the scope of SCS 0106)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	57.1 Ω - 2.6 jΩ
Return Loss	-23.0 dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.346 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.  
The dipole is made of standard straight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is set according to the Standard.  
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------

Certificate No: D450V3-1103\_Apr21 Page 4 of 6

**DASY5 Validation Report for Head TSL**

Test Laboratory: SPEAG, Zurich, Switzerland Date: 21.04.2021

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UID 0 - CW; Frequency: 450 MHz  
Medium parameters used: f = 450 MHz, α = 0.87 S/m; α<sub>0</sub> = 43.1; ρ = 1000 kg/m<sup>3</sup>  
Phantom section: Flat Section  
Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)


**DASY52 Configuration:**

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA4 Sn654; Calibrated: 26.06.2020
- Phantom: ELJ v4.0; Type: QDOVA001BB; Serial: TP:1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

**Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:**  
Measurement grid: dx=5mm, dy=5mm, dz=5mm  
Reference Value = 39.18 V/m; Power Drift = -0.08 dB  
Peak SAR (extrapolated) = 1.76 W/kg  
SAR(1 g) = 1.14 W/kg; SAR(10 g) = 0.767 W/kg  
Smallest distance from peaks to all points: 3 dB below: Larger than measurement grid  
Ratio of SAR at M2 to SAR at M1 = 64.9%  
Maximum value of SAR (measured) = 1.53 W/kg

0 dB = 1.53 W/kg = 1.85 dBW/kg

Certificate No: D450V3-1103\_Apr21 Page 5 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

## 1.3 D750V3 - SN 1188

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-82306633-2112 Fax: +86-10-82306633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn</p> <p>Client: <b>SGS-CN</b> Certificate No: <b>Z22-60103</b></p> <p><b>CALIBRATION CERTIFICATE</b></p> <p>Object: <b>D750V3 - SN: 1188</b></p> <p>Calibration Procedure(s): <b>FF-Z11-003-01</b> Calibration Procedures for dipole validation kits</p> <p>Calibration date: <b>March 28, 2022</b></p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity&lt;70%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>102277</td> <td>24-Sep-21 (CITL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP88</td> <td>104291</td> <td>24-Sep-21 (CITL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX30V4</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CITL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4439C</td> <td>MY49071430</td> <td>13-Jan-22 (CITL No.J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110973</td> <td>14-Jan-22 (CITL No.J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: <b>Zhao Jing</b> SAR Test Engineer</p> <p>Reviewed by: <b>Lin Hao</b> SAR Test Engineer</p> <p>Approved by: <b>Qi Dianyuan</b> SAR Project Leader</p> <p>Issued: April 3, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60103 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	102277	24-Sep-21 (CITL No.J21X08326)	Sep-22	Power sensor NRP88	104291	24-Sep-21 (CITL No.J21X08326)	Sep-22	Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CITL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4439C	MY49071430	13-Jan-22 (CITL No.J22X00409)	Jan-23	Network Analyzer E5071C	MY46110973	14-Jan-22 (CITL No.J22X00409)	Jan-23	<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-82306633-2079 Fax: +86-10-82306633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn</p> <p><b>Glossary:</b></p> <p>TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.yz N/A: not applicable or not measured</p> <p><b>Calibration is Performed According to the Following Standards:</b></p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices-Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 KDB 865684, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>b) DKB 865684, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>c) DASY4/5 System Handbook</p> <p><b>Methods Applied and Interpretation of Parameters:</b></p> <ul style="list-style-type: none"> <li>Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.</li> <li>Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.</li> <li>Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.</li> <li>Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.</li> <li>SAR measured: SAR measured at the stated antenna input power.</li> <li>SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.</li> <li>SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</li> </ul> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: Z22-60103 Page 2 of 6</p>																												
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	102277	24-Sep-21 (CITL No.J21X08326)	Sep-22																																																										
Power sensor NRP88	104291	24-Sep-21 (CITL No.J21X08326)	Sep-22																																																										
Reference Probe EX30V4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																										
DAE4	SN 1556	12-Jan-22(CITL-SPEAG.No.Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4439C	MY49071430	13-Jan-22 (CITL No.J22X00409)	Jan-23																																																										
Network Analyzer E5071C	MY46110973	14-Jan-22 (CITL No.J22X00409)	Jan-23																																																										
<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-82306633-2079 Fax: +86-10-82306633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn</p> <p><b>Measurement Conditions</b></p> <p>DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>V52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 6 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>750 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p><b>Head TSL parameters</b></p> <p>The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>42.0</td> <td>0.90 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>41.4 ± 0.6 %</td> <td>0.89 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt;1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p><b>SAR result with Head TSL</b></p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm<sup>2</sup> (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.07 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>8.27 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm<sup>2</sup> (10 g) of Head TSL</td> <th>Condition</th> <th></th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.37 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>5.48 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: Z22-60103 Page 3 of 6</p>	DASY Version	DASY52	V52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	15 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 6 mm		Frequency	750 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition		SAR measured	250 mW input power	2.07 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition		SAR measured	250 mW input power	1.37 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)	<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-82306633-2079 Fax: +86-10-82306633-2504 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn</p> <p><b>Appendix (Additional assessments outside the scope of CNAS L0570)</b></p> <p><b>Antenna Parameters with Head TSL</b></p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>53.60 - 1.13jΩ</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-28.7dB</td> </tr> </tbody> </table> <p><b>General Antenna Parameters and Design</b></p> <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>0.947 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p><b>Additional EUT Data</b></p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> <p>Certificate No: Z22-60103 Page 4 of 6</p>	Impedance, transformed to feed point	53.60 - 1.13jΩ	Return Loss	-28.7dB	Electrical Delay (one direction)	0.947 ns	Manufactured by	SPEAG
DASY Version	DASY52	V52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	15 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 6 mm																																																												
Frequency	750 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	42.0	0.90 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.4 ± 0.6 %	0.89 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	2.07 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	8.27 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	1.37 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	5.48 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	53.60 - 1.13jΩ																																																												
Return Loss	-28.7dB																																																												
Electrical Delay (one direction)	0.947 ns																																																												
Manufactured by	SPEAG																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

**TTL Speaq** Calibration Laboratory  
 Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304631-2079 Fax: +86-10-42304631-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

**DASY5 Validation Report for Head TSL** Date: 2022-03-29  
 Test Laboratory: CTTL, Beijing, China  
 DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3 - SN: 1188  
 Communication System: UID 0, CW; Frequency: 750 MHz; Duty Cycle: 1:1  
 Medium parameters used:  $f = 750 \text{ MHz}$ ;  $\sigma = 0.888 \text{ S/m}$ ;  $\epsilon_r = 41.36$ ;  $\rho = 1000 \text{ kg/m}^3$   
 Phantom section: Right Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.31, 10.31, 10.31) @ 750 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 S01556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (20kg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm  
 Reference Value = 55.06 V/m; Power Drift = 0.00 dB  
 Peak SAR (extrapolated) = 3.07 W/kg  
 SAR(1 g) = 2.07 W/kg; SAR(10 g) = 1.37 W/kg  
 Smallest distance from peaks to all points 3 dB below = 18.9 mm  
 Ratio of SAR at M2 to SAR at M1 = 67.1%  
 Maximum value of SAR (measured) = 2.74 W/kg

Certificate No: Z22-60103 Page 5 of 6

**TTL Speaq** Calibration Laboratory  
 Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304631-2079 Fax: +86-10-42304631-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60103 Page 6 of 6

## 1.4 D835V2 - SN 4d114

**TTL Speaq** Calibration Laboratory  
 Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191  
 Tel: +86-10-42304631-2079 Fax: +86-10-42304631-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

Client: **SGS-CN** Certificate No: **Z22-60104**

**CALIBRATION CERTIFICATE**

Object: D835V2 - SN: 4d114  
 Calibration Procedure(s): FF-Z11-003-01  
 Calibration Procedures for dipole validation kits  
 Calibration date: March 31, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL, No.J21X08326)	Sep-22
Power sensor NRPBS	104291	24-Sep-21 (CTTL, No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG, No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG, No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL, No.J22X00409)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL, No.J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer  
 Reviewed by: Lin Hao, SAR Test Engineer  
 Approved by: Qi Dianyuan, SAR Project Leader

Issued: April 6, 2022  
 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60104 Page 1 of 6

**TTL Speaq** Calibration Laboratory  
 Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304631-2079 Fax: +86-10-42304631-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**CAICT**

**Glossary:**  
 TSL: tissue simulating liquid  
 ConvF: sensitivity in TSL / NORMx.yz  
 N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**  
 a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020  
 b) KDB 685864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**  
 c) DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60104 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com  
 No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com



In Collaboration with  
**TTL Speag**  
 CALIBRATION LABORATORY

CAICT

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Measurement Conditions**  
 DASYS system configuration, as far as not given on page 1.

DASY Version	DASY52	VS2 10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

**Head TSL parameters**  
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	41.0 ± 8 %	0.91 mho/m ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.37 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	9.40 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	1.54 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	6.12 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60104 Page 3 of 6

In Collaboration with  
**TTL Speag**  
 CALIBRATION LABORATORY

CAICT

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Appendix (Additional assessments outside the scope of CNAS L0570)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	48.70 - j22jΩ
Return Loss	-25.3dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.307 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60104 Page 4 of 6

In Collaboration with  
**TTL Speag**  
 CALIBRATION LABORATORY

CAICT

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**DASY5 Validation Report for Head TSL**  
 Test Laboratory: CTTL, Beijing, China Date: 2022-03-21

**DUT: Dipole 835 MHz; Type: DR35V2; Serial: DR35V2 - SN: 4d114**  
 Communication System: UID 0, CW; Frequency: 835 MHz; Duty Cycle: 1:1  
 Medium parameters used: f = 835 MHz; σ = 0.907 S/m; ε<sub>r</sub> = 40.98; ρ = 1000 kg/m<sup>3</sup>  
 Phantom section: Right Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

**DASY5 Configuration:**

- Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA E4 Sn 1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
 Reference Value = 57.88 V/m; Power Drift = 0.04 dB  
 Peak SAR (extrapolated) = 3.56 W/kg  
 SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg  
 Smallest distance from peaks to all points 3 dB below = 15.8 mm  
 Ratio of SAR at M2 to SAR at M1 = 66.2%  
 Maximum value of SAR (measured) = 3.17 W/kg

0 dB = 3.17 W/kg = 5.01 dBW/kg

Certificate No: Z22-60104 Page 5 of 6

In Collaboration with  
**TTL Speag**  
 CALIBRATION LABORATORY

CAICT

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60104 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No. 10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

## 1.5 D900V2 - SN 1d079

<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;">   <small>In Collaboration with</small>  <b>TTL Calibration Laboratory</b>  <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China</small>  <small>Tel: +86-10-62036323-2117</small>  <small>E-mail: cti@china.com</small>  <small>http://www.caict.ac.cn</small> </div> <div style="text-align: center;">   <small>中国合格评定</small>  <b>CNAS</b>  <small>校准</small>  <small>CACT  <small>校准</small>  <small>CMS 1070</small> </small></div> </div> <div style="text-align: right; margin-top: 10px;"> <small>Client</small> <b>SGS-CN</b>      <small>Certificate No:</small> <b>Z22-60184</b> </div> <div style="border: 1px solid black; padding: 5px; margin-top: 5px;"> <p><b>CALIBRATION CERTIFICATE</b></p> <p><small>Object</small>                    <b>D900V2 - SN: 1d079</b></p> <p><small>Calibration Procedure(s)</small>                    <b>FF-Z11-003-01</b>  <small>Calibration Procedures for dipole validation kits</small></p> <p><small>Calibration date:</small>                    <b>June 7, 2022</b></p> <p><small>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</small></p> <p><small>All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity&lt;70%.</small></p> <p><small>Calibration Equipment used (M&amp;TE critical for calibration)</small></p> <table border="1" style="width: 100%; border-collapse: collapse; font-size: 8px;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DV4</td> <td>SN 7464</td> <td>26-Jan-22(SPEAG No.EK3-7464_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse; font-size: 8px;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MV42071430</td> <td>13-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MV48110673</td> <td>14-Jan-22 (CTTL No.J22X00409)</td> <td>Jan-23</td> </tr> </tbody> </table> <div style="margin-top: 5px;"> <table style="width: 100%; border-collapse: collapse; font-size: 8px;"> <tr> <td style="width: 30%;"><small>Calibrated by:</small></td> <td style="width: 40%;"><small>Name</small>                    <b>Zhao Jing</b>                    <small>Function</small>                    <b>SAR Test Engineer</b>                    <small>Signature</small></td> <td style="width: 30%;"></td> </tr> <tr> <td><small>Reviewed by:</small></td> <td><small>Name</small>                    <b>Lin Hao</b>                    <small>Function</small>                    <b>SAR Test Engineer</b></td> <td></td> </tr> <tr> <td><small>Approved by:</small></td> <td><small>Name</small>                    <b>Qi Diqiyuan</b>                    <small>Function</small>                    <b>SAR Project Leader</b></td> <td></td> </tr> </table> <p style="text-align: right; font-size: 8px;"><small>Issued: June 13, 2022</small></p> <p style="font-size: 8px;"><small>The calibration certificate shall not be reproduced except in full without written approval of the laboratory.</small></p> </div> </div>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No.EK3-7464_Jan22)	Jan-23	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MV42071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23	Network Analyzer E5071C	MV48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23	<small>Calibrated by:</small>	<small>Name</small> <b>Zhao Jing</b> <small>Function</small> <b>SAR Test Engineer</b> <small>Signature</small>		<small>Reviewed by:</small>	<small>Name</small> <b>Lin Hao</b> <small>Function</small> <b>SAR Test Engineer</b>		<small>Approved by:</small>	<small>Name</small> <b>Qi Diqiyuan</b> <small>Function</small> <b>SAR Project Leader</b>		<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="text-align: center;">   <small>In Collaboration with</small>  <b>TTL Calibration Laboratory</b>  <small>ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China</small>  <small>Tel: +86-10-62036323-2117</small>  <small>E-mail: cti@china.com</small>  <small>http://www.caict.ac.cn</small> </div> <div style="text-align: center;"> </div> </div> <div style="margin-top: 10px;"> <p><small>Glossary:</small></p> <p>TSL                    <small>tissue simulating liquid</small></p> <p>ConvF                    <small>sensitivity in TSL / NORMx,y,z</small></p> <p>N/A                    <small>not applicable or not measured</small></p> </div> <div style="margin-top: 5px;"> <p><small>Calibration is Performed According to the Following Standards:</small></p> <p>a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020</p> <p>b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p><small>Additional Documentation:</small></p> <p>c) DASy4/5 System Handbook</p> <p><small>Methods Applied and Interpretation of Parameters:</small></p> <ul style="list-style-type: none"> <li>• <small>Measurement Conditions:</small> Further details are available from the Validation Report at the end of the certificate. All figures stated in this certificate are valid at the frequency indicated.</li> <li>• <small>Antenna Parameters with TSL:</small> The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.</li> <li>• <small>Feed Point Impedance and Return Loss:</small> These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.</li> <li>• <small>Electrical Delay:</small> One-way delay between the SMA connector and the antenna feed point. No uncertainty required.</li> <li>• <small>SAR measured:</small> SAR measured at the stated antenna input power.</li> <li>• <small>SAR normalized:</small> SAR as measured, normalized to an input power of 1 W at the antenna connector.</li> <li>• <small>SAR for nominal TSL parameters:</small> The measured TSL parameters are used to calculate the nominal SAR result.</li> </ul> </div> <div style="border: 1px solid black; padding: 5px; margin-top: 5px; font-size: 8px;"> <p><small>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</small></p> </div>
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																							
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																							
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																							
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No.EK3-7464_Jan22)	Jan-23																																							
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No.Z22-60007)	Jan-23																																							
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																							
Signal Generator E4438C	MV42071430	13-Jan-22 (CTTL No.J22X00409)	Jan-23																																							
Network Analyzer E5071C	MV48110673	14-Jan-22 (CTTL No.J22X00409)	Jan-23																																							
<small>Calibrated by:</small>	<small>Name</small> <b>Zhao Jing</b> <small>Function</small> <b>SAR Test Engineer</b> <small>Signature</small>																																									
<small>Reviewed by:</small>	<small>Name</small> <b>Lin Hao</b> <small>Function</small> <b>SAR Test Engineer</b>																																									
<small>Approved by:</small>	<small>Name</small> <b>Qi Diqiyuan</b> <small>Function</small> <b>SAR Project Leader</b>																																									



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300    t(86-512) 5735 5888    f(86-512) 57370818    www.sgsgroup.com.cn  
 中国·江苏·昆山开发区伟业路10号    邮编: 215300    t(86-512) 57355888    f(86-512) 57370818    sgs.china@sgs.com



In Collaboration with  
**CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2117 Fax: +86-10-42306033-2118  
E-mail: cti@chinaetl.com http://www.caict.ac.cn

**DASY5 Validation Report for Head TSL** Date: 2022-06-07

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 14079

Communication System: UTD 0, CW; Frequency: 900 MHz; Duty Cycle: 1:1

Medium parameters used:  $f = 900$  MHz;  $\sigma = 0.98$  S/m;  $\epsilon_r = 42.05$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Right Section

Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(9.72, 9.72) @ 900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronic: DA44 - SN1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5.1C (2dmg probe kit); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7) (7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 59.81 V/m; Power Drift = -0.01 dB

Peak SAR (extrapolated) = 4.20 W/kg

SAR(1 g) = 2.78 W/kg; SAR(10 g) = 1.78 W/kg

Smallest distance from peaks to all points 3 dB below = 16 mm

Ratio of SAR at M2 to SAR at M1 = 65.8%

Maximum value of SAR (measured) = 3.71 W/kg

Certificate No: Z22-60184 Page 6 of 6

In Collaboration with  
**CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2117 Fax: +86-10-42306033-2118  
E-mail: cti@chinaetl.com http://www.caict.ac.cn

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60184 Page 6 of 6

## 1.6 D1800V2 - SN 2d170

In Collaboration with  
**CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2117 Fax: +86-10-42306033-2118  
E-mail: cti@chinaetl.com http://www.caict.ac.cn

中国认可  
国际互认  
CALIBRATION  
CNAS L870

Client: **SGS-CN** Certificate No: **Z22-60105**

**CALIBRATION CERTIFICATE**

Object: **D1800V2 - SN: 2d170**

Calibration Procedure(s): **FF-Z11-003-01**  
Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21 (SPEAG No.JX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22 (CTTL-SPEAG No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23
Network Analyzer E5071C	MY46110973	14-Jan-22 (CTTL No.J22X00406)	Jan-23

Calibrated by: **Zhao Jing** SAR Test Engineer

Reviewed by: **Lin Hao** SAR Test Engineer

Approved by: **Qi Dianyan** SAR Project Leader

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

In Collaboration with  
**CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42306033-2079 Fax: +86-10-42306033-2104  
E-mail: cti@chinaetl.com http://www.caict.ac.cn

**Glossary:**

TSL: tissue simulating liquid

ConvF: sensitivity in TSL / NORMx,y,z

N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 865684, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 8888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@chinaeui.com http://www.chinaeui.com

**Measurement Conditions**  
DASY5 system configuration, as far as not given on page 1.

DASY5 Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1800 MHz ± 1 MHz	

**Head TSL parameters**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.8 ± 8 %	1.41 mho/m ± 8 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	9.73 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	38.9 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.11 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	20.4 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60105 Page 3 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@chinaeui.com http://www.chinaeui.com

**Appendix (Additional assessments outside the scope of CNAS L0570)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	47.90-2.54jΩ
Return Loss	-29.4dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.116 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60105 Page 4 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@chinaeui.com http://www.chinaeui.com

**DASY5 Validation Report for Head TSL** Date: 2022-03-31  
Test Laboratory: CTTL, Beijing, China  
DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: 2d170  
Communication System: UID 0, CW; Frequency: 1800 MHz; Duty Cycle: 1:1  
Medium parameters used: f = 1800 MHz; σ = 1.411 S/m; ε = 40.62; ρ = 1000 kg/m<sup>3</sup>  
Phantom section: Right Section  
Measurement Standard: DASY5 (IEEE/ANSI C63.19-2007)  
DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(8.34, 8.34, 8.34) @ 1800 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
Reference Value = 98.14 V/m; Power Drift = 0.03 dB  
Peak SAR (extrapolated) = 18.2 W/kg  
SAR(1 g) = 9.73 W/kg; SAR(10 g) = 5.11 W/kg  
Smallest distance from peaks to all points 3 dB below = 10 mm  
Ratio of SAR at M2 to SAR at M1 = 54%  
Maximum value of SAR (measured) = 15.2 W/kg

Certificate No: Z22-60105 Page 5 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY and **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@chinaeui.com http://www.chinaeui.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60105 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com  
No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

## 1.7 D1900V2 - SN 5d136

<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com</p> <p>Client: <b>SGS-CN</b> Certificate No: <b>Z22-60185</b></p> <h3>CALIBRATION CERTIFICATE</h3> <p>Object: D1900V2 - SN: 5d136</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (23±)°C and humidity &lt;70%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP6S</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EXSDV4</td> <td>SN 7484</td> <td>28-Jan-22 (SPEAG No. EX3-7484_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1656</td> <td>12-Jan-22 (CTTL-SPEAG No. Z22-90007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by, Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY48071430</td> <td>13-Jan-22 (CTTL No. J22X00409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyser ES071C</td> <td>MY48110073</td> <td>14-Jan-22 (CTTL No. J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing, SAR Test Engineer</p> <p>Reviewed by: Lin Hao, SAR Test Engineer</p> <p>Approved by: Qi Diaryuan, SAR Project Leader</p> <p>Issued: June 13, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: Z22-60185 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Power sensor NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22	Reference Probe EXSDV4	SN 7484	28-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23	DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-90007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23	Network Analyser ES071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23	<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com</p> <p>Glossary: TSL: tissue simulating liquid ComF: sensitivity in TSL / NORMx.y.z NA: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1526: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Additional Documentation: c) DASY4/S System Handbook</p> <p>Methods Applied and Interpretation of Parameters: • Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. • Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. • Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. • Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. • SAR measured: SAR measured at the stated antenna input power. • SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. • SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</p> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: Z22-60185 Page 2 of 6</p>																												
Primary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Power sensor NRP6S	104291	24-Sep-21 (CTTL No. J21X08326)	Sep-22																																																										
Reference Probe EXSDV4	SN 7484	28-Jan-22 (SPEAG No. EX3-7484_Jan22)	Jan-23																																																										
DAE4	SN 1656	12-Jan-22 (CTTL-SPEAG No. Z22-90007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by, Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY48071430	13-Jan-22 (CTTL No. J22X00409)	Jan-23																																																										
Network Analyser ES071C	MY48110073	14-Jan-22 (CTTL No. J22X00406)	Jan-23																																																										
<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com</p> <p>Measurement Conditions DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>1900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th>Nominal Head TSL parameters</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td></td> <td>22.0 °C</td> <td>40.0</td> <td>1.40 mS/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.9 ± 6 %</td> <td>1.39 mS/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt;1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th>95 % Wkg</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>9.65 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>40.0 W/kg ± 18.8 % (k=2)</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th>95 % Wkg</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.18 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>20.8 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: Z22-60185 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	1900 MHz ± 1 MHz		Nominal Head TSL parameters	Temperature	Permittivity	Conductivity		22.0 °C	40.0	1.40 mS/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 6 %	1.39 mS/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm² (1 g) of Head TSL	Condition	95 % Wkg	SAR measured	250 mW input power	9.65 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm² (10 g) of Head TSL	Condition	95 % Wkg	SAR measured	250 mW input power	5.18 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)	<p>Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42204633-2117 E-mail: vt@china.ttl.com</p> <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>51.20 ± 7.56Ω</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-22.4dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>1.109 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> <p>Certificate No: Z22-60185 Page 4 of 6</p>	Impedance, transformed to feed point	51.20 ± 7.56Ω	Return Loss	-22.4dB	Electrical Delay (one direction)	1.109 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	1900 MHz ± 1 MHz																																																												
Nominal Head TSL parameters	Temperature	Permittivity	Conductivity																																																										
	22.0 °C	40.0	1.40 mS/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.9 ± 6 %	1.39 mS/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm² (1 g) of Head TSL	Condition	95 % Wkg																																																											
SAR measured	250 mW input power	9.65 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	40.0 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm² (10 g) of Head TSL	Condition	95 % Wkg																																																											
SAR measured	250 mW input power	5.18 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	20.8 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	51.20 ± 7.56Ω																																																												
Return Loss	-22.4dB																																																												
Electrical Delay (one direction)	1.109 ns																																																												
Manufactured by	SPEAG																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No. 10 Weiyao Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com



In Collaboration with  
**CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62586317  
E-mail: cti@ctilab.com

**DASY5 Validation Report for Head TSL**  
Date: 2022-04-07  
Test Laboratory: CCTL, Beijing, China  
DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 54136  
Communication System: UTD 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1  
Medium parameters used:  $f = 1900 \text{ MHz}$ ;  $\sigma = 1.385 \text{ S/m}$ ;  $\epsilon_r = 39.85$ ;  $\rho = 1000 \text{ kg/m}^3$   
Phantom section: Right Section  
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
DASY5 Configuration:  

- Probe: EX3DV4 - SN7464; ConvF(8, 1.8, 8, 1.8) @ 1900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5\_1C (20kg probe 0H); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X.14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube D: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
Reference Value = 99.99 V/m; Power Drift = 0.04 dB  
Peak SAR (extrapolated) = 18.6 W/kg  
SAR(1g) = 9.95 W/kg; SAR(10g) = 5.18 W/kg  
Smallest distance from peaks to all points 3 dB below = 9.2 mm  
Ratio of SAR at M2 to SAR at M1 = 54.1%

Maximum value of SAR (measured) = 15.6 W/kg

0 dB = 15.6 W/kg = 11.93 dBW/kg

Certificate No: Z22-60185 Page 5 of 6

In Collaboration with  
**CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62586317  
E-mail: cti@ctilab.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60185 Page 6 of 6

## 1.8 D2000V2 - SN 1041

In Collaboration with  
**CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62586317  
E-mail: cti@ctilab.com

Client: **SGS-CN** Certificate No: **Z22-60186**

**CALIBRATION CERTIFICATE**

Object: D2000V2 - SN: 1041

Calibration Procedure(s): FF-Z11-003-01  
Calibration Procedures for dipole validation kits

Calibration date: June 8, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (8). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22±)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CCTL No.J21X06326)	Sep-22
Power sensor NRP5S	104291	24-Sep-21 (CCTL No.J21X06326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No EX3-7464-Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CCTL-SPEAG No Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CCTL No.Z22X0409)	Jan-23
Network Analyzer E5071C	MY48110673	14-Jan-22 (CCTL No.Z22X0406)	Jan-23

Calibrated by: Zhao Jing SAR Test Engineer

Reviewed by: Lin Hao SAR Test Engineer

Approved by: Qi Dianyan SAR Project Leader

Issued: June 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60186 Page 1 of 6

In Collaboration with  
**CAICT**  
CALIBRATION LABORATORY

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62586317  
E-mail: cti@ctilab.com

**Glossary:**

TSL: Issue simulating liquid sensitivity in TSL / NORMx.y.z

N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60186 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

In Collaboration with **TTL S P E A G** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-4239683-2117  
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

**Measurement Conditions**  
 DASYS system configuration, as far as not given on page 1.

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2000 MHz ± 1 MHz	

**Head TSL parameters**  
 The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	40.2 ± 0.5 %	1.39 mho/m ± 0.6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.4 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	41.8 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	5.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.3 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60186 Page 3 of 6

In Collaboration with **TTL S P E A G** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-4239683-2117  
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

**Appendix (Additional assessments outside the scope of CNAS L6570)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	48.4Ω ± 0.74jΩ
Return Loss	-34.9dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.088 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, g-ral and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60186 Page 4 of 6

In Collaboration with **TTL S P E A G** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-4239683-2117  
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

**DASY5 Validation Report for Head TSL** Date: 2022-06-06

Test Laboratory: CCTL, Beijing, China  
 DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041  
 Communication System: LUD 0; CW; Frequency: 2000 MHz; Duty Cycle: 1:1  
 Medium parameters used: f = 2000 MHz; σ = 1.392 S/m; ε<sub>r</sub> = 40.21; ρ = 1000 kg/m<sup>3</sup>  
 Phantom section: Right Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
 DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(R,2, 8.2, 8.2) @ 2000 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA64 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52: S2.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
 Reference Value = 103.4 V/m; Power Drift = 0.03 dB  
 Peak SAR (extrapolated) = 19.6 W/kg  
 SAR(1 g) = 10.4 W/kg; SAR(10 g) = 5.3 W/kg  
 Smallest distance from peaks to all points 3 dB below = 9.1 mm  
 Ratio of SAR at M2 to SAR at M1 = 53.0%

Maximum value of SAR (measured) = 16.3 W/kg

0 dB = 16.3 W/kg = 12.12 dBW/kg

Certificate No: Z22-60186 Page 5 of 6

In Collaboration with **TTL S P E A G** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-4239683-2117  
 E-mail: ott@china.ttl.com http://www.caict.ac.cn

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60186 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com  
 No. 10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

## 1.9 D2300V2 - SN 1096

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn</p>																																	
Client: <b>SGS-CN</b>	Certificate No: <b>Z22-60106</b>																																
<b>CALIBRATION CERTIFICATE</b> Object: <b>D2300V2 - SN 1096</b> Calibration Procedure(s): <b>FF-Z11-003-01 Calibration Procedures for dipole validation kits</b> Calibration date: <b>March 31, 2022</b> This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity<70%. Calibration Equipment used (M&TE critical for calibration)																																	
<table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>108277</td> <td>24-Sep-21 (CTTL No.J21X08328)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08328)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3/DVA</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EK3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table>	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22	Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EK3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	<table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table>	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																														
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08328)	Sep-22																														
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08328)	Sep-22																														
Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EK3-7307_May21)	May-22																														
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																														
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																														
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23																														
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23																														
Calibrated by: <b>Zhao Jing</b> SAR Test Engineer Reviewed by: <b>Lin Hao</b> SAR Test Engineer Approved by: <b>Qi Diaryuan</b> SAR Project Leader Issued: April 6, 2022 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.																																	
Certificate No: Z22-60106	Page 1 of 6																																

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn</p>																			
<b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1	<table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2300 MHz ± 1 MHz</td> <td></td> </tr> </table>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2300 MHz ± 1 MHz	
DASY Version	DASY52	52.10.4																	
Extrapolation	Advanced Extrapolation																		
Phantom	Triple Flat Phantom 5.1C																		
Distance Dipole Center - TSL	10 mm	with Spacer																	
Zoom Scan Resolution	dx, dy, dz = 5 mm																		
Frequency	2300 MHz ± 1 MHz																		
<b>Head TSL parameters</b> The following parameters and calculations were applied.	<table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.5</td> <td>1.67 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>39.8 ± 6 %</td> <td>1.70 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt;1.0 °C</td> <td>—</td> <td>—</td> </tr> </tbody> </table>		Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	—	—		
	Temperature	Permittivity	Conductivity																
Nominal Head TSL parameters	22.0 °C	39.5	1.67 mho/m																
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.8 ± 6 %	1.70 mho/m ± 6 %																
Head TSL temperature change during test	<1.0 °C	—	—																
<b>SAR result with Head TSL</b>	<table border="1"> <thead> <tr> <th>SAR averaged over 1 cm<sup>3</sup> (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>12.4 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>49.2 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm<sup>3</sup> (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.88 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>23.4 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table>	SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition		SAR measured	250 mW input power	12.4 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition		SAR measured	250 mW input power	5.88 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)
SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition																		
SAR measured	250 mW input power	12.4 W/kg																	
SAR for nominal Head TSL parameters	normalized to 1W	49.2 W/kg ± 18.8 % (k=2)																	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition																		
SAR measured	250 mW input power	5.88 W/kg																	
SAR for nominal Head TSL parameters	normalized to 1W	23.4 W/kg ± 18.7 % (k=2)																	
Certificate No: Z22-60106	Page 3 of 6																		

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn</p>	
<b>Glossary:</b> TSL: Issue simulating liquid ConvF: sensitivity in TSL / NCRMx,y,z N/A: not applicable or not measured	<b>Calibration is Performed According to the Following Standards:</b> a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz" <b>Additional Documentation:</b> c) DASY4/5 System Handbook <b>Methods Applied and Interpretation of Parameters:</b> <ul style="list-style-type: none"> <li>Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.</li> <li>Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.</li> <li>Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.</li> <li>Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.</li> <li>SAR measured: SAR measured at the stated antenna input power.</li> <li>SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.</li> <li>SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</li> </ul>
The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.	
Certificate No: Z22-60106	Page 2 of 6

<p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2504 E-mail: cti@chinaul.com http://www.chinaul.cn</p>					
<b>Appendix (Additional assessments outside the scope of CNAS L0570)</b> <b>Antenna Parameters with Head TSL</b>	<table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>49.2Ω -4.56jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-26.6dB</td> </tr> </table>	Impedance, transformed to feed point	49.2Ω -4.56jΩ	Return Loss	-26.6dB
Impedance, transformed to feed point	49.2Ω -4.56jΩ				
Return Loss	-26.6dB				
<b>General Antenna Parameters and Design</b>	<table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.083 ns</td> </tr> </table>	Electrical Delay (one direction)	1.083 ns		
Electrical Delay (one direction)	1.083 ns				
After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small and caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.					
<b>Additional EUT Data</b>	<table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table>	Manufactured by	SPEAG		
Manufactured by	SPEAG				
Certificate No: Z22-60106	Page 4 of 6				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 | (86-512) 5735 5888 | (86-512) 57370818 | [www.sgs.com](http://www.sgs.com)  
 中国·江苏·昆山开发区伟业路10号 | 邮编: 215300 | (86-512) 57355888 | (86-512) 57370818 | [sgs.china@sgs.com](mailto:sgs.china@sgs.com)



**TTL Speaq Calibration Laboratory** | **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**DASY5 Validation Report for Head TSL** | Date: 2022-03-31

Test Laboratory: CTTL, Beijing, China  
 DUT: Dipole 2300 MHz; Type: D2300V2; Serial: D2300V2 - SN: 1096  
 Communication System: UTD 0, CW; Frequency: 2300 MHz; Duty Cycle: 1:1  
 Medium parameters used:  $f = 2300$  MHz;  $\sigma = 1.702$  S/m;  $\epsilon = 39.77$ ;  $\rho = 1000$  kg/m<sup>3</sup>  
 Phantom section: Right Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(8.01, 8.01) @ 2300 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sst1556; Calibrated: 2022-01-12
- Phantom: MFP V5.1C (2ldag probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 S2.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
 Reference Value = 102.7 V/m; Power Drift = 0.00 dB  
 Peak SAR (extrapolated) = 24.8 W/kg  
 SAR(1 g) = 12.4 W/kg; SAR(10 g) = 5.88 W/kg  
 Smallest distance from peaks to all points 3 dB below = 9 mm  
 Ratio of SAR at M2 to SAR at M1 = 50.4%  
 Maximum value of SAR (measured) = 20.3 W/kg

0 dB = 20.3 W/kg = 13.07 dBW/kg

Certificate No: Z22-60106 | Page 1 of 6

**TTL Speaq Calibration Laboratory** | **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60106 | Page 4 of 6

## 1.10 D2450V2 - SN 817

**TTL Speaq Calibration Laboratory** | **CAICT** | **CNAS** | **中国合格评定国家认可委员会**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** | Certificate No: **Z22-60107**

**CALIBRATION CERTIFICATE**

Object: D2450V2 - SN: 817  
 Calibration Procedure(s): FF-Z11-003-01  
 Calibration Procedures for dipole validation kits  
 Calibration date: April 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08320)	Sep-22
Reference Probe EX3DV4	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer E5071C	MY46110873	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer  
 Reviewed by: Lin Hao, SAR Test Engineer  
 Approved by: Qi Dianyan, SAR Project Leader

Signature: [Signatures]  
 Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60107 | Page 1 of 6

**TTL Speaq Calibration Laboratory** | **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62304633-2079 Fax: +86-10-62304633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Glossary:**

- TSL: tissue simulating liquid
- ConvF: sensitivity in TSL / NORMx,y,z
- N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 855664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60107 | Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 | (86-512) 57355888 | (86-512) 57370818 | [www.sgs.com](http://www.sgs.com)  
 中国·江苏·昆山开发区伟业路10号 | 邮编: 215300 | (86-512) 57355888 | (86-512) 57370818 | [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**Measurement Conditions**  
DASY system configuration, as far as not given on page 1

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	2450 MHz ± 1 MHz	

**Head TSL parameters**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	39.2	1.80 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 0 %	1.79 mho/m ± 0 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL**

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	13.2 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	53.0 W/kg ± 18.8 % (k=2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	6.15 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	24.7 W/kg ± 18.7 % (k=2)

Certificate No: Z22-60107 Page 3 of 6

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**Appendix (Additional assessments outside the scope of CNAS L0570)**

**Antenna Parameters with Head TSL**

Impedance, transformed to feed point	52.10 ± 3.20Ω
Return Loss	-28.5dB

**General Antenna Parameters and Design**

Electrical Delay (one direction)	1.066 ns
----------------------------------	----------

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard serringid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

**Additional EUT Data**

Manufactured by	SPEAG
-----------------	-------

Certificate No: Z22-60107 Page 4 of 6

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**DASY5 Validation Report for Head TSL** Date: 2022-04-01

Test Laboratory: CTTL, Beijing, China  
DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817  
Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1  
Medium parameters used: f = 2450 MHz; σ = 1.79 S/m; ε = 39.52; ρ = 1000 kg/m<sup>3</sup>  
Phantom section: Right Section  
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA14 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP\_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration** Zoom Scan (7x7x7) (7x7x7)/Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm  
Reference Value = 104.6 V/m; Power Drift = -0.03 dB  
Peak SAR (extrapolated) = 27.0 W/kg  
SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg  
Smallest distance from peaks to all points 3 dB below = 8.9 mm  
Ratio of SAR at M2 to SAR at M1 = -49.2%  
Maximum value of SAR (measured) = 22.1 W/kg

0 dB = 22.1 W/kg = 13.44 dBW/kg

Certificate No: Z22-60107 Page 1 of 6

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504  
E-mail: cti@china.ttl.com http://www.china.ttl.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60107 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com



## 1.11 D2600V2 - SN 1158

<div style="display: flex; justify-content: space-between;"> </div> <p>Client: <b>SGS-CN</b> Certificate No: <b>Z22-60108</b></p> <h3 style="text-align: center;">CALIBRATION CERTIFICATE</h3> <p>Object: <b>D2600V2 - SN: 1158</b></p> <p>Calibration Procedure(s): <b>FF-Z11-003-01</b> Calibration Procedures for dipole validation kits</p> <p>Calibration date: <b>March 31, 2022</b></p> <p>This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity&lt;70%.</p> <p>Calibration Equipment used (M&amp;TE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>108277</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S</td> <td>104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3/DVA</td> <td>SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4</td> <td>SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY49071430</td> <td>13-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C</td> <td>MY46110673</td> <td>14-Jan-22 (CTTL No.J22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Calibrated by:</th> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td></td> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyuan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p style="text-align: right;">Issued: April 6, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p style="font-size: small;">Certificate No: Z22-60108 Page 1 of 6</p>	Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22	Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22	DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23	Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration	Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23	Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23	Calibrated by:	Name	Function	Signature		Zhao Jing	SAR Test Engineer		Reviewed by:	Lin Hao	SAR Test Engineer		Approved by:	Qi Dianyuan	SAR Project Leader		<div style="display: flex; justify-content: space-between;"> </div> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2964 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p>Glossary: TSL: tissue simulating liquid ConvF: sensitivity in TSL / NORMx.y.z N/A: not applicable or not measured</p> <p><b>Calibration is Performed According to the Following Standards:</b> a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p><b>Additional Documentation:</b> c) DASY4/S System Handbook</p> <p><b>Methods Applied and Interpretation of Parameters:</b></p> <ul style="list-style-type: none"> <li>Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.</li> <li>Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.</li> <li>Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.</li> <li>Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.</li> <li>SAR measured: SAR measured at the stated antenna input power.</li> <li>SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.</li> <li>SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.</li> </ul> <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.</p> </div> <p style="font-size: small;">Certificate No: Z22-60108 Page 2 of 6</p>												
Primary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																										
Power Meter NRP2	108277	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Power sensor NRP8S	104291	24-Sep-21 (CTTL No.J21X08326)	Sep-22																																																										
Reference Probe EX3/DVA	SN 7307	26-May-21(SPEAG.No.EX3-7307_May21)	May-22																																																										
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG.No.Z22-60007)	Jan-23																																																										
Secondary Standards	ID #	Cal Date (Calibrated by Certificate No.)	Scheduled Calibration																																																										
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No.J22X00406)	Jan-23																																																										
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No.J22X00406)	Jan-23																																																										
Calibrated by:	Name	Function	Signature																																																										
	Zhao Jing	SAR Test Engineer																																																											
Reviewed by:	Lin Hao	SAR Test Engineer																																																											
Approved by:	Qi Dianyuan	SAR Project Leader																																																											
<div style="display: flex; justify-content: space-between;"> </div> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2964 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p><b>Measurement Conditions</b> DASY system configuration, as far as not given on page 1.</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2600 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p><b>Head TSL parameters</b> The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>39.0</td> <td>1.96 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>38.7 ± 6 %</td> <td>1.96 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>&lt;1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p><b>SAR result with Head TSL</b></p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>54.8 W/kg ± 18.8 % (k=2)</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>24.6 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> <p style="font-size: small;">Certificate No: Z22-60108 Page 3 of 6</p>	DASY Version	DASY52	52.10.4	Extrapolation	Advanced Extrapolation		Phantom	Triple Flat Phantom 5.1C		Distance Dipole Center - TSL	10 mm	with Spacer	Zoom Scan Resolution	dx, dy, dz = 5 mm		Frequency	2600 MHz ± 1 MHz			Temperature	Permittivity	Conductivity	Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m	Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 6 %	Head TSL temperature change during test	<1.0 °C	---	---	SAR averaged over 1 cm² (1 g) of Head TSL	Condition		SAR measured	250 mW input power	13.7 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)	SAR averaged over 10 cm² (10 g) of Head TSL	Condition		SAR measured	250 mW input power	6.12 W/kg	SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)	<div style="display: flex; justify-content: space-between;"> </div> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2964 E-mail: cti@china.ttl.com http://www.chinatit.com</p> <p><b>Appendix (Additional assessments outside the scope of CNAS L0570)</b></p> <p><b>Antenna Parameters with Head TSL</b></p> <table border="1"> <thead> <tr> <th>Impedance, transformed to feed point</th> <th>49.90- 6.49jΩ</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-23.8dB</td> </tr> </tbody> </table> <p><b>General Antenna Parameters and Design</b></p> <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>1.053 ns</th> </tr> </thead> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p><b>Additional EUT Data</b></p> <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> <p style="font-size: small;">Certificate No: Z22-60108 Page 4 of 6</p>	Impedance, transformed to feed point	49.90- 6.49jΩ	Return Loss	-23.8dB	Electrical Delay (one direction)	1.053 ns	Manufactured by	SPEAG
DASY Version	DASY52	52.10.4																																																											
Extrapolation	Advanced Extrapolation																																																												
Phantom	Triple Flat Phantom 5.1C																																																												
Distance Dipole Center - TSL	10 mm	with Spacer																																																											
Zoom Scan Resolution	dx, dy, dz = 5 mm																																																												
Frequency	2600 MHz ± 1 MHz																																																												
	Temperature	Permittivity	Conductivity																																																										
Nominal Head TSL parameters	22.0 °C	39.0	1.96 mho/m																																																										
Measured Head TSL parameters	(22.0 ± 0.2) °C	38.7 ± 6 %	1.96 mho/m ± 6 %																																																										
Head TSL temperature change during test	<1.0 °C	---	---																																																										
SAR averaged over 1 cm² (1 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	13.7 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	54.8 W/kg ± 18.8 % (k=2)																																																											
SAR averaged over 10 cm² (10 g) of Head TSL	Condition																																																												
SAR measured	250 mW input power	6.12 W/kg																																																											
SAR for nominal Head TSL parameters	normalized to 1W	24.6 W/kg ± 18.7 % (k=2)																																																											
Impedance, transformed to feed point	49.90- 6.49jΩ																																																												
Return Loss	-23.8dB																																																												
Electrical Delay (one direction)	1.053 ns																																																												
Manufactured by	SPEAG																																																												



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300    t(86-512) 5735 5888    f(86-512) 57370818    www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号    邮编: 215300    t(86-512) 57355888    f(86-512) 57370818    sgs.china@sgs.com



**TTL Speaq** CALIBRATION LABORATORY  
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**DASY5 Validation Report for Head TSL** Date: 2022-03-31  
 Test Laboratory: CTTL, Beijing, China  
 DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN: 1158  
 Communication System: UID 0, CW; Frequency: 2600 MHz; Duty Cycle: 1:1  
 Medium parameters used:  $f = 2600$  MHz;  $\sigma = 1.955$  S/m;  $\epsilon_r = 38.68$ ;  $\rho = 1000$  kg/m<sup>3</sup>  
 Phantom section: Right Section  
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)  
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.5, 7.5, 7.5) @ 2600 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP-V5.1C (2dkg probe fill); Type: QD 000 P51 Cx; Serial: 1062
- DASY52.52.10.4(1535); SEMCAD X 14.6.14(7501)

**Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm**  
 Reference Value = 103.3 V/m; Power Drift = 0.04 dB  
 Peak SAR (extrapolated) = 29.0 W/kg  
 SAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.12 W/kg  
 Smallest distance from peaks to all points 3 dB below = 8.9 mm  
 Ratio of SAR at M2 to SAR at M1 = 47.5%  
 Maximum value of SAR (measured) = 23.4 W/kg

0 dB = 23.4 W/kg = 13.69 dBW/kg

Certificate No: Z22-60108 Page 5 of 6

**TTL Speaq** CALIBRATION LABORATORY  
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Impedance Measurement Plot for Head TSL**

Certificate No: Z22-60108 Page 6 of 6

## 1.12 D5GHzV2 - SN 1095

**TTL Speaq** CALIBRATION LABORATORY  
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60187**

**CALIBRATION CERTIFICATE**

Object: D5GHzV2 - SN: 1095

Calibration Procedure(s): FF-Z11-003-01  
 Calibration Procedures for dipole validation kits

Calibration date: June 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity < 70%.

Calibration Equipment used (MTE critical for calibration)

Primary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Power Meter NRP2	106277	24-Sep-21 (CTTL No. J211008326)	Sep-22
Power sensor NRP8S	104291	24-Sep-21 (CTTL No. J211008326)	Sep-22
Reference Probe EX3DV4	SN 7464	26-Jan-22(SPEAG No. EX3-7464_Jan22)	Jan-23
DAE4	SN 1556	12-Jan-22(CTTL-SPEAG No. Z22-60007)	Jan-23

Secondary Standards	ID #	Cal Date (Calibrated by: Certificate No.)	Scheduled Calibration
Signal Generator E4438C	MY49071430	13-Jan-22 (CTTL No. J22X00406)	Jan-23
Network Analyzer E5071C	MY46110673	14-Jan-22 (CTTL No. J22X00406)	Jan-23

Calibrated by: Zhao Jing, SAR Test Engineer  
 Reviewed by: Lin Hao, SAR Test Engineer  
 Approved by: Qi Dianyan, SAR Project Leader

Issued: June 6, 2022  
 This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

**TTL Speaq** CALIBRATION LABORATORY  
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
 Tel: +86-10-62104633-2079 Fax: +86-10-62104633-2504  
 E-mail: cti@china.ttl.com http://www.china.ttl.com

**Glossary:**

TSL: Issue simulating liquid  
 ConvF: sensitivity in TSL; NORMx.y.z  
 N/A: not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 665664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/G System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 2 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No. 10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com

In Collaboration with  
**TTL**  
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62382117  
E-mail: csl@ttsl.com.cn

**Measurement Conditions**  
DASY system configuration, as for as not given on page 1.

DASY Version	DASY52	52.10.4
Extrapolation	Advanced Extrapolation	
Phantom	Triple Flat Phantom 5.1C	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy = 4 mm, dz = 1.4 mm	Graded Ratio = 1.4 (Z direction)
Frequency	5200 MHz ± 1 MHz 5200 MHz ± 1 MHz 5500 MHz ± 1 MHz 5800 MHz ± 1 MHz	

**Head TSL parameters at 5200MHz**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.0	4.66 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.4 ± 6 %	4.82 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL at 5200MHz**

SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	7.79 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	17.6 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.22 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.1 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 3 of 10

In Collaboration with  
**TTL**  
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62382117  
E-mail: csl@ttsl.com.cn

**Head TSL parameters at 5300MHz**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.9	4.76 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	35.2 ± 6 %	4.73 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL at 5300MHz**

SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.94 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	79.1 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.27 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.6 W/kg ± 24.2 % (k=2)

**Head TSL parameters at 5500MHz**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.6	4.96 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.8 ± 6 %	4.94 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL at 5500MHz**

SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.29 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	82.6 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.34 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	23.3 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 4 of 10

In Collaboration with  
**TTL**  
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62382117  
E-mail: csl@ttsl.com.cn

**Head TSL parameters at 5600MHz**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.5	5.07 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.7 ± 6 %	5.05 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL at 5600MHz**

SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	8.12 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	80.8 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.30 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	22.9 W/kg ± 24.2 % (k=2)

**Head TSL parameters at 5800MHz**  
The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	35.3	5.27 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	34.4 ± 6 %	5.25 mho/m ± 6 %
Head TSL temperature change during test	<1.0 °C	---	---

**SAR result with Head TSL at 5800MHz**

SAR averaged over 1 cm <sup>2</sup> (1 g) of Head TSL	Condition	
SAR measured	100 mW input power	7.71 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	76.7 W/kg ± 24.4 % (k=2)
SAR averaged over 10 cm <sup>2</sup> (10 g) of Head TSL	Condition	
SAR measured	100 mW input power	2.16 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.8 W/kg ± 24.2 % (k=2)

Certificate No: Z22-60187 Page 5 of 10

In Collaboration with  
**TTL**  
CALIBRATION LABORATORY

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China  
Tel: +86-10-62382117  
E-mail: csl@ttsl.com.cn

**Appendix (Additional assessments outside the scope of CNAS L0570)**

**Antenna Parameters with Head TSL at 5200MHz**

Impedance, transformed to feed point	46.10-5.03jΩ
Return Loss	-23.6dB

**Antenna Parameters with Head TSL at 5300MHz**

Impedance, transformed to feed point	47.80-2.42jΩ
Return Loss	-28.5dB

**Antenna Parameters with Head TSL at 5500MHz**

Impedance, transformed to feed point	50.30-4.26jΩ
Return Loss	-27.4dB

**Antenna Parameters with Head TSL at 5600MHz**

Impedance, transformed to feed point	54.50-4.80jΩ
Return Loss	-24.0dB

**Antenna Parameters with Head TSL at 5800MHz**



Impedance, transformed to feed point	51.50-5.61jΩ
Return Loss	-24.9dB

Certificate No: Z22-60187 Page 6 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300    t(86-512) 5735 5888    f(86-512) 57370818    www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号    邮编: 215300    t(86-512) 57355888    f(86-512) 57370818    sgs.china@sgs.com

<p>In Collaboration with <b>TTL</b> <b>speaq</b> <b>CAICT</b>  <b>CALIBRATION LABORATORY</b></p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China          Tel: +86-10-62302117          E-mail: <a href="mailto:ott@tstnet.com">ott@tstnet.com</a> <a href="http://www.caict.ac.cn">http://www.caict.ac.cn</a></p> <p><b>General Antenna Parameters and Design</b></p> <p>Electrical Delay (one direction): 1.101 ns</p> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.          No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p><b>Additional EUT Data</b></p> <p>Manufactured by: SPEAG</p> <p>Certificate No: Z22-60187 Page 7 of 10</p>	<p>In Collaboration with <b>TTL</b> <b>speaq</b> <b>CAICT</b>  <b>CALIBRATION LABORATORY</b></p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China          Tel: +86-10-62302117          E-mail: <a href="mailto:ott@tstnet.com">ott@tstnet.com</a> <a href="http://www.caict.ac.cn">http://www.caict.ac.cn</a></p> <p><b>DASY5 Validation Report for Head TSL</b></p> <p>Test Laboratory: CTTL, Beijing, China          Date: 2022-06-01</p> <p><b>DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095</b></p> <p>Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1          Medium parameters used: f = 5200 MHz; <math>\sigma = 4.62</math> S/m; <math>\epsilon_r = 35.19</math>; <math>\rho = 1000</math> kg/m<sup>3</sup>          Medium parameters used: f = 5300 MHz; <math>\sigma = 4.73</math> S/m; <math>\epsilon_r = 35.19</math>; <math>\rho = 1000</math> kg/m<sup>3</sup>          Medium parameters used: f = 5500 MHz; <math>\sigma = 4.939</math> S/m; <math>\epsilon_r = 34.83</math>; <math>\rho = 1000</math> kg/m<sup>3</sup>          Medium parameters used: f = 5600 MHz; <math>\sigma = 5.051</math> S/m; <math>\epsilon_r = 34.88</math>; <math>\rho = 1000</math> kg/m<sup>3</sup>          Medium parameters used: f = 5800 MHz; <math>\sigma = 5.247</math> S/m; <math>\epsilon_r = 34.42</math>; <math>\rho = 1000</math> kg/m<sup>3</sup>          Phantom section: Right Section          Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)          DASY5 Configuration:</p> <ul style="list-style-type: none"> <li>Probe: EX3DV4 - SN7484; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26</li> <li>Sensor-Surface: 1.4mm (Mechanical Surface Detection)</li> <li>Electronics: DA64 Sn1556; Calibrated: 2022-01-12</li> <li>Phantom: MFP_V5.1C (20deg probe tilt); Type: GD Q00 P51 Cx; Serial: 1062</li> <li>DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)</li> </ul> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b>          Reference Value = 60.80 V/m; Power Drift = -0.06 dB          Peak SAR (extrapolated) = 29.8 W/kg          SAR(1 g) = 7.73 W/kg; SAR(10 g) = 2.22 W/kg          Smallest distance from peaks to all points 3 dB below = 7.2 mm          Ratio of SAR at M2 to SAR at M1 = 66.8%          Maximum value of SAR (measured) = 18.3 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b>          Reference Value = 61.08 V/m; Power Drift = -0.07 dB          Peak SAR (extrapolated) = 31.5 W/kg          SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg          Smallest distance from peaks to all points 3 dB below = 7.2 mm          Ratio of SAR at M2 to SAR at M1 = 65.5%          Maximum value of SAR (measured) = 19.0 W/kg</p> <p>Certificate No: Z22-60187 Page 8 of 10</p>
<p>In Collaboration with <b>TTL</b> <b>speaq</b> <b>CAICT</b>  <b>CALIBRATION LABORATORY</b></p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China          Tel: +86-10-62302117          E-mail: <a href="mailto:ott@tstnet.com">ott@tstnet.com</a> <a href="http://www.caict.ac.cn">http://www.caict.ac.cn</a></p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b>          Reference Value = 61.92 V/m; Power Drift = -0.08 dB          Peak SAR (extrapolated) = 34.7 W/kg          SAR(1 g) = 8.28 W/kg; SAR(10 g) = 2.34 W/kg          Smallest distance from peaks to all points 3 dB below = 7.2 mm          Ratio of SAR at M2 to SAR at M1 = 63.9%          Maximum value of SAR (measured) = 20.2 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b>          Reference Value = 65.08 V/m; Power Drift = -0.07 dB          Peak SAR (extrapolated) = 35.2 W/kg          SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg          Smallest distance from peaks to all points 3 dB below = 7.2 mm          Ratio of SAR at M2 to SAR at M1 = 62.5%          Maximum value of SAR (measured) = 19.1 W/kg</p> <p><b>Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm</b>          Reference Value = 62.13 V/m; Power Drift = -0.06 dB          Peak SAR (extrapolated) = 34.8 W/kg          SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg          Smallest distance from peaks to all points 3 dB below = 7.2 mm          Ratio of SAR at M2 to SAR at M1 = 61.6%          Maximum value of SAR (measured) = 18.7 W/kg</p>  <p>Certificate No: Z22-60187 Page 9 of 10</p>	<p>In Collaboration with <b>TTL</b> <b>speaq</b> <b>CAICT</b>  <b>CALIBRATION LABORATORY</b></p> <p>Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China          Tel: +86-10-62302117          E-mail: <a href="mailto:ott@tstnet.com">ott@tstnet.com</a> <a href="http://www.caict.ac.cn">http://www.caict.ac.cn</a></p> <p><b>Impedance Measurement Plot for Head TSL</b></p>  <p>Certificate No: Z22-60187 Page 10 of 10</p>



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com



## 2 DAE4 - SN 1245

<p style="text-align: center;"><b>s p e a g</b></p> <p style="font-size: small;">Schmid &amp; Partner Engineering AG Zugstrasse 10, 8004 Zurich, Switzerland Phone +41 44 248 9700, Fax +41 44 248 9770 www.sgs.com, info@speag.com</p> <p style="text-align: center; color: red;"><b>IMPORTANT NOTICE</b></p> <p><b>USAGE OF THE DAE4</b></p> <p>The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:</p> <p><b>Battery Exchange:</b> The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.</p> <p><b>Shipping of the DAE:</b> Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.</p> <p><b>E-Stop Failures:</b> Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failures, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.</p> <p><b>Repair:</b> Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.</p> <p><b>DASY Configuration Files:</b> Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.</p> <p><b>Important Note:</b> Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.</p> <p><b>Important Note:</b> Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.</p> <p><b>Important Note:</b> To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.</p> <p style="font-size: x-small;">TN_EH160306AE_DAE4.docx <span style="float: right;">07.03.2019</span></p>	<p style="text-align: center;"><b>Calibration Laboratory of Schmid &amp; Partner Engineering AG</b> Zugstrasse 10, 8004 Zurich, Switzerland</p> <p style="text-align: center;"> </p> <p style="font-size: x-small;">Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <p>Client: <b>SGS-CN (Auden)</b> Certificate No.: <b>DAE4-1245_May22</b></p> <p style="text-align: center;"><b>CALIBRATION CERTIFICATE</b></p> <p>Object: <b>DAE4 - SD 000 D04 BM - SN: 1245</b></p> <p>Calibration procedure(s): <b>QA CAL-06 v30</b> Calibration procedure for the data acquisition electronics (DAE)</p> <p>Calibration date: <b>May 30, 2022</b></p> <p style="font-size: x-small;">This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility, environment temperature (22 ± 3)°C and humidity &lt; 70%.</p> <p style="font-size: x-small;">Calibration Equipment used (M&amp;E critical for calibration)</p> <table border="1" style="width: 100%; font-size: x-small;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Kelway Multimeter Type 2001</td> <td>SN: 0810276</td> <td>31-Aug-21 (No:31368)</td> <td>Aug-22</td> </tr> </tbody> </table> <table border="1" style="width: 100%; font-size: x-small;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Auto DAE Calibration Unit</td> <td>SE LWS 003 AA 1001</td> <td>24-Jan-22 (in house check)</td> <td>In house check: Jan-23</td> </tr> <tr> <td>Calibrator class V0.1</td> <td>SE LWS 100 AA 1002</td> <td>24-Jan-22 (in house check)</td> <td>In house check: Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: <b>Dominique Shelten</b> (Name), <b>Laboratory Technician</b> (Function), <i>[Signature]</i> (Signature)</p> <p>Approved by: <b>Steven Kuhn</b> (Name), <b>Technical Manager</b> (Function), <i>[Signature]</i> (Signature)</p> <p style="font-size: x-small;">This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: May 30, 2022</p> <p style="font-size: x-small;">Certificate No: DAE4-1245_May22 <span style="float: right;">Page 1 of 5</span></p>	Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration	Kelway Multimeter Type 2001	SN: 0810276	31-Aug-21 (No:31368)	Aug-22	Secondary Standards	ID #	Check Date (in house)	Scheduled Check	Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check: Jan-23	Calibrator class V0.1	SE LWS 100 AA 1002	24-Jan-22 (in house check)	In house check: Jan-23
Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration																		
Kelway Multimeter Type 2001	SN: 0810276	31-Aug-21 (No:31368)	Aug-22																		
Secondary Standards	ID #	Check Date (in house)	Scheduled Check																		
Auto DAE Calibration Unit	SE LWS 003 AA 1001	24-Jan-22 (in house check)	In house check: Jan-23																		
Calibrator class V0.1	SE LWS 100 AA 1002	24-Jan-22 (in house check)	In house check: Jan-23																		
<p style="text-align: center;"><b>Calibration Laboratory of Schmid &amp; Partner Engineering AG</b> Zugstrasse 10, 8004 Zurich, Switzerland</p> <p style="text-align: center;"> </p> <p style="font-size: x-small;">Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p style="text-align: right;">Accreditation No.: SCS 0108</p> <p><b>Glossary</b></p> <p>DAE: data acquisition electronics</p> <p>Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.</p> <p><b>Methods Applied and Interpretation of Parameters</b></p> <ul style="list-style-type: none"> <li><b>DC Voltage Measurement:</b> Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.</li> <li><b>Connector angle:</b> The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.</li> <li>The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.             <ul style="list-style-type: none"> <li><b>DC Voltage Measurement Linearity:</b> Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.</li> <li><b>Common mode sensitivity:</b> Influence of a positive or negative common mode voltage on the differential measurement.</li> <li><b>Channel separation:</b> Influence of a voltage on the neighbor channels not subject to an input voltage.</li> <li><b>AD Converter Values with inputs shorted:</b> Values on the internal AD converter corresponding to zero input voltage</li> <li><b>Input Offset Measurement:</b> Output voltage and statistical results over a large number of zero voltage measurements.</li> <li><b>Input Offset Current:</b> Typical value for information; Maximum channel input offset current, not considering the input resistance.</li> <li><b>Input resistance:</b> Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement.</li> <li><b>Low Battery Alarm Voltage:</b> Typical value for information. Below this voltage, a battery alarm signal is generated.</li> <li><b>Power consumption:</b> Typical value for information. Supply currents in various operating modes.</li> </ul> </li> </ul> <p style="font-size: x-small;">Certificate No: DAE4-1245_May22 <span style="float: right;">Page 2 of 5</span></p>	<p><b>DC Voltage Measurement</b></p> <p>AD - Converter Resolution nominal</p> <p>High Range: 1LSB = 6.1µV, full range = -10...+30 mV</p> <p>Low Range: 1LSB = 61µV, full range = -1...+3mV</p> <p>DASY measurement parameters: Auto Zero-Time: 3 sec; Measuring time: 3 sec</p> <table border="1" style="width: 100%; font-size: x-small;"> <thead> <tr> <th>Calibration Factors</th> <th>X</th> <th>Y</th> <th>Z</th> </tr> </thead> <tbody> <tr> <td>High Range</td> <td>405.265 ± 0.02% (k=2)</td> <td>403.974 ± 0.02% (k=2)</td> <td>406.092 ± 0.02% (k=2)</td> </tr> <tr> <td>Low Range</td> <td>3.99534 ± 1.50% (k=2)</td> <td>3.99508 ± 1.50% (k=2)</td> <td>4.01015 ± 1.50% (k=2)</td> </tr> </tbody> </table> <p><b>Connector Angle</b></p> <table border="1" style="width: 100%; font-size: x-small;"> <tr> <td>Connector Angle to be used in DASY system</td> <td>30.0° ± 1°</td> </tr> </table> <p style="font-size: x-small;">Certificate No: DAE4-1245_May22 <span style="float: right;">Page 3 of 5</span></p>	Calibration Factors	X	Y	Z	High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)	Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)	Connector Angle to be used in DASY system	30.0° ± 1°						
Calibration Factors	X	Y	Z																		
High Range	405.265 ± 0.02% (k=2)	403.974 ± 0.02% (k=2)	406.092 ± 0.02% (k=2)																		
Low Range	3.99534 ± 1.50% (k=2)	3.99508 ± 1.50% (k=2)	4.01015 ± 1.50% (k=2)																		
Connector Angle to be used in DASY system	30.0° ± 1°																				



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300    t(86-512) 5735 5888    f(86-512) 57370818    www.sgsgroup.com.cn  
 中国·江苏·昆山开发区伟业路10号    邮编: 215300    t(86-512) 57355888    f(86-512) 57370818    sgs.china@sgs.com

**Appendix (Additional assessments outside the scope of SCS0108)**

**1. DC Voltage Linearity**

High Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	19994.45	1.52	0.00
Channel X + Input	20004.58	2.22	0.01
Channel X - Input	-20000.14	1.12	-0.01
Channel Y + Input	199994.72	1.58	0.00
Channel Y + Input	20001.22	-1.00	-0.00
Channel Y - Input	-20003.05	-1.87	0.01
Channel Z + Input	199992.84	0.19	0.00
Channel Z + Input	20003.09	0.58	0.00
Channel Z - Input	-20001.73	-0.27	0.00

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2001.91	0.41	0.02
Channel X + Input	202.54	0.65	0.32
Channel X - Input	-197.86	0.07	-0.04
Channel Y + Input	2002.05	0.58	0.03
Channel Y + Input	201.27	-0.57	-0.28
Channel Y - Input	-199.23	-0.06	0.03
Channel Z + Input	2001.36	0.08	0.00
Channel Z + Input	200.09	-1.53	-0.76
Channel Z - Input	-199.89	-1.57	0.79

**2. Common mode sensitivity**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Common mode Input Voltage (mV)	High Range Average Reading (µV)	Low Range Average Reading (µV)
Channel X	-3.87	-7.69
-200	9.12	7.79
Channel Y	-8.68	-9.28
-200	8.52	6.36
Channel Z	-5.36	-5.80
-200	3.58	3.08

**3. Channel separation**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)	
Channel X	200	-	4.07	-3.14
Channel Y	200	9.36	-	4.27
Channel Z	200	10.11	7.14	-

Certificate No: DAE4-1245\_May22 Page 4 of 5

**4. AD-Converter Values with inputs shorted**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	15984	17040
Channel Y	16562	15768
Channel Z	16035	15658

**5. Input Offset Measurement**

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10MΩ

	Average (µV)	min. Offset (µV)	max. Offset (µV)	Std. Deviation (µV)
Channel X	1.00	-0.15	1.93	0.45
Channel Y	-0.18	-1.28	0.94	0.45
Channel Z	-0.58	-2.61	0.58	0.60

**6. Input Offset Current**

Nominal input circuitry offset current on all channels: <math>-25\mu A</math>

**7. Input Resistance (Typical values for information)**

	Zeroing (Ω)	Measuring (MΩ)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

**8. Low Battery Alarm Voltage (Typical values for information)**

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

**9. Power Consumption (Typical values for information)**

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	-0.01	+8	+14
Supply (- Vcc)	-0.01	-8	-9

Certificate No: DAE4-1245\_May22 Page 5 of 5

### 3 EX3DV4 - SN 7346

Calibration Laboratory of Schmid & Partner Engineering AG  
Zugstrasse 61, 8048 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **Auden** Certificate No: **EX3-7346\_Mar22**

**CALIBRATION CERTIFICATE**

Object: **EX3DV4 - SN 7346**

Calibration procedure(s): **QA CAL-01 v6; QA CAL-14 v6; QA CAL-23 v5; QA CAL-25 v7**  
Calibration procedure for dosimetric E-field probes

Calibration date: **March 30, 2022**

This calibration certificate documents the traceability to national standards, which make the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the stated laboratory facility, environment temperature (22 ± 3)°C and humidity <math>= 70\%</math>

Calibration Equipment used (MATE critical for calibration):

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NMP	SR 10078	09-Apr-21 (No. 211-0201-0202)	Apr-22
Power sensor NMP-291	SR 10324	09-Apr-21 (No. 211-0201)	Apr-22
Power sensor NMP-291	SR 10324	09-Apr-21 (No. 211-0202)	Apr-22
Reference 20dB attenuator	SR C2252 (20)	09-Apr-21 (No. 211-0204)	Apr-22
DASA	SR 460	13-Oct-21 (No. DAE4-468_0401)	Oct-22
Reference Probe ES302	SR 3013	07-Oct-21 (No. ES3-3013_0401)	Dec-22

Secondary Standards

ID	Check Date (in house)	Scheduled Check	
Power meter E4415B	SR G24128304	08-Apr-22 (in house check Jun-23)	In house check Jun-22
Power sensor E4412A	SR MY4149897	08-Apr-22 (in house check Jun-23)	In house check Jun-22
Power sensor E4412A	SR 40011016	08-Apr-22 (in house check Jun-23)	In house check Jun-22
RF generator HP 8646C	SR US340101700	04-Apr-22 (in house check Jun-23)	In house check Jun-22
Network Analyzer E8363A	SR US41809477	31-Mar-14 (in house check Oct-20)	In house check Oct-22

Calibrated by: **Benjamin** Function: **Laboratory Technician** Signature: *[Signature]*

Approved by: **Benjamin** Function: **Quality Manager** Signature: *[Signature]*

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: EX3-7346\_Mar22 Page 1 of 24

Calibration Laboratory of Schmid & Partner Engineering AG  
Zugstrasse 61, 8048 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **Auden** Certificate No: **EX3-7346\_Mar22**

**CALIBRATION CERTIFICATE**

Object: **EX3DV4 - SN 7346**

Calibration procedure(s): **QA CAL-01 v6; QA CAL-14 v6; QA CAL-23 v5; QA CAL-25 v7**  
Calibration procedure for dosimetric E-field probes

Calibration date: **March 30, 2022**

This calibration certificate documents the traceability to national standards, which make the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the stated laboratory facility, environment temperature (22 ± 3)°C and humidity <math>= 70\%</math>

Calibration Equipment used (MATE critical for calibration):

Primary Standards	ID	Cal Date (Certificate No.)	Scheduled Calibration
Power meter NMP	SR 10078	09-Apr-21 (No. 211-0201-0202)	Apr-22
Power sensor NMP-291	SR 10324	09-Apr-21 (No. 211-0201)	Apr-22
Power sensor NMP-291	SR 10324	09-Apr-21 (No. 211-0202)	Apr-22
Reference 20dB attenuator	SR C2252 (20)	09-Apr-21 (No. 211-0204)	Apr-22
DASA	SR 460	13-Oct-21 (No. DAE4-468_0401)	Oct-22
Reference Probe ES302	SR 3013	07-Oct-21 (No. ES3-3013_0401)	Dec-22

Secondary Standards

ID	Check Date (in house)	Scheduled Check	
Power meter E4415B	SR G24128304	08-Apr-22 (in house check Jun-23)	In house check Jun-22
Power sensor E4412A	SR MY4149897	08-Apr-22 (in house check Jun-23)	In house check Jun-22
Power sensor E4412A	SR 40011016	08-Apr-22 (in house check Jun-23)	In house check Jun-22
RF generator HP 8646C	SR US340101700	04-Apr-22 (in house check Jun-23)	In house check Jun-22
Network Analyzer E8363A	SR US41809477	31-Mar-14 (in house check Oct-20)	In house check Oct-22

Calibrated by: **Benjamin** Function: **Laboratory Technician** Signature: *[Signature]*

Approved by: **Benjamin** Function: **Quality Manager** Signature: *[Signature]*

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: EX3-7346\_Mar22 Page 2 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 | (86-512) 5735 5888 | (86-512) 57370818 | www.sgsgroup.com.cn  
中国·江苏·昆山开发区伟业路10号 | 邮编: 215300 | (86-512) 57355888 | (86-512) 57370818 | sgs.china@sgs.com

EX3DV4- SN:7346 March 30, 2022

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

**Basic Calibration Parameters**

Norm. (uV/VmV) <sup>1</sup>	Sensor X	Sensor Y	Sensor Z	Unc. (k=2)
DCP (mV/V) <sup>2</sup>	0.48	0.47	0.61	± 10.1 %
	101.4	106.0	106.9	

**Calibration Results for Modulation Response**

UID	Communication System Name	A	B	C	D	VR	Max	Max
		dB	dB	dB	dB	mV	dir.	Used
		dB	dB	dB	dB	mV		(N=2)
0	CW	X: 0.00	0.00	1.00	0.00	143.5	± 3.0 %	
		Y: 0.00	0.00	1.00	0.00	139.3		
		Z: 0.00	0.00	1.00	0.00	139.0		
10035-AAA	Pulse Waveform (200Hz, 10%)	X: 3.33	68.90	11.66	10.00	66.0	± 3.5 %	± 9.6 %
		Y: 4.03	79.70	12.35	10.00	66.0		
		Z: 1.63	61.25	6.76	10.00	66.0		
10035-AAA	Pulse Waveform (200Hz, 20%)	X: 3.00	79.65	11.31	6.99	66.0	± 2.4 %	± 9.6 %
		Y: 11.31	81.32	14.72	10.00	66.0		
		Z: 9.83	69.00	5.11	10.00	66.0		
10035-AAA	Pulse Waveform (200Hz, 40%)	X: 7.41	79.85	12.51	3.98	66.0	± 2.7 %	± 9.6 %
		Y: 26.03	81.62	15.51	10.00	66.0		
		Z: 0.18	138.38	0.01	10.00	66.0		
10035-AAA	Pulse Waveform (200Hz, 60%)	X: 2.27	75.13	9.54	2.22	120.0	± 1.1 %	± 9.6 %
		Y: 20.90	91.58	16.29	10.00	120.0		
		Z: 7.54	126.51	16.67	10.00	120.0		
10037-AAA	QPSK Waveform, 1 MHz	X: 1.47	64.88	13.82	1.00	150.0	± 4.2 %	± 9.6 %
		Y: 1.56	66.27	14.65	0.00	150.0		
		Z: 0.45	65.88	11.05	10.00	150.0		
10038-AAA	QPSK Waveform, 10 MHz	X: 1.56	66.27	14.65	0.00	150.0	± 1.1 %	± 9.6 %
		Y: 2.06	67.33	13.38	10.00	150.0		
		Z: 2.1	64.75	13.38	10.00	150.0		
10036-AAA	64-QAM Waveform, 100 MHz	X: 2.63	68.51	18.25	3.01	150.0	± 1.0 %	± 9.6 %
		Y: 2.74	70.83	19.01	10.00	150.0		
		Z: 1.70	64.72	15.99	10.00	150.0		
10038-AAA	64-QAM Waveform, 40 MHz	X: 3.38	66.39	15.25	0.00	150.0	± 2.0 %	± 9.6 %
		Y: 3.42	68.81	16.01	10.00	150.0		
		Z: 2.70	65.72	14.74	10.00	150.0		
10041-AAA	WiLAN CCDF, 64-QAM, 20MHz	X: 65.35	15.77	0.00	10.00	150.0	± 3.6 %	± 9.6 %
		Y: 4.70	65.54	15.41	10.00	150.0		
		Z: 3.63	66.16	15.28	10.00	150.0		

Note: For details on UID parameters see Appendix.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

\* The uncertainties of Norm. X, Y, Z do not affect the E<sub>1</sub> field uncertainty under T11, see Pages 5 and 6.  
 \* Numerical simulation parameter, uncertainty not required.  
 \* Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7346\_Mar22 Page 3 of 34

EX3DV4- SN:7346 March 30, 2022

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

**Sensor Model Parameters**

T1	C2	a	T1	T2	T3	T4	T5	T6
IP	IP	V <sup>1</sup>	ms.V <sup>1</sup>	ms.V <sup>1</sup>	ms	V <sup>1</sup>	V <sup>1</sup>	V <sup>1</sup>
X	39.2	291.80	35.10	5.63	5.02	1.42	0.12	1.01
Y	37.1	270.84	34.12	6.29	0.00	5.01	1.82	0.05
Z	9.7	69.74	33.37	4.96	0.00	4.94	0.61	0.00

**Other Probe Parameters**

Sensor Arrangement	Triangular
Connector Angle (°)	-166.1
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	237 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.

Certificate No: EX3-7346\_Mar22 Page 4 of 34

EX3DV4- SN:7346 March 30, 2022

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

**Calibration Parameter Determined in Head Tissue Simulating Media**

f (MHz)	Relative Permittivity <sup>1</sup>	Conductivity (S/m) <sup>2</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>3</sup>	Depth <sup>4</sup> (mm)	Unc. (k=2)
750	41.9	0.69	10.56	10.56	10.56	0.55	0.85	± 12.0 %
835	41.5	0.90	10.12	10.12	10.12	0.42	0.96	± 12.0 %
900	41.5	0.97	10.10	10.10	10.10	0.53	0.80	± 12.0 %
1450	40.5	1.20	9.26	9.26	9.26	0.50	0.80	± 12.0 %
1750	40.1	1.37	8.83	8.83	8.83	0.34	0.86	± 12.0 %
1900	40.0	1.40	8.48	8.48	8.48	0.35	0.95	± 12.0 %
2000	40.0	1.40	8.35	8.35	8.35	0.34	0.86	± 12.0 %
2300	39.5	1.67	7.86	7.86	7.86	0.39	0.90	± 12.0 %
2450	39.2	1.80	7.63	7.63	7.63	0.41	0.90	± 12.0 %
2600	39.0	1.96	7.33	7.33	7.33	0.44	0.90	± 12.0 %
3300	38.2	2.71	7.15	7.15	7.15	0.30	1.35	± 13.1 %
3500	37.9	2.91	7.14	7.14	7.14	0.30	1.35	± 13.1 %
3700	37.7	3.12	6.85	6.85	6.85	0.30	1.35	± 13.1 %
3900	37.5	3.32	6.71	6.71	6.71	0.40	1.60	± 13.1 %
4100	37.2	3.53	6.58	6.58	6.58	0.40	1.60	± 13.1 %
4200	37.1	3.63	6.30	6.30	6.30	0.40	1.70	± 13.1 %
4400	36.9	3.84	6.24	6.24	6.24	0.40	1.70	± 13.1 %
4600	36.7	4.04	6.11	6.11	6.11	0.40	1.70	± 13.1 %
4800	36.4	4.25	6.08	6.08	6.08	0.40	1.80	± 13.1 %
4900	36.3	4.40	5.84	5.84	5.84	0.40	1.80	± 13.1 %
5200	36.0	4.66	5.25	5.25	5.25	0.40	1.80	± 13.1 %
5300	35.9	4.78	5.12	5.12	5.12	0.40	1.80	± 13.1 %
5500	35.6	4.98	4.85	4.85	4.85	0.40	1.80	± 13.1 %
5800	35.5	5.07	4.70	4.70	4.70	0.40	1.80	± 13.1 %
5900	35.3	5.27	4.75	4.75	4.75	0.40	1.80	± 13.1 %

<sup>1</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the Conf. uncertainty at calibration frequency and the uncertainty for the measured frequency and the uncertainty for the target tissue parameters.  
<sup>2</sup> At frequencies 0-10 GHz, the validity of tissue parameters (ρ and σ) can be related to ± 10% if liquid compensation formula is applied to measured data values. The uncertainty is the RSS of the Conf. uncertainty for indicated target tissue parameters.  
<sup>3</sup> AlphaDepth are determined during calibration. SPS-AD warns that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz, below ± 2% for frequencies between 3-6 GHz, and below ± 4% for frequencies between 6-10 GHz at any distance larger than half the probe diameter from the boundary.  
<sup>4</sup> AlphaDepth are determined during calibration. SPS-AD warns that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe diameter from the boundary.

Certificate No: EX3-7346\_Mar22 Page 5 of 34

EX3DV4- SN:7346 March 30, 2022

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

**Calibration Parameter Determined in Head Tissue Simulating Media**

f (MHz)	Relative Permittivity <sup>1</sup>	Conductivity (S/m) <sup>2</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>3</sup>	Depth <sup>4</sup> (mm)	Unc. (k=2)
6500	34.5	6.07	5.30	5.30	5.30	0.20	2.50	± 18.6 %

<sup>1</sup> Frequency validity above 60 MHz is ± 700 MHz. The uncertainty is the RSS of the Conf. uncertainty at calibration frequency and the uncertainty for the measured frequency and the uncertainty for the target tissue parameters.  
<sup>2</sup> At frequencies 0-10 GHz, the validity of tissue parameters (ρ and σ) can be related to ± 10% if liquid compensation formula is applied to measured data values. The uncertainty is the RSS of the Conf. uncertainty for indicated target tissue parameters.  
<sup>3</sup> AlphaDepth are determined during calibration. SPS-AD warns that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz, below ± 2% for frequencies between 3-6 GHz, and below ± 4% for frequencies between 6-10 GHz at any distance larger than half the probe diameter from the boundary.  
<sup>4</sup> AlphaDepth are determined during calibration. SPS-AD warns that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe diameter from the boundary.


Certificate No: EX3-7346\_Mar22 Page 6 of 34



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiyu Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com





Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
 No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 57370818 www.sgs.com.cn  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 57355888 f(86-512) 57370818 sgs.china@sgs.com



EX30V4-SN-7346 March 30, 2022

Appendix: Modulation Calibration Parameters

UID	Rev	Communication System Name	Group	FAR (%)	Unc (%)
0		CAV	CW	0.00	4.71
10015	CA	SAR Validation (Square, 100MHz, 10ms)	Test	10.00	0.00
10011	CAB	LIMIT-FDD (WCDMA)	WCDMA	2.91	0.67
10012	CAB	IEEE 802.11a W/F 2.4 GHz (DSSS, 1 Mbps)	WLAN	1.87	0.57
10013	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 6 Mbps)	WLAN	9.46	0.97
10021	DAC	GSMA-FDD (TDMA, GSM)	GSM	9.39	0.97
10023	DAC	GSMA-FDD (TDMA, GSM, TW)	GSM	9.27	0.97
10024	DAC	GSMA-FDD (TDMA, GSM, TW 0-1)	GSM	6.56	0.97
10025	DAC	EDGE-FDD (TDMA, BPSK, TW 0)	GSM	12.62	0.97
10026	DAC	EDGE-FDD (TDMA, BPSK, TW 0-1)	GSM	9.45	0.97
10027	DAC	GSMA-FDD (TDMA, GSM, TW 0-1)	GSM	4.80	0.97
10028	DAC	GSMA-FDD (TDMA, GSM, TW 0-2)	GSM	3.55	0.97
10029	DAC	EDGE-FDD (TDMA, BPSK, TW 0-1)	GSM	7.79	0.97
10030	CAA	IEEE 802.15.1 Bluetooth (GFSK, DSSS)	Bluetooth	5.20	0.97
10031	CAA	IEEE 802.15.1 Bluetooth (GFSK, DSSS)	Bluetooth	1.87	0.97
10032	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, DSSS)	Bluetooth	1.16	0.97
10033	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, DSSS)	Bluetooth	7.74	0.97
10034	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, DSSS)	Bluetooth	4.53	0.97
10035	CAA	IEEE 802.15.1 Bluetooth (P4-QPSK, DSSS)	Bluetooth	2.83	0.97
10036	CAA	IEEE 802.15.1 Bluetooth (8-PSK, DSSS)	Bluetooth	8.01	0.97
10037	CAA	IEEE 802.15.1 Bluetooth (8-PSK, DSSS)	Bluetooth	4.77	0.97
10038	CAA	IEEE 802.15.1 Bluetooth (8-PSK, DSSS)	Bluetooth	4.33	0.97
10039	CAB	CDMA2000 (1XRTT, RCT)	CDMA2000	4.57	0.97
10040	CAB	IS-136 (IS-136 FDD/FDMA, P4-QPSK, Full Rate)	AMPS	7.78	0.97
10044	CAA	IS-97A/ETIA-833 (FDD/FDMA, Full Rate)	AMPS	0.80	0.97
10045	CAB	IS-97A/ETIA-833 (FDD/FDMA, Full Rate)	AMPS	13.80	0.97
10046	CAB	IS-97A/ETIA-833 (FDD/FDMA, Full Rate)	AMPS	19.79	0.97
10049	CAA	IS-136 (IS-136 FDD/FDMA, P4-QPSK, DSSS, 3/12)	IS-136	11.01	0.97
10056	CAA	LIMIT-FDD (TD-SCDMA, 1.28 Mbps)	TD-SCDMA	11.01	0.97
10058	CAB	EDGE-FDD (TDMA, BPSK, TW 0-1)	GSM	6.52	0.97
10059	CAB	IEEE 802.11a W/F 2.4 GHz (DSSS, 2 Mbps)	WLAN	2.12	0.97
10060	CAB	IEEE 802.11a W/F 2.4 GHz (DSSS, 5.5 Mbps)	WLAN	2.83	0.97
10061	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, 11 Mbps)	WLAN	9.46	0.97
10062	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 6 Mbps)	WLAN	8.68	0.97
10063	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 3 Mbps)	WLAN	8.63	0.97
10064	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 12 Mbps)	WLAN	9.09	0.97
10065	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 18 Mbps)	WLAN	9.09	0.97
10066	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 24 Mbps)	WLAN	9.38	0.97
10067	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 36 Mbps)	WLAN	10.12	0.97
10068	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 48 Mbps)	WLAN	10.44	0.97
10069	CAB	IEEE 802.11n W/F 5 GHz (OFDM, 54 Mbps)	WLAN	10.58	0.97
10071	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 9 Mbps)	WLAN	6.82	0.97
10072	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 12 Mbps)	WLAN	9.42	0.97
10073	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 18 Mbps)	WLAN	9.34	0.97
10074	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 24 Mbps)	WLAN	10.30	0.97
10075	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 36 Mbps)	WLAN	10.77	0.97
10076	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 48 Mbps)	WLAN	10.84	0.97
10077	CAB	IEEE 802.11g W/F 2.4 GHz (DSSS, OFDM, 54 Mbps)	WLAN	11.00	0.97
10081	CAB	CDMA2000 (1XRTT, RCT)	CDMA2000	3.97	0.97
10082	CAB	IS-136 (IS-136 FDD/FDMA, P4-QPSK, Full Rate)	AMPS	4.77	0.97
10090	DAC	GSMA-FDD (TDMA, GSM, TW 0-4)	GSM	6.56	0.97
10091	CAB	LIMIT-FDD (WCDMA)	WCDMA	3.98	0.97
10098	CAB	LIMIT-FDD (HSPA, Subnet 2)	WCDMA	3.98	0.97
10099	DAC	EDGE-FDD (TDMA, BPSK, TW 0-4)	GSM	9.55	0.97

Certificate No: EX3-7346\_Mar22 Page 11 of 24

EX30V4-SN-7346 March 30, 2022

10100	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-FDD	5.87	0.97
10101	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-FDD	6.42	0.97
10102	CAE	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-FDD	8.60	0.97
10103	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.29	0.97
10104	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.97	0.97
10105	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.01	0.97
10106	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	LTE-FDD	5.80	0.97
10107	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	0.97
10110	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	LTE-FDD	5.75	0.97
10111	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	LTE-FDD	6.44	0.97
10112	CAG	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	LTE-FDD	6.59	0.97
10113	CAG	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	LTE-FDD	6.62	0.97
10114	CAD	IEEE 802.11a (HT Overhead, 1.5 Mbps, BPSK)	WLAN	8.30	0.97
10115	CAD	IEEE 802.11a (HT Overhead, 6 Mbps, 16-QAM)	WLAN	8.46	0.97
10116	CAD	IEEE 802.11a (HT Overhead, 12 Mbps, 64-QAM)	WLAN	8.15	0.97
10117	CAD	IEEE 802.11n (HT Mixed, 1.5 Mbps, BPSK)	WLAN	8.07	0.97
10118	CAD	IEEE 802.11n (HT Mixed, 6 Mbps, 16-QAM)	WLAN	8.30	0.97
10119	CAD	IEEE 802.11n (HT Mixed, 12 Mbps, 64-QAM)	WLAN	8.13	0.97
10140	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	LTE-FDD	6.49	0.97
10141	CAE	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	LTE-FDD	6.51	0.97
10142	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	LTE-FDD	5.73	0.97
10143	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-FDD	6.35	0.97
10144	CAE	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-FDD	6.65	0.97
10145	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	LTE-FDD	5.76	0.97
10146	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.41	0.97
10147	CAF	LTE-FDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.72	0.97
10148	CAF	LTE-FDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-FDD	6.43	0.97
10150	CAE	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	LTE-FDD	6.60	0.97
10151	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	LTE-TDD	9.28	0.97
10152	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	LTE-TDD	9.96	0.97
10153	CAG	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	LTE-TDD	10.05	0.97
10154	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-FDD	6.40	0.97
10155	CAG	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-FDD	6.43	0.97
10156	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-FDD	5.79	0.97
10157	CAG	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-FDD	6.42	0.97
10158	CAG	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-FDD	6.62	0.97
10159	CAG	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-FDD	6.68	0.97
10160	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-FDD	5.82	0.97
10161	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-FDD	6.43	0.97
10162	CAE	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-FDD	6.59	0.97
10166	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	5.48	0.97
10167	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.19	0.97
10168	CAF	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.79	0.97
10170	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-FDD	5.73	0.97
10171	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	6.49	0.97
10172	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-FDD	6.49	0.97
10173	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-TDD	9.48	0.97
10174	CAG	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-TDD	10.15	0.97
10175	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	LTE-FDD	5.72	0.97
10176	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	LTE-FDD	6.52	0.97
10177	CAE	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	LTE-FDD	6.52	0.97
10178	CAF	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6.52	0.97
10179	CAF	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	6.56	0.97
10180	CAG	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64-QAM)	LTE-FDD	6.50	0.97
10181	CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	LTE-FDD	5.73	0.97

Certificate No: EX3-7346\_Mar22 Page 12 of 24

EX30V4-SN-7346 March 30, 2022

10182	CAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	LTE-FDD	6.52	0.97
10183	AAE	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	6.50	0.97
10184	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-FDD	5.73	0.97
10185	CAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-FDD	6.50	0.97
10186	AAE	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-FDD	6.50	0.97
10187	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	LTE-FDD	5.87	0.97
10188	CAF	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-FDD	6.52	0.97
10189	AAE	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-FDD	6.50	0.97
10190	CAD	IEEE 802.11n (HT Overhead, 6 Mbps, 16-QAM)	WLAN	8.21	0.97
10191	CAD	IEEE 802.11n (HT Overhead, 12 Mbps, 64-QAM)	WLAN	8.10	0.97
10192	DAC	IEEE 802.11n (HT Mixed, 39 Mbps, 16-QAM)	WLAN	8.13	0.97
10193	CAD	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	WLAN	8.27	0.97
10219	CAF	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	WLAN	8.03	0.97
10220	CAD	IEEE 802.11n (HT Mixed, 14.4 Mbps, 16-QAM)	WLAN	8.13	0.97
10221	CAD	IEEE 802.11n (HT Mixed, 28.8 Mbps, 64-QAM)	WLAN	8.27	0.97
10222	CAD	IEEE 802.11n (HT Mixed, 57.6 Mbps, BPSK)	WLAN	8.06	0.97
10223	CAD	IEEE 802.11n (HT Mixed, 86.4 Mbps, 16-QAM)	WLAN	8.49	0.97
10224	CAD	IEEE 802.11n (HT Mixed, 172.8 Mbps, 64-QAM)	WLAN	8.68	0.97
10225	CAB	LIMIT-FDD (HSPA)	WCDMA	5.97	0.97
10226	CAB	LIMIT-FDD (HSPA)	WCDMA	5.97	0.97
10228	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.48	0.97
10229	CAB	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.28	0.97
10230	CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.22	0.97
10231	CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-TDD	9.48	0.97
10232	CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.19	0.97
10233	CAG	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	LTE-TDD	9.28	0.97
10234	CAG	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16-QAM)	LTE-TDD	9.48	0.97
10235	CAG	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64-QAM)	LTE-TDD	10.25	0.97
10236	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	LTE-TDD	9.21	0.97
10237	CAG	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	LTE-TDD	9.48	0.97
10238	CAF	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	LTE-FDD	10.25	0.97
10239	CAF	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	LTE-FDD	9.21	0.97
10241	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.82	0.97
10242	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	LTE-TDD	9.86	0.97
10243	CAB	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	LTE-TDD	9.48	0.97
10244	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	LTE-TDD	10.08	0.97
10245	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	LTE-TDD	9.91	0.97
10246	CAD	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	LTE-TDD	9.30	0.97
10247	CAD	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	LTE-TDD	9.81	0.97
10248	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	LTE-TDD	10.09	0.97
10249	CAG	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	LTE-TDD	9.29	0.97
10250	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	LTE-TDD	9.81	0.97
10251	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	LTE-TDD	10.17	0.97
10252	CAG	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	LTE-TDD	9.24	0.97
10253	CAF	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	LTE-TDD	9.80	0.97
10254	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	LTE-TDD	10.14	0.97
10255	CAF	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	LTE-TDD	9.20	0.97
10256	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	LTE-TDD	9.96	0.97
10257	CAB	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	LTE-TDD	10.08	0.97
10258	CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-TDD	9.34	0.97
10259	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	LTE-TDD	9.98	0.97
10260	CAD	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	LTE-TDD	9.98	0.97





Table with columns: Item No., Description, Standard, Result, and Date. Includes items like 10414 AAA, 10415 AAA, 10416 AAA, etc.

Certificate No. EX3-7346\_Mar22 Page 15 of 24

Table with columns: Item No., Description, Standard, Result, and Date. Includes items like 10489 AAF, 10490 AAF, 10491 AAF, etc.

Certificate No. EX3-7346\_Mar22 Page 16 of 24

Table with columns: Item No., Description, Standard, Result, and Date. Includes items like 10547 AAC, 10548 AAC, 10549 AAC, etc.

Certificate No. EX3-7346\_Mar22 Page 17 of 24

Table with columns: Item No., Description, Standard, Result, and Date. Includes items like 10625 AAC, 10626 AAC, 10627 AAC, etc.

Certificate No. EX3-7346\_Mar22 Page 18 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at https://www.sgs.com/en/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doecbeck@sgs.com. No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 (86-512) 5735 5888 (86-512) 57370818 www.sgsgroup.com.cn 中国·江苏·昆山开发区伟业路10号 邮编: 215300 (86-512) 57355888 (86-512) 57370818 sgs.china@sgs.com





Table with columns: Item No., Standard, Location, and Result. Includes header information for EX37346 dated March 30, 2022.

Table with columns: Item No., Standard, Location, and Result. Includes header information for EX37346 dated March 30, 2022.

Table with columns: Item No., Standard, Location, and Result. Includes header information for EX37346 dated March 30, 2022.

Table with columns: Item No., Standard, Location, and Result. Includes header information for EX37346 dated March 30, 2022.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at https://www.sgs.com/en/Terms-and-Conditions. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 (86-512) 57355888 (86-512) 57370818 www.sgsgroup.com.cn

EX3D14-SN 7346		March 30, 2022	
15985	AAA	50 NR DL (CP-CPQM, TM 3.1, 40 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.54 ± 9.8 %
15986	AAA	50 NR DL (CP-CPQM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.50 ± 9.8 %
15987	AAA	50 NR DL (CP-CPQM, TM 3.1, 60 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.53 ± 9.8 %
15988	AAA	50 NR DL (CP-CPQM, TM 3.1, 70 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.38 ± 9.8 %
15989	AAA	50 NR DL (CP-CPQM, TM 3.1, 80 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.33 ± 9.8 %
15990	AAA	50 NR DL (CP-CPQM, TM 3.1, 90 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.52 ± 9.8 %

<sup>1</sup> Linearity is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3D14-SN 7346 March 30, 2022

15985	AAA	50 NR DL (CP-CPQM, TM 3.1, 40 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.54 ± 9.8 %
15986	AAA	50 NR DL (CP-CPQM, TM 3.1, 50 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.50 ± 9.8 %
15987	AAA	50 NR DL (CP-CPQM, TM 3.1, 60 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.53 ± 9.8 %
15988	AAA	50 NR DL (CP-CPQM, TM 3.1, 70 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.38 ± 9.8 %
15989	AAA	50 NR DL (CP-CPQM, TM 3.1, 80 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.33 ± 9.8 %
15990	AAA	50 NR DL (CP-CPQM, TM 3.1, 90 MHz, 64-QAM, 30 kHz)	SG NR FR1 TDD 9.52 ± 9.8 %

Continued on: EX3-D14\_S22 Page 27 of 28

Certificate No: EX3-7346\_Mar22 Page 24 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)  
No.10 Weiyue Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512) 5735 5888 f(86-512) 5737 0818 www.sgs.com.cn  
中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512) 5735 5888 f(86-512) 5737 0818 sgs.china@sgs.com

#### 4 Impedance and return loss

Dipole CLA150 SN 4025				
Head Liquid				
Date of Measurement	Return Loss(dB)	$\Delta$ %	Impedance ( $\Omega$ )	$\Delta\Omega$
2021/4/26	-31.4	/	47.8	/
Dipole D450V3 SN 1103				
Head Liquid				
Date of Measurement	Return Loss(dB)	$\Delta$ %	Impedance ( $\Omega$ )	$\Delta\Omega$
2021/4/21	-23	/	57.1	/



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

No.10 Weiye Road, Development Zone, Kunshan, Jiangsu, China 215300 t(86-512)5735 5888 f(86-512)57370818 [www.sgs.com](http://www.sgs.com)  
 中国·江苏·昆山开发区伟业路10号 邮编: 215300 t(86-512)57355888 f(86-512)57370818 [sgs.china@sgs.com](mailto:sgs.china@sgs.com)