


CLIP LABELS



WARNING- SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION



WARNING- SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION



WARNING- SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION



WARNING- SUBSTITUTION OF
COMPONENTS MAY VOID
FCC CERTIFICATION



CASE LABELS



EEx ia IIC T3
(Tamb=40°C)



EEx ia IIC T4
(Tamb=50°C)

WARNING- BATTERIES MUST
ONLY BE CHANGED IN AREAS
KNOWN TO BE NON HAZARDOUS
GROUP IIC GASES REQUIRE
USE OF BOOT MO4437
FCC ID: OXOGASMANI



EEx ia IIC T4
(Tamb=50°C)

WARNING- BATTERIES MUST
ONLY BE CHANGED IN AREAS
KNOWN TO BE NON HAZARDOUS
GROUP IIC GASES REQUIRE
USE OF BOOT MO4437
FCC ID: OXOGASMANI

31

35

USED ON	STOCK No.
ISS. 3	

This drawing is the property of CROWCON Detection Inst. Ltd and must not be loaned, copied in whole or in part or used for manufacturing or testing purposes without their consent. Copyright Reserved.		MATERIAL	SEE DRG	TOLERANCES	GEN= ±1 ANG= ±1°	2 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY
3	1408	6/00				
2	1389	3/00	DO NOT SCALE DRAWING. IF IN DOUBT ASK. DRAWN TO B.S. 308/3939	FINISH CLEAN	DIMS mm SHT SIZE A3	TITLE GASMAN II TX CENELEC CERTIFICATION LABELS
1	—	3/00	PROJECT No	DRN DWW	CHKO APPR	ORIGINAL SCALE 2:1 DRG. No. GASMAN-2677-A3 ISS. 3