

User Manual Addendum

Gasman II Portable Gas Detector

(This is a draft addendum supplied for content evaluation only)

Crowcon Detection Instruments Limited
2 Blacklands Way
Abingdon Business Park
Abingdon
Oxfordshire
OX14 1DY
United Kingdom

Contents

1

Contents	Page 1
Introduction	Page 2
- Determining if the Low-Power Transmitter Option is fitted	Page 2
- Preparation for Use	Page 2
- Use of a Gasman II fitted with the Low-Power Transmitter option	Page 2
Operation	Page 2
- Checking the Transmitter Function	Page 2
FCC Notice	Page 2

Introduction

This addendum relates to the Low-Power Transmitter option for the Gasman II range of portable gas detectors and should be read in conjunction with the main User Manual, Crowcon document number M07-118.

Determining if the Low-Power Transmitter Option is fitted

If a Gasman II is fitted with the Low-Power Transmitter option it will be fitted with a label including the words "Complies with FCC part 15," and "FCC ID number xxx.123" where 'xxx.123' are for example only. It will also have a model/serial number in the range 25009/00001 to 25011/99999.

Preparation for Use

Use of the device must be in accordance with the recognised standards of the appropriate authority in the company concerned. Prior to use please ensure local regulations and site procedures are followed.

Use of a Gasman II fitted with the Low-Power Transmitter option

Use is unaffected by the presence of the Low-Power Transmitter option.

Operation

Checking the Transmitter Function

In addition to the checks outlined in the Crowcon instruction manual M07-118, the following check must be made at start-up: The Vibrating Pager Unit should be turned on and its self-test vibration detected to verify its correct operation. The Gasman II transmitter should then be turned on and the vibration detected once again to show the correct operation of the Gasman II transmitter and Vibrating Pager Unit as a system. Failure to carry out this test may result in an alarm condition not being detected by the Vibrating Pager Unit, and hence the vibration version of alarm being unavailable to the user, in the event of gas being detected.

FCC Notice

This device complies with part 15 of the U.S. Federal Communications Commission (FCC) Rules. Operation is subject to two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

NOTE

Modifications to the Gasman II fitted with the Low-Power Transmitter option not expressly approved by Crowcon could void the users authority to operate the equipment.