

DRG. No.	GMP-3935-A4	ISS. 2	This drawing is the property of CROWCON Detection Inst. Ltd. and must not be loaned, copied in whole or in part or used for manufacturing or tendering purposes without their consent. Copyright Reserved		DO NOT SCALE DRAWING. IF IN DOUBT ASK. DRAWN TO BS.308/3939
USED ON	STOCK No.				

TEXT 1.5 HIGH

UL LOGO

TEXT 1.0 HIGH

SCALE 2:1

35.5

UL

19.5

CROWCON DETECTION INSTRUMENTS LTD  
VIBRATING ALARM

CLASSIFIED BY UNDERWIRTS LABORATORIES  
INC ONLY AS TO INTRINSIC SAFETY FOR USE  
IN HAZARDOUS LOCATIONS DIV 1; CLASS 1;  
A C & D, TEMP CLASS T4 FCC ID OX0VPU

WARNINGS: READ THE INSTRUCTION MANUAL BEFORE  
USE. SUBSTITUTION OF COMPONENTS MAY IMPAIR  
INTRINSIC SAFETY. NOT CLASSIFIED FOR USE IN  
ATMOSPHERES CONTAINING MORE THAN 21% OXYGEN

## NOTES

1. BLACK CHARACTERS ON ORANGE BACKGROUND
2. UL LOGO TO BE 6.2 DIA (CIRCLE DIA)
3. CHARACTERS TO BE UNDERSURFACE PRINTED
4. LABEL TO BE ADHESIVE BACKED

			MATERIAL POLYCARBONATE	TOLERANCES GEN= ANG=	 <b>CROWCON</b> 2 Blacklands Way, Abingdon Business Park, Abingdon, Oxfordshire OX14 1DY			
2	1389	4/00			FINISH	CLEAN	DIMS mm	
1		01/00	PROJECT No.			SHT SIZE A4	TITLE VIBRATING ALARM UL LABEL	
ISSUE	D.C.I. No.	DATE	DRN DWW	CHKD	APPR	ORIGINAL SCALE 1:1		
						DRG. No.	GMP-3935-A4	ISS. 2