

Annex E. Calibration Certificates

| ID | Device | Type/Model | Serial Number | Manufacturer | Calibration Certificate |
|---------|----------------------------------|------------|---------------|--------------|-------------------------|
| 003-007 | Dosimetric E-field Probe | EX3DV4 | 7465 | SPEAG | See attachment |
| 070-000 | 2450MHz System Validation Dipole | D2450V2 | 937 | SPEAG | See attachment |
| 068-000 | 5GHz System Validation Dipole | D5GHzV2 | 1164 | SPEAG | See attachment |

Dipole calibration

According to the KDB 865664 D01, a dipole must be calibrated using a fully validated SAR system according to the tissue dielectric parameters and SAR probe calibration frequency required for device testing. However, instead of the typical annual calibration recommended by measurement standards, longer calibration intervals of up to three years may be considered when it is demonstrated that the SAR target, impedance and return loss of a dipole have remain stable according to the following requirements.

1. When the most recent return-loss result, measured at least annually, deviates by less than 20% from the previous measurement (i.e. value in dB \times 0.2) and meeting the required 20 dB minimum return-loss requirement.
2. When the most recent measurement of the real and imaginary parts of the impedance, measured at least annually, deviates by less than 5 Ω from the previous measurement

The below results show the latest return loss and impedance measurements for each dipole performed by the lab:

| Dipole ID #070-000 | | | |
|-------------------------|------------------|------------------------|------------|
| Dipole 2450MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -29.7 | 50.90 + 3.20 j | 2020-05-12 |
| Last | -34.8 | 51.10 + 1.50 j | 2021-11-17 |
| Dipole ID #068-000 | | | |
| Dipole 5200MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -31.5 | 50.0 – 2.6 j | 2021-05-18 |
| Dipole 5300MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -31.0 | 50.7 +2.7 j | 2021-05-18 |
| Dipole 5500MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -29.9 | 49.0 + 3.0 j | 2021-05-18 |
| Dipole 5600MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -23.6 | 53.2 +6.0 j | 2021-05-18 |
| Dipole 5800MHz Body TSL | | | |
| | Return Loss [dB] | Impedance [Ω] | Date |
| Initial Calibration | -20.9 | 53.8 +8.6 j | 2021-05-18 |