



*Expanding the scope of GI*

November 4, 2020

Telefication B.V.  
Attn: Dept. FCC TCB  
Edisonstraat 12A  
6902 PK ZEVENAAR  
The Netherlands

Subject: Application for Class II Permissive Change of certification of a Bravo Reflux Capsule,  
FCC ID: PHZ-BRAVO100

Dear Gentlemen,

Please find attached our application for Class II Permissive Change of certification of Bravo Reflux Capsule,  
FCC ID: PHZ-BRAVO100, prepared in accordance with FCC CFR 47 Part 15.231.  
The following changes were implemented in approved Bravo Reflux Capsule, FCC ID: PHZ-BRAVO100

1. Replacement of reed switch by a NVE magnetic sensor and electrical switch on capsule PCA.
2. Capsule RF transistor was declared EOL and replaced with a new one with the same parameters.
3. Capacitor Vdd filtering 100 pF was changed to 150 pF.
4. A LED was added on the PCA to indicate capsule's status during the assembly process with no layout change.

No impact in the RF performance of the device after above mentioned changes.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Avishag Metzer".

Avishag Metzer

RA Program Manager