

FCC Test Report

Product Name	“Chiline” Pulse Oximeter and Respiratory Monitoring System
Model No.	S68
FCC ID	POTS68

Applicant	Inventec Appliances Corp.
Address	37 Wugong 5th road, New Taipei Industrial Park, Wugu District, New Taipei City, Taiwan

Date of Receipt	Feb. 02, 2021
Issued Date	Sep. 11, 2021
Report No.	2120100R-E3032110108
Report Version	V1.0



The test results relate only to the samples tested.

The test results shown in the test report are traceable to the national/international standard through the calibration report of the equipment and evaluated measurement uncertainty herein.

This report must not be used to claim product endorsement by TAF or any agency of the government.

The test report shall not be reproduced without the written approval of DEKRA Testing and Certification Co., Ltd.

Measurement uncertainties evaluated for each testing system and associated connections are given here to provide the system information for reference. Compliance determinations do not take into account measurement uncertainties for each testing system, but are based on the results of the compliance measurement.

Test Report

Issued Date: Sep. 11, 2021

Report No.: 2120100R-E3032110108




Product Name	“Chiline” Pulse Oximeter and Respiratory Monitoring System
Applicant	Inventec Appliances Corp.
Address	37 Wugong 5th road, New Taipei Industrial Park, Wugu District, New Taipei City, Taiwan
Manufacturer	Inventec Appliances Corp.
Model No.	S68
FCC ID	POTS68
EUT Rated Voltage	DC 5V by USB or DC 3.7V by battery
EUT Test Voltage	DC 5V by USB
Trade Name	Chiline
Applicable Standard	FCC CFR Title 47 Part 15 Subpart C ANSI C63.4: 2014, ANSI C63.10: 2013
Test Result	Complied

Documented By :




(Senior Project Specialist / Joanne Lin)

Tested By :



(Senior Engineer / Ivan Chuang)

Approved By :



(Senior Engineer / Jack Hsu)

TABLE OF CONTENTS

Description	Page
1. GENERAL INFORMATION	5
1.1. EUT Description.....	5
1.2. Tested System Details.....	7
1.3. Configuration of Tested System	8
1.4. EUT Exercise Software	8
1.5. Test Facility	9
1.6. List of Test Equipment.....	9
1.7. Uncertainty	11
2. CONDUCTED EMISSION	12
2.1. Test Setup	12
2.2. Limits.....	12
2.3. Test Procedure	13
2.4. Test Result of Conducted Emission.....	14
3. PEAK POWER OUTPUT	18
3.1. Test Setup	18
3.2. Limit	18
3.3. Test Procedure	18
3.4. Test Result of Peak Power Output.....	19
4. RADIATED EMISSION	20
4.1. Test Setup	20
4.2. Limits.....	21
4.3. Test Procedure	22
4.4. Test Result of Radiated Emission.....	24
5. RF ANTENNA CONDUCTED TEST	27
5.1. Test Setup	27
5.2. Limits.....	27
5.3. Test Procedure	27
5.4. Test Result of RF Antenna Conducted Test.....	28
6. BAND EDGE	29
6.1. Test Setup	29
6.2. Limit	30
6.3. Test Procedure	30
6.4. Test Result of Band Edge	32
7. 6DB BANDWIDTH	34
7.1. Test Setup	34
7.2. Limits.....	34
7.3. Test Procedure	34
7.4. Test Result of 6dB Bandwidth.....	35
8. POWER DENSITY	37
8.1. Test Setup	37
8.2. Limits.....	37
8.3. Test Procedure	37
8.4. Test Result of Power Density	38
9. DUTY CYCLE	40
9.1. Test Setup	40
9.2. Test Procedure	40
9.3. Test Result of Duty Cycle.....	41
10. EMI REDUCTION METHOD DURING COMPLIANCE TESTING	42

Appendix 1: EUT Test Photographs

Appendix 2: EUT Detailed Photographs

Revision History

Report No.	Version	Description	Issued Date
2120100R-E3032110108	V1.0	Initial issue of report.	Sep. 11, 2021

1. GENERAL INFORMATION

1.1. EUT Description

Product Name	“Chiline” Pulse Oximeter and Respiratory Monitoring System
Trade Name	Chiline
Model No.	S68
FCC ID	POTS68
Frequency Range	2402-2480 MHz
Channel Number	40CH
Type of Modulation	GFSK
Antenna Type	Chip
Channel Control	Auto
Antenna Gain	Refer to the Antenna List
Power Adapter	MFR: ADAPTER TECH, M/N: ATM012T-W050VU Input: AC 100-240V~50-60Hz, 0.32-0.19A Output: 5.0V=2.0A, 10.0W
USB Cable (Optional)	UTE / 61401.12A40, Non-shielded, 1m
L-Size SpO ₂ Probe (Optional)	Chiline / PA68
XS-Size SpO ₂ Probe (Optional)	Chiline / PB68
S-Size SpO ₂ Probe (Optional)	Chiline / PC68

Antenna List

No.	Manufacturer	Part No.	Antenna Type	Peak Gain
1	YAGEO	ANT3216LL11R2400A	Chip	3.68 dBi for 2.4GHz

Note: The antenna of EUT is conforming to FCC 15.203.

Center Frequency of Each Channel:

Channel	Frequency	Channel	Frequency	Channel	Frequency	Channel	Frequency
Channel 00:	2402 MHz	Channel 01:	2404 MHz	Channel 02:	2406 MHz	Channel 03:	2408 MHz
Channel 04:	2410 MHz	Channel 05:	2412 MHz	Channel 06:	2414 MHz	Channel 07:	2416 MHz
Channel 08:	2418 MHz	Channel 09:	2420 MHz	Channel 10:	2422 MHz	Channel 11:	2424 MHz
Channel 12:	2426 MHz	Channel 13:	2428 MHz	Channel 14:	2430 MHz	Channel 15:	2432 MHz
Channel 16:	2434 MHz	Channel 17:	2436 MHz	Channel 18:	2438 MHz	Channel 19:	2440 MHz
Channel 20:	2442 MHz	Channel 21:	2444 MHz	Channel 22:	2446 MHz	Channel 23:	2448 MHz
Channel 24:	2450 MHz	Channel 25:	2452 MHz	Channel 26:	2454 MHz	Channel 27:	2456 MHz
Channel 28:	2458 MHz	Channel 29:	2460 MHz	Channel 30:	2462 MHz	Channel 31:	2464 MHz
Channel 32:	2466 MHz	Channel 33:	2468 MHz	Channel 34:	2470 MHz	Channel 35:	2472 MHz
Channel 36:	2474 MHz	Channel 37:	2476 MHz	Channel 38:	2478 MHz	Channel 39:	2480 MHz

Note:

1. The EUT is a “Chiline” Pulse Oximeter and Respiratory Monitoring System with a built-in 2.4 GHz WLAN and Bluetooth V5.1 transceiver, this report for Bluetooth V5.1.
2. Bluetooth V5.1 only 1Mbps.
3. The radiation measurements are performed in X, Y, Z axis positioning. Only the worst case is shown in the report.
4. These tests were conducted on a sample for the purpose of demonstrating compliance of Bluetooth V5.1 transmitter with Part 15 Subpart C Paragraph 15.247 for spread spectrum devices.
5. Regarding to the operation frequency, the lowest, middle and highest frequency are selected to perform the test.

Test Mode	Mode 1: Transmit Mode 2: Charge Mode
-----------	---

1.2. Tested System Details

The types for all equipment, plus descriptions of all cables used in the tested system (including inserted cards) are:


For Transmit					
Product		Manufacturer	Model No.	Serial No.	Power Cord
1	Notebook PC	DELL	Latitude E5440	HG26TZ1	N/A
2	Test Fixture	Chiline	CP2102 USB to TTL	N/A	N/A
3	Test Fixture	Chiline	FT232 USB UART BOARD	N/A	N/A

For Charge Mode					
Product		Manufacturer	Model No.	Serial No.	Power Cord
1	Power Adapter	ADAPTER TECH	ATM012T-W050VU	N/A	N/A

For Transmit		
Signal Cable Type		Signal cable Description
A	USB Cable	Non-shielded, 1.55m
B	Signal Cable	Non-shielded, 0.2m
C	USB Cable	Shielded, 1.8m

For Charge Mode		
Signal Cable Type		Signal cable Description
A	USB Cable	Non-shielded, 1m

1.3. Configuration of Tested System



1.4. EUT Exercise Software

1. Setup the EUT as shown in Section 1.3.
2. Execute software “cybluetooth V0.1.55.1” on the EUT.
3. Configure the test mode, the test channel, and the data rate.
4. Press “OK” to start the continuous Transmit.
5. Verify that the EUT works properly.

1.5. Test Facility

Ambient conditions in the laboratory:

Performed Item	Items	Required	Actual
Conducted Emission	Temperature (°C)	10~40 °C	33 °C
	Humidity (%RH)	10~90 %	56.5 %
Radiated Emission	Temperature (°C)	10~40 °C	25 °C
	Humidity (%RH)	10~90 %	64 %
Conductive	Temperature (°C)	10~40 °C	24.7 °C
	Humidity (%RH)	10~90 %	54.7 %

USA : FCC Registration Number: TW1134

Canada : CAB Identifier Number: TW3023 / Company Number: 26930

Site Description : Accredited by TAF
Accredited Number: 3023

Test Laboratory : DEKRA Testing and Certification Co., Ltd
Address : No. 5-22, Ruishukeng Linkou District, New Taipei City, 24451, Taiwan
Performed Location : No. 26, Huaya 1st Rd., Guishan Dist., Taoyuan City 333411, Taiwan, R.O.C.
Phone number : +886-3-275-7255
Fax number : +886-3-327-8031
Email address : info.tw@dekra.com
Website : <http://www.dekra.com.tw>

1.6. List of Test Equipment

For Conduction measurements / SH1

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	EMI Test Receiver	R&S	ESR7	101601	2021.06.19	2022.06.18
X	Two-Line V-Network	R&S	ENV216	101306	2021.04.08	2022.04.07
X	Two-Line V-Network	R&S	ENV216	101307	2021.05.04	2022.05.03
X	Coaxial Cable	DEKRA	RG400_BNC	RF001	2021.05.24	2022.05.23

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : DEKRA Testing System V2.0.

For Conducted measurements / SH2

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	Spectrum Analyzer	R&S	FSV30	103466	2020.12.28	2021.12.27
X	Peak Power Analyzer	KEYSIGHT	8900B	MY51000539	2021.06.07	2022.06.06
X	Power Sensor	KEYSIGHT	N1923A	MY59240002	2021.05.17	2022.05.16
X	Power Sensor	KEYSIGHT	N1923A	MY59240003	2021.05.17	2022.05.16

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : DEKRA Conduction Test System V9.0.5.

For Radiated measurements / 966-3

	Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Due. Date
X	Loop Antenna	AMETEK	HLA6121	56736	2021.04.14	2022.04.13
X	Bi-Log Antenna	SCHWARZBECK	VULB9168	9168-678	2020.09.04	2021.09.03
X	Horn Antenna	ETS-Lindgren	3117	00203761	2020.11.23	2021.11.22
	Horn Antenna	Com-Power	AH-840	101087	2021.06.16	2022.06.15
X	Pre-Amplifier	EMCI	EMC001330	980302	2021.07.26	2022.07.25
X	Pre-Amplifier	EMCI	EMC051835SE	980312	2021.02.24	2022.02.23
X	Pre-Amplifier	EMCI	EMC05820SE	980308	2020.09.18	2021.09.17
	Pre-Amplifier	EMCI	EMC184045SE	980369	2021.04.27	2022.04.26
X	Filter	MICRO TRONICS	BRM50702	G251	2020.09.17	2021.09.16
	Filter	MICRO TRONICS	BRM50716	G188	2020.09.17	2021.09.16
X	EMI Test Receiver	R&S	ESR	102793	2020.12.17	2021.12.16
X	Spectrum Analyzer	R&S	FSV3044	101113	2021.02.03	2022.02.02
X	Coaxial Cable	SGH, EMCI	HA800, SGH18	HY2103-001C	2021.03.03	2022.03.02
X	Mircoflex Cable	HUBER SUHNER	SUCOFLEX 102	MY3381/2	2021.06.25	2022.06.24

Note:

1. All equipments are calibrated every one year.
2. The test instruments marked with "X" are used to measure the final test results.
3. Test Software version : AUDIX e3 V9

1.7. Uncertainty

Uncertainties have been calculated according to the DEKRA internal document, and is described in each test chapter of this report.


The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor of $k=2$, providing a level of confidence of approximately 95%.

Measurement uncertainties evaluated for each testing system and associated connections are given here to provide the system information for reference. Compliance determinations do not take into account measurement uncertainties for each testing system, but are based on the results of the compliance measurement.

Test item	Uncertainty	
Conducted Emission	±3.42dB	
Peak Power Output	Power Meter ±0.89dB	Spectrum Analyzer ±2.06dB
Radiated Emission	Under 1GHz ±4.05 dB	Above 1GHz ±3.73 dB
RF antenna conducted test	±2.06dB	
Band Edge	Under 1GHz ±4.05 dB	Above 1GHz ±3.73 dB
6dB Bandwidth	±1544.74Hz	
Power Density	±2.06dB	
Duty Cycle	±2.31msec	

2. Conducted Emission

2.1. Test Setup



2.2. Limits

FCC Part 15 Subpart C Paragraph 15.207 (dBuV) Limit		
Frequency MHz	Limits	
	QP	AV
0.15 - 0.50	66-56	56-46
0.50-5.0	56	46
5.0 - 30	60	50

Remarks: In the above table, the tighter limit applies at the band edges.

2.3. Test Procedure

The EUT and Peripherals are connected to the main power through a line impedance stabilization network (L.I.S.N.). This provides a 50 ohm /50uH coupling impedance for the measuring equipment. The peripheral devices are also connected to the main power through a LISN that provides a 50ohm /50uH coupling impedance with 50ohm termination. (Please refer to the block diagram of the test setup and photographs.)

Both sides of A.C. line are checked for maximum conducted interference. In order to find the maximum emission, the relative positions of equipment and all the interface cables must be changed according to ANSI C63.4: 2014 on conducted measurement.


Conducted emissions were investigated over the frequency range from 0.15MHz to 30MHz using a receiver bandwidth of 9kHz.

The EUT was setup to ANSI C63.4, 2014; tested to DTS test procedure of FCC KDB-558074 for compliance to FCC 47CFR Subpart C requirements.

2.4. Test Result of Conducted Emission

Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
 Test Item : Conducted Emission Test
 Test date : 2021/06/17
 Test Mode : Mode 1: Transmit (2440MHz)

L 1




No	Frequency (MHz)	Emission Level (dBuV)	Limit (dBuV)	Margin (dB)	Reading Level (dBuV)	Correct Factor (dB)	Detector Type
1	0.158	39.61	65.58	-25.97	29.95	9.66	QP
2	0.158	19.36	55.58	-36.22	9.70	9.66	AV
3	0.159	39.36	65.52	-26.16	29.70	9.66	QP
4	0.159	19.19	55.52	-36.33	9.53	9.66	AV
5	0.166	38.45	65.14	-26.69	28.79	9.66	QP
6	0.166	17.19	55.14	-37.95	7.53	9.66	AV
7	0.176	37.22	64.67	-27.45	27.57	9.65	QP
8	0.176	17.72	54.67	-36.95	8.07	9.65	AV
9	0.396	29.78	57.94	-28.16	20.12	9.66	QP
*10	0.396	27.63	47.94	-20.31	17.97	9.66	AV
11	0.506	30.88	56.00	-25.12	21.22	9.66	QP
12	0.506	13.73	46.00	-32.27	4.07	9.66	AV

Remark:

1. "*" means this data is the worst emission level; "!" means this data is over limit.
2. Emission Level=Reading Level + Correct Factor(Correct Factor=LISN Factor+Cable Loss).
3. Margin=Emission Level-Limit

Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
 Test Item : Conducted Emission Test
 Test date : 2021/06/17
 Test Mode : Mode 1: Transmit (2440MHz)

N




No	Frequency (MHz)	Emission Level (dBuV)	Limit (dBuV)	Margin (dB)	Reading Level (dBuV)	Correct Factor (dB)	Detector Type
*1	0.152	40.02	65.90	-25.88	30.35	9.67	QP
2	0.152	18.95	55.90	-36.95	9.28	9.67	AV
3	0.170	37.67	64.97	-27.30	28.00	9.67	QP
4	0.170	17.66	54.97	-37.31	7.99	9.67	AV
5	0.170	38.12	64.96	-26.84	28.45	9.67	QP
6	0.170	17.77	54.96	-37.19	8.10	9.67	AV
7	0.186	35.99	64.22	-28.23	26.32	9.67	QP
8	0.186	17.66	54.22	-36.56	7.99	9.67	AV
9	0.216	32.81	62.98	-30.17	23.14	9.67	QP
10	0.216	13.43	52.98	-39.55	3.76	9.67	AV
11	0.250	29.52	61.75	-32.23	19.85	9.67	QP
12	0.250	11.29	51.75	-40.46	1.62	9.67	AV

Remark:

1. "*" means this data is the worst emission level; "!" means this data is over limit.
2. Emission Level=Reading Level + Correct Factor(Correct Factor=LISN Factor+Cable Loss).
3. Margin=Emission Level-Limit

Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
 Test Item : Conducted Emission Test
 Test date : 2021/06/18
 Test Mode : Mode 2: Charge Mode

L 1




No	Frequency (MHz)	Emission Level (dBuV)	Limit (dBuV)	Margin (dB)	Reading Level (dBuV)	Correct Factor (dB)	Detector Type
1	0.306	22.37	60.09	-37.72	12.72	9.65	QP
2	0.306	3.07	50.09	-47.02	-6.58	9.65	AV
3	0.459	33.17	56.71	-23.54	23.51	9.66	QP
4	0.459	16.19	46.71	-30.52	6.53	9.66	AV
*5	0.503	37.65	56.00	-18.35	27.99	9.66	QP
6	0.503	18.15	46.00	-27.85	8.49	9.66	AV
7	10.276	7.37	60.00	-52.63	-2.52	9.89	QP
8	10.276	-2.68	50.00	-52.68	-12.57	9.89	AV
9	11.529	1.03	60.00	-58.97	-8.88	9.91	QP
10	11.529	-2.38	50.00	-52.38	-12.29	9.91	AV
11	12.612	0.69	60.00	-59.31	-9.23	9.92	QP
12	12.612	-2.27	50.00	-52.27	-12.19	9.92	AV

Remark:

1. "*" means this data is the worst emission level; "!" means this data is over limit.
2. Emission Level=Reading Level + Correct Factor(Correct Factor=LISN Factor+Cable Loss).
3. Margin=Emission Level-Limit

Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
 Test Item : Conducted Emission Test
 Test date : 2021/06/18
 Test Mode : Mode 2: Charge Mode (2440MHz)

N




No	Frequency (MHz)	Emission Level (dBuV)	Limit (dBuV)	Margin (dB)	Reading Level (dBuV)	Correct Factor (dB)	Detector Type
1	0.201	29.69	63.56	-33.87	20.02	9.67	QP
2	0.201	11.86	53.56	-41.70	2.19	9.67	AV
3	0.224	30.61	62.67	-32.06	20.94	9.67	QP
4	0.224	8.60	52.67	-44.07	-1.07	9.67	AV
5	0.249	29.01	61.78	-32.77	19.34	9.67	QP
6	0.249	8.36	51.78	-43.42	-1.31	9.67	AV
7	0.463	33.64	56.64	-23.00	23.97	9.67	QP
8	0.463	17.70	46.64	-28.94	8.03	9.67	AV
*9	0.506	39.07	56.00	-16.93	29.40	9.67	QP
10	0.506	19.62	46.00	-26.38	9.95	9.67	AV
11	0.528	36.57	56.00	-19.43	26.90	9.67	QP
12	0.528	19.04	46.00	-26.96	9.37	9.67	AV

Remark:

1. "*" means this data is the worst emission level; "!" means this data is over limit.
2. Emission Level=Reading Level + Correct Factor(Correct Factor=LISN Factor+Cable Loss).
3. Margin=Emission Level-Limit

3. Peak Power Output

3.1. Test Setup



3.2. Limit

The maximum peak power shall be less 1Watt.

3.3. Test Procedure

The EUT was tested according to C63.10:2013 for compliance to FCC 47CFR 15.247 requirements. The maximum peak conducted output power using C63.10:2013 Section 11.9.1.3 PKPM1 Peak power meter method.

3.4. Test Result of Peak Power Output


Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
Test Item : Peak Power Output
Test Mode : Mode 1: Transmit
Test Date : 2021/08/24

Channel No.	Frequency (MHz)	Measurement (dBm)	Required Limit	Result
Channel 00	2402.00	-2.04	1 Watt= 30 dBm	Pass
Channel 19	2440.00	-1.92	1 Watt= 30 dBm	Pass
Channel 39	2480.00	-2.00	1 Watt= 30 dBm	Pass


4. Radiated Emission

4.1. Test Setup


Radiated Emission Under 30MHz



Radiated Emission Below 1GHz



Radiated Emission Above 1GHz



4.2. Limits

➤ General Radiated Emission Limits

Emissions radiated outside of the specified frequency bands, except for harmonics, shall be attenuated by at least 20dB below the level of the fundamental or to the general radiated emission limits in paragraph 15.209, whichever is the lesser attenuation.

FCC Part 15 Subpart C Paragraph 15.209 Limits		
Frequency MHz	Field strength (microvolts/meter)	Measurement distance (meter)
0.009-0.490	2400/F(kHz)	300
0.490-1.705	24000/F(kHz)	30
1.705-30	30	30
30-88	100	3
88-216	150	3
216-960	200	3
Above 960	500	3

- Remarks:
1. RF Voltage (dBuV) = 20 log RF Voltage (uV)
 2. In the Above Table, the tighter limit applies at the band edges.
 3. Distance refers to the distance in meters between the measuring instrument antenna and the closed point of any part of the device or system.

4.3. Test Procedure

The EUT was setup according to ANSI C63.10: 2013 and tested according to C63.10:2013 Section 11.12.1 for compliance to FCC 47CFR 15.247 requirements.

Measuring the frequency range below 1GHz, the EUT is placed on a turn table which is 0.8 meter above ground, when measuring the frequency range above 1GHz, the EUT is placed on a turn table which is 1.5 meter above ground.

The turn table is rotated 360 degrees to determine the position of the maximum emission level.

The EUT was positioned such that the distance from antenna to the EUT was 3 meters.

The antenna is scanned between 1 meter and 4 meters to find out the maximum emission level. This is repeated for both horizontal and vertical polarization of the antenna. In order to find the maximum emission, all of the interface cables were manipulated according to ANSI C63.10: 2013 on radiated measurement.

The resolution bandwidth below 30MHz setting on the field strength meter is 9kHz and 30MHz~1GHz is 120kHz and above 1GHz is 1MHz.

Radiated emission measurements below 30MHz are made using Chip Antenna and 30MHz~1GHz are made using broadband Bilog antenna and above 1GHz are made using Horn Antennas.

The measurement is divided into the Preliminary Measurement and the Final Measurement.

The suspected frequencies are searched for in Preliminary Measurement with the measurement antenna kept pointed at the source of the emission both in azimuth and elevation, with the polarization of the antenna oriented for maximum response. The antenna is pointed at an angle towards the source of the emission, and the EUT is rotated in both height and polarization to maximize the measured emission. The emission is kept within the illumination area of the 3 dB bandwidth of the antenna. The measurement frequency range from 9kHz - 10th Harmonic of fundamental was investigated.

RBW and VBW Parameter setting:

According to C63.10 Section 11.12.2.4 Peak measurement procedure.

RBW = as specified in Table 1.

$VBW \geq 3 \times RBW$.

Table 1 —RBW as a function of frequency

Frequency	RBW
9-150 kHz	200-300 Hz
0.15-30 MHz	9-10 kHz
30-1000 MHz	100-120 kHz
> 1000 MHz	1 MHz

According to C63.10 Section 11.12.2.5 Average measurement procedure.

RBW = 1MHz.

VBW = 10Hz, when duty cycle $\geq 98 \%$




$VBW \geq 1/T$, when duty cycle $< 98 \%$





(T refers to the minimum transmission duration over which the transmitter is on and is transmitting at its maximum power control level for the tested mode of operation.)



2.4GHz band	Duty Cycle (%)	T (ms)	1/T (Hz)	VBW (Hz)
BLE	64.56	0.3913	2556	3000

Note: Duty Cycle Refer to Section 9.

4.4. Test Result of Radiated Emission








5. RF Antenna Conducted Test

5.1. Test Setup



5.2. Limits

According to FCC Section 15.247(d). In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or measurement, provided the transmitter demonstrates compliance with the peak conducted power limits. If the transmitter complies with the conducted power limits based on the use of RMS averaging over a time interval, the attenuation required under this paragraph shall be 30 dB instead of 20 dB.

5.3. Test Procedure

The EUT was tested according to C63.10:2013 Section 11.11 for compliance to FCC 47CFR 15.247 requirements.

Set RBW = 100 kHz, Set VBW > RBW, scan up through 10th harmonic.

5.4. Test Result of RF Antenna Conducted Test

Product : “Chiline” Pulse Oximeter and Respiratory Monitoring System
 Test Item : RF Antenna Conducted Test
 Test Mode : Mode 1: Transmit
 Test Date : 2021/08/24

Figure Channel 00:




Figure Channel 19:





Figure Channel 39:




Note: The above test pattern is synthesized by multiple of the frequency range.


6. Band Edge

6.1. Test Setup

RF Conducted Measurement



RF Radiated Measurement:



6.2. Limit

According to FCC Section 15.247(d). In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or measurement, provided the transmitter demonstrates compliance with the peak conducted power limits. If the transmitter complies with the conducted power limits based on the use of RMS averaging over a time interval, the attenuation required under this paragraph shall be 30 dB instead of 20 dB. Attenuation below the general limits specified in Section 15.209(a) is not required. In addition, radiated emissions which fall in the restricted bands, as defined in Section 15.205(a), must also comply with the radiated emission limits specified in Section 15.209(a) (see Section 15.205(c)).

6.3. Test Procedure

The EUT was setup according to ANSI C63.10, 2013 and tested according to C63.10:2013 Section 11.12.1 for compliance to FCC 47CFR 15.247 requirements.

The EUT is placed on a turn table which is 1.5 meter above ground. The turn table is rotated 360 degrees to determine the position of the maximum emission level. The EUT was positioned such that the distance from antenna to the EUT was 3 meters.

The antenna is scanned from 1 meter to 4 meters to find out the maximum emission level. This is repeated for both horizontal and vertical polarization of the antenna. In order to find the maximum emission, all of the interface cables were manipulated according to ANSI C63.10:2013 on radiated measurement.

RBW and VBW Parameter setting:

According to C63.10 Section 11.12.2.4 Peak measurement procedure.

RBW = as specified in Table 1.

$VBW \geq 3 \times RBW$.

Table 1 —RBW as a function of frequency

Frequency	RBW
9-150 kHz	200-300 Hz
0.15-30 MHz	9-10 kHz
30-1000 MHz	100-120 kHz
> 1000 MHz	1 MHz

According to C63.10 Section 11.12.2.5 Average measurement procedure.

RBW = 1MHz.

VBW = 10Hz, when duty cycle $\geq 98 \%$





$VBW \geq 1/T$, when duty cycle $< 98 \%$





(T refers to the minimum transmission duration over which the transmitter is on and is transmitting at its maximum power control level for the tested mode of operation.)

2.4GHz band	Duty Cycle (%)	T (ms)	1/T (Hz)	VBW (Hz)
BLE	64.56	0.3913	2556	3000

Note: Duty Cycle Refer to Section 9.


6.4. Test Result of Band Edge





7. 6dB Bandwidth

7.1. Test Setup



7.2. Limits

The minimum bandwidth shall be at least 500 kHz.

7.3. Test Procedure


The EUT was setup according to ANSI C63.4, 2014; tested according to ANSI C63.10 Section 11.8 for compliance to FCC 47CFR 15.247 requirements.

7.4. Test Result of 6dB Bandwidth

Product : “Chiline” Pulse Oximeter and Respiratory Monitoring System
 Test Item : 6dB Bandwidth Data
 Test Mode : Mode 1: Transmit


Channel No.	Frequency (MHz)	Measurement Level (kHz)	Required Limit (kHz)	Result
00	2402	720	>500	Pass
19	2440	725	>500	Pass
39	2480	720	>500	Pass

Figure Channel 00:




Date: 24.AUG.2021 10:16:00

Figure Channel 19:



Date: 24.AUG.2021 10:19:00


Figure Channel 39:



Date: 24.AUG.2021 10:23:47

8. Power Density

8.1. Test Setup



8.2. Limits

The transmitted power density averaged over any 1 second interval shall not be greater +8dBm in any 3kHz bandwidth.

8.3. Test Procedure

The EUT was setup according to ANSI C63.10, 2013; tested according to DTS test procedure of KDB 558074 for compliance to FCC 47CFR 15.247 requirements.


The maximum power spectral density using C63.10 Section 11.10.2 Method PKPSD (peak PSD)

8.4. Test Result of Power Density

Product : "Chiline" Pulse Oximeter and Respiratory Monitoring System
 Test Item : Power Density Data
 Test Mode : Mode 1: Transmit


Channel No.	Frequency (MHz)	Measure Level (dBm)	Limit (dBm)	Result
00	2402	-2.66	≤ 8dBm	Pass
19	2440	-2.58	≤ 8dBm	Pass
39	2480	-2.89	≤ 8dBm	Pass

Figure Channel 00:




Date: 24.AUG.2021 10:16:21

Figure Channel 19:



Date: 24.AUG.2021 10:19:22


Figure Channel 39:



Date: 24.AUG.2021 10:24:08

9. Duty Cycle

9.1. Test Setup



9.2. Test Procedure

The EUT was setup according to ANSI C63.10 2013; tested according to ANSI C63.10 2013 for compliance to FCC 47CFR 15.247 requirements.

9.3. Test Result of Duty Cycle

Product : “Chiline” Pulse Oximeter and Respiratory Monitoring System
 Test Item : Duty Cycle
 Test Mode : Mode 1: Transmit


Duty Cycle Formula:

$$\text{Duty Cycle} = \text{Ton} / (\text{Ton} + \text{Toff})$$

$$\text{Duty Factor} = 10 \text{ Log} (1/\text{Duty Cycle})$$

Results:

2.4GHz band	Ton (ms)	Ton + Toff (ms)	Duty Cycle (%)	Duty Factor (dB)
BLE	0.3913	0.6061	64.56	1.90



Date: 28.AUG.2021 01:22:55

10. EMI Reduction Method During Compliance Testing

No modification was made during testing.