

ATTACHMENT R – PROBE CALIBRATION DATA

Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

Additional Conversion Factors for Dosimetric E-Field Probe

Type:

ET3DV6

Serial Number:

1608

Place of Assessment:

Zurich

Date of Assessment:


May 6, 2002

Probe Calibration Date:

August 10, 2001

Schmid & Partner Engineering AG hereby certifies that conversion factor(s) of this probe have been evaluated on the date indicated above. The assessment was performed using the FDTD numerical code SEMCAD of Schmid & Partner Engineering AG. Since the evaluation is coupled with measured conversion factors, it has to be recalculated yearly, i.e., following the re-calibration schedule of the probe. The uncertainty of the numerical assessment is based on the extrapolation from measured value at 900 MHz or at 1800 MHz.

Assessed by:



Dosimetric E-Field Probe ET3DV6 SN:1608Conversion factor ($\pm$ standard deviation)**835 MHz**

ConvF

 $6.7 \pm 8\%$ **$\epsilon_r = 41.5 \pm 5\%$** **$\sigma = 0.90 \pm 5\%$ mho/m**

(head tissue)