

Philips Medizin Systeme Böblingen GmbH  
Hewlett-Packard-Str. 2  
71034 Böblingen  
Germany

Declaration about Hardware – IntelliVue CL NBP Pod 865216

To Whom It May Concern:  
Product: IntelliVue CL NBP Pod 865216  
FCC ID: PQC-CLNBPBV2

We hereby declare that Product: IntelliVue CL NBP Pod 865216 uses the same SRR Hardware and Software as Product: 2.4 GHz Transceiver - SRR Module BV3 (FCC ID: PQC-SRRBV3). No RF related modifications (redesign, component replacements, pcb redesign) have been made.

Based on this fact conducted signal measurements in FCC part 15.247 reports (report No. 1-5420/12-01-11-A, dated 2013-10-17) issued for Product/Model: 2.4 GHz Transceiver/ SRR Module BV3 also apply to Product: IntelliVue CL NBP Pod 865216.

If you have any questions feel free to contact us at the address shown below.

Sincerely



---

Hansjörg Geywitz (Signature)  
Hewlett-Packard-Str. 2  
71034 Böblingen, Germany  
Tel: 07031-4630  
Email: hansjoerg.geywitz@philips.com