

RF Exposure Statement

Philips Medical Systems hereby declares that the

Wireless medical telemetry transmitter module,
MDL4851 (FCC ID: PQC-MDL4851)

is compliant to 47 CFR Part 15B and 15C, §1.1310 and §2.1093 of the FCC rules.

As per the FCC/TCB training of October 15, 2004, this statement of compliance is sufficient for devices with sourced based time average power of less than 1 mW. This applies to both TCB and FCC filings.

The output power for this device based on sourced based time average power is less than 1 mW. The duty cycle is 1.6 %.

Zety Billard, Regulatory Engineer

Submitted 22-November-2006