

## Philips Medical Systems

---

May 23, 2007

Federal Communications Commission  
Authorization and Evaluation Division  
Washington, DC 20554

RE: Grant ID: PQC-MDL4851

To whom it may concern:

This is a request for a Class II Permissive Change under Section 15.204 [c][1] for the Instrument telemetry module, Model MDL4851. The current grant covers two antennas, the Radial-Larsen Dipole Whip, Philips part number M4842-61400, and the Multi-band (aka Tri-band), Philips part number M8100-66490. We would like to include a third antenna to this grant, the dual-band antenna, Philips part number M3002-66493. This additional internal antenna is offered as a solution to space constraints.

This module incorporates the same radio technology and implementation as that used in other Philips devices. This device differs in that it is modular, intended to be incorporated into certain Philips patient monitors, thereby providing wireless communication to existing healthcare facility infrastructure. The module complies with FCC Rules as a self-contained entity; it does not rely on external mounting or shielding or other input/output mechanisms to achieve compliance. The device was designed in accordance with FCC Public Notice DA 00-1407, 26-Jun-2000, "Unlicensed Modular Transmitter Approval". The module is not being offered for sale as an OEM component. It is for incorporation into Philips Medical Systems products only.

Thank you in advance for your consideration,



Barry Wyshogrod  
Regulatory Engineer  
Philips Medical Systems

