



775 Montague Expressway
Milpitas, CA 95035
Tel: 408-526-1188
Fax: 408-526-1088
Email: TCB@siemic.com

FCC CLASS II PERMISSIVE CHANGE REQUEST LETTER

Reason for Amendment (current / obsolete)	Revision History		Approved Date
	From	To	
Initial Release (current)	1.0	1.0	Feb-20-2012



FCC Class II Permissive Change Request Letter

Date: 6/5/2015

To FCC:

RE: FCC Permissive II Change Request for Company: Philips Medical Systems North America Co. **FCC ID:** PQC-MX40WL3

We are submitting an application for a class II permissive change to the FCC approval of the Company name: Philips Medical Systems North America Co., product description: SDIO Wireless Module (FCC: PQC-MX40WL3, Original Grant Date: 03/02/2015). The transmitter module itself has not changed. Here are the changes:

- 1) BT is disabled by software.
- 2) Adding new antennas (Philips P/N: 453564553001r1444, and 453564252521r1444).
- 3) Increase transmit power for 802.11g & 802.11n-HT20.

Sincerely,

A handwritten signature in black ink, reading "Delroy Smith", is positioned below the "Sincerely," text.

Client's signature

Delroy Smith / Principal Scientist & Project Leader
3000 Minuteman Road (MS 4304), Andover, MA 018 10-1032, United States
delroy.smith@philips.com / PH: (978)659-3382