



NFCBV1-A NFC Module User Manual

Model: NFCBV1-A

FCC-ID: PQC-NFCBV1

IC: 3549C-NFCBV1

For OEM integration only – device cannot be sold to general public. Therefore, we will not supply a user manual.

We will ask OEM to include the following notices required by FCC/IC on the product and in the user manual.

Manual Requirements according 15.105

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Manual Requirements according 15.19 / RSS-GEN

This device complies with Part 15 of the FCC Rules and with Industry Canada licence-exempt RSS standard(s).

Operation is subject to the following two conditions:

- this device may not cause harmful interference, and
- this device must accept any interference received, including interference that may cause undesired operation.



Philips Medizin Systeme Böblingen GmbH

Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202

Geschäftsführer: Dr. Dieter Haase (Sprecher), Dr. Peter Ziese, Matthijs de Groot

Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, www.philips.com/healthcare



Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes:

- l'appareil ne doit pas produire de brouillage, et
- l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

Manual Requirements according 15.21

Changes or modifications made to this equipment not expressly approved by (manufacturer name) may void the FCC authorization to operate this equipment.

RF Exposure Information according 2.1091 / 2.1093 / KDB 447498 / RSS-102

This equipment complies with FCC and IC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance of 20 cm between the radiator and your body.

This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

Cet équipement est conforme aux limites d'exposition aux rayonnements IC établies pour un environnement non contrôlé. Cet équipement doit être installé et utilisé avec un minimum de 20 cm de distance entre la source de rayonnement et votre corps.

Ce transmetteur ne doit pas être placé au même endroit ou utilisé simultanément avec un autre transmetteur ou antenne.



Philips Medizin Systeme Böblingen GmbH

Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202

Geschäftsführer: Dr. Dieter Haase (Sprecher), Dr. Peter Ziese, Matthijs de Groot

Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, www.philips.com/healthcare