



9 November 2020

Federal Communication Commission  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

Certification and Engineering Bureau  
Industry Canada  
Spectrum Engineering Branch  
3701 Carling Avenue, Building 94  
Ottawa, Ontario K2H 8S2

### **PCII Request / Re-assessment**

CTG Base Station (Model Id. 866074), FCC ID: PQC-OBRSBV1 granted on November 18, 2013 and  
IC: 3549C-OBRSBV1 granted on November 18, 2013

TO WHOM IT MAY CONCERN

Pursuant to CFR § 2.1043 and RSP-100, Philips Medizin Systeme Böblingen GmbH hereby requests  
a Permissive Change Class II / Re-assessment.

Reason:

Due to material discontinuance, the RF switch which toggles between RX and TX has to be  
replaced (drop-in replacement).



Philips Medizin Systeme Böblingen GmbH  
Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202  
Geschäftsführer: Dr. Dieter Haase (Sprecher), Dr. Peter Ziese, Matthijs de Groot  
Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, [www.philips.com/healthcare](http://www.philips.com/healthcare)


# PHILIPS

Modifications:

**Model Id. 866074:**

Ref. Des	Current	New
U820	Renesas UPG2214TB-A	CEL CG2179M2-C4

Sincerely,



Christian Noguera  
Quality & Regulatory Affairs

Philips Medizin Systeme Böblingen GmbH  
Hewlett-Packard-Str. 2 71034 Böblingen, Germany  
Tel.: +49 7031 463-2316 Fax.: +49 7031 463-2442

E-Mail: [christian.noguera@philips.com](mailto:christian.noguera@philips.com)

