



9 November 2020

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

PCII Request / Re-assessment

Wireless Transducer (Model Id. 866075, 866076, 866077), FCC ID: PQC-OBRTBV1 granted on July 11, 2013 and IC: 3549C-OBRTBV1 granted on July 11, 2013

TO WHOM IT MAY CONCERN

Pursuant to CFR § 2.1043 and RSP-100, Philips Medizin Systeme Böblingen GmbH hereby requests a Permissive Change Class II / Re-assessment.

Reason:

Due to material discontinuance, the RF switch which toggles between RX and TX has to be replaced (drop-in replacement).



Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Strasse 2, 71034 Böblingen, Germany, Tel +49(0)7031 463-0, Fax +49(0)7031 463-2202
Geschäftsführer: Dr. Dieter Haase (Sprecher), Dr. Peter Ziese, Matthijs de Groot
Registergericht Stuttgart Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen, www.philips.com/healthcare


PHILIPS

Modifications:

Model Ids. 866075, 866076, 866077:

Ref. Des	Current	New
U220	Renesas UPG2214TB-A	CEL CG2179M2-C4

Sincerely,



Christian Noguera
Quality & Regulatory Affairs

Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Str. 2 71034 Böblingen, Germany
Tel.: +49 7031 463-2316 Fax.: +49 7031 463-2442

E-Mail: christian.noguera@philips.com

