

Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

Declaration on Electrically Identical Models

TO WHOM IT MAY CONCERN

We, Philips Medizin Systeme Böblingen GmbH, declare on our sole responsibility that the new model IntelliVue MP5T with FCC ID: PQC-SRRBV1 and IC: 3549C-SRRBV1 is identical in hardware and software to the certified model

IntelliVue MP5 under FCC ID: PQC-SRRBV1 and IC: 3549C-SRRBV1


The only differences between the new model and the original model are:

- Compared to the MP5 the following measurements were removed from the MP5T: ECG, invasive pressure, CO2, and temperature.
- The following interface capabilities were removed: LAN, WLAN, IntelliVue Instrument Telemetry, and MSL.

We attest that above changing are not relevant for any RF behavior subject to regulatory items. For this reason all radio tests performed on the IntelliVue MP5 are also representative for the IntelliVue MP5T and therefore all test reports remain applicable.

If you have any questions, please feel free to contact us at the address shown below

Sincerely,



Markus Stächa
Sr. Regulatory Affairs Engineer
Telephone: +49 7031 463 2480

