

Philips Medizin Systeme Böblingen GmbH

---

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

September 04, 2008

TO WHOM IT MAY CONCERN


Limited Modular Approval Request

The Short Range Radio module consists of a transmitter and a receiver in the 2.4 GHz band.

1. It does not have its own RF shielding.
2. It has digital input buffers.
3. It has its own power supply regulation.
4. The module has been tested with two onboard antennas which are permanently attached. It has no antenna plugs.
5. The module has been tested in a stand-alone configuration.
6. The module will be labelled with its own FCC ID number.
7. The module complies with all applicable FCC rules. Instructions for maintaining compliance are provided in documents accompanying the host devices in which the module will be incorporated (Instructions for Use or Service Guide).
8. The module complies with applicable SAR and RF exposure requirements.

If you have any questions, please feel free to contact us at the address shown below

Sincerely,



Markus Stacha  
Sr. Regulatory Affairs Engineer



Philips Medizin Systeme Böblingen GmbH  
Hewlett-Packard-Str. 2, 71034 Böblingen, Germany  
Telefon +49 (0) 7031 463-0  
Telefax +49 (0) 7031 463-2202  
Geschäftsführer: Michael Dreher (Sprecher), Dr. Werner Haas  
Registergericht Böblingen Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen