

Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Aug 23, 2010

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

Declaration on Electrically Identical Models

TO WHOM IT MAY CONCERN

We, Philips Medizin Systeme Boeblingen GmbH declare on our sole responsibility that the new model

Short Range Radio Module with FCC ID: PQC-SRRBV1 and
Dual SRR Module 865221 and Dual SRR Module 865240 with IC: 3549C-SRRBV1

is identical in hardware and software to the certified model

Short Range Radio Module with FCC ID: PQC-SRRBV1 and
Short Range Radio Module with IC: 3549C-SRRBV1.

The only differences between the new model and the original model are:

- New host device main board on which the SRR Module circuitry is placed (new host device: CL Transmitter 865221),
- Change of chip antennas from 2 x Würth P/N 7488920245 to 1 x GIGAANT P/N 3030A5645 and 1 x Fractus P/N FR05-S1-N-0-104.

We attest that above changing are not relevant for any RF behavior subject to regulatory items.

If you have any questions, please feel free to contact us at the address shown below



Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Str. 2, 71034 Böblingen, Germany
Telefon +49 (0) 7031 463-0
Telefax +49 (0) 7031 463-2202
Geschäftsführer: Michael Dreher (Sprecher), Dr. Werner Haas
Registergericht Böblingen Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen