

## Philips Medizin Systeme Böblingen GmbH

---

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission  
Equipment Authorization Division, Application Processing Branch  
7435 Oakland Mills Road  
Columbia, MD 21048

Aug 23, 2010

TO WHOM IT MAY CONCERN

Request for Permissive Change Class II  
FCC ID: PQC- SRRBV1

Pursuant to CFR § 2.1043 and RSP-100, Philips Medizin Systeme Boeblingen GmbH hereby requests a Permissive Change Class II for its application with FCC ID: PQC- SRRBV1 granted on 23-Mar-2009.

Approved model: Short Range Radio Module

Modifications:

- New host device main board on which the SRR Module circuitry is placed (new host devices: CL Transmitter 865221 and Monitor MX800 865240),
- Change of chip antennas from 2 x Würth P/N 7488920245 to 1 x GIGAANT P/N 3030A5645 and 1 x Fractus P/N FR05-S1-N-0-104.

The following new exhibits will be uploaded:

- Test report on radiated measurements
- External Photos

Sincerely,

(Signature)

Name: Markus Stacha

Company: Philips Medizin Systeme Boeblingen GmbH

Address: Hewlett-Packard-Str. 2  
71034 Boeblingen / Germany

Phone: +49(0)7031 463-2840

Fax: +49(0)7031 463-2442

Email: markus.stacha@philips.com

