

Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

Declaration on Electrically Identical Models

TO WHOM IT MAY CONCERN:

We, Philips Medizin Systeme Boeblingen GmbH, declare on our sole responsibility that IntelliVue MP2 is the marketing variant of the IntelliVue X2, which provides reduced functionality. Both models, the MP2 and the X2, have the same hardware and software. Compared to the X2, the MP2 model lacks the capability to communicate directly with another IntelliVue patient monitor.

We attest that above changing is not relevant to any RF behavior subject to regulatory items. For this reason, all radio tests performed on the X2 are also representative for the MP2 and all test reports remain applicable.

If you have any questions, please feel free to contact me at the address shown below

Sincerely,



Name	Andreas Suchi Sr. Regulatory Affairs Engineer
Company	Philips Medizin Systeme Boeblingen GmbH
Address	Hewlett-Packard-Str. 2 71034 Boeblingen, Germany
Telephone	+49(0)7031 463-1291
Fax	+49(0)7031 463-2442
E-mail	markus.stacha@philips.com

