

Philips Medizin Systeme Böblingen GmbH

Postfach 1471, 71004 Böblingen, Germany

Federal Communication Commission
Equipment Authorization Division, Application Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Industry Canada
Certification & Engineering Bureau
3701 Carling Ave. Bldg. 94
Ottawa, On K2H 8S2

Aug 23, 2010

TO WHOM IT MAY CONCERN

Request for Permissive Change Class II
FCC ID: PQC-WLANBV1, IC: 3549C-WLANBV1

Pursuant to CFR § 2.1043 and RSP-100, Philips Medizin Systeme Boeblingen GmbH hereby requests a Permissive Change Class II for its application with FCC ID: PQC-WLANBV1 and IC: 3549C-WLANBV1 granted on 03/06/2009.

Approved model: M3002-66480 WLAN Module

Modifications:

- no hardware changes
- new antenna P/N 453564175981, Rev. 1011
- add portable platform

The following new exhibits will be uploaded:

- Test report on radiated measurements
- RF Exposure Info
- External Photos

Sincerely,

(Signature)

Name: Markus Stacha

Company: Philips Medizin Systeme Boeblingen GmbH

Address: Hewlett-Packard-Str. 2
71034 Boeblingen / Germany

Phone: +49(0)7031 463-2840

Fax: +49(0)7031 463-2442

Email: markus.stacha@philips.com



Philips Medizin Systeme Böblingen GmbH
Hewlett-Packard-Str. 2, 71034 Böblingen, Germany
Telefon +49 (0) 7031 463-0
Telefax +49 (0) 7031 463-2202
Geschäftsführer: Michael Dreher (Sprecher), Dr. Werner Haas
Registergericht Böblingen Reg.-Nr. HRB 245187, Sitz der Gesellschaft: Böblingen