

June 7, 2010

Federal Communications Commission
Washington, DC 20554

RE: Request for Limited Modular Approval for PQC-WMTS-ITS2

To Whom It May Concern:


Philips Medical Systems of Andover, Massachusetts hereby submits the enclosed application for Equipment Authorization under 47 CFR Part 95, Subpart H for a wireless medical telemetry transmitter module, whose FCC grant ID would be PQC-WMTS-ITS2.

This module incorporates the same radio technology and implementation as that used in the Philips TRx4841A. This device differs in that it is modular, intended to be incorporated into certain Philips patient monitors, thereby providing wireless communication to existing infrastructure in healthcare facilities. The module complies with FCC Rules as a self-contained entity; it does not rely on external mounting or shielding or other input/output mechanisms to achieve compliance to either Parts 15 or 95 of the FCC Rules. The module is not being offered for sale as an OEM component, it is for incorporation into Philips Medical Systems products only.

Upon receipt of this grant, the grant ID will be applied to the device. Furthermore, the instruction to the user concerning frequency coordination (as per 47 CFR 95.1109) will appear on the module. Products which incorporate this device will be labeled "Contains PQC-WMTS-ITS2", and their associated Instructions for Use will also advise the user concerning frequency coordination (47 CFR 95.1109).

Thank you in advance for your consideration of this application.

Sincerely,



Terrence J. Sweeney
VP – Global Clinical Affairs